

## JRC TECHNICAL REPORTS

# Evaluation of the Laboratory Comparison Exercise for SO<sub>2</sub>, CO, O<sub>3</sub>, NO and NO<sub>2</sub> 19-23 October 2015, Ispra

*European Commission  
harmonisation programme for  
Air Quality Measurements*

Barbiere M., Lagler F., Borowiak A.

2015



This publication is a Technical report by the Joint Research Centre, the European Commission's in-house science service. It aims to provide evidence-based scientific support to the European policy-making process. The scientific output expressed does not imply a policy position of the European Commission. Neither the European Commission nor any person acting on behalf of the Commission is responsible for the use which might be made of this publication.

**Contact information**

Maurizio Barbieri

Address: Joint Research Centre

Via Enrico Fermi 2749, TP 120

21027 Ispra (VA)

ITALY

E-mail: maurizio.barbieri@jrc.ec.europa.eu

Tel.: +39 0332 783057

**JRC Science Hub**

<https://ec.europa.eu/jrc>

JRC102049

EUR 28047 EN

PDF	ISBN 978-92-79-46959-6	ISSN 1831-9424	doi:10.2790/805192
Print	ISBN 978-92-79-46958-9	ISSN 1018-5593	doi:10.2790/92143

Luxembourg: Publications Office of the European Union, 2016

© European Union, 2016

Reproduction is authorised provided the source is acknowledged.

All images © European Union 2016

How to cite: Maurizio Barbieri, Friedrich Lagler, Annette Borowiak; Evaluation of the Laboratory Comparison exercise for SO<sub>2</sub>, CO, O<sub>3</sub>, NO and NO<sub>2</sub> 19-23 October 2015, Ispra; EUR 28047 EN; doi:10.2790/805192

In collaboration with:

Nikolay Panayotov, Milena Parvanova, Tomas Gaizutis, Juozas Molis, Gregory Abbou, Christophe Debert, Jan Silhavy, Miroslav Vokoun, Christakis Papadopoulos, Christos Kiza, Pilar Morillo, Maria Rodriguez, Ljiljana Novakovic, Slavisa Mitrovic, François Mathé, Emmanuel Tison, Leif Marsteen and Franck Dauge.

	<b>NAME</b>	<b>VERSION</b>	<b>DATE</b>
<b>AUTHOR</b>	M. BARBIERE, F. LAGLER, A. BOROWIAK	DRAFT 1	16/02/2016
<b>REVIEW</b>	N. JENSEN	DRAFT 2	27/05/2016
<b>APPROVAL</b>	E. VIGNATI	1.0	08/07/2016

## Abstract

Within the harmonisation programme of Air Quality monitoring in Europe the European Reference Laboratory of Air Pollution (ERLAP) is organizing Inter-Laboratory Comparison Exercises (IE) in the facility of Ispra (Italy).

From the 19 to the 23 October 2015 in Ispra (Italy), eight Laboratories of AQUILA (Network of European Air Quality Reference Laboratories) and one of French regional network (AIRPARIF) met for a comparison exercise to evaluate their proficiency in the analysis of inorganic gaseous pollutants. In order to cover the prescription of the European Directive 2008/50/EC [1] and its the recent amendment 2015/1480/EC [42] about air quality, the following pollutants were measured: sulphur dioxide (SO<sub>2</sub>), nitrogen dioxide (NO<sub>2</sub>), nitrogen monoxide (NO), carbon monoxide (CO) and ozone (O<sub>3</sub>).

The proficiency evaluation, where each participant's bias was compared to two criteria, provides information on compliance with Data Quality Objectives (DQO) and measurement capabilities of the National Air Quality Laboratories to the European Commission (EC) and can be used by participants in their laboratory's quality system.

On the basis of criteria imposed by the EC, 75.6% of the results reported by the participants was satisfactory both in terms of measured values and reported uncertainties. Part of the results (18.1%) had acceptable measured values, but the reported uncertainties were either too high (10.8%) or too small (7.3%). Against the usual tendency during this IE a great number of results (2.2%) were unsatisfactory for both the value and the uncertainty.

Comparability of results among AQUILA participants at the highest concentration level, excluding outliers, is acceptable for almost all pollutants measurements. CO and NO<sub>2</sub> show a deviation from the objectives.

# Contents

ABSTRACT .....	4
CONTENTS .....	5
LIST OF TABLES .....	6
LIST OF FIGURES .....	7
ABBREVIATIONS .....	8
MATHEMATICAL SYMBOLS.....	8
1. INTRODUCTION .....	9
2. INTER-LABORATORY ORGANISATION.....	11
2.1. PARTICIPANTS.....	11
2.2. PREPARATION OF TEST MIXTURES .....	13
3. THE EVALUATION OF LABORATORY'S MEASUREMENT PROFICIENCY .....	14
3.1. Z' SCORES .....	14
3.2. E <sub>N</sub> SCORES.....	18
4. PERFORMANCE CHARACTERISTICS OF INDIVIDUAL LABORATORIES .....	25
4.1. CONVERTER EFFICIENCIES OF NO <sub>2</sub> -TO-NO FOR NO <sub>x</sub> ANALYSERS .....	25
5. DISCUSSION .....	27
6. CONCLUSIONS .....	29
7. REFERENCES.....	31
<i>Annex A. Assigned values .....</i>	<i>34</i>
<i>Annex B. The results of the IE.....</i>	<i>37</i>
<i>Annex C: The precision of standardised measurement methods .....</i>	<i>56</i>
<i>Annex C. Results evaluation for consistency and outlier test.....</i>	<i>62</i>
<i>Annex D. Accreditation certificate .....</i>	<i>63</i>
.....	63

## List of tables

TABLE 1: LIST OF PARTICIPATING ORGANISATIONS .....	11
TABLE 2: LIST OF INSTRUMENTS USED BY PARTICIPANTS.....	12
TABLE 3: SEQUENCE PROGRAMME OF GENERATED TEST GASES WITH INDICATIVE POLLUTANT CONCENTRATIONS .....	13
TABLE 4: THE STANDARD DEVIATION FOR PROFICIENCY ASSESSMENT ( $\sigma_p$ ) .....	15
TABLE 5: QUESTIONABLE AND UNSATISFACTORY RESULTS ACCORDING TO $Z'$ .....	15
TABLE 6: UNSATISFACTORY RESULTS ACCORDING TO THE $E_N$ .....	19
TABLE 7: EFFICIENCY OF NO <sub>2</sub> -TO-NO CONVERTERS.....	26
TABLE 8: GENERAL ASSESSMENT OF PROFICIENCY RESULTS .....	28
TABLE 9: FLAGS SUMMARY.....	29
TABLE 10: $Z'$ SCORES SUMMARY .....	30
TABLE 11: VALIDATION OF ASSIGNED VALUES (X).....	35
TABLE 12: REPORTED VALUES FOR SO <sub>2</sub> RUN 0 .....	37
TABLE 13: REPORTED VALUES FOR SO <sub>2</sub> RUN 1 .....	38
TABLE 14: REPORTED VALUES FOR SO <sub>2</sub> RUN 2 .....	38
TABLE 15: REPORTED VALUES FOR SO <sub>2</sub> RUN 3 .....	39
TABLE 16: REPORTED VALUES FOR SO <sub>2</sub> RUN 4 .....	39
TABLE 17: REPORTED VALUES FOR SO <sub>2</sub> RUN 5 .....	40
TABLE 18: REPORTED VALUES FOR CO RUN 0.....	41
TABLE 19: REPORTED VALUES FOR CO RUN 1.....	41
TABLE 20: REPORTED VALUES FOR CO RUN 2 .....	42
TABLE 21: REPORTED VALUES FOR CO RUN 3 .....	42
TABLE 22: REPORTED VALUES FOR CO RUN 4 .....	43
TABLE 23: REPORTED VALUES FOR CO RUN 5 .....	43
TABLE 24: REPORTED VALUES FOR O <sub>3</sub> RUN 0 .....	44
TABLE 25: REPORTED VALUES FOR O <sub>3</sub> RUN 1 .....	44
TABLE 26: REPORTED VALUES FOR O <sub>3</sub> RUN 2 .....	45
TABLE 27: REPORTED VALUES FOR O <sub>3</sub> RUN 3 .....	45
TABLE 28: REPORTED VALUES FOR O <sub>3</sub> RUN 4 .....	46
TABLE 29: REPORTED VALUES FOR O <sub>3</sub> RUN 5 .....	46
TABLE 30: REPORTED VALUES FOR NO RUN 0.....	47
TABLE 31: REPORTED VALUES FOR NO RUN 1.....	47
TABLE 32: REPORTED VALUES FOR NO RUN 2.....	48
TABLE 33: REPORTED VALUES FOR NO RUN 3.....	48
TABLE 34: REPORTED VALUES FOR NO RUN 4.....	49
TABLE 35: REPORTED VALUES FOR NO RUN 5.....	49
TABLE 36: REPORTED VALUES FOR NO RUN 6.....	50
TABLE 37: REPORTED VALUES FOR NO RUN 7.....	50
TABLE 38: REPORTED VALUES FOR NO RUN 8.....	51
TABLE 39: REPORTED VALUES FOR NO RUN 9.....	51
TABLE 40: REPORTED VALUES FOR NO RUN 10.....	52
TABLE 41: REPORTED VALUES FOR NO <sub>2</sub> RUN 0 .....	53
TABLE 42: REPORTED VALUES FOR NO <sub>2</sub> RUN 2 .....	53
TABLE 43: REPORTED VALUES FOR NO <sub>2</sub> RUN 4 .....	54
TABLE 44: REPORTED VALUES FOR NO <sub>2</sub> RUN 6 .....	54
TABLE 45: REPORTED VALUES FOR NO <sub>2</sub> RUN 8 .....	55
TABLE 46: REPORTED VALUES FOR NO <sub>2</sub> RUN 10 .....	55
TABLE 47: CRITICAL VALUES OF T USED IN THE REPEATABILITY R AND R EVALUATION .....	56
TABLE 48: THE R AND R OF SO <sub>2</sub> STANDARD MEASUREMENT METHOD.....	57
TABLE 49: THE R AND R OF CO STANDARD MEASUREMENT METHOD.....	58
TABLE 50: THE R AND R OF O <sub>3</sub> STANDARD MEASUREMENT METHOD .....	59
TABLE 51: THE R AND R OF NO STANDARD MEASUREMENT METHOD .....	60
TABLE 52: THE R AND R OF NO <sub>2</sub> STANDARD MEASUREMENT METHOD.....	61
TABLE 53: "GENUINE" STATISTICAL OUTLIERS ACCORDING TO GRUBB'S ONE OUTLYING OBSERVATION TEST .....	62
TABLE 54: STRAGGLERS ACCORDING TO GRUBB'S ONE OBSERVATION TEST .....	62

## List of figures

FIGURE 1: Z' SCORES EVALUATIONS OF SO <sub>2</sub> MEASUREMENTS .....	16
FIGURE 2: Z' SCORES EVALUATIONS OF CO MEASUREMENTS .....	16
FIGURE 3: Z' SCORES EVALUATIONS OF O <sub>3</sub> MEASUREMENTS .....	17
FIGURE 4: Z' SCORES EVALUATIONS OF NO MEASUREMENTS .....	17
FIGURE 5: Z' SCORES EVALUATIONS OF NO <sub>2</sub> MEASUREMENTS .....	18
FIGURE 6: BIAS OF PARTICIPANT'S SO <sub>2</sub> MEASUREMENT RESULTS .....	20
FIGURE 7: BIAS OF PARTICIPANT'S CO MEASUREMENT RESULTS .....	21
FIGURE 8: BIAS OF PARTICIPANT'S O <sub>3</sub> MEASUREMENT RESULTS .....	22
FIGURE 9: BIAS OF PARTICIPANT'S NO MEASUREMENT RESULTS .....	23
FIGURE 10: BIAS OF PARTICIPANT'S NO <sub>2</sub> MEASUREMENT RESULTS .....	24
FIGURE 11: BIAS OF PARTICIPANT'S NO <sub>2</sub> MEASUREMENTS FOR RUN NUMBERS 1, 3, 5, 7 AND 9 .....	25
FIGURE 12: THE DECISION DIAGRAM FOR GENERAL ASSESSMENT OF PROFICIENCY RESULTS .....	27
FIGURE 13: REPORTED VALUES FOR SO <sub>2</sub> RUN 0 .....	37
FIGURE 14: REPORTED VALUES FOR SO <sub>2</sub> RUN 1 .....	38
FIGURE 15: REPORTED VALUES FOR SO <sub>2</sub> RUN 2 .....	38
FIGURE 16: REPORTED VALUES FOR SO <sub>2</sub> RUN 3 .....	39
FIGURE 17: REPORTED VALUES FOR SO <sub>2</sub> RUN 4 .....	39
FIGURE 18: REPORTED VALUES FOR SO <sub>2</sub> RUN 5 .....	40
FIGURE 19: REPORTED VALUES FOR CO RUN 0 .....	41
FIGURE 20: REPORTED VALUES FOR CO RUN 1 .....	41
FIGURE 21: REPORTED VALUES FOR CO RUN 2 .....	42
FIGURE 22: REPORTED VALUES FOR CO RUN 3 .....	42
FIGURE 23: REPORTED VALUES FOR CO RUN 4 .....	43
FIGURE 24: REPORTED VALUES FOR CO RUN 5 .....	43
FIGURE 25: REPORTED VALUES FOR O <sub>3</sub> RUN 0 .....	44
FIGURE 26: REPORTED VALUES FOR O <sub>3</sub> RUN 1 .....	44
FIGURE 27: REPORTED VALUES FOR O <sub>3</sub> RUN 2 .....	45
FIGURE 28: REPORTED VALUES FOR O <sub>3</sub> RUN 3 .....	45
FIGURE 29: REPORTED VALUES FOR O <sub>3</sub> RUN 4 .....	46
FIGURE 30: REPORTED VALUES FOR O <sub>3</sub> RUN 5 .....	46
FIGURE 31: REPORTED VALUES FOR NO RUN 0 .....	47
FIGURE 32: REPORTED VALUES FOR NO RUN 1 .....	47
FIGURE 33: REPORTED VALUES FOR NO RUN 2 .....	48
FIGURE 34: REPORTED VALUES FOR NO RUN 3 .....	48
FIGURE 35: REPORTED VALUES FOR NO RUN 4 .....	49
FIGURE 36: REPORTED VALUES FOR NO RUN 5 .....	49
FIGURE 37: REPORTED VALUES FOR NO RUN 6 .....	50
FIGURE 38: REPORTED VALUES FOR NO RUN 7 .....	50
FIGURE 39: REPORTED VALUES FOR NO RUN 8 .....	51
FIGURE 40: REPORTED VALUES FOR NO RUN 9 .....	51
FIGURE 41: REPORTED VALUES FOR NO RUN 10 .....	52
FIGURE 42: REPORTED VALUES FOR NO <sub>2</sub> RUN 0 .....	53
FIGURE 43: REPORTED VALUES FOR NO <sub>2</sub> RUN 2 .....	53
FIGURE 44: REPORTED VALUES FOR NO <sub>2</sub> RUN 4 .....	54
FIGURE 45: REPORTED VALUES FOR NO <sub>2</sub> RUN 6 .....	54
FIGURE 46: REPORTED VALUES FOR NO <sub>2</sub> RUN 8 .....	55
FIGURE 47: REPORTED VALUES FOR NO <sub>2</sub> RUN 10 .....	55
FIGURE 48: THE R AND r OF SO <sub>2</sub> STANDARD MEASUREMENT METHOD AS A FUNCTION OF CONCENTRATION .....	57
FIGURE 49: THE R AND r OF CO STANDARD MEASUREMENT METHOD AS A FUNCTION OF CONCENTRATION .....	58
FIGURE 50: THE R AND r OF O <sub>3</sub> STANDARD MEASUREMENT METHOD AS A FUNCTION OF CONCENTRATION .....	59
FIGURE 51: THE R AND r OF NO STANDARD MEASUREMENT METHOD AS A FUNCTION OF CONCENTRATION .....	60
FIGURE 52: THE R AND r OF NO <sub>2</sub> STANDARD MEASUREMENT METHOD AS A FUNCTION OF CONCENTRATION .....	61

## Abbreviations

AQUILA	Network of National Reference Laboratories for Air Quality
BIPM	Bureau International des Poids et Mesures
CCQM	Consultative Committee for Amount of Substance: Metrology in Chemistry and Biology
CO	Carbon monoxide
CRMs	Certified Reference Materials
DQO	Data Quality Objective
ERLAP	European Reference Laboratory of Air Pollution
EC	European Commission
GPT	Gas Phase Titration
IE	Inter-laboratory Comparison Exercise
ISO	International Organisation for Standardization
JRC	Joint Research Centre
NIST	National Institute of Standards and Technology
NO	Nitrogen monoxide
NO <sub>2</sub>	Nitrogen dioxide
NO <sub>x</sub>	The oxides of nitrogen, the sum of NO and NO <sub>2</sub>
NRL	National Reference Laboratory
O <sub>3</sub>	Ozone
SO <sub>2</sub>	Sulphur dioxide
WHO-CC	World Health Organisation Collaborating Centre for Air Quality Management and Air Pollution Control, Berlin

## Mathematical symbols

<i>symbol</i>	<i>explanation</i>
$\alpha$	converter efficiency (EN 14211)
$E_n$	$E_n$ scores statistic (ISO 13528)
$r$	repeatability limit (ISO 5725)
$R$	reproducibility limit (ISO 5725)
$\sigma_p$	standard deviation for proficiency assessment (ISO 13528)
$x^*$	robust average (Annex C ISO 13528)
$s^*$	robust standard deviation (Annex C ISO 13528)
$s_r$	repeatability standard deviation (ISO 5725)
$s_R$	reproducibility standard deviation (ISO 5725)
$U_{X'}$	expanded uncertainty of the assigned/reference value (ISO 13528)
$U_{xi}$	expanded uncertainty of the participant's value
$u_{X'}$	standard uncertainty of the assigned/reference value (ISO 13528)
$X$	assigned/reference value (ISO 13528)
$x_i$	average of three values reported by the participant $i$ (for particular parameter and concentration level) (ISO 5725)
$x_{i,j}$	$j$ is the reported value of participant $i$ (for particular parameter and concentration level) (ISO 5725)
$z'$	$z'$ scores statistic (ISO 13528)



## 1. Introduction

The Directive 2008/50/EC [1] on ambient air quality and cleaner air for Europe sets a framework for a harmonized air quality assessment in Europe.

One important objective of the Directive [1] is that the ambient air quality shall be assessed on the basis of common methods and criteria. It concerns with the air pollutants SO<sub>2</sub>, NO<sub>2</sub>, NO, particulate matter, lead, benzene, CO and O<sub>3</sub>. Among other things it specifies the reference methods for measurements and DQOs for the accuracy of measurements.

The EC has supported the development and publication of reference measurement methods for CO [2], SO<sub>2</sub> [3], NO-NO<sub>2</sub> [4] and O<sub>3</sub> [5] as European standards. Appropriate calibration methods [6], [7] and [8] have been standardised by the International Organisation for Standardisation (ISO).

As foreseen in the Air Quality Directive, the European Reference Laboratory of Air Pollution (ERLAP) of the Directorate Energy, Transport & Climate the Joint Research Centre (JRC) regularly organises IEs to assess and improve the status of comparability of measurements of National Reference Laboratories (NRLs) of the Member States of the European Union.

The World Health Organisation Collaborating Centre for Air Quality Management and Air Pollution Control (WHO CC) in Berlin is carrying out similar activities since 1994 [9] [10], [24], [31], [33], [35], [38] and [43] but with a view to obtaining harmonized air quality data for health related studies. Their programme is integrated within the WHO EURO area, which includes public health institutes and other national institutes — especially from the Central and Eastern Europe, the Caucasus and countries in Central Asia.

In 2004, it was decided to bring together the efforts of both the JRC-ERLAP and the WHO CC and to coordinate activities as much as possible, with a view to optimizing resources and securing better international harmonisation.

The following report deals with the IE that took place from 19 to the 23 October 2015 in Ispra (Italy).

Since 1990 ERLAP has been organizing IEs aiming at evaluating the comparability of measurements carried out by NRLs and promoting information exchange among the expert laboratories. Currently, a more systematic approach has been adopted, in accordance with the Network of National Reference Laboratories for Air Quality (AQUILA) [11], aiming both at providing an alert mechanism for the purposes of the EC legislation and at supporting the implementation of quality schemes by NRLs.

The methodology for the organisation of IEs was developed by ERLAP in collaboration with AQUILA and is described in a paper on the organisation of laboratory comparison exercises for gaseous air pollutants [12].

This evaluation scheme was adopted by AQUILA in December 2008 and has since been applied to all IEs. It contains common criteria to alert the EC of possible performance failures which do not rely solely on the uncertainty claimed by participants. The evaluation scheme implements the  $z'$  method [13] with the uncertainty requirements for calibration gases stated in the European standards [2], [3], [4] and [5], which are consistent with the DQOs of European directives.

According to the above mentioned document, NRLs with an overall unsatisfactory performance in the  $z'$  evaluation (one unsatisfactory or two questionable results per parameter) ought to repeat their participation in the following IE in order to demonstrate remediation measures [12]. In addition, considering that the evaluation scheme should be useful to participants for accreditation according to ISO/IEC 17025 [44], they are requested to include their measurement uncertainty. Hence, the results

of participants (measurement values and uncertainties) are compared to the assigned values applying the  $E_n$  scores method [13].

Beside the proficiency of participating laboratories, the repeatability ( $r$ ) and reproducibility ( $R$ ) of standardised measurement methods [14], [15] and [16] are evaluated as well. These group evaluations are useful indicators of trends in measurement quality over different IE.

## 2. Inter-laboratory organisation

The IE was announced in June 2015 to the members of the AQUILA network and the WHO CC representative. Registration was opened in September 2015 and closed at the beginning of October 2015.

Participants were required to bring their own measurement instruments, data acquisition equipment and travelling standards (to be used for calibrations or checks during the IE).

Participants were invited to arrive on Monday, 19<sup>th</sup> of October 2015, for the installation of their equipment. The calibration of NO<sub>x</sub> and O<sub>3</sub> analysers was carried out on Tuesday morning and the generation of NO<sub>x</sub> and O<sub>3</sub> gas mixtures started at 11:00.

The calibration of SO<sub>2</sub> and CO analysers was carried out on Wednesday afternoon and the generation of CO and SO<sub>2</sub> gas mixtures started at 20:00.

The test gases generation and measurements finished on Thursday at 9:00.

### 2.1. Participants

All participants were either organisations dealing with the routine ambient air monitoring or institutions involved in environmental or public health protection. The national representatives came from Bulgaria, Republic of Lithuania, France, the Czech Republic, Cyprus, Republic of Serbia and Norway.

Country	Laboratory	Code
Bulgaria	Executive Environmental Agency (EEA)	A
Republic of Lithuania	Environmental Protection Agency (AAA)	B
France	AirParif (AIRPARIF)	C
Czech Republic	Czech Hydrometereological Institute (CHMI)	D
Cyprus	Dept. of Labour Inspection (DLI)	E
Spain	Instituto De Salud Carlos III (ISCIII)	F
European Commission	European Reference Laboratory for Air Pollution (ERLAP)	G
Republic of Serbia	Serbian Environmental Protection Agency (SEPA)	H
France	LCSQA – Mines de Douai (LCSQA)	I
Norway	Norwegian Institute for Air Research (NILU)	L

**Table 1: list of participating organisations**

Table 2 reports the manufacturer and model of the instrumentation used by each participant during the IE, included those used in the calculation of the assigned values.

As a whole, the instrumentation was manufactured by four different companies for all parameters analysed.

The list contains the information reported by participants and by no means can it be considered as an implicit or explicit endorsement of the organisers to any specific type of instrumentation.

Gas	Lab code	Instrument
SO <sub>2</sub>	A	Horiba, 2009, APSA 370
	B	Horiba, Apsa 370, 2011
	C	SO2 analyzer: Thermo electron 43C (2001)
	D	Thermo Environmental Instruments, Inc., 1997, model 43C
	E	Ecotech, Year: 2005, EC9850B with an external pump
	F	THERMO 43i
	G	Thermo Scientific, 2009, 43i-TLE
	H	Teledyn API; 2011; T100
	I	API, 2008, UV fluorescence, model 100E
	L	Teledyne API, 2005, API 100E
NO <sub>x</sub>	A	Horiba, 2009, analyzer NOx, APNA 370
	B	Horiba, Apna 370, 2011
	C	NO analyzer: Environnement SA AC32M (2002) / NO2 analyzer: Environnement SA AS32M (2015)
	D	Thermo Environmental Instruments, Inc., 1997, model 42C
	E	Ecotech, Year: 2005, EC 9841B with an external pump
	F	TELEDYNE API 200E
	G	Thermo Electrom Corporation, 1999, 42C
	H	Teledyn API; 2011; T200
	I	Thermo Environmental Instruments Inc., 2000, model 42C
	L	Teledyne API, 2005, API 200E
CO	A	Horiba, 2009, APMA 370
	B	Horiba, Apma 370, 2011
	C	CO analyzer: Thermo electron 48C (2001)
	D	Thermo Environmental Instruments, Inc., 1997, model 48C
	E	Ecotech, Year: 2005, EC9830B with an external pump
	F	TELEDYNE API T300
	G	Horiba Model APMA-370, 2010
	H	Teledyn API; 2011; T300
	I	HORIBA, 2013, non dispersive IR spectroscopy, APMA 370
	L	Teledyne API, 2005, API 300E
O <sub>3</sub>	A	Horiba, 2008, APOA 370
	B	Horiba, APOA 370, 2011
	C	O3 analyzer: Environnement SA O342e (2015)
	D	Thermo Environmental Instruments, Inc., 1997, model 49C
	E	Thermo Electron Corporation, Year: 2006, Model 49i Ozone Analyzer
	F	THERMO 49i
	G	Thermo Scientific 49-PS , 2014
	H	Teledyn API; 2011; T400
	I	Thermo Environmental Instruments Inc., 2000, model 49 C
	L	Teledyne API, 2005, API 400E

**Table 2: list of instruments used by participants**

## 2.2. Preparation of test mixtures

The ERLAP IE facility has been described in several reports [17] and [18]. During this IE, gas mixtures were prepared for SO<sub>2</sub>, CO, O<sub>3</sub>, NO and NO<sub>2</sub> at concentration levels around limit values, critical levels and assessment thresholds set by the European Air Quality Directive [1].

The test mixtures were prepared by the dilution of gases from cylinders containing high concentrations of NO, SO<sub>2</sub> or CO using thermal mass flow controllers [8]. O<sub>3</sub> was added using an ozone generator and NO<sub>2</sub> was produced applying the gas phase titration (GPT) method [19] in a condition of NO excess.

The participants were required to report three 30-min mean measurements for each concentration level (run) in order to evaluate the repeatability of standardised measurement methods. Zero concentration levels were generated for 1 hour and one 30-min mean measurement was reported. The sequence programme of generated test gases is given in Table 3.

day	start time	duration	parameter	installation	calibration	Zero Air	NO	NO <sub>2</sub>	O <sub>3</sub>	CO	SO <sub>2</sub>
		h				nmol/mol	nmol/mol	nmol/mol	nmol/mol	mmol/mol	nmol/mol
1st	9:00	5	/	X							
2nd	8:00	3	/		X						
2nd	11:00	1	NO-NO <sub>2</sub> -O <sub>3</sub>			0					
2nd	12:00	2	NO-NO <sub>2</sub>				600				
2nd	14:00	2	NO-NO <sub>2</sub>				450	150			
2nd	16:00	2	O <sub>3</sub>						150		
2nd	18:00	2	NO-NO <sub>2</sub>				210				
2nd	20:00	2	NO-NO <sub>2</sub>				100	110			
2nd	22:00	2	O <sub>3</sub>						120		
3rd	0:00	2	NO-NO <sub>2</sub>				80				
3rd	2:00	2	NO-NO <sub>2</sub>				20	60			
3rd	4:00	2	O <sub>3</sub>						55		
3rd	6:00	2	NO-NO <sub>2</sub>				390				
3rd	8:00	2	NO-NO <sub>2</sub>				300	90			
3rd	10:00	2	O <sub>3</sub>						90		
3rd	12:00	2	NO-NO <sub>2</sub>				50				
3rd	14:00	2	NO-NO <sub>2</sub>				30	20			
3rd	16:00	2	O <sub>3</sub>						15		
3rd	< 18:00	2	calibration		X						
3rd	20:00	1	CO-SO <sub>2</sub>			0					
3rd	21:00	2	CO-SO <sub>2</sub>							8.5	130
3rd	23:00	2	CO-SO <sub>2</sub>							3.5	70
4th	1:00	1	CO-SO <sub>2</sub>	Zero Air not reported						0	0
4th	2:00	2	CO-SO <sub>2</sub>							1.0	5
4th	4:00	2	CO-SO <sub>2</sub>							5.0	30
4th	6:00	2	CO-SO <sub>2</sub>							2.0	10
4th	8:00	1				0					
4th	9:00	END									

**Table 3: sequence programme of generated test gases with indicative pollutant concentrations**

### 3. The evaluation of laboratory's measurement proficiency

To evaluate the participants measurement proficiency the methodology described in ISO 13528 [13] was applied. It was agreed among the AQUILA members to consider the measurement results of ERLAP as the assigned/reference values for the whole IE [12].

The traceability of ERLAP's measurement results and the method applied to validate them are presented in Annex A. In the following proficiency evaluations, the uncertainty of test gas homogeneity (Annex A) was added to the uncertainties of ERLAP's measurement results.

All data reported by participating laboratories are presented in Annex B.

As it is described in the position paper [12], the proficiency of the participants was assessed by calculating two performance indicators.

The first performance indicator ( $z'$  scores) tests whether the difference between the participant's measured value and the assigned/reference value remains within the limits of a common criterion.

The second performance indicator ( $E_n$  scores) tests if the difference between the participant's measured values and assigned/reference value remains within the limits of a criterion, which is calculated individually for each participant, from the uncertainty of the participants measurement result and the uncertainty of the assigned/reference value.

#### 3.1. $z'$ scores

The  $z'$  scores statistic is calculated according to ISO 13528 [13] as:

$$z' = \frac{x_i - X}{\sqrt{\sigma_p^2 + u_x^2}} = \frac{x_i - X}{\sqrt{(a \cdot X + b)^2 + u_x^2}} \quad \text{Equation 1}$$

Where ' $x_i$ ' is a participant's average value for each run, ' $X$ ' is the assigned/reference value, ' $\sigma_p$ ' is the 'standard deviation for proficiency assessment' and ' $u_x$ ' is the standard uncertainty of the assigned value. For ' $a$ ' and ' $b$ ' see Table 4.

In the European standards [2], [3], [4] and [5] the uncertainties for calibration gases used in ongoing quality control are prescribed. In fact, it is stated that the maximum permitted expanded uncertainty for calibration gases is 5% and that 'zero gas' shall not give an instrument reading higher than the detection limit. As one of the tasks of NRLs is to supply calibration gas mixtures, the 'standard deviation for proficiency assessment' ( $\sigma_p$ ) [13] is calculated in a fitness-for-purpose manner from requirements given in European standards.

Over the whole measurement range,  $\sigma_p$  is calculated by linear interpolation between 2.5% at the calibration point (75% of calibration range) and the limit of detection at zero concentration level. The limits of detection of studied measurement methods were evaluated from the data of previous IE. The linear function parameters of  $\sigma_p$  are given in Table 4.

Gas	$\sigma_p = a \cdot c + b$	
	a	b
		nmol/mol
SO <sub>2</sub>	0.022	1
CO	0.024	100
O <sub>3</sub>	0.020	1
NO	0.024	1
NO <sub>2</sub>	0.020	1

**Table 4: The standard deviation for proficiency assessment ( $\sigma_p$ )**

$\sigma_p$  is a linear function of concentration (c) with parameters: slope (a) and intercept (b).

The assessment of results in the  $z'$  evaluation is made according to the following criteria:

- $|z'| \leq 2$  are considered satisfactory.
- $2 < |z'| \leq 3$  are considered questionable.
- $|z'| > 3$  are considered unsatisfactory. Scores falling in this range are very unusual and are taken as evidence that has occurred an anomaly that should be investigated and corrected. A list of questionable and unsatisfactory results is reported in the following table (Table 5).

Parameter	Lab. code	Lab. value	$z'$ evaluation
NO2 _4	A	108.9	questionable
CO _1	B	9.64	unsatisfactory
CO _2	B	4.004	questionable
CO _4	B	5.7	questionable
NO2 _2	B	138.2	unsatisfactory
NO2 _8	B	82.67	unsatisfactory
NO2 _4	E	108.2	questionable
NO _2	F	489.3	questionable
NO _4	F	109.1	questionable
NO _6	F	23.08	questionable
CO _1	I	4.696	unsatisfactory
CO _2	I	2.014	unsatisfactory
CO _3	I	0.757	questionable
CO _4	I	2.882	unsatisfactory
CO _5	I	1.253	unsatisfactory
NO _10	I	24.91	questionable
NO _4	I	93.38	questionable
NO _5	I	71.99	questionable
NO _9	I	43.33	questionable
NO2 _6	I	55.53	questionable


**Table 5: Questionable and unsatisfactory results according to  $z'$**

The results of the  $z'$  evaluation are presented in bar plots (Figure 1 to Figure 5) in which the  $z'$  of each participant are grouped together, and assessment criteria are presented as  $z' = \pm 2$  and  $z' = \pm 3$  lines.



**Figure 1: z' scores evaluations of SO<sub>2</sub> measurements**

Scores are given for each participant and each tested concentration level (run). Run number order (with nominal concentration) is: 0 (0 nmol/mol), 1 (130 nmol/mol), 2 (70 nmol/mol), 3 (5 nmol/mol), 4 (30 nmol/mol), 5 (10 nmol/mol). The assessment criteria are presented as  $z' = \pm 2$  (blue line) and  $z' = \pm 3$  (red line). They represent the limits for the questionable and unsatisfactory results.



**Figure 2: z' scores evaluations of CO measurements**


Scores are given for each participant and each tested concentration level (run). Run number order (with nominal concentration) is: 0 (0  $\mu\text{mol/mol}$ ), 1 (8,5  $\mu\text{mol/mol}$ ), 2 (3,5  $\mu\text{mol/mol}$ ), 3 (1  $\mu\text{mol/mol}$ ), 4 (5  $\mu\text{mol/mol}$ ), 5 (2  $\mu\text{mol/mol}$ ). The assessment criteria are presented as  $z' = \pm 2$  (blue line) and  $z' = \pm 3$  (red line). They represent the limits for the questionable and unsatisfactory results.






**Figure 3: z' scores evaluations of O<sub>3</sub> measurements**

Scores are given for each participant and each concentration level (run). Run number order (with nominal concentration) is: 0 (0 nmol/mol), 1 (150 nmol/mol), 2 (120 nmol/mol), 3 (55 nmol/mol), 4 (90 nmol/mol), 5 (15 nmol/mol). The assessment criteria are presented as  $z' = \pm 2$  (blue line) and  $z' = \pm 3$  (red line). They represent the limits for the questionable and unsatisfactory results.



**Figure 4: z' scores evaluations of NO measurements**

Scores are given for each participant and each tested concentration level (run). Run number order (with nominal concentration) is: 0 (0 nmol/mol), 1 (600 nmol/mol), 2 (450 nmol/mol), 3 (210 nmol/mol), 4 (100 nmol/mol), 5 (80 nmol/mol), 6 (20 nmol/mol), 7 (390 nmol/mol), 8 (300 nmol/mol), 9 (50 nmol/mol), 10 (30 nmol/mol). The assessment criteria are presented as  $z' = \pm 2$  (blue line) and  $z' = \pm 3$  (red line). They represent the limits for the questionable and unsatisfactory results.



**Figure 5: z' scores evaluations of NO<sub>2</sub> measurements**

Scores are given for each participant and each concentration level (run). Run number order (with nominal concentration) is: 0 (0 nmol/mol), 1 (150 nmol/mol), 2 (110 nmol/mol), 3 (60 nmol/mol), 4 (90 nmol/mol), 5 (20 nmol/mol). The assessment criteria are presented as  $z' = \pm 2$  (blue line) and  $z' = \pm 3$  (red line). They represent the limits for the questionable and unsatisfactory results.

### 3.2. $E_n$ scores

The normalised deviations [13] ( $E_n$ ) were calculated according to:

$$E_n = \frac{x_i - X}{\sqrt{U_{x_i}^2 + U_X^2}} \quad \text{Equation 2}$$


where 'X' is the assigned/reference value with an expanded uncertainty ' $U_X$ ' and ' $x_i$ ' is the participant's average value with an expanded uncertainty ' $U_{x_i}$ '. The values with  $|E_n| \leq 1$  are satisfactory.

From Figure 6 to Figure 10 the bias of each participant ( $x_i - X$ ) is plotted and error bars are used to show the value of the denominator of Equation 2. These plots represent also the  $E_n$  evaluations where, considering the  $E_n$  criteria ( $|E_n| \leq 1$ ), all results with error bars touching or crossing the x-axis are satisfactory. Reported standard uncertainties (Annex B) that are bigger than "standard deviation for proficiency assessments" ( $\sigma_p$ , Table 4) are considered not fit-for-purpose and are denoted with '\*' in the x-axis of each figure. The  $E_n$  evaluation showed some unsatisfactory results as reported in Table 6.

This high number of results with an  $E_n$  value above 1 (unsatisfactory) could be explained observing the uncertainties submitted by the participants that are rather small. It could be interesting to check the calculation form used by the participants to calculate the uncertainty.

Parameter	Lab. code	Lab. value	Lab. U	Unit	E <sub>n</sub> evaluation
NO2 _4	A	108.9	3.9	nmol/mol	unsatisfactory
NO2 _6	A	58.52	2.1	nmol/mol	unsatisfactory
NO2 _10	A	18	0.64	nmol/mol	unsatisfactory
SO2 _4	A	29.43	1.06	nmol/mol	unsatisfactory
CO _1	B	9.64	1	µmol/mol	unsatisfactory
NO2 _2	B	138.2	16.1	nmol/mol	unsatisfactory
NO2 _8	B	82.67	9.72	nmol/mol	unsatisfactory
NO _2	C	455.8	4.94	nmol/mol	unsatisfactory
NO _7	C	393.5	4.86	nmol/mol	unsatisfactory
NO2 _4	E	108.2	5.02	nmol/mol	unsatisfactory
NO2 _6	E	58.08	2.76	nmol/mol	unsatisfactory
NO2 _8	E	88.46	4.1	nmol/mol	unsatisfactory
CO _3	F	1.088	0.03	µmol/mol	unsatisfactory
CO _4	F	5.213	0.12	µmol/mol	unsatisfactory
NO _1	F	637.4	12.8	nmol/mol	unsatisfactory
NO _2	F	489.3	9.82	nmol/mol	unsatisfactory
NO _4	F	109.1	2.2	nmol/mol	unsatisfactory
NO _6	F	23.08	0.64	nmol/mol	unsatisfactory
NO _7	F	413.8	8.3	nmol/mol	unsatisfactory
NO _8	F	324.4	6.5	nmol/mol	unsatisfactory
NO _10	F	31.87	0.68	nmol/mol	unsatisfactory
NO2 _6	F	59.1	1.92	nmol/mol	unsatisfactory
NO2 _10	F	18.71	0.66	nmol/mol	unsatisfactory
SO2 _1	F	137.4	2.3	nmol/mol	unsatisfactory
SO2 _2	F	74.45	1.2	nmol/mol	unsatisfactory
SO2 _1	H	126.3	5.69	nmol/mol	unsatisfactory
CO _1	I	4.696	0.3	µmol/mol	unsatisfactory
CO _2	I	2.014	0.33	µmol/mol	unsatisfactory
CO _3	I	0.757	0.22	µmol/mol	unsatisfactory
CO _4	I	2.882	0.25	µmol/mol	unsatisfactory
CO _5	I	1.253	0.18	µmol/mol	unsatisfactory
NO _3	I	206.1	6.6	nmol/mol	unsatisfactory
NO _4	I	93.38	3.2	nmol/mol	unsatisfactory
NO _5	I	71.99	2.7	nmol/mol	unsatisfactory
NO _6	I	16.45	1.4	nmol/mol	unsatisfactory
NO _8	I	296	9.4	nmol/mol	unsatisfactory
NO _9	I	43.33	2.1	nmol/mol	unsatisfactory
NO _10	I	24.91	1.5	nmol/mol	unsatisfactory
NO2 _6	I	55.53	2.2	nmol/mol	unsatisfactory
CO _5	L	2.171	0.1	µmol/mol	unsatisfactory


**Table 6: Unsatisfactory results according to the E<sub>n</sub>**



**Figure 6: Bias of participant's SO<sub>2</sub> measurement results**


Expanded uncertainty of bias for each run is presented as an error bar. The results with error bars touching or crossing the x-axis are satisfactory. For each evaluation the run number (numbers 0 to 5) together with the participants rounded run average (nmol/mol) is given. The '\*' mark indicates reported standard uncertainties bigger than  $\sigma_p$ .





**Figure 7: Bias of participant's CO measurement results**


Expanded uncertainty of bias for each run is presented as an error bar. Results with error bars touching or crossing the x-axis are satisfactory. For each evaluation the run number (numbers 0 to 5) together with the participants rounded run average ( $\mu\text{mol/mol}$ ) is given. The '\*' mark indicates reported standard uncertainties bigger than  $\sigma_p$ .



**Figure 8: Bias of participant's O<sub>3</sub> measurement results**


Expanded uncertainty of bias for each run is presented as an error bar. Results with error bars touching or crossing the x-axis are satisfactory. For each evaluation the run number (numbers 0 to 5) together with the participants rounded run average (nmol/mol) is given. The '\*' mark indicates reported standard uncertainties bigger than  $\sigma_p$ .





**Figure 9: Bias of participant's NO measurement results**

Expanded uncertainty of bias for each run is presented as an error bar. Results with error bars touching or crossing the x-axis are satisfactory. For each evaluation the run number (numbers 0 to 10) together with the participants rounded run average (nmol/mol) is given. The '\*' mark indicates reported standard uncertainties bigger than  $\sigma_p$ .




**Figure 10: Bias of participant's NO<sub>2</sub> measurement results**

Expanded uncertainty of bias is presented as an error bar for NO<sub>2</sub> run numbers 0, 2, 4, 6, 8 and 10 (see Table 3). Results with error bars touching or crossing the x-axis are satisfactory. For each evaluation the run number together with the participants rounded run average (nmol/mol) is given. The '\*' mark indicates reported standard uncertainties bigger than  $\sigma_p$ .



## 4. Performance characteristics of individual laboratories

Individual participant's bias was evaluated and is presented in Chapter 3.2 (Figure 6 to Figure 10). Since the results of NO<sub>2</sub> runs 1, 3, 5, 7 and 9 were not treated in proficiency evaluation, the bias of these runs are presented in Figure 11.



**Figure 11: Bias of participant's NO<sub>2</sub> measurements for run numbers 1, 3, 5, 7 and 9**

Within these test gas mixtures there is no GPT to produce NO<sub>2</sub> (see Table 3). For each evaluation the run number together with the participants rounded run average (nmol/mol) is given.

### 4.1. Converter efficiencies of NO<sub>2</sub>-to-NO for NO<sub>x</sub> analysers

Since NO and NO<sub>2</sub> test gases were produced by GPT it is possible to evaluate the efficiency of the NO<sub>2</sub>-to-NO converter of each participant's NO<sub>x</sub> analyser. The evaluation takes each participant's NO and NO<sub>2</sub> measurements before and after oxidation by O<sub>3</sub>. However, possible minor instabilities in the preparation of the test gas mixtures were not taken into account. The converter efficiency ( $\alpha$ ) is calculated using Equation 3 [4]:

$$\alpha = \frac{[NO_2]_i - [NO_2]_{i-1}}{[NO]_{i-1} - [NO]_i} \cdot 100\% \quad \text{Equation 3}$$

Ideal value for  $\alpha$  is 100%.

Lab code	NO <sub>2</sub> nmol/mol	$\alpha$ %
A	150	98.3
A	110	93.9
A	60	95.2
A	90	98.0
A	20	83.8
B	150	100.2
B	110	99.0
B	60	100.2
B	90	99.9
B	20	100.0
C	150	99.6
C	110	103.2
C	60	105.1
C	90	100.3
C	20	103.8
D	150	100.7
D	110	100.0
D	60	100.4
D	90	100.1
D	20	100.2
E	150	99.4
E	110	99.2
E	60	99.0
E	90	98.3
E	20	99.7
F	150	103.2
F	110	102.0
F	60	102.1
F	90	101.4
F	20	104.0

IE code	NO <sub>2</sub> nmol/mol	$\alpha$ %
G	150	99.5
G	110	100.6
G	60	100.0
G	90	100.5
G	20	100.2
H	150	100.9
H	110	101.4
H	60	101.2
H	90	99.3
H	20	100.6
I	150	95.6
I	110	98.6
I	60	100.1
I	90	99.6
I	20	95.5
L	150	101.0
L	110	99.0
L	60	100.2
L	90	99.1
L	20	100.3


**Table 7: efficiency of NO<sub>2</sub>-to-NO converters**

The evaluation of Equation 3 for each participant at different concentration levels is given in Table 7.

## 5. Discussion

A decision diagram was developed for a general assessment of the quality of each result (Figure 12). It results in seven categories (1 to 7). The general comments for each category are the following:

- **1:** measurement result is completely satisfactory.
- **2:** measurement result is satisfactory ( $z'$  satisfactory and  $E_n$  ok) but the reported uncertainty is too high.
- **3:** measured value is satisfactory ( $z'$  satisfactory) but the reported uncertainty is underestimated ( $E_n$  not ok).
- **4:** measurement result is questionable ( $z'$  questionable) but due to a high uncertainty being reported it can be considered valid ( $E_n$  ok).
- **5:** measurement result is questionable ( $z'$  questionable and  $E_n$  not ok).
- **6:** measurement result is unsatisfactory ( $z'$  unsatisfactory) but due to a high uncertainty being reported it can be considered valid ( $E_n$  ok).
- **7:** measurement result is unsatisfactory ( $z'$  unsatisfactory and  $E_n$  not ok).



**Figure 12: The decision diagram for general assessment of proficiency results.**

The results of the IE were assigned to categories according to the diagram given in Figure 12 and are presented in the following Table 8.

	run num	Ref. conc. level	IE code								
			A	B	C	D	E	F	H	I	L
CO (µmol/mol)	0	0.003	1	2	2	1	1	3	1	1	1
	1	8.544	1	7	1	1	1	1	1	7	1
	2	3.556	1	4	1	1	1	1	1	7	1
	3	1.038	1	2	2	1	1	3	1	5	1
	4	5.062	1	4	1	1	1	3	1	7	1
	5	2.039	1	2	1	1	1	1	1	7	3
NO (nmol/mol)	0	0.09	1	1	1	1	1	1	1	1	1
	1	618.80	1	2	1	1	1	3	1	1	1
	2	464.08	1	2	3	1	1	5	1	1	1
	3	215.56	1	2	1	1	1	1	1	3	1
	4	100.72	1	2	2	1	1	5	1	5	1
	5	80.53	1	2	2	1	1	1	1	5	1
	6	18.78	1	1	2	1	1	5	1	3	1
	7	402.30	1	2	3	1	1	3	1	1	1
	8	309.51	1	2	1	1	1	3	1	3	1
	9	50.24	1	2	2	1	1	1	1	5	1
	10	29.67	1	2	2	1	1	3	1	5	1
NO <sub>2</sub> (nmol/mol)	0	0.00	1	1	1	1	1	1	1	1	1
	2	157.66	1	7	1	1	1	1	1	2	1
	4	116.11	5	2	2	1	5	1	1	1	1
	6	62.23	3	2	2	1	3	3	1	5	1
	8	94.89	1	7	2	1	3	1	1	2	1
	10	20.83	3	2	2	1	1	3	1	1	1
O <sub>3</sub> (nmol/mol)	0	0.02	1	1	1	1	1	1	1	1	1
	1	153.17	1	2	1	1	1	1	1	1	1
	2	123.80	1	2	1	1	1	1	1	1	1
	3	55.46	1	2	1	1	1	1	1	1	1
	4	91.53	1	2	1	1	1	1	1	1	1
	5	16.91	1	1	1	1	1	1	1	1	1
SO <sub>2</sub> (nmol/mol)	0	0.10	1	1	1	1	1	1	1	1	1
	1	133.94	1	2	1	1	1	3	3	1	1
	2	72.36	1	2	1	1	1	3	1	1	1
	3	5.31	1	1	1	1	1	1	1	1	1
	4	31.18	3	1	1	1	1	1	1	1	1
	5	10.43	1	1	1	1	1	1	1	1	1

**Table 8: general assessment of proficiency results**

## 6. Conclusions

The proficiency evaluation scheme has provided an assessment of the participants measured values and their evaluated uncertainties.

In terms of the criteria imposed by the European Directive ( $\sigma_p$ ) 75.6% of the results reported (see Table 9) by AQUILA laboratories fall into category '1' and are satisfactory both in terms of measured values and of evaluated uncertainties. Among the remaining results the majority presented satisfactory measured values, but the evaluated uncertainties were either too high, category '2' (10.8%), or too small, category '3' (7.3%). Two results were found to be questionable for the  $z'$  and valid for the  $E_n$  (0.6% in category '4'). Eleven results were found questionable for the  $z'$  and not valid for the  $E_n$  (3.5% in category '5'). Seven results were found to be unsatisfactory for both indicators (2.2% in category '7').

IE	Site	Categories %						
		1	2	3	4	5	6	7
Apr-08	Ispra (IT)	68.4	18.1	7.3	1.0	1.0	2.6	1.6
Oct-08 (I)	Ispra (IT)	37.9	40.8	14.2	0.6	3.6	1.0	1.9
Oct-08 (II)	Ispra (IT)	34.3	38.9	23.7	1.0	2.0	0.0	0.0
Sep-09	Langen (DE)	60.8	29.9	3.1	4.1	1.0	1.0	0.0
Oct-09	Ispra (IT)	85.0	5.7	7.5	0.4	1.4	0.0	0.0
Jun-10	Ispra (IT)	84.6	8.1	4.4	0.7	2.3	0.0	0.0
Sep-11	Ispra (IT)	86.1	7.9	5.4	0.0	0.3	0.0	0.3
Oct-11 (I)	Ispra (IT)	78.6	12.5	7.6	0.0	1.3	0.0	0.0
Oct-11 (II)	Langen (DE)	59.4	39.9	0.0	0.7	0.0	0.0	0.0
Jun-12	Ispra (IT)	92.2	0.5	7.3	0.0	0.0	0.0	0.0
Sep-13	Langen (DE)	75.7	20.9	2.0	0.0	1.4	0.0	0.0
Sep-13	Ispra (IT)	89.4	7.3	3.3	0.0	0.0	0.0	0.0
Oct-13	Ispra (IT)	86.8	8.9	3.6	0.4	0.4	0.0	0.0
May-14	Ispra (IT)	81.8	15.2	1.1	0.0	0.7	0.0	1.1
Oct-15	Langen (DE)	73.2	23.9	0.7	1.4	0.0	0.7	0.0
Oct-15-(I)	Ispra (IT)	89.2	7.6	1.9	0.3	0.6	0.0	0.3
Oct-15-(II)	Ispra (IT)	75.6	10.8	7.3	0.6	3.5	0.0	2.2

**Table 9: Flags summary**

As in previous IE, the adopted criteria for high concentrations were the standard deviations for proficiency assessment, deriving from the European Standards' uncertainty requirements.

The reproducibility standard deviation ( $s_R$ ) obtained at this (0) and previous IE, [20], [21], [22], [23], [24], [25], [33], [34], [35], [36], [37], [38], [39], [40], [41] and [43] is comparable to the mentioned criteria. On the other hand, the uncertainty criteria for zero levels were those set in AQUILA's position paper [12].

In the present IE the results in category '1' are lower than the last few years of exercises in Ispra. The increased number of values in category '5' and the highest number ever obtained of poor quality results in category '7' has to be underlined.

In this IE, 93.7% of the results in the z' evaluations (Table 10) were satisfactory, 13 results were found to be questionable (4.1%) and 7 unsatisfactory (2.2%).

<i>IE</i>	<i>Site</i>	<i>Satisfactory (%)</i>	<i>Questionable (%)</i>	<i>Unsatisfactory (%)</i>
June/05	Ispra (IT)	94.7	2.3	3.0
June/07	Ispra (IT)	97.8	1.9	0.3
October/07	Essen (DE)	93.2	4.6	2.2
April/08	Ispra (IT)	93.8	2.1	4.1
October 2008_1	Ispra (IT)	92.9	4.2	2.9
October 2008_2	Ispra (IT)	97.0	3.0	0.0
September/09	Langen (DE)	94.3	4.7	0.9
October/09	Ispra (IT)	98.2	1.8	0.0
June/10	Ispra (IT)	97.0	3.0	0.0
September/11	Ispra (IT)	99.4	0.3	0.3
October/11	Ispra (IT)	98.7	1.3	0.0
October/11	Langen (DE)	99.3	0.7	0.0
June/12	Ispra (IT)	100.0	0.0	0.0
September/13	Langen (DE)	98.6	1.4	0.0
September/13	Ispra (IT)	100.0	0.0	0.0
October/13	Ispra (IT)	99.3	0.7	0.0
May/14	Ispra (IT)	98.1	0.7	1.1
October/15	Langen (DE)	97.9	1.4	0.7
October/15_I	Ispra (IT)	98.8	0.9	0.3
October/15_II	Ispra (IT)	93.7	4.1	2.2

**Table 10: z' scores summary**

Comparability of results among AQUILA participants at the highest concentration level, excluding outliers, is acceptable for almost all pollutants measurements.

The relative reproducibility (R) limits, at the highest studied concentration levels, are 8.8% for SO<sub>2</sub>, 14.8% for CO, 3.9% for O<sub>3</sub>, for NO 4.9% and for NO<sub>2</sub> 14.2% almost all within the objective derived from criteria imposed by the EC ( $\sigma_p$  see Table 4). NO<sub>2</sub> shows a deviation from the objectives already at the level of 50 ppb while CO deviates at 5.5ppm.

During this IE the performance of all NRL was generally satisfactory. Only laboratory I had an unsatisfactory performance for CO due to instrument problem occurred during the calibration.

## 7. References

- [1] Directive 2008/50/EC of the European Parliament and of the Council of 21 May 2008 on ambient air quality and cleaner air for Europe, L 152, 11.06.2008
- [2] EN 14626:2012, Ambient air quality - Standard method for the measurement of the concentration of carbon monoxide by non-dispersive infrared spectroscopy
- [3] EN 14212:2012, Ambient air quality - Standard method for the measurement of the concentration of sulphur dioxide by ultraviolet fluorescence
- [4] EN 14211:2012, Ambient air quality - Standard method for the measurement of the concentration of nitrogen dioxide and nitrogen monoxide by chemiluminescence
- [5] EN 14625:2012, Ambient air quality - Standard method for the measurement of the concentration of ozone by ultraviolet photometry
- [6] ISO 6143:2001, Gas analysis - Comparison methods for determining and checking the composition of calibration gas mixtures
- [7] ISO 6144:2003, Gas analysis - Preparation of calibration gas mixtures - Static volumetric method
- [8] ISO 6145-7:2001, Gas analysis - Preparation of calibration gas mixtures using dynamic volumetric methods - Part 7: Thermal mass-flow controllers
- [9] Mücke H.-G., (2008), Air quality management in the WHO European Region – Results of a quality assurance and control programme on air quality monitoring (1994-2004), Environment International, EI-01718
- [10] Mücke H.-G., et al. (2000), European Intercomparison workshop on air quality monitoring vol.4 – Measuring NO, NO<sub>2</sub>, O<sub>3</sub> and SO<sub>2</sub> – Air Hygiene Report 13, WHO Collaboration Centre for Air Quality Management and Air Pollution Control, ISSN 0938 - 9822
- [11] <http://ies.jrc.ec.europa.eu/aquila-project/aquila-homepage.html>
- [12] AQUILA POSITION PAPER N. 37, (2008) Protocol for intercomparison exercise. Organisation of intercomparison exercises for gaseous air pollution for EU national air quality reference laboratories and laboratories of the WHO EURO region  
[https://ec.europa.eu/jrc/sites/jrcsh/files/aquila-n\\_37-intercomparison-exercise%20protocol%202008.pdf](https://ec.europa.eu/jrc/sites/jrcsh/files/aquila-n_37-intercomparison-exercise%20protocol%202008.pdf)
- [13] ISO 13528:2015, Statistical methods for use in proficiency testing by interlaboratory comparisons
- [14] ISO 5725-1:1994, Accuracy (trueness and precision) of measurement methods and results – Part 1: General principles and definitions
- [15] ISO 5725-2:1994, Accuracy (trueness and precision) of measurement methods and results – Part 2: Basic method for the determination of repeatability and reproducibility of a standard measurement method
- [16] ISO 5725-6:1994, Accuracy (trueness and precision) of measurement methods and results - Part 6: Use in practice of accuracy values

- [17] Harmonisation of Directive 92/72/EEC on air pollution by ozone, E. De Saeger et al., EUR 17662, 1997
- [18] De Saeger E. et al., European comparison of Nitrogen Dioxide calibration methods, EUR 17661, 1997
- [19] ISO 15337:2009, Ambient air - Gas phase titration - Calibration of analysers for ozone
- [20] Kapus M. et al. (2009). The evaluation of the Intercomparison Exercise for SO<sub>2</sub>, CO, O<sub>3</sub>, NO and NO<sub>2</sub> carried out in June 2007 in Ispra. JRC scientific and technical reports. EUR 23804.
- [21] Kapus M. et al. (2009). The evaluation of the Intercomparison Exercise for SO<sub>2</sub>, CO, O<sub>3</sub>, NO and NO<sub>2</sub> - April 2008. JRC scientific and technical reports. EUR 23805.
- [22] Kapus M. et al. (2009). The evaluation of the Intercomparison Exercise for SO<sub>2</sub>, CO, O<sub>3</sub>, NO and NO<sub>2</sub> 6-9 October 2008. JRC scientific and technical reports. EUR 23806.
- [23] Kapus M. et al. (2009). The evaluation of the Intercomparison Exercise for SO<sub>2</sub>, CO, O<sub>3</sub>, NO and NO<sub>2</sub> 13-16 October 2008. JRC scientific and technical reports. EUR 23807.
- [24] Belis C. A. et al. (2010). The evaluation of the Interlaboratory comparison Exercise for SO<sub>2</sub>, CO, O<sub>3</sub>, NO and NO<sub>2</sub> Langen 20-25 September 2009. EUR 24376.
- [25] Belis C. A. et al. (2010). The evaluation of the Interlaboratory comparison Exercise for SO<sub>2</sub>, CO, O<sub>3</sub>, NO and NO<sub>2</sub> 19-22 October 2009. EUR 24476.
- [26] Viallon J. et al 2009 Metrologia 46 08017. Final report, on-going key comparison BIPM.QM-K1: Ozone at ambient level, comparison with JRC, 2008. doi: 10.1088/0026-1394/46/1A/08017
- [27] Viallon, J., et al. (2006), International comparison CCQM-P28: Ozone at ambient level, Metrologia, 43, Tech. Suppl., 08010, doi:10.1088/0026-1394/43/1A/08010
- [28] Tanimoto, H., et al. (2006), Intercomparison of ultraviolet photometry and gas-phase titration techniques for ozone reference standards at ambient levels, Journal of Geophysical Research, vol. 111, D16313, doi:10.1029/2005JD006983
- [29] GUM Workbench, The Tool for Expression of Uncertainty of Measurements
- [30] VDI 2449 Part3: 2001, Measurement methods test criteria- General method for the determination of the uncertainty of calibratable measurement methods.
- [31] Mücke H-G, et al. (1996). European Intercomparison Workshops on Air Quality Monitoring. Vol. 2 – Measuring of CO, NO, NO<sub>2</sub> and O<sub>3</sub> – Air Hygiene Report 9. Berlin, Germany: WHO Collaborating Centre for Air Quality Management and Air Pollution Control; ISSN 0938-9822.
- [32] ISO 17043:2010, Conformity assessment - General requirements for proficiency testing
- [33] Barbieri M. et al. (2011). The evaluation of the Interlaboratory Comparison Exercise for SO<sub>2</sub>, CO, O<sub>3</sub>, NO and NO<sub>2</sub> Ispra 14-17 June 2010, EUR 24943



- [34] Barbieri M. et al. (2012) Evaluation of the Laboratory Comparison Exercise for SO<sub>2</sub>, CO, O<sub>3</sub>, NO and NO<sub>2</sub>, 11<sup>th</sup>-14<sup>th</sup> June 2012 Ispra, EUR 25536.
- [35] Barbieri M. et al. (2012) Evaluation of the Laboratory Comparison Exercise for SO<sub>2</sub>, CO, O<sub>3</sub>, NO and NO<sub>2</sub>, Langen 23<sup>rd</sup>-28<sup>th</sup> October 2011, EUR 25387.
- [36] Barbieri M. et al. (2012) Evaluation of the Laboratory Comparison Exercise for SO<sub>2</sub>, CO, O<sub>3</sub>, NO and NO<sub>2</sub>, 03<sup>rd</sup>-06<sup>th</sup> October 2011 Ispra, EUR 25386.
- [37] Barbieri M. et al. (2012) Evaluation of the Laboratory Comparison Exercise for SO<sub>2</sub>, CO, O<sub>3</sub>, NO and NO<sub>2</sub>, 26<sup>th</sup>-29<sup>th</sup> September 2011 Ispra, EUR 25385.
- [38] Barbieri M., Lagler F., Mücke H.G., Wirtz K. and Stummer V. (2014) Evaluation of the Laboratory Comparison Exercise for NO, NO<sub>2</sub>, SO<sub>2</sub>, CO, and O<sub>3</sub> Langen (D) 1st-6th September 2013, EUR 26578.
- [39] Barbieri M., Lagler F., (2014) Evaluation of the Laboratory Comparison Exercise for SO<sub>2</sub>, CO, O<sub>3</sub>, NO and NO<sub>2</sub> 30<sup>st</sup> September-3<sup>rd</sup> October 2013 Ispra, EUR 26604.
- [40] Barbieri M., Lagler F., (2014) Evaluation of the Laboratory Comparison Exercise for SO<sub>2</sub>, CO, O<sub>3</sub>, NO and NO<sub>2</sub> 7<sup>st</sup>-10<sup>th</sup> October 2013 Ispra, EUR 26639.
- [41] Barbieri M., Lagler F., (2014) Evaluation of the Laboratory Comparison Exercise for SO<sub>2</sub>, CO, O<sub>3</sub>, NO and NO<sub>2</sub> 19<sup>th</sup>-22<sup>nd</sup> May 2014 Ispra, EUR 27199.
- [42] COMMISSION DIRECTIVE (EU) 2015/1480 of 28 August 2015 (L226/4) amending several annexes to Directives 2004/107/EC and 2008/50/EC of the European Parliament and of the Council laying down the rules concerning reference methods, data validation and location of sampling points for the assessment of ambient air quality.
- [43] Barbieri M., Lagler F., Mücke H.G., Wirtz K. and Stummer V. (2014) Evaluation of the Laboratory Comparison Exercise for NO, NO<sub>2</sub>, SO<sub>2</sub>, CO, and O<sub>3</sub> Langen (D) 4th-9th October 2015, EUR 27918.
- [44] ISO/IEC 17025:2005, General requirements for the competence of testing and calibration laboratories

## **Annex A. Assigned values**

The assigned values of tested concentration levels (run) were derived from ERLAPs measurements which are calibrated against the certified reference values of CRMs and are traceable to international standards. In this perspective the assigned values are reference values as defined in the ISO 13528 [13].

To foster its reference function ERLAP is regularly participating in key comparisons of the Gas Analysis Working Group within the framework of the BIPM's CCQM.

During this IE ERLAP's SO<sub>2</sub>, CO and NO analysers were calibrated according to the methodology described in the ISO 6143 [6]. Reference gas mixtures were produced from the primary reference materials (produced and certified by NMi Van Swinden Laboratorium) by a dynamic dilution method using mass flow controllers [8]. All flows were measured with a certified molbloc/molbox1 system. For O<sub>3</sub> measurements, the analyzers were calibrated using the JRC SRP42 primary standard (constructed by NIST) which has been compared to BIPM primary standard [26]. The photometer absorption cross section uncertainty (1.06%) was included in the uncertainty budget [27], [28].

The reference gas mixture and the calibration experiment evaluation were carried out using two computer applications, the "GUM WORKBENCH" [29] and "B-least" [30] respectively. For extending calibration from the NO to NO<sub>2</sub> channel of NO<sub>x</sub> analyser the GPT test was performed to establish the efficiency of the NO<sub>2</sub>-converter.

ERLAP's measurement results were validated in comparison to the group statistics ( $\bar{x}^*$  and  $s^*$ ) for every parameter and concentration level of the IE. These statistics are calculated from participants, applying the robust method described in the Annex C of the ISO 13528 [13]. The validation takes into account ERLAP's measurement result ( $X$ ) and its standard uncertainty ( $u_X$ ) as given in Equation 4 [13].

$$\frac{|\bar{x}^* - X|}{\sqrt{\frac{(1,25 \cdot s^*)^2}{p} + u_X^2}} < 2 \quad \text{Equation 4}$$

Where ' $\bar{x}^*$ ' and ' $s^*$ ' represent the robust average and the robust standard deviation, respectively, and ' $p$ ' is the number of participants.

In Table 11 all inputs for Equation 4 are given and all of ERLAP's measurement results are confirmed to be valid.

As a group evaluation,  $\bar{x}^*$  and  $s^*$  were calculated (applying the procedure described in Annex C of ISO 13528) for each run, and are presented in the following tables.

run	unit	X	uX	x*	s*	p	val.
NO_0	nmol/mol	0.09	0.71	0.07	0.14	10	OK
NO_1	nmol/mol	618.80	3.54	620.33	7.69	10	OK
NO_2	nmol/mol	464.08	2.70	466.84	4.71	10	OK
NO_3	nmol/mol	215.56	1.41	214.96	4.13	10	OK
NO_4	nmol/mol	100.72	0.92	101.58	2.09	10	OK
NO_5	nmol/mol	80.53	0.85	79.58	1.84	10	OK
NO_6	nmol/mol	18.78	0.73	19.24	1.19	10	OK
NO_7	nmol/mol	402.30	2.37	400.38	7.92	10	OK
NO_8	nmol/mol	309.51	1.88	308.62	5.78	10	OK
NO_9	nmol/mol	50.24	0.77	49.27	1.41	10	OK
NO_10	nmol/mol	29.67	0.74	29.42	0.99	10	OK
NO2_0	nmol/mol	0.00	0.71	-0.03	0.16	10	OK
NO2_1	nmol/mol	3.10	2.30	2.70	0.65	10	OK
NO2_2	nmol/mol	157.66	2.50	154.92	4.65	10	OK
NO2_3	nmol/mol	0.85	1.05	0.65	0.50	10	OK
NO2_4	nmol/mol	116.11	1.29	112.90	3.72	10	OK
NO2_5	nmol/mol	0.25	0.78	0.12	0.54	10	OK
NO2_6	nmol/mol	62.23	0.87	59.97	2.32	10	OK
NO2_7	nmol/mol	1.58	1.60	1.04	0.69	10	OK
NO2_8	nmol/mol	94.89	1.70	92.29	3.36	10	OK
NO2_9	nmol/mol	0.23	0.74	0.09	0.25	10	OK
NO2_10	nmol/mol	20.83	0.75	19.49	1.28	10	OK
CO_0	μmol/mol	0.00	0.01	0.02	0.03	10	OK
CO_1	μmol/mol	8.54	0.04	8.57	0.12	10	OK
CO_2	μmol/mol	3.56	0.02	3.59	0.05	10	OK
CO_3	μmol/mol	1.04	0.01	1.05	0.03	10	OK
CO_4	μmol/mol	5.06	0.03	5.11	0.06	10	OK
CO_5	μmol/mol	2.04	0.01	2.07	0.04	10	OK
O3_0	nmol/mol	0.02	0.22	0.06	0.12	10	OK
O3_1	nmol/mol	153.17	1.10	151.82	1.49	10	OK
O3_2	nmol/mol	123.80	0.89	123.36	0.85	10	OK
O3_3	nmol/mol	55.46	0.40	55.55	0.32	10	OK
O3_4	nmol/mol	91.53	0.66	91.39	0.55	10	OK
O3_5	nmol/mol	16.91	0.23	17.07	0.46	10	OK
SO2_0	nmol/mol	0.10	0.50	0.07	0.19	10	OK
SO2_1	nmol/mol	133.94	0.96	132.48	3.50	10	OK
SO2_2	nmol/mol	72.36	0.67	71.63	2.06	10	OK
SO2_3	nmol/mol	5.31	0.51	5.12	0.29	10	OK
SO2_4	nmol/mol	31.18	0.54	30.86	1.00	10	OK
SO2_5	nmol/mol	10.43	0.51	10.16	0.40	10	OK

**Table 11: validation of assigned values (X)**

In comparison to the robust averages ( $x^*$ ) with taking into account the standard uncertainties of assigned values ( $uX$ ), and robust standard deviations ( $s^*$ ) as denoted by Equation 4.

The homogeneity of test gas was evaluated from measurements at the beginning and end of the distribution line. From the relative differences between these measurements, average and standard deviation were calculated. The uncertainty of test gas, due to lack of homogeneity, was also calculated as the sum of the squares of these average and standard deviations.

$$u_{X'}^2 = u_X^2 + (X \cdot u_{\text{homogeneity}})^2 \quad \text{Equation 5}$$

The upper and lower limits of bias due to homogeneity were evaluated as being smaller than 0.5% which constitutes the relative standard uncertainty of 0.3% of each concentration level. The standard uncertainties of assigned/reference values ( $u_{X'}$ ) were calculated with Equation 5 and used in the proficiency evaluations of Chapter 2.

## Annex B. The results of the IE

This annex reports the participant's results, presented both in tables and in graphs. For all mixture concentration that is generated (run), participants were asked to report three results each representing 30 minutes measurement ( $x_{ij}$ ).

This annex presents the reported data and their combined uncertainty  $u(x_i)$  and expanded uncertainty  $U(x_i)$  expressed in mol/mol units.

For all of the runs except concentration levels 0, the average ( $\bar{x}_i$ ) and the standard deviation ( $s_i$ ) of each participant are also presented.

The assigned value is indicated in the graphs with the red line and the individual laboratories expanded uncertainties ( $U_{x_i}$ ) are indicated with error bars.

### Reported values for SO<sub>2</sub>

values	laboratories									
	A	B	C	D	E	F	G	H	I	L
$x_{i, 1}$	0.20	0.00	-0.40	0.03	-0.15	0.26	0.10	0.16	1.01	0.00
$u(x_i)$	0.02	0.78	0.58	0.58	0.18	0.00	0.50	0.70	0.65	0.43
$U(x_i)$	0.04	1.56	1.15	1.20	0.36	0.01	1.01	1.40	1.30	0.86

Table 12: Reported values for SO<sub>2</sub> run 0





Figure 13: Reported values for SO<sub>2</sub> run 0

values	laboratories									
	A	B	C	D	E	F	G	H	I	L
xi, 1	129.50	136.30	131.60	131.20	128.85	136.87	133.77	125.77	132.60	133.14
xi, 2	129.60	136.60	132.30	132.60	129.21	137.44	134.00	126.22	133.10	133.95
xi, 3	129.80	136.70	132.40		129.36	138.02	134.05	127.04	132.90	134.05
xi	129.63	136.53	132.10	131.90	129.14	137.44	133.94	126.34	132.86	133.71
si	0.15	0.20	0.43	0.99	0.26	0.57	0.14	0.64	0.25	0.49
u(xi)	2.33	5.46	1.41	1.28	2.24	1.15	0.87	2.85	2.00	1.31
U(xi)	4.66	10.92	2.82	2.60	4.48	2.30	1.75	5.69	4.00	2.62


**Table 13: Reported values for SO<sub>2</sub> run 1**



**Figure 14: Reported values for SO<sub>2</sub> run 1**

values	laboratories									
	A	B	C	D	E	F	G	H	I	L
xi, 1	68.50	73.80	71.10	71.60	69.75	74.59	72.33	68.70	72.85	72.16
xi, 2	69.70	73.80	71.20	71.60	69.83	74.32	72.39	68.74	72.77	72.37
xi, 3	70.20	73.90	71.20		70.00	74.45	72.37	68.79	71.90	71.96
xi	69.46	73.83	71.16	71.60	69.86	74.45	72.36	68.74	72.50	72.16
si	0.87	0.05	0.05	0.00	0.12	0.13	0.03	0.04	0.52	0.20
u(xi)	1.25	3.07	0.81	0.85	1.21	0.60	0.64	1.71	1.15	0.71
U(xi)	2.50	6.14	1.62	1.70	2.43	1.20	1.27	3.43	2.30	1.42


**Table 14: Reported values for SO<sub>2</sub> run 2**



**Figure 15: Reported values for SO<sub>2</sub> run 2**

values	laboratories									
	A	B	C	D	E	F	G	H	I	L
xi, 1	4.85	5.30	4.70	5.32	5.02	5.12	5.31	5.16	6.31	4.94
xi, 2	4.90	5.40	4.60	5.23	5.00	5.21	5.31	5.15	6.03	4.74
xi, 3	4.90	5.30	4.60		5.00	5.19	5.30	5.09	5.99	4.64
xi	4.88	5.33	4.63	5.27	5.00	5.17	5.30	5.13	6.11	4.77
si	0.02	0.05	0.05	0.06	0.01	0.04	0.00	0.03	0.17	0.15
u(xi)	0.11	0.84	0.77	0.58	0.18	0.05	0.51	0.28	0.65	0.43
U(xi)	0.22	1.68	1.54	1.20	0.36	0.10	1.01	0.56	1.30	0.86


**Table 15: Reported values for SO<sub>2</sub> run 3**



**Figure 16: Reported values for SO<sub>2</sub> run 3**

values	laboratories									
	A	B	C	D	E	F	G	H	I	L
xi, 1	29.30	31.80	30.30	30.80	30.46	31.74	31.15	29.73	32.13	30.84
xi, 2	29.50	31.80	30.30	30.90	30.46	31.64	31.19	29.78	31.80	30.94
xi, 3	29.50	31.80	30.30		30.45	31.77	31.20	29.77	32.45	31.14
xi	29.43	31.80	30.30	30.85	30.45	31.71	31.18	29.76	32.12	30.97
si	0.11	0.00	0.00	0.07	0.00	0.06	0.02	0.02	0.32	0.15
u(xi)	0.53	1.55	0.67	0.64	0.53	0.26	0.53	1.13	0.75	0.43
U(xi)	1.06	3.10	1.34	1.30	1.06	0.52	1.07	2.26	1.50	0.86


**Table 16: Reported values for SO<sub>2</sub> run 4**



**Figure 17: Reported values for SO<sub>2</sub> run 4**

values	laboratories									
	A	B	C	D	E	F	G	H	I	L
xi, 1	9.70	10.50	9.70	10.40	9.99	10.31	10.46	9.97	11.70	9.98
xi, 2	9.75	10.50	9.70	10.34	10.02	10.26	10.43	9.93	11.49	9.98
xi, 3	9.85	10.50	9.70		9.96	10.25	10.39	9.94	11.30	9.68
xi	9.76	10.50	9.70	10.37	9.99	10.27	10.42	9.94	11.49	9.88
si	0.07	0.00	0.00	0.04	0.03	0.03	0.03	0.02	0.20	0.17
u(xi)	0.16	0.93	0.74	0.59	0.18	0.08	0.51	0.51	0.65	0.43
U(xi)	0.32	1.86	1.48	1.20	0.37	0.16	1.02	1.01	1.30	0.86

**Table 17: Reported values for SO<sub>2</sub> run 5**




**Figure 18: Reported values for SO<sub>2</sub> run 5**



## Reported values for CO

	laboratories									
values	A	B	C	D	E	F	G	H	I	L
$x_i, 1$	0.015	0.000	-0.050	0.005	0.041	0.057	0.003	0.035	0.052	-0.001
$u(x_i)$	0.001	0.234	0.120	0.017	0.030	0.001	0.009	0.061	0.060	0.049
$U(x_i)$	0.002	0.468	0.230	0.035	0.060	0.002	0.017	0.122	0.120	0.098


**Table 18: Reported values for CO run 0**



**Figure 19: Reported values for CO run 0**

	laboratories									
values	A	B	C	D	E	F	G	H	I	L
$x_i, 1$	8.680	9.640	8.500	8.515	8.476	8.700	8.533	8.605	4.695	8.502
$x_i, 2$	8.705	9.640	8.530	8.565	8.473	8.717	8.546	8.605	4.704	8.496
$x_i, 3$	8.755	9.640	8.550		8.469	8.717	8.554	8.607	4.689	8.503
$\bar{x}_i$	8.713	9.640	8.527	8.540	8.473	8.711	8.544	8.606	4.696	8.500
$s_i$	0.038	0.000	0.025	0.035	0.004	0.010	0.011	0.001	0.008	0.004
$u(x_i)$	0.157	0.500	0.040	0.041	0.166	0.100	0.034	0.197	0.152	0.173
$U(x_i)$	0.314	1.000	0.090	0.082	0.332	0.200	0.069	0.394	0.304	0.347


**Table 19: Reported values for CO run 1**



**Figure 20: Reported values for CO run 1**

values	laboratories									
	A	B	C	D	E	F	G	H	I	L
xi, 1	3.625	4.010	3.560	3.565	3.581	3.613	3.556	3.567	2.008	3.704
xi, 2	3.635	4.001	3.560	3.570	3.577	3.633	3.556	3.567	2.012	3.701
xi, 3	3.645	4.001	3.560		3.575	3.620	3.556	3.565	2.023	3.702
xi	3.635	4.004	3.560	3.568	3.578	3.622	3.556	3.566	2.014	3.702
si	0.010	0.005	0.000	0.004	0.003	0.010	0.000	0.001	0.008	0.002
u(xi)	0.066	0.294	0.090	0.023	0.070	0.042	0.016	0.097	0.164	0.075
U(xi)	0.132	0.588	0.190	0.046	0.141	0.084	0.032	0.194	0.328	0.151


**Table 20: Reported values for CO run 2**



**Figure 21: Reported values for CO run 2**

values	laboratories									
	A	B	C	D	E	F	G	H	I	L
xi, 1	1.055	1.180	1.030	1.043	1.039	1.083	1.039	1.043	0.761	1.125
xi, 2	1.058	1.180	1.030	1.044	1.042	1.087	1.038	1.042	0.757	1.125
xi, 3	1.075	1.180	1.020		1.039	1.093	1.037	1.045	0.752	1.126
xi	1.063	1.180	1.027	1.043	1.040	1.088	1.038	1.043	0.757	1.125
si	0.011	0.000	0.006	0.001	0.002	0.005	0.001	0.002	0.005	0.001
u(xi)	0.019	0.238	0.130	0.018	0.035	0.013	0.009	0.058	0.110	0.049
U(xi)	0.038	0.476	0.260	0.036	0.070	0.026	0.018	0.116	0.220	0.098


**Table 21: Reported values for CO run 3**



**Figure 22: Reported values for CO run 3**

values	laboratories									
	A	B	C	D	E	F	G	H	I	L
xi, 1	5.155	5.700	5.090	5.078	5.060	5.200	5.061	5.129	2.884	5.119
xi, 2	5.165	5.700	5.090	5.088	5.064	5.213	5.063	5.128	2.882	5.113
xi, 3	5.175	5.700	5.090		5.069	5.227	5.062	5.128	2.879	5.097
xi	5.165	5.700	5.090	5.083	5.064	5.213	5.062	5.128	2.882	5.110
si	0.010	0.000	0.000	0.007	0.005	0.014	0.001	0.001	0.003	0.011
u(xi)	0.093	0.348	0.070	0.028	0.099	0.060	0.021	0.124	0.124	0.104
U(xi)	0.186	0.696	0.140	0.056	0.199	0.120	0.043	0.248	0.248	0.208


**Table 22: Reported values for CO run 4**



**Figure 23: Reported values for CO run 4**

values	laboratories									
	A	B	C	D	E	F	G	H	I	L
xi, 1	2.055	2.310	2.060	2.070	2.074	2.097	2.039	2.042	1.253	2.168
xi, 2	2.065	2.310	2.060	2.074	2.071	2.100	2.039	2.038	1.251	2.166
xi, 3	2.070	2.310	2.050		2.073	2.073	2.038	2.040	1.254	2.180
xi	2.063	2.310	2.057	2.072	2.073	2.090	2.039	2.040	1.253	2.171
si	0.008	0.000	0.006	0.003	0.002	0.015	0.001	0.002	0.002	0.008
u(xi)	0.037	0.254	0.120	0.020	0.041	0.025	0.011	0.076	0.092	0.049
U(xi)	0.074	0.508	0.230	0.040	0.083	0.050	0.023	0.152	0.184	0.098

**Table 23: Reported values for CO run 5**




**Figure 24: Reported values for CO run 5**

## Reported values for O<sub>3</sub>

	laboratories									
values	A	B	C	D	E	F	G	H	I	L
$x_i, 1$	0.20	-0.10	0.10	0.03	0.17	0.01	0.02	0.16	-0.60	0.10
$u(x_i)$	0.02	0.94	0.58	0.85	0.51	0.75	0.22	0.60	1.00	0.60
$U(x_i)$	0.04	1.88	1.15	1.70	1.02	1.50	0.44	1.21	2.00	1.20


**Table 24: Reported values for O<sub>3</sub> run 0**



**Figure 25: Reported values for O<sub>3</sub> run 0**

	laboratories									
values	A	B	C	D	E	F	G	H	I	L
$x_i, 1$	149.40	150.12	152.30	149.10	150.96	151.78	152.80	151.26	155.63	150.00
$x_i, 2$	150.25	151.08	153.40	151.60	151.48	153.04	153.27	151.98	155.93	151.20
$x_i, 3$	150.75	151.39	153.60		151.66	152.77	153.43	151.95	156.27	151.60
$\bar{x}_i$	150.13	150.86	153.10	150.35	151.36	152.53	153.16	151.73	155.94	150.93
$s_i$	0.68	0.66	0.70	1.76	0.36	0.66	0.32	0.40	0.32	0.83
$u(x_i)$	2.71	6.88	1.91	1.92	2.26	2.38	1.00	3.42	2.60	3.00
$U(x_i)$	5.42	13.76	3.83	3.80	4.51	4.76	1.99	6.83	5.20	6.00


**Table 25: Reported values for O<sub>3</sub> run 1**



**Figure 26: Reported values for O<sub>3</sub> run 1**

values	laboratories									
	A	B	C	D	E	F	G	H	I	L
xi, 1	121.80	122.62	123.30	122.60	122.56	124.53	123.66	123.20	125.63	122.40
xi, 2	122.20	122.94	124.10	123.50	122.83	124.85	123.83	123.56	126.26	122.90
xi, 3	122.55	123.04	124.20		122.88	124.58	123.90	123.60	126.26	123.10
xi	122.18	122.86	123.86	123.05	122.75	124.65	123.79	123.45	126.05	122.80
si	0.37	0.21	0.49	0.63	0.17	0.17	0.12	0.22	0.36	0.36
u(xi)	2.20	5.63	1.62	1.64	1.85	2.06	0.80	2.78	2.15	2.44
U(xi)	4.40	11.26	3.24	3.30	3.70	4.12	1.61	5.57	4.30	4.88


**Table 26: Reported values for O<sub>3</sub> run 2**



**Figure 27: Reported values for O<sub>3</sub> run 2**

values	laboratories									
	A	B	C	D	E	F	G	H	I	L
xi, 1	54.50	55.33	55.00	55.60	55.28	56.17	55.37	55.55	55.93	55.40
xi, 2	54.75	55.54	55.30	55.80	55.39	56.45	55.46	55.79	56.19	55.60
xi, 3	55.00	55.65	55.50		55.50	56.56	55.54	55.78	56.08	55.50
xi	54.75	55.50	55.26	55.70	55.39	56.39	55.45	55.70	56.06	55.50
si	0.25	0.16	0.25	0.14	0.11	0.20	0.08	0.13	0.13	0.10
u(xi)	0.98	2.68	1.23	1.06	0.94	1.35	0.37	1.48	1.20	1.10
U(xi)	1.96	5.36	2.45	2.10	1.88	2.70	0.73	2.96	2.40	2.20


**Table 27: Reported values for O<sub>3</sub> run 3**



**Figure 28: Reported values for O<sub>3</sub> run 3**

values	laboratories									
	A	B	C	D	E	F	G	H	I	L
xi, 1	90.25	91.09	90.80	91.00	90.69	92.54	91.42	91.15	93.23	90.50
xi, 2	90.50	91.30	91.70	91.90	91.01	92.73	91.57	91.55	93.37	91.00
xi, 3	90.55	91.51	91.80		91.10	92.91	91.61	91.57	93.53	90.90
xi	90.43	91.30	91.43	91.45	90.93	92.72	91.53	91.42	93.37	90.80
si	0.16	0.21	0.55	0.63	0.21	0.18	0.10	0.23	0.15	0.26
u(xi)	1.63	4.23	1.32	1.34	1.41	1.73	0.60	2.41	1.65	1.80
U(xi)	3.26	8.46	2.64	2.70	2.83	3.46	1.19	4.82	3.30	3.60


**Table 28: Reported values for O<sub>3</sub> run 4**



**Figure 29: Reported values for O<sub>3</sub> run 4**

values	laboratories									
	A	B	C	D	E	F	G	H	I	L
xi, 1	16.50	17.13	16.30	17.10	17.28	17.59	16.93	17.33	16.43	17.50
xi, 2	16.65	17.24	16.50	17.20	17.25	17.65	16.92	17.37	16.66	17.60
xi, 3	16.80	17.24	16.50		17.26	17.57	16.89	17.36	16.46	17.50
xi	16.65	17.20	16.43	17.15	17.26	17.60	16.91	17.35	16.51	17.53
si	0.15	0.06	0.11	0.07	0.01	0.04	0.02	0.02	0.12	0.05
u(xi)	0.30	1.22	1.20	0.87	0.51	0.94	0.22	0.90	0.90	0.60
U(xi)	0.60	2.44	2.40	1.80	1.02	1.87	0.44	1.81	1.80	1.20

**Table 29: Reported values for O<sub>3</sub> run 5.**




**Figure 30: Reported values for O<sub>3</sub> run 5**

## Reported values for NO

values	laboratories									
	A	B	C	D	E	F	G	H	I	L
$x_i, 1$	0.20	0.30	0.00	0.01	-0.03	0.00	0.09	0.15	-0.20	0.20
$u(x_i)$	0.02	0.90	0.58	0.58	0.45	0.01	0.71	0.60	0.61	0.40
$U(x_i)$	0.04	1.80	1.15	1.20	0.91	0.02	1.42	1.20	1.30	0.80


**Table 30: Reported values for NO run 0**



**Figure 31: Reported values for NO run 0**

values	laboratories									
	A	B	C	D	E	F	G	H	I	L
$x_i, 1$	611.50	626.80	609.20	616.90	615.86	636.35	618.65	619.03	633.82	621.29
$x_i, 2$	612.60	627.40	610.50	618.90	616.55	637.23	618.95	618.53	634.21	621.99
$x_i, 3$	613.20	627.90	610.80		616.88	638.71	618.81	619.45	634.11	622.78
$\bar{x}_i$	612.43	627.36	610.16	617.90	616.43	637.43	618.80	619.00	634.04	622.02
$s_i$	0.86	0.55	0.85	1.41	0.52	1.19	0.15	0.46	0.20	0.74
$u(x_i)$	11.03	36.32	2.71	4.40	10.96	6.41	3.02	13.84	9.95	13.05
$U(x_i)$	22.06	72.64	5.42	8.80	21.92	12.82	6.03	27.67	19.90	26.10

**Table 31: Reported values for NO run 1**




**Figure 32: Reported values for NO run 1**



values	laboratories									
	A	B	C	D	E	F	G	H	I	L
xi, 1	462.20	472.70	457.00	465.40	467.42	489.93	464.73	470.08	468.60	469.95
xi, 2	461.90	472.20	455.30	464.00	466.20	489.23	464.00	469.43	466.39	468.86
xi, 3	460.80	471.80	455.00		466.96	488.65	463.52	469.16	465.01	468.56
xi	461.63	472.23	455.76	464.70	466.86	489.27	464.08	469.55	466.66	469.12
si	0.73	0.45	1.07	0.99	0.61	0.64	0.60	0.47	1.81	0.73
u(xi)	8.30	27.34	2.47	3.33	8.31	4.91	2.32	10.59	7.40	9.84
U(xi)	16.60	54.68	4.94	6.70	16.61	9.82	4.64	21.18	14.80	19.69


**Table 32: Reported values for NO run 2**



**Figure 33: Reported values for NO run 2**

values	laboratories									
	A	B	C	D	E	F	G	H	I	L
xi, 1	214.60	219.30	209.80	216.10	211.29	219.80	215.24	214.05	206.05	216.85
xi, 2	214.90	219.50	210.50	216.60	212.11	221.10	215.52	214.05	206.15	216.75
xi, 3	215.00	219.50	210.00		212.26	221.48	215.92	214.68	206.06	216.85
xi	214.83	219.43	210.10	216.35	211.88	220.79	215.56	214.26	206.08	216.81
si	0.20	0.11	0.36	0.35	0.52	0.88	0.34	0.36	0.05	0.05
u(xi)	3.86	12.73	4.45	1.64	3.78	2.27	1.25	4.84	3.30	4.55
U(xi)	7.72	25.46	8.90	3.30	7.56	4.54	2.51	9.68	6.60	9.10


**Table 33: Reported values for NO run 3**



**Figure 34: Reported values for NO run 3**

values	laboratories									
	A	B	C	D	E	F	G	H	I	L
x <sub>i,1</sub>	100.80	102.90	97.30	100.90	103.02	108.96	100.89	103.68	94.15	102.95
x <sub>i,2</sub>	100.70	102.70	97.00	100.30	102.84	109.45	100.71	103.60	93.11	102.45
x <sub>i,3</sub>	100.35	102.60	96.70		102.66	108.89	100.55	103.37	92.89	102.15
$\bar{x}_i$	100.61	102.73	97.00	100.60	102.84	109.10	100.71	103.55	93.38	102.51
s <sub>i</sub>	0.23	0.15	0.30	0.42	0.18	0.30	0.17	0.16	0.67	0.40
u(x <sub>i</sub> )	1.80	6.00	5.22	0.92	1.85	1.10	0.88	2.47	1.60	2.15
U(x <sub>i</sub> )	3.60	12.00	10.43	1.90	3.70	2.20	1.75	4.93	3.20	4.30


**Table 34: Reported values for NO run 4**



**Figure 35: Reported values for NO run 4**

values	laboratories									
	A	B	C	D	E	F	G	H	I	L
x <sub>i,1</sub>	78.70	81.50	76.50	80.15	78.21	80.81	80.22	79.18	71.82	80.35
x <sub>i,2</sub>	78.95	81.90	77.00	80.50	78.47	81.26	80.63	79.51	72.01	80.75
x <sub>i,3</sub>	79.00	82.00	77.00		78.54	81.76	80.73	79.59	72.13	81.15
$\bar{x}_i$	78.88	81.80	76.83	80.32	78.40	81.27	80.52	79.42	71.98	80.75
s <sub>i</sub>	0.16	0.26	0.28	0.24	0.17	0.47	0.27	0.21	0.15	0.40
u(x <sub>i</sub> )	1.41	4.81	5.36	0.81	1.42	0.86	0.81	1.89	1.35	1.69
U(x <sub>i</sub> )	2.82	9.62	10.72	1.60	2.85	1.72	1.63	3.78	2.70	3.39


**Table 35: Reported values for NO run 5**



**Figure 36: Reported values for NO run 5**

values	laboratories									
	A	B	C	D	E	F	G	H	I	L
xi, 1	18.90	19.60	18.00	18.86	19.77	22.72	18.88	20.77	16.54	19.91
xi, 2	18.75	19.40	18.00	18.65	19.97	23.01	18.73	20.58	16.46	19.71
xi, 3	18.65	19.40	18.00		19.68	23.50	18.72	20.71	16.34	19.91
xi	18.76	19.46	18.00	18.75	19.80	23.07	18.77	20.68	16.44	19.84
si	0.12	0.11	0.00	0.14	0.14	0.39	0.09	0.09	0.10	0.11
u(xi)	0.34	1.43	5.79	0.59	0.58	0.32	0.72	1.02	0.70	0.42
U(xi)	0.68	2.86	11.57	1.20	1.17	0.64	1.45	2.03	1.40	0.84


**Table 36: Reported values for NO run 6**



**Figure 37: Reported values for NO run 6**

values	laboratories									
	A	B	C	D	E	F	G	H	I	L
xi, 1	396.50	408.40	392.70	403.10	393.84	413.24	402.01	399.60	389.00	402.84
xi, 2	397.00	408.70	393.70	403.20	394.25	413.87	402.44	400.54	390.13	402.74
xi, 3	397.50	408.90	394.00		394.08	414.31	402.45	400.09	390.29	403.34
xi	397.00	408.66	393.46	403.15	394.05	413.80	402.30	400.07	389.80	402.97
si	0.50	0.25	0.68	0.07	0.20	0.53	0.25	0.47	0.70	0.32
u(xi)	7.15	23.66	2.43	2.90	7.01	4.15	2.04	9.02	6.20	8.45
U(xi)	14.30	47.32	4.86	5.80	14.02	8.30	4.09	18.04	12.40	16.91


**Table 37: Reported values for NO run 7**



**Figure 38: Reported values for NO run 7**

values	laboratories									
	A	B	C	D	E	F	G	H	I	L
xi, 1	305.50	314.70	301.00	309.70	305.17	324.12	309.39	309.56	296.49	311.74
xi, 2	306.55	314.70	300.50	309.80	305.18	324.12	309.57	308.83	295.68	311.14
xi, 3	306.90	314.60	300.70		305.58	324.84	309.57	308.88	295.82	311.24
xi	306.31	314.66	300.73	309.75	305.31	324.36	309.51	309.09	295.99	311.37
si	0.72	0.05	0.25	0.07	0.23	0.41	0.10	0.40	0.43	0.32
u(xi)	5.50	18.23	3.97	2.26	5.44	3.25	1.64	6.97	4.70	6.53
U(xi)	11.00	36.46	7.94	4.60	10.87	6.50	3.27	13.94	9.40	13.07


**Table 38: Reported values for NO run 8**



**Figure 39: Reported values for NO run 8**

values	laboratories									
	A	B	C	D	E	F	G	H	I	L
xi, 1	48.85	51.10	47.30	50.10	48.32	50.33	50.22	48.50	43.16	49.68
xi, 2	48.90	51.00	48.00	50.10	48.22	50.50	50.26	48.44	43.61	49.98
xi, 3	49.00	51.10	48.70		48.34	50.38	50.25	48.50	43.23	50.18
xi	48.91	51.06	48.00	50.10	48.29	50.40	50.24	48.48	43.33	49.94
si	0.07	0.05	0.70	0.00	0.06	0.08	0.02	0.03	0.24	0.25
u(xi)	0.87	3.08	5.57	0.68	0.91	0.51	0.75	1.42	1.05	1.05
U(xi)	1.74	6.16	11.13	1.40	1.81	1.02	1.51	2.85	2.10	2.10


**Table 39: Reported values for NO run 9**



**Figure 40: Reported values for NO run 9**

values	laboratories									
	A	B	C	D	E	F	G	H	I	L
xi, 1	28.50	30.40	28.20	29.56	29.08	31.66	29.67	29.34	24.91	29.87
xi, 2	28.75	30.40	28.30	29.59	29.01	31.87	29.67	29.43	24.91	29.87
xi, 3	29.00	30.30	28.70		29.10	32.07	29.68	29.41	24.91	30.27
xi	28.75	30.36	28.40	29.57	29.06	31.86	29.67	29.39	24.91	30.00
si	0.25	0.05	0.26	0.02	0.04	0.20	0.00	0.04	0.00	0.23
u(xi)	0.51	1.96	5.71	0.62	0.70	0.34	0.73	1.44	0.75	0.63
U(xi)	1.02	3.92	11.42	1.30	1.40	0.68	1.46	2.88	1.50	1.26

**Table 40: Reported values for NO run 10**




**Figure 41: Reported values for NO run 10**

## Reported values for NO<sub>2</sub>

	laboratories									
values	A	B	C	D	E	F	G	H	I	L
$x_i, 1$	0.20	-0.30	0.00	0.04	-0.07	-0.38	0.00	0.08	-0.40	0.10
$u(x_i)$	0.02	0.90	0.58	0.58	0.40	0.01	0.71	0.60	0.90	0.49
$U(x_i)$	0.04	1.80	1.15	1.20	0.80	0.02	1.42	1.20	1.80	0.98


**Table 41: Reported values for NO<sub>2</sub> run 0**



**Figure 42: Reported values for NO<sub>2</sub> run 0**

	laboratories									
values	A	B	C	D	E	F	G	H	I	L
$x_i, 1$	151.50	137.50	157.20	157.48	150.79	153.47	157.20	153.31	160.27	154.74
$x_i, 2$	151.80	138.20	158.70	158.82	150.65	154.53	157.66	153.99	161.85	155.24
$x_i, 3$	152.00	138.80	159.00		150.59	155.92	158.13	153.54	162.40	154.74
$\bar{x}_i$	151.76	138.16	158.30	158.15	150.67	154.64	157.66	153.61	161.50	154.90
$s_i$	0.25	0.65	0.96	0.94	0.10	1.22	0.46	0.34	1.10	0.28
$u(x_i)$	2.73	8.04	3.66	2.46	3.46	2.57	2.45	3.48	10.50	3.26
$U(x_i)$	5.46	16.08	7.32	4.90	6.93	5.14	4.90	6.96	21.00	6.51


**Table 42: Reported values for NO<sub>2</sub> run 2**



**Figure 43: Reported values for NO<sub>2</sub> run 2**

values	laboratories									
	A	B	C	D	E	F	G	H	I	L
xi, 1	108.50	109.50	117.30	116.36	108.35	113.49	115.92	112.97	111.30	113.10
xi, 2	109.00	109.60	118.00	116.36	108.33	113.09	116.07	113.10	111.95	113.90
xi, 3	109.25	109.70	118.00		108.05	113.98	116.34	113.40	111.98	113.70
xi	108.91	109.60	117.76	116.36	108.24	113.52	116.11	113.15	111.74	113.56
si	0.38	0.10	0.40	0.00	0.16	0.44	0.21	0.22	0.38	0.41
u(xi)	1.95	6.39	3.87	1.85	2.51	1.83	1.24	2.70	2.30	2.39
U(xi)	3.90	12.78	7.74	3.70	5.02	3.66	2.48	5.39	4.60	4.77


**Table 43: Reported values for NO<sub>2</sub> run 4**



**Figure 44: Reported values for NO<sub>2</sub> run 4**

values	laboratories									
	A	B	C	D	E	F	G	H	I	L
xi, 1	58.25	59.60	63.00	62.01	58.20	58.95	62.16	59.55	55.47	60.53
xi, 2	58.55	59.90	63.00	62.29	57.98	59.40	62.27	59.73	55.55	60.72
xi, 3	58.75	59.90	63.00		58.06	58.94	62.27	59.80	55.58	60.53
xi	58.51	59.80	63.00	62.15	58.08	59.09	62.23	59.69	55.53	60.59
si	0.25	0.17	0.00	0.19	0.11	0.26	0.06	0.12	0.05	0.11
u(xi)	1.05	3.56	4.17	1.11	1.38	0.96	0.85	1.75	1.10	1.27
U(xi)	2.10	7.12	8.33	2.20	2.76	1.92	1.70	3.51	2.20	2.55

**Table 44: Reported values for NO<sub>2</sub> run 6**




**Figure 45: Reported values for NO<sub>2</sub> run 6**



values	laboratories									
	A	B	C	D	E	F	G	H	I	L
xi, 1	90.00	82.50	96.00	95.18	88.49	92.09	94.67	91.93	93.63	91.93
xi, 2	90.15	82.70	96.00	95.87	88.52	92.84	95.01	91.82	94.15	92.53
xi, 3	90.50	82.80	96.00		88.38	92.41	94.98	91.59	94.67	92.13
xi	90.21	82.66	96.00	95.52	88.46	92.44	94.88	91.78	94.15	92.19
si	0.25	0.15	0.00	0.48	0.07	0.37	0.18	0.17	0.52	0.30
u(xi)	1.63	4.86	3.98	1.57	2.05	1.49	1.68	2.19	6.70	1.94
U(xi)	3.26	9.72	7.97	3.20	4.10	2.98	3.35	4.37	13.40	3.88


**Table 45: Reported values for NO<sub>2</sub> run 8**



**Figure 46: Reported values for NO<sub>2</sub> run 8**

values	laboratories									
	A	B	C	D	E	F	G	H	I	L
xi, 1	18.00	19.20	21.00	20.81	19.07	18.96	20.81	19.36	18.11	20.17
xi, 2	18.00	19.20	21.00	20.87	19.26	18.57	20.85	19.12	17.99	19.78
xi, 3	18.00	19.10	21.00		19.14	18.60	20.83	19.27	17.91	19.78
xi	18.00	19.16	21.00	20.84	19.15	18.71	20.83	19.25	18.00	19.91
si	0.00	0.05	0.00	0.04	0.09	0.21	0.02	0.12	0.10	0.22
u(xi)	0.32	1.42	4.40	0.66	0.57	0.33	0.74	0.94	1.25	0.49
U(xi)	0.64	2.84	8.80	1.40	1.15	0.66	1.49	1.89	2.50	0.98

**Table 46: Reported values for NO<sub>2</sub> run 10**



**Figure 47: Reported values for NO<sub>2</sub> run 10**

## ***Annex C: The precision of standardised measurement methods***

For the main purpose of monitoring trends between different IE undertaken by ERLAP the precision of standardised SO<sub>2</sub>, CO, O<sub>3</sub> and NO<sub>x</sub> measurement methods [2], [3], [4] and [5] as implemented by NRLs was evaluated.

The applied methodology is described in ISO 5725-Part 1 [14], Part 2 [15] and Part 6 [16]. The precision experiment has involved a total of seven laboratories, the actual number of labs ( $p_j$ ) varying from run to run (Table 47). Six concentration levels (for run 0 only one value is requested so repeatability cannot be evaluated) were tested for O<sub>3</sub>, CO, SO<sub>2</sub> and NO<sub>2</sub>, and eleven for NO. Outlier tests were performed and the results are reported in Annex D.

The repeatability standard deviation ( $s_r$ ) was calculated in accordance with ISO 5725-6 as the square root of average within-laboratory variance. The repeatability ( $r$ ) limit is calculated using Equation 6 [16]. It represents the biggest difference between two test results found on an identical test gas by one laboratory using the same apparatus within the shortest feasible time interval, which should not have been exceeded on average more than once in 20 cases in the normal and correct operation of method.

$$r = t_{95\%,\nu} \cdot \sqrt{2} \cdot s_r \quad \text{Equation 6}$$

The reproducibility standard deviation ( $s_R$ ) was calculated in accordance with ISO 5725-6 as the square root of the sum of repeatability and between-laboratory variance. The reproducibility limit ( $R$ ) is calculated using Equation 7 [16]. It represents the biggest difference between two measurements on an identical test gas reported by two laboratories, which should not occur on average more than once in 20 cases in the normal and correct operation of method.

$$R = t_{95\%,\nu} \cdot \sqrt{2} \cdot s_R \quad \text{Equation 7}$$

The ' $s_r$ ' was evaluated with ( $p_j \cdot (3-1)$ ) degrees of freedom ( $\nu$ ) and ' $s_R$ ' with ( $p_j-1$ ) degrees of freedom. The critical range student factors ( $t_{\alpha,\nu}$ ) are reported in Table 47.


parameter	run	$p_j$	t critical value 95% for r	t critical value 95% for R
CO	1,2,3,4,5	10	2.086	2.262
NO	1,2,3,4,5,6,7,8,9,10	10	2.086	2.262
NO <sub>2</sub>	2,4,6,8,10	10	2.086	2.262
O <sub>3</sub>	1,2,3,4,5	10	2.086	2.262
SO <sub>2</sub>	1,2,3,4,5	10	2.086	2.262

**Table 47: Critical values of t used in the repeatability r and R evaluation**

The  $r$  and  $R$  limits of measurement methods are presented from Table 48 to Table 52 and from Figure 48 to Figure 52. The  $R$  from common criteria ( $R$  (from  $\sigma_p$ ) calculated by substituting  $S_R$  in Equation 7 with a 'standard deviation for proficiency assessment' (Table 4) is also reported. Comparison between  $R$  and  $R$  (from  $\sigma_p$ ) serves to indicate that  $\sigma_p$  is realistic ([13] 6.3.1) or from another point of view, that the general methodology implemented by NRLs is appropriate for  $\sigma_p$ . The green ( $R$ ) and blue ( $r$ ) lines are representing a good performance if they run below the red line which represents the data quality objective of the IE.

SO <sub>2</sub> data (nmol/mol) without outliers			
group average	repeatability limit : $r$	reproducibility limit : $R$	reproducibility limit (relative)
0.1		1.2	
5.2	0.3	1.4	
10.2	0.3	1.8	
30.9	0.4	3.1	
71.6	1.1	6.5	
132.4	1.2	11.7	8.8%


**Table 48: The  $R$  and  $r$  of SO<sub>2</sub> standard measurement method**



**Figure 48: The  $R$  and  $r$  of SO<sub>2</sub> standard measurement method as a function of concentration**

CO data (μmol/mol) without outliers			
group average	repeatability limit: r	reproducibility limit: R	reproducibility limit (relative)
0.016		0.102	
1.076	0.015	0.179	
2.105	0.021	0.311	
3.653	0.017	0.503	
5.192	0.023	0.709	
8.714	0.052	1.288	14.8%


**Table 49: The R and r of CO standard measurement method**



**Figure 49: The R and r of CO standard measurement method as a function of concentration**

O <sub>3</sub> data (nmol/mol) without outliers			
group average	repeatability limit: r	reproducibility limit: R	reproducibility limit (relative)
0.0		0.7	
17.1	0.2	1.4	
55.6	0.5	1.6	
91.6	0.8	3.1	
123.6	0.9	3.9	
152.2	1.7	5.9	3.9%


**Table 50: The R and r of O<sub>3</sub> standard measurement method**



**Figure 50: The R and r of O<sub>3</sub> standard measurement method as a function of concentration**

NO data (nmol/mol) without outliers			
group average	repeatability limit: r	reproducibility limit: R	reproducibility limit (relative)
0.1		0.5	
19.4	0.5	6.0	
29.2	0.5	6.2	
48.7	0.8	7.5	
78.9	0.9	9.9	
101.4	1.0	14.3	
214.4	1.2	15.0	
308.6	1.1	26.6	
400.2	1.4	25.0	
468.4	2.6	30.2	
622.0	2.1	30.4	4.9%


**Table 51: The R and r of NO standard measurement method**



**Figure 51: The R and r of NO standard measurement method as a function of concentration**

NO <sub>2</sub> data (nmol/mol) without outliers			
group average	repeatability limit: r	reproducibility limit: R	reproducibility limit (relative)
-0.1		0.7	
19.3	0.4	3.5	
59.6	0.5	7.3	
91.4	0.8	13.1	
112.5	1.0	10.6	
153.5	2.1	21.8	14.2%

**Table 52: The R and r of NO<sub>2</sub> standard measurement method**



**Figure 52: The R and r of NO<sub>2</sub> standard measurement method as a function of concentration**



## Annex C. Results evaluation for consistency and outlier test

The precision evaluation (Annex C) focuses on data that are as much as possible the reflection of the everyday work of NRLs and thus represents the comparability of participant's standard operating procedures.

For that reason a procedure for the detection of exceptional errors (error during typing, slip in performing the measurement or the calculation, wrong averaging interval, malfunction of instrumentation, etc.) was applied. This procedure carried out tests for data consistency and statistical outliers as described in ISO 5725-2.

Laboratories showing some form of statistical inconsistency were requested to investigate the cause of discrepancies.

Laboratories were allowed to correct their results if they identified of exceptional errors. Subsequently, data were considered definitive and "Grubb's one outlying observation test" was performed.

For runs where outliers were detected, these were removed and "Grubb's one outlying observation test" was repeated until no more outliers were observed. Statistical outliers obtained at this stage are not considered to be due to extraordinary errors but to a significant difference in each participant's standard operating procedure.

During this IE the statistical process analysis has identified some outliers that are presented in the table below (Table 53). Laboratory "I" had problem with the calibration process of CO with the consequence of underestimation during the IE.

Laboratory	Parameter	Run	Value	Failing test	Confidence level
I	CO	1	4.696	G1 min	1%, 5%
I	CO	2	2.014	G1 min	1%, 5%
I	CO	3	0.756	G1 min	1%, 5%
I	CO	4	2.881	G1 min	1%, 5%
I	CO	5	1.252	G1 min	1%, 5%
I	O3	0	-0.6	G1 min	1%, 5%

**Table 53: "Genuine" statistical outliers according to Grubb's one outlying observation test**

The precision of standardised measurement methods reported in 0 are calculated using the database without outliers.

According to Grubb's test, results that have a confidence level between 1 and 5% are considered straggler and should be specifically checked.

In order to provide the participants with useful information for judging their performance, the stragglers are reported in the following table.

Laboratory	parameter	run	value	G1min_5%	G1max_5%
B	NO2	2	138.16	Straggler	OK
F	NO	2	489.27	OK	straggler
I	NO	5	71.98	Straggler	OK
I	NO	9	43.33	Straggler	OK
I	NO	10	24.91	Straggler	OK
I	SO2	0	1.01	OK	straggler
I	SO2	3	6.11	OK	straggler
I	SO2	5	11.49	OK	straggler

**Table 54: Stragglers according to Grubb's one observation test**

## Annex D. Accreditation certificate

This annex provides the accreditation certificate for ISO/CEI 17025 [44] of ERLAP, who organised this inter-laboratory comparison and delivered the assigned value.



### CERTIFICATO DI ACCREDITAMENTO Accreditation Certificate

Accreditamento n°  
Accreditation n°

1362

Rev. 0

Si dichiara che  
We declare that

**European Reference Laboratory for Air Pollution  
(ERLAP) Air and Climate Unit - Institute for  
Environment and Sustainability - Joint Research  
Centre - European Commission**

Sede:  
Via E. Fermi 2749 - 21027 Ispra VA

è conforme ai requisiti  
della norma  
  
meets the requirements  
of the standard

UNI CEI EN ISO/IEC 17025:2005 "Requisiti generali per la competenza dei  
Laboratori di prova e taratura"  
  
EN ISO/IEC 17025:2005 "General Requirements for the Competence of Testing  
and Calibration Laboratories" standard

quale  
as

**Laboratorio di Prova  
Testing Laboratory**

L'accreditamento attesta la competenza tecnica del Laboratorio relativamente allo scopo riportato nelle schede allegate al presente certificato. Le schede possono variare nel tempo. I requisiti gestionali della ISO/IEC 17025:2005 (sezione 4) sono scritti in un linguaggio idoneo all'attività dei Laboratori di Prova, sono conformi ai principi della ISO 9001:2008 ed allineati con i suoi requisiti applicabili.  
Il presente certificato non è da ritenersi valido se non accompagnato dalle schede allegate e può essere sospeso o revocato in qualsiasi momento nel caso di inadempienza accertata da parte di ACCREDIA.  
La validità dell'accreditamento può essere verificata sul sito WEB ([www.accredia.it](http://www.accredia.it)) o richiesta direttamente ai singoli Dipartimenti.

The accreditation certifies the technical competence of the laboratory limited to the scope detailed in the attached Enclosure. The scope may vary in the time. The management system requirements in ISO/IEC 17025:2005 (Section 4) are written in a language relevant to Testing Laboratories operations and meet the principles of ISO 9001:2008 and are aligned with its pertinent requirements.  
The present certificate is valid only if associated to the annexed schedule, and can be suspended or withdrawn at any time in the event of non fulfilment as ascertained by ACCREDIA.  
The in force status of the accreditation may be checked in the WEB site ([www.accredia.it](http://www.accredia.it)) or on direct request to appointed Department.

Data di 1<sup>a</sup> emissione  
1st issue date  
2013-06-19

Data di modifica  
Modification date  
2013-06-19

Data di scadenza  
Expiring date  
2017-06-18

  
Il Direttore Generale  
The General Director  
(Dr. Filippo Trifiletti)

  
Il Direttore di Dipartimento  
Department Director  
(Dr. Paolo Bianco)

  
Il Presidente  
The President  
(Cav. del Lav. Federico Grazioli)



<b>European Reference Laboratory for Air Pollution (ERLAP)</b> <b>Air and Climate Unit - Institute for Environment and Sustainability - Joint Research Centre - European Commission</b>  Via E. Fermi 2749 21027 Ispra VA	Numero di accreditamento: <b>1362 Sede A</b>
	Revisione: <b>0</b> Data: <b>22/07/2013</b>
	Scheda <b>1</b> di <b>1</b> PA1779AR0.pdf

**ELENCO PROVE ACCREDITATE - CATEGORIA: 0**

**Synthetic mixture gas**

<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
carbon monoxide (0-86 mmol/mol)	EN 14626:2012
nitrogen oxides (NO: 0-962 nmol/mol; NO <sub>2</sub> : 0-261 nmol/mol)	EN 14211:2012
ozone (0-250 nmol/mol)	EN 14625:2012
sulphur dioxide (0-376 nmol/mol)	EN 14212:2012

*Legenda*

En= norma europea

ACCREDIA  
Il Direttore del Dipartimento  
(Dr. Paolo Bianco)

**Bianco  
Paolo**

Firmato digitalmente da Bianco Paolo  
 DN: c=IT, o=ACCREDIA/10566361001,  
 cn=Bianco Paolo,  
 serialNumber=IT8NCPLAS2M23L219N,  
 givenName=Paolo, sn=Bianco,  
 dnQualifier=11004771, title=Direttore  
 Dipartimento Laboratori di prova  
 Data: 2013.07.22 13:37:04 +02'00'

Europe Direct is a service to help you find answers to your questions about the European Union  
Free phone number (\*): 00 800 6 7 8 9 10 11  
(\*) Certain mobile telephone operators do not allow access to 00 800 numbers or these calls may be billed.

A great deal of additional information on the European Union is available on the Internet.  
It can be accessed through the Europa server <http://europa.eu>

#### **How to obtain EU publications**

Our publications are available from EU Bookshop (<http://bookshop.europa.eu>),  
where you can place an order with the sales agent of your choice.

The Publications Office has a worldwide network of sales agents.  
You can obtain their contact details by sending a fax to (352) 29 29-42758.

## JRC Mission

As the science and knowledge service of the European Commission, the Joint Research Centre's mission is to support EU policies with independent evidence throughout the whole policy cycle.



**EU Science Hub**

[ec.europa.eu/jrc](https://ec.europa.eu/jrc)



@EU\_ScienceHub



EU Science Hub - Joint Research Centre



Joint Research Centre



EU Science Hub

