

JRC TECHNICAL REPORTS

Evaluation of the 2016 ENV57/MetroERM measurement comparison on simulated airborne particulates: ^{137}Cs , ^{134}Cs and ^{131}I in air filters

Timotheos Altitzoglou
Petya Malo

2017



This publication is a Technical report by the Joint Research Centre (JRC), the European Commission's science and knowledge service. It aims to provide evidence-based scientific support to the European policymaking process. The scientific output expressed does not imply a policy position of the European Commission. Neither the European Commission nor any person acting on behalf of the Commission is responsible for the use that might be made of this publication.

Contact information

Name: Timotheos Altitzoglou

Address: European Commission, Joint Research Centre, Directorate for Nuclear Safety and Security

Email: timotheos.altitzoglou@ec.europa.eu

Tel.: +32 14 571266

JRC Science Hub

<https://ec.europa.eu/jrc>

JRC105477

EUR 28431 EN

PDF ISBN 978-92-79-65276-9 ISSN 1831-9424 doi:10.2760/030337

Luxembourg: Publications Office of the European Union, 2017

© European Atomic Energy Community, 2017

The reuse of the document is authorised, provided the source is acknowledged and the original meaning or message of the texts are not distorted. The European Commission shall not be held liable for any consequences stemming from the reuse.

How to cite this report: Timotheos Altitzoglou, Petya Malo; Evaluation of the 2016 ENV57/MetroERM measurement comparison on simulated airborne particulates: ^{137}Cs , ^{134}Cs and ^{131}I in air filters; EUR 28431 EN; doi:10.2760/030337

All images © European Atomic Energy Community, 2017

Contents

Acknowledgements	1
Abstract	3
1. Introduction	5
2. The 2016 ENV57 MetroERM measurement comparison	7
2.1 Description of the sample	7
2.2 Participating laboratories	7
2.3 Reporting of results	8
2.4 Timetable of the ILC	8
3. The Reference value	9
3.1 Standardisation of the spike solution	9
3.1.1 ^{137}Cs	10
3.1.2 ^{134}Cs	11
3.1.3 ^{131}I	11
3.2 Dilutions	12
3.3 Spiking of the air filters	14
3.4 Reference values	14
3.5 Quality control measurements	15
3.6 Influence of discrete spiked activity distribution on the counting efficiency... ..	15
4. Questionnaire on sampling method, air filter used and measurement conditions	20
5. Questionnaire on analytical and measurement procedures	21
6. Evaluation and comparison of the results	22
6.1 Identification of outliers and normal distribution test	22
6.2 Scores and evaluation criteria	23
6.2.1 Percentage difference	23
6.2.2 E_n number	24
6.2.3 Compatibility test	24
6.2.4 PomPlot	24
7. Results	26
7.1 ^{137}Cs	26
7.1.1 Reported results	26
7.1.2 Evaluation	31
7.1.3 Parameters possibly influencing the results	35
7.1.4 Comparison to the 2003 and the 2014 EC ILC	36
7.2 ^{134}Cs	38
7.2.1 Reported results	38
7.2.2 Evaluation	43

7.2.3 Parameters possibly influencing the results: Coincidence summing correction.....	46
7.3 ¹³¹ I	48
7.3.1 Reported results	48
7.3.2 Evaluation.....	54
7.3.3 Parameters possibly influencing the results	58
7.3.3.1 Decay during measurement.....	58
7.3.3.2 Coincidence summing correction	59
7.3.3.3 Radioactivity transferred to the protective plastic bag	59
7.3.4 Conclusions.....	61
8. Conclusions	63
References.....	69
List of abbreviations and definitions	71
List of figures	73
List of tables	75
ANNEXES.....	78
Annex 1: <i>List of participating laboratories per country (in alphabetic order)</i>	78
Annex 2: <i>Communication to the participants</i>	87
2A. <i>Expression of interest</i>	87
2B. <i>Invitation e-mail and accompanying letter</i>	89
2C. <i>Registration information (e-mail)</i>	91
2D. <i>Spiked air filter dispatch (e-mail) and accompanying letter</i>	92
2E. <i>Information to participants during the measurement exercise</i>	95
2F. <i>Reporting results</i>	96
2G. <i>Communication of preliminary results to the participants</i>	97
2H. <i>Invitation to the Workshop</i>	101
Annex 3: <i>First questionnaire sent to the participants to collect information on air filters and sampling</i>	106
Annex 4: <i>Responses of the participants to the questionnaire for collecting information on air filters and sampling</i>	109
Annex 5: <i>Type, material, size and form of the air filters used by the participants</i>	112
Annex 6: <i>Sequence of spiked filters and quality control sources prepared for the 1026 ENV57/MetroERM ILC exercise</i>	115
Annex 7: <i>Examples of spiked filters</i>	117
Annex 8: <i>Quality control filters and sources; measurement results and their deviation from the reference activity values</i>	118
Annex 9: <i>Reporting form and second questionnaire for submitting the measurement results</i>	124
Annex 10: <i>Responses to the second questionnaire accompanying the reporting form</i>	129

Annex 11: <i>Coincidence summing correction factors as applied by the participants</i>	131
Annex 12: <i>Uncertainty budget as reported by the participating laboratories</i>	133

Acknowledgements

The authors would like to express their thanks to all participants from the 67 laboratories in 29 countries for their contribution to the successful completion of this work. This interlaboratory comparison has been conducted in the frame of the ENV57 MetroERM EMRP project. The authors would like to thank Marc De Cort, JRC leader of this project and Stefan Neumeier of PTB, coordinator of the entire MetroERM project for their continuous interest and support to this work. Finally, they appreciate the comments and suggestions of Mikael Hult, Head of the Radionuclide Metrology Sector of the JRC Unit G.2-Standards for Nuclear safety, Security and Safeguards, reviewing the manuscript.

This work was supported by the ENV57 MetroERM project under the European Metrology Research programme (EMRP). The EMRP is jointly funded by the EMRP participating countries within EURAMET and the European Union.

Abstract

In 2016, the European Commission (EC) Directorate-General Joint Research Centre (JRC) Geel site (former Institute for Reference Materials and Measurements (IRMM)) organized an interlaboratory comparison (ILC) exercise on the measurement of ^{137}Cs , ^{134}Cs and ^{131}I in air filters. Similar exercises, although EC ILC, were organized in 2014 (Altitzoglou and Máté, 2016) and in 2003 (Wätjen et al., 2007); in both exercises the measurand was ^{137}Cs only.

The 2016 ILC, which is the subject of this report, was conducted in the frame of the EMRP project with code ENV57 and entitled MetroERM "Metrology for radiological early warning networks in Europe" project with the aim to optimise the metrological foundation of measurements for monitoring airborne radioactivity and promote pan-European harmonisation in data reliability.

The JRC participates in the ENV57 MetroERM project having the responsibility to carry out a number of tasks. Work Package 3 (WP3) addresses the traceability of dose rate and airborne radioactivity measurements and explicitly supports the process of harmonisation of radiological data from early warning networks in Europe by systematic investigations, comparison exercises and the publication of recommendations. Task 3.2 of WP3 entitled "Traceability management for airborne radioactivity" aims in developing traceable reference materials and standard sources in the form of large-area spiked aerosol filters and conducting a laboratory comparison exercise to quantify the performance of the airborne radioactivity measuring field stations.

This report describes the full life cycle of the above mentioned comparison among 67 European laboratories monitoring radioactivity in the environment. JRC provided the comparison samples, which were prepared individually for each laboratory using a gravimetrically diluted solution of ^{137}Cs , ^{134}Cs and ^{131}I . Reference values traceable to the International System of Units (SI) and the International Reference System (SIR) for gamma-ray emitting radionuclides were determined at the JRC. The samples were made by gravimetrically dispensing the appropriate activity amounts, close to those the laboratories routinely measure, on blank air filters provided by the participants. A robust evaluation of the individual performance using three different approaches, percentage difference ($D_{\%}$), E_n numbers and PomPlots, is presented. Finally, for the laboratories which have participated in the 2003 and 201 exercises, their performance evolution in measuring ^{137}Cs in air filters over the years, is examined.

All 67 participating laboratories reported valid results. The majority of the laboratories reported reliable measurement results for ^{137}Cs and ^{134}Cs ; 56 (84%) out of the 67 participants reported values with a percentage difference from the reference value within the $\pm 20\%$ range. Furthermore, 42 (63%) of the laboratories fulfilled the criterion of the compatibility test based on E_n numbers for ^{137}Cs and 36 (54%) for ^{134}Cs . As the calculation of the E_n numbers takes into account the uncertainties on both the measured activity and the reference value, the lower scores for the E_n numbers reveals that the uncertainty estimation is not adequate in many laboratories and there is a need to improve their application of uncertainty propagation. An observation for the ^{134}Cs results is a negative bias on the reported results compared to the reference values, which might be partly attributed to a non-adequate summing correction applied by some laboratories.

The evaluation of the performance of the laboratories on ^{131}I was complicated by two effects. First, in a number of spiked air filters a fraction of the ^{131}I activity was transferred to the protective plastic bag. Second, the integrity of the ^{131}I activity content of many spiked air filters was compromised to a variable degree. Nevertheless, 20 (30%) laboratories reported results for ^{131}I with a percentage difference from the reference value within the $\pm 20\%$ range and 9 (13%) with compatible E_n numbers. This performance is not satisfactory, but the difficulties encountered have to be taken into account. A major output of this project is the

identification of the need of additional studies on how to perform reliable measurements of radioiodine on air filters.

Overall, the majority of laboratories master their gamma-ray spectrometry analytical procedures, including the counting efficiency calibration of the detection systems and the corrections for coincidence summing. More attention has to be paid to the elaboration of the uncertainty to obtain correct and realistic uncertainties on the reported results.

1. Introduction

According to the Articles 35/36 of the Euratom Treaty (Euratom, 2012) and the Commission Recommendation 473/2000 (2000/473/Euratom, 2000) derived from the Euratom Treaty, the Member States (MS) of the European Union (EU) have the legal obligation to inform the European Commission (EC) on a regular basis on the radioactivity levels in their environment (drinking water, soil, air and mixed diet). In order to obtain more information on the MS's measurement methods and on the quality of their reported values for the radioactivity levels determined in their environment, the EC has established the International Comparison Scheme for Radioactivity Environmental Monitoring (ICS-REM) (Wätjen, 2008).

The JRC-Geel (from 1993 to 2016 named "Institute for Reference Materials and Measurements" (IRMM) and earlier CBNM "Central Bureau for Nuclear Measurements") is one of the seven institutes of the European Commission's Directorate-General Joint Research Centre (JRC). In the frame of ICS-REM, the Directorate for Nuclear Safety and Security and in particular the Radionuclide Metrology Sector of the Standards for Safety, Security and Safeguard Unit of the JRC, organises on request of the Directorate-General for Energy (DG ENER) the EC Interlaboratory Comparisons (EC ILCs) since 2003. The aim of the EC ILCs is not only to evaluate the results submitted by the participants but also to provide help and advice to the participating MS laboratories via workshops and meetings on how to improve the measurements and methods applied by them. During the last decade, the test materials used in EC ILCs have included air filters (2003), soil samples (2010) and foodstuff samples, such as milk powder (2005), bilberry powder (2011), mineral water (2008) drinking water (2012) and ^{137}Cs in air filters (2014). The approach of JRC-Geel in organising the comparisons is sketched in *Figure 1*.




Fig. 1. Key comparisons of CCRI(II) and traceability of the reference values for samples provided by JRC for the interlaboratory comparisons amongst monitoring laboratories (KCRV = Key Comparison Reference Value).

The EU Member States are obliged to report radiological monitoring data of airborne radioactivity to the EC according to the Council Decision 600/87 (87/600/Euratom,

1987) and the Commission Recommendation 473/2000, (2000/473/Euratom, 2000). These measurements are collected by the JRC in the European Union Radiological Data Exchange Platform (EURDEP, 2016), which makes the non-validated radiological monitoring data available in nearly real-time for the countries reporting to the system. For the moment there are about 5000 operational stations to monitor the airborne radioactivity and dose rate within the European early warning network and EURDEP, but the number of stations with sampling equipment for radioactive particulates in air is significantly lower (~240) than those with instrumentation for dose rate measurements. Furthermore, the stations with sampling equipment use a wide variety of different methods, instruments and air filters.

During a radiological emergency with trans-boundary implications in Europe, the European Commission will issue recommendations to EU Member States based on data from national early warning networks. Therefore, metrologically sound monitoring data of ambient dose rate and airborne radionuclide activity concentrations, co-ordinated with data from international radiological networks, are a prerequisite for adequate environmental radiation monitoring in Europe.

The ENV57 MetroERM "Metrology for radiological early warning networks in Europe" project (MetroERM, 2016) in the frame of the European Metrology Research Program (EMRP) aims at optimising the metrological foundation of measurements (devices and methods) for monitoring airborne radioactivity and promoting pan-European harmonisation in data reliability for area dose rate measurements which are input to the European Radiological Data Exchange Platform (EURDEP) and other monitoring networks. It will provide the unique possibility to comprehensively address the harmonisation of the radiological early warning networks in Europe - the largest and most comprehensive environmental radiation monitoring system worldwide.

The JRC participates in the ENV57 MetroERM project, having the responsibility to carry out a number of tasks. Work Package 3 (WP3) addresses the traceability of dose rate and airborne radioactivity measurements and explicitly supports the process of harmonisation of radiological data from early warning networks in Europe by systematic investigations, comparison exercises and the publication of recommendations.

Task 3.2 of WP3 entitled "Traceability management for airborne radioactivity" aims at developing traceable reference materials and standard sources in the form of large-area spiked aerosol filters and conducting a laboratory comparison exercise to quantify airborne radioactivity field station performance. Deliverables D3.2.1 through D3.2.8 are directly linked to the present intercomparison exercise, of which the JRC undertook the organization and execution. Similar exercises, although EC ILC, were organized in 2014 (Altzitzoglou and Máté, 2016) and in 2003 (Wätjen et al., 2007).

Initially, a questionnaire was conceived and distributed to all European network operators to collect information relevant to the preparation of the exercise. From the analysis of the answers, it was decided to conduct the ILC using ^{137}Cs , ^{134}Cs and ^{131}I as measurands, a change from the two previous similar exercises where ^{137}Cs only was present in the distributed samples.

The aim of this ILC (or Proficiency Test (PT)) exercise was to quantify the performance of field stations measuring the airborne radioactivity. The participating laboratories provided blank air filters of the type they routinely use, which the JRC spiked gravimetrically with known amount of radioactivity individually for each laboratory and returned to them on February 22 and 23, 2016. For this exercise, the filters were spiked with ^{137}Cs , ^{134}Cs and ^{131}I and with activities close to those the laboratories routinely measure, i.e. in the range from 0.2 to 5.0 Bq for ^{137}Cs , from 0.4 to 6.1 Bq for ^{134}Cs and from 0.6 to 8.7 Bq for ^{131}I .

The ILC was accompanied by a second questionnaire to collect information about sample collection, measurement and analysis, as routinely carried out by the

participating laboratories. The deadline to report the measurement results and submit the answers to the second questionnaire was 29 March 2016.

The aim of the present ILC was to obtain an overview of the quality of the results reported, of the application of the measurement methods by the participating laboratories and of any changes that have occurred since the last similar exercises in 2014 and 2003.

This report presents in detail all phases of the ILC organized and conducted in 2016, the description of the intercomparison sample preparation at JRC, the analytical methods used by the laboratories, the treatment of the data reported by the participants and, finally, the evaluation and comparison of the results to the reference values. A robust evaluation of the performance of individual laboratories was performed using three different approaches: percentage difference ($D_{\%}$), E_n numbers (ISO, 2005a), and 'PomPlots' (Pommé, 2016).

2. The 2016 ENV57 MetroERM measurement comparison

2.1 Description of the sample

<i>Nature:</i>	blank air filters provided by the participating laboratories spiked with a known amount of ^{137}Cs , ^{134}Cs and ^{131}I at JRC-Geel
<i>Reference date:</i>	1 March 2016 0:00 UTC
<i>Activity levels:</i>	air filters spiked with activity levels similar to those which the laboratories routinely measure, and in every case, activity levels above the reported detection limit (^{137}Cs activity ranging from 0.2 up 5.0 Bq, ^{134}Cs activity ranging from 0.4 up 6.1 Bq, ^{131}I activity ranging from 0.6 to 8.7 Bq)
<i>Shipping:</i>	spiked air filters sealed in double plastic bags were sent via regular mail to the participating laboratories

2.2 Participating laboratories

In December 2015, the JRC invited laboratories routinely measuring radioactivity in exposed air filters to participate at the 2016 ENV57/MetroERM interlaboratory comparison.

The laboratories participating in this ILC were 67 European laboratories from 29 countries, of which 61 from 26 EU Member States (MS) and 2 from Norway, 1 from Switzerland and 2 from Turkey. Of these laboratories, 57 had participated in the 2014 EC ILC and 26 in the 2003 EC ILC. Seven of the participating laboratories are also partners in the ENV57 MetroERM project.

All laboratories that registered to the ENV57/MetroERM ILC reported their results for all three radionuclides. The list of all participating laboratories is given in *Annex 1*. Since the anonymity is a requirement in this ILC, the identity of the laboratories is not shown in this compilation of the results. The laboratory numbers used throughout the data evaluation in this report are not related to the order of listing the participants in *Annex 1*, but to the Laboratory Code communicated only to the corresponding laboratory.

2.3 Reporting of results

The results for the activity per filter as well as the associated combined uncertainty with coverage factor k had to be reported in the unit Bq.

The reporting of the results together with a questionnaire was performed via an online reporting system, operated by JRC-Geel. Participants were asked to answer all relevant questions regarding the measurement procedure they followed. Information given in this questionnaire was essential for the evaluation of the ILC results. Moreover, it allowed to find possible sources of difficulties and to get an overview of the methods used by the laboratories.

The reference date for all results was 1 March 2016 0:00 UTC.

Table 1. Recommended nuclear data for the three radionuclides of interest (standard uncertainties in parenthesis ($k=1$)) (DDEP, 2016).

Nuclide	Half-life (d)	Energy (keV)	Photons per 100 disintegrations
^{137}Cs	10975 (29)	661.7	84.99 (20)
^{134}Cs	754.0 (5)	563.2	8.342 (15)
		569.3	15.368 (21)
		604.7	97.63 (8)
		795.9	85.47 (9)
		802.0	8.694 (16)
^{131}I	8.0233 (19)	284.3	6.14 (6)
		364.5	81.2 (5)
		637.0	7.12 (7)

2.4 Timetable of the ILC

26 November 2015:	expression of interest
11 December 2015:	invitation letter was sent to the interested laboratories
8 January 2016:	nominated laboratories sent the blank air filters, together with information on their routine measurement conditions
February 2016:	spiked air filters were prepared
22-23 February 2016:	spiked air filters were sent to the participants via express mail (DHL) together with relevant information on the ILC
29 March 2016:	laboratories submitted their results and answered a second questionnaire with information on the air filter measurements
7-8 April 2016:	workshop organized for the participants and the preliminary results presented
20 September 2016:	preliminary results were officially sent to participants

All registered communication with the participants related to this ILC can be found in the *Annexes* of the present report.

3. The Reference value

3.1 Standardisation of the spike solution

For the present ILC exercise three radionuclides had been chosen as measurands, ^{137}Cs , ^{134}Cs and ^{131}I . These three could be considered the major radionuclides to monitor following a nuclear accident, like that in Chernobyl. The determination of ^{137}Cs is simple and straightforward, as it needs no specific corrections. The ^{134}Cs had been selected for its significant coincidence summing effect, which calls for appropriate corrections. Finally, ^{131}I is interesting as it is seldom included in ILCs. It is short lived, easy to measure and it does not present significant coincidence summing effect. Despite its pronounced volatility, it had been chosen to be included in this exercise and it will check for proper application of the decay correction, the correction for decay during measurement and the proper handling of the air filter to avoid loss of the volatile iodine.

An additional difficulty lies in the low to very low activity levels (close to detection limit) and the possible deviation from the routine counting geometry and thus counting efficiency due to the imperfect spiking homogeneity of the filters with the radioactive solution.

The standardisation of the solutions of the three radionuclides used for spiking of the air filters was performed at the JRC-Geel by liquid scintillation counting (LSC). Gamma-ray spectrometry and counting with an ionization chamber were used as secondary methods and as an additional link to older traceable standard solutions.

For the LSC, the efficiency tracing method developed by CIEMAT/NIST (Grau Malonda and Garcia-Toraño, 1982) was used. The principle of the CIEMAT/NIST efficiency tracing method is a combination of theoretical calculations of the radionuclide beta particle counting efficiency and an experimental determination of correction factors with the help of a tracer radionuclide, ^3H in this case. For the CIEMAT/NIST efficiency tracer method, the computer code CN2005 (Günther, 2002) was employed to calculate the radionuclide counting efficiencies. As tracer, the IRMM tritiated water standard was used (Spennol and Denecke, 1964; Makepeace et al, 1998).

All sources were prepared gravimetrically using a Mettler AX26 (Mettler-Toledo, Greifensee, Switzerland) mass comparator, calibrated using traceable weights. To prepare the LSC sources, aliquots of the radioactive solution were gravimetrically dispensed using the pycnometer method (Sibbens and Altitoglou, 2007; Campion, 1975) into 20-mL low-potassium glass LSC vials containing 15 mL of UltimaGold® (PerkinElmer, Boston, MA, USA) liquid scintillation (LS) cocktail, mixed with 1 mL of deionized water. In a similar way, for the gamma-ray spectrometric measurements, the radioactive solution was dispensed by means of a pycnometer onto a plastic foil supported by a 34-mm thin stainless steel annulus and covered by another plastic foil after drying completely.

The LSC sources were measured using a Packard 3100 Tri-Carb TR/AB (PerkinElmer, Boston, MA, USA) liquid scintillation counter and a Wallac Quantulus 1220 (PerkinElmer, Boston, MA, USA) LSC. For the gamma-ray measurements, two high-purity germanium (HPGe) detector systems were used, one with a 36% relative efficiency co-axial detector (Detector A) and the other with a 92% low-background co-axial detector (Detector B) (Canberra Industries, Inc., Meriden, CT, USA). The first

detector was housed in a 10-cm thick Pb shield of circular cross-section, lined with 1 mm Cd and 1 mm Cu; the inner 2 cm of the Pb shield was made of high radiopurity Pb. The latter was housed in a 5-cm thick Pb shield of square cross-section. Both detectors were connected to commercial analogue electronics.

The gamma-ray spectra analysis was performed using the GammaVision-32 software program (ORTEC, Oak Ridge, TN, USA) and the data analysis for both the gamma-ray analysis and the LSC analysis was done with custom made spreadsheets. *Figure 2* shows a typical gamma-ray spectrum of all three radionuclides of interest.




Fig. 2. Typical gamma-ray spectrum of a mixture of the three radionuclides of interest, ^{137}Cs , ^{134}Cs and ^{131}I . The most abundant peaks are indicated.

3.1.1 ^{137}Cs

Seven sources were gravimetrically prepared from the first dilution B1 of the original mother solution (A1) of ^{137}Cs and 4 from the second dilution (C1). The amount of radioactive solution in each source ranged from 12 to 45 mg for B1 and from 35 to 57 mg for C1. All sources were measured 17 times (for 20 minutes each per run) using the Packard LSC and 6 times (for 15 minutes each per run) using the Quantulus LSC over a period of one and a half month. The search for possible impurities and especially for ^{134}Cs by high resolution gamma ray spectrometry was negative.

The activity concentration of the mother solution was found to be $3.50 (3) \text{ MBq g}^{-1}$ on the reference date 1 March 2016 0:00 UTC. As usually, the numbers in parentheses are the numerical values of the combined standard uncertainties u_c expressed in the unit of the quoted result.

Traceability was guaranteed by using the exact same method as for participation at the ongoing comparison BIPM.RI(II)-K1.Cs-137 (Ratel et al., 2005a), following the standardisation of ^{137}Cs for the EC ILC organised in 2003; the standardised solution of

^{137}Cs of the 2003 campaign was submitted together with the results of the described standardisation at JRC-Geel as entry into the International Reference System (SIR) of the Bureau International des Poids et Mesures (BIPM) in order to establish traceability of the activity values of the spiking solutions and subsequently of the reference values of spiked activities on filters. The results of the measurements in the SIR at BIPM, are confirming traceability. The degree of equivalence for our laboratory with respect to the reference value ($D_i=(x_i-x_{\text{Ref}})$) was -0.19 MBq with an uncertainty of 0.32 MBq, which means that the JRC result was 0.73% lower than the reference value with an uncertainty on this difference of 1.16% ($k=2$).

In addition, 3 sources were prepared with 9 to 36 mg from the dilution B1 and 2 with 11 and 43 mg from the dilution C1 for gamma-ray spectrometry. The sources were measured at two different distances from the detectors, a close one and a farther one, against similar sources prepared during the 2003 EC ILC campaign and the 2014 EC ILC campaign and the discrepancy was determined to be less than 1%.

3.1.2 ^{134}Cs

Seven sources were gravimetrically prepared from the first dilution B1 of the original mother solution (A1) of ^{134}Cs and 4 from the second dilution (C1). The amount of radioactive solution in each source ranged from 10 to 45 mg for B1 and from 20 to 54 mg for C1. All sources were measured 31 times (for 20 minutes each per run) using the Packard LSC and 6 times (for 15 minutes each per run) using the Quantulus LSC over a period of one and a half month. No impurities were detected in the solution by high resolution gamma-ray measurements checking for possible impurities and especially for ^{137}Cs .

The activity concentration of the mother solution was found to be 35.5 (4) kBq g⁻¹ on the reference date 1 March 2016 0:00 UTC. As usually, the numbers in parentheses are the numerical values of the combined standard uncertainties u_c expressed in the unit of the quoted result.

Traceability was guaranteed by using the exact same method as for participation at the ongoing comparison BIPM.RI(II)-K1.Cs-134 (Ratel et al., 2005b; Ratel et al., 2007), following the standardisation of ^{134}Cs in 2003; the standardised solution of ^{134}Cs of 2003 was submitted together with the results of the described standardisation at JRC-IRMM as entry into the International Reference System (SIR) of the Bureau International des Poids et Mesures (BIPM) in order to establish traceability of the activity values of the spiking solutions and subsequently of the reference values of spiked activities on filters. The degree of equivalence for our laboratory with respect to the reference value ($D_i=(x_i-x_{\text{Ref}})$) was -69 kBq with an uncertainty of 75 kBq, which means that the JRC result was 0.68% lower than the reference value with an uncertainty on this difference of 0.74% ($k=2$).

In addition, 3 sources were prepared with 11 to 39 mg from the dilution B1 and 2 with 32 and 41 mg from the dilution C1 for gamma-ray spectrometry. The sources were measured at two different distances from the detectors, a close one and a farther one, against similar sources prepared during the 2003 EC ILC campaign and the discrepancy was determined to be less than 1%. Additional measurements using an ionisation chamber gave results higher by 0.7% compared to those obtained by LSC and gamma-ray spectrometry.

3.1.3 ^{131}I

Ten sources were gravimetrically prepared from the original mother solution (A1) of ^{131}I with an amount of radioactive solution in each source ranging from 13 to 55 mg. All sources were measured 16 times (for 20 minutes each per run) using the Packard

LSC and 4 times (for 10 minutes each per run) using the Quantulus LSC over a period of one and a half month.

The activity concentration of the mother solution was found to be 6.12 (6) kBq g⁻¹ on the reference date 1 March 2016 0:00 UTC. As usually, the numbers in parentheses are the numerical values of the combined standard uncertainties u_c expressed in the unit of the quoted result.

A further confirmation of the JRC result was obtained by gamma-ray spectrometry measurements. Three point sources were prepared with 23 to 43 mg from the mother solution A1 for gamma-ray spectrometry. The sources were measured at two different distances from the detectors, a close one and a further one. The results agree well with the activity concentration result obtained by LSC within the corresponding uncertainties, as the discrepancy was determined to be less than 1.5%. It is worth noticing that the activity concentration on the certificate by the Czech Metrology Institute (CMI), supplier of the material, was 202.4(2) kBq g⁻¹ on 20-01-2016, or 4.3% lower than that measured at JRC-Geel.

3.2 Dilutions

In order to approximate the activity level measured at each participating laboratory under routine conditions with the appropriate amount of spiked ¹³⁷Cs, ¹³⁴Cs and ¹³¹I on the filters, five different diluted solutions (E1, E2A, E3, E4, E5) were prepared following the preparation scheme shown in *Figure 3*. The standardised mother solution

Table 2. Dilutions and dilution factors with their combined standard uncertainties u_c (in parentheses).

Solution	Code	Dilution factor D		
		¹³⁷ Cs	¹³⁴ Cs	¹³¹ I
Mother solution	A1	1	1	1
Dilution	B1	98.24 (2)	1	1
Dilution	C1	887.7 (2)	10.0218 (9)	1
Dilution	D1	9092 (2)	75.368 (7)	9.1542 (3)
Dilution	D2A	11111 (6) ×10 ¹	657.3 (2)	93.91 (3)
Dilution	D3	12500 (7) ×10 ¹	751.2 (2)	92.08 (3)
Dilution	D4	7171 (7) ×10 ²	3633 (2)	386.6 (2)
Dilution	D5	947 (1) ×10 ³	4432 (2)	476.4 (3)
Dilution	E1	9571 (2) ×10 ¹	793.5 (1)	96.37 (1)
Dilution	E2A	8969 (5) ×10 ²	5306 (2)	758.1 (3)
Dilution	E3	7519 (4) ×10 ²	4519 (1)	553.9 (2)
Dilution	E4	3990 (4) ×10 ³	20215 (8)	2151 (1)
Dilution	E5	842 (1) ×10 ⁴	3939 (2) ×10 ¹	4233 (2)

(A1) of ^{137}Cs was first diluted via two intermediate dilutions (B1 and C1). The diluent was a solution of $50 \mu\text{g mL}^{-1} \text{Cs}^+$ (as CsCl) in 0.1M HCl . The standardised mother solution (A1) of ^{134}Cs was first diluted via one intermediate dilution (B1). The diluent was equally a solution of $50 \mu\text{g mL}^{-1} \text{Cs}^+$ (as CsCl) in 0.1M HCl .

Then, aliquots from the C1 dilution of ^{137}Cs , from the B1 of ^{134}Cs and the mother solution (A1) of ^{131}I were gravimetrically added to 5 vials containing given amounts of a solution of $25 \text{ mg g}^{-1} \text{KI}$ and $25 \text{ mg g}^{-1} \text{Na}_2\text{S}_2\text{O}_3$ to produce the dilutions D1, D2A, D3, D4 and D5. Therefore, the D series dilutions contain all three radionuclides at various levels. Furthermore, methylene blue, used to visualise the spikes on the filters, was gravimetrically added to each of the dilutions.

Finally, the D series dilutions were further diluted by adding amounts of a solution containing $25 \text{ mg g}^{-1} \text{KI}$ and $25 \text{ mg g}^{-1} \text{Na}_2\text{S}_2\text{O}_3$ to produce the dilutions E1, E2A, E3, E4 and E5. Those final dilutions were used to spike the air filters.

Table 2 lists the solutions and dilution factors including the addition of methylene blue. The numbers in parentheses are the numerical values of the combined standard uncertainties u_c expressed in the unit of the quoted result. In Figure 3 the dilution scheme is shown schematically together with the approximate dilution factors and the activity concentration of each of the solutions prepared.




Fig. 3. Preparation scheme of the ^{137}Cs , ^{134}Cs and ^{131}I dilutions.

It should be noted that all dilutions were prepared gravimetrically for utmost traceability and small uncertainty on the dilution factors. In addition, quantitative sources were prepared from all dilutions for quality control by both liquid scintillation counting and gamma-ray spectrometry. The results of these measurements for the last two dilution steps D and E are presented in *Annex 8* and confirm the gravimetric dilution factors, with the exception of ^{131}I , which will be discussed later. The scatter of the results is large and it is due to counting statistics; many sources had to be measured in short time. In addition no optimization was done for the counting efficiency for the filter geometry.

3.3 Spiking of the air filters

On the basis of the information provided by the participating laboratories in the relevant questionnaire, the ^{137}Cs , ^{134}Cs and ^{131}I activity per filter and the spiking pattern was determined for each filter individually. The amount of the radioactivity spiked onto each filter was chosen to resemble the activity routinely measured by the corresponding laboratory in a whole filter (or set of filters if it measures several at the same time). In cases where the laboratory declared to usually measure (sets of) filters with activities below detection limit, an activity higher than the declared detection limit was distributed.

The JRC Radionuclide Metrology Sector prepared 67 air filters (*Annex 6*) by depositing on each of them gravimetrically an amount from one of the standard solutions E1, E2A, E3, E4 and E5, containing ^{137}Cs , ^{134}Cs and ^{131}I . Each participating laboratory in this ILC exercise had sent two blank filters of the type it is routinely using and after spiking one of them at JRC with the ^{137}Cs , ^{134}Cs and ^{131}I solution, the filters were returned to the participants in order to be measured according to their routine procedure. The second blank filter was kept in reserve. The types, materials, dimensions and shapes of the air filters are given in *Annex 5*.

Since a uniform distribution of spikes was not feasible, depending on the filter size about 10 to 100 droplets were dispensed gravimetrically using a pycnometer, in a more or less symmetric pattern on the air filter. In any case, the spiked spots could easily be distinguished due to their methylene blue color, which allowed accounting for their discrete distribution when preparing the filters for measurement or when calculating counting efficiency corrections.

Where the filter was large enough, the filter was folded up in a way that the active part came into the centre of the pack such that any substance falling off the filter during transport by regular mail would still be caught in the surrounding filter and thus would not be lost. In the case of small filters this solution was not possible, therefore, the laboratories were asked to also measure the empty plastic bag in which the samples were shipped (or to measure the filter inside the bag) in order to verify that no losses from the filter had occurred.

Various spiked filters are shown in *Annex 7*. In many of the filters (e.g. those made of polypropylene or glass fibres) spiking became difficult due to the hydrophobic nature of the filters. The drops of radioactive solution were not absorbed into the filter material, therefore, the spiking had to be followed by several hours of drying at room temperature.

3.4 Reference values

The reference activity values, i.e. the spiked activities on the filters, were calculated using the activity concentration of the mother solution determined by primary standardisation and the gravimetrically determined dilution factors (*Table 2*) of the spiking standard solutions E1, E2A, E3, E4 and E5. The mass and activity of the dispensed solution on each filter was determined once per filter by weighing the

pycnometer before and after depositing the total number of drops on the corresponding filter. *Tables 5, 10 and 14* show the reference values A_0 for the deposited ^{137}Cs , ^{134}Cs and ^{131}I activity, respectively, on each filter and its combined standard uncertainty u_c . The standard uncertainty of activity includes the uncertainty contributions from the primary standardisation, the dilutions and the weighings of the filter spiked aliquots. The lowest activity of ^{137}Cs , ^{134}Cs and ^{131}I spiked on an air filter was 0.2, 0.4 and 0.6 Bq respectively and the highest values were 5.0, 6.1 and 8.7 Bq, respectively. The reference date for reporting the activity (and likewise of the reference values) is 1 March 2016 0:00 UTC. The numbers in parenthesis are the expanded uncertainties U with coverage factor $k=1$.

3.5 Quality control measurements

For the purpose of quality control, sources were prepared at each stage of the preparation of the dilutions and the spiking of the filters. Especially for the filter spiking control, 15 additional filters (Whatman 42 ashless $\varnothing 70$ mm) and 7 point sources were produced. *Table A-6* in *Annex 6* lists in sequence all filter samples prepared for the participating laboratories as well as all control samples. It shows clearly, how well the quality control samples interleave the rest of the samples prepared.

The 15 spiked control filters and the 7 point sources were measured by gamma-ray spectrometry at the JRC-Geel, using the two HPGe detectors mentioned in *Section 3.1*. Some of the sources were measured by both detectors. The point sources were measured at a close distance (2 mm) from the detector window and the counting efficiency calibration was performed with similar standard point sources. The filters were measured in their protective plastic bags and placed directly on the detector window. The acquisition time ranged from 4 to 64 hours. No attempt was made for further geometry corrections, as the results were treated in a relative way, just to confirm the integrity of the dilutions. Due to the large number of quality control sources to measure, the low spiked activity and the limited time, the counting statistics were not optimal. In conclusion, the measurement results confirm the activity concentrations determined from the gravimetric dilution and spiking procedure for ^{137}Cs and ^{134}Cs . However, for ^{131}I are not satisfactory and they will be discussed later.

The measured activity, its ratio to the reference value and its relative deviation from the reference activity, calculated from the mass of the spiked solution and its activity concentration and expressed in percent, are given for the quality control filters in *Tables A-7a* and *A-7b*, while those for the point sources in *Tables A-8a* and *A-8b* of *Annex 8*. The uncertainty is the expanded uncertainty combining both the uncertainty on the measured and reference values. All results are calculated for the reference date of 1 March 2016 0:00 UTC. *Figure A-1* of *Annex 8* shows the ratio of the measured to the reference activity for the quality control filters and for all three radionuclides. *Figure A-2* shows the corresponding ratio for the point sources.

3.6 Influence of discrete spiked activity distribution on the counting efficiency

The activity distribution collected by air aspiration on air filters is presumably uniform (in the sense of continuous) and assumed mostly homogeneous, although exceptions may exist. The spiked air filters prepared for the present ILC campaign have been prepared by discretely depositing aliquots of a standard radioactive solution on the filters. Therefore, the activity distribution on the spiked filters is discrete and symmetric, but not continuous. When considering a small part of the filter, one can thus claim that the distribution is inhomogeneous. The more drops deposited, the

closer to a homogeneous distribution. This inhomogeneity is expected to affect the counting efficiency, but there are more parameters influencing the counting efficiency, which we will make an attempt to quantify.

A second parameter is the size of the deposited drops. Although the deposited solution mass, and thus activity, per drop is rather constant, depending on the quality and material composition of each filter the drop spreads to a smaller or larger area.

A third parameter for the determination of the spiked activity on the air filters is the counting geometry. For the smaller filters the usual geometry is placing the filter in front of the detector, either on the detector window or at a certain distance from the detector window. For larger filters the usual way is to fold in a suitable manner the filter, or even press it to form a pellet and then place it in front of the detector. Unless the folding brings for some specific reason most of the active spots closer or further from the detector, it is expected that folding the filters distributes uniformly the active spots and consequently improves the homogeneity of the activity distribution.

A fourth parameter could be the size (active diameter) of the counting detector, giving an advantage to larger detectors, as in that case, because of the subtended solid angle, the active area of the filter is better covered by the detector.

In order to study the influence of those parameters on the counting efficiency for the measurement of the spiked air filters, a number of Monte Carlo simulations have been performed using the GEOLEP proprietary computer code (Solé, 1990; Lépy et al., 2010).

The counting efficiency has been calculated for three different HPGe detectors (of different Ge crystal sizes) and for 3 different filter sizes spiked in different patterns and with different number (9 or 16) and size of active spots (0.3, 0.7 and 1.0 mm in diameter). The size of the detectors modelled was $\varnothing 58.5 \times 53.5$ mm (Detector A), $\varnothing 77 \times 78$ mm (Detector B) and $\varnothing 80 \times 30$ mm (Detector C). The dimensions, matrix composition and density of the air filters, as well as the dimensions of the detectors have been used as model inputs to the Monte Carlo code. The simulations assumed that the gamma-ray emissions were isotropic and uncorrelated. The uncertainty of the Monte Carlo simulations was in all cases better than 1%.

The filters in *Figure 4* depict the different cases which have been modelled and *Tables 3a* and *3b* give more details, as well as the results of the Monte Carlo simulations. The last column in the tables gives the percent deviation of the calculated counting efficiency for the spiked air filters from that of an air filter of the same size but with a homogeneous activity distribution.

From *Tables 3a* and *3b* and the simulated counting efficiency results it is shown that for active filter areas smaller than or equal to the detector active areas, the discrepancies of the counting efficiency between discrete and homogeneous activity distribution is not large and, in the studied cases, below 5%.

However, the worst case scenario was that of Detector C and the larger filter size of 130-mm diameter and a 110-mm diameter active area, which resulted in an overestimation of the efficiency by 35.6% or an underestimation of the activity by the same percentage, most probably because the active spots are placed closer to the center of the detector, whereas on a homogeneous exposed filter the active area on the filter extends further from the center of the detector. If more spots are added towards the circumference (red spots in *Figure 4c*) and the calculations are repeated, the underestimation of the activity is reduced to 10.7%. If we compare the efficiency result to that of a 100-mm diameter active area (more realistic as the spots do not reach the edge of the 100-mm circle) then we will find it only 0.6% lower than that for the homogeneous source.

From these calculations, it can be concluded that the most critical parameter to simulate an exposed air filter, is not the quasi-homogeneous distribution of the active

spots, or the size of the spots, or their number. Instead, it is crucial to cover with active spots the entire area, which is normally loaded with activity in a real exposed filter. Activity deposited more to the center of the filter will increase the apparent counting efficiency and give higher activity results (if the efficiency is not appropriately corrected). Too much activity placed closer to the circumference will lower the apparent counting efficiency and give lower activity results.




Fig. 4. Different cases of spiked air filters, which have been modelled for Monte Carlo simulations to calculate counting efficiencies with some typical detectors.

- a) \varnothing 70-mm filter with 9 active spots
- b) \varnothing 110-mm filter with 16 active spots and
- c) \varnothing 130-mm filter with 16 active spots and (including the red dots) 24 active spots

Table 3a. Deviation of the counting efficiency of a spiked filter from that of a filter with homogeneous distribution of the activity as calculated using the Monte Carlo code GEOLEP (see text). Calculations are performed for two different detectors (A and B), two different sizes of the homogeneous active area ($\varnothing 50$ and 60 mm) and three different diameters of the 9 active spots ($\varnothing 3$, 7 and 10 mm).

Detector and Ge crystal size (dia.xheight in mm)	Measurement geometry	Filter diameter (mm)	Homogeneous active area diameter (mm)	Number of active spots on the spiked filter	Diameter of active spots on the spiked filter (mm)	Efficiency deviation from the homogeneous distribution (%)
Detector A $\varnothing 58.5 \times 53.5$	on detector window	70	50	9	3	-4.27
					7	-4.50
					10	-5.09
			60	9	3	6.62
					7	6.38
					10	5.71
	at 10 mm	70	50	9	3	-2.92
					7	-3.16
					10	-3.63
			60	9	3	4.67
					7	4.42
					10	3.91
Detector B $\varnothing 77 \times 78$	on detector window	70	50	9	3	-2.54
					7	-2.67
					10	-2.98
			60	9	3	4.01
					7	3.88
					10	3.54
	at 10 mm	70	50	9	3	-1.96
					7	-2.13
					10	-2.25
			60	9	3	2.86
					7	2.68
					10	2.56

Table 3b. Deviation of the counting efficiency of a spiked filter from that of a filter with homogeneous distribution of the activity as calculated using the Monte Carlo code GEOLEP (see text). Calculations are performed for two different detectors (A and B), two different sizes of the homogeneous active area ($\varnothing 90$ and 110 mm) and three different diameters of the 16 active spots ($\varnothing 3$, 7 and 10 mm). A third detector (C) was introduced as well and the counting efficiency for filters with homogeneous active area of $\varnothing 110$ mm and $\varnothing 100$ mm was simulated for 16 and 24 active spots (see *Figure 4c*).

Detector and Ge crystal size (dia.xheight in mm)	Measurement geometry	Filter diameter (mm)	Homogeneous active area diameter (mm)	Number of active spots on the spiked filter	Diameter of active spots on the spiked filter (mm)	Efficiency deviation from the homogeneous distribution (%)
Detector A $\varnothing 58.5 \times 53.5$	on detector window	110	90	16	3	-5.63
					7	-5.83
					10	-5.82
		110	100	16	3	4.24
					7	4.02
					10	4.03
	at 10 mm	110	90	16	3	-4.53
					7	-4.61
					10	-4.52
			100	16	3	3.94
					7	3.86
					10	3.96
Detector B $\varnothing 77 \times 78$	on detector window	110	90	16	3	-5.55
					7	-5.66
					10	-5.55
		110	100	16	3	4.94
					7	4.82
					10	4.95
	at 10 mm	110	90	16	3	-3.82
					7	-3.89
					10	-3.82
			100	16	3	4.08
					7	4.01
					10	4.08
Detector C $\varnothing 80 \times 30$	on detector window	130	110	16	15	35.6
				24	15	10.7
			100	24	15	-0.6

4. Questionnaire on sampling method, air filter used and measurement conditions

In the preparative phase of the exercise the participants were asked to fill in a questionnaire (*Annex 3*) in order to receive air filters spiked with activity levels similar to those which the laboratories routinely measure and to allow them to measure following their usual procedure and measurement geometry. The questionnaire focused on two main fields: (1) general information and (2) measurement details. The questionnaire, besides general information, required information on the sampling method, the air filter used and the measurement conditions. The compilation of the answers provided by the participants to the questionnaire is described in this chapter and in *Annex 4*.

The first part of the questionnaire collected general information about the organisations and laboratories. The questions were grouped into four major topics: 1) contact details, 2) identity of laboratory, 3) accreditation and 4) previous participation in EC ILCs. On the basis of the answers to the question "What is the type of your laboratory?" it could be concluded that almost all (63 out of 67) participating laboratories monitor radioactivity in the environment routinely. Regarding accreditation, two thirds of the laboratories are accredited or certified primarily for gamma-ray spectrometry measurements and one fifth of the laboratories are authorised by the government or responsible body for radioactivity measurements in the environment. Furthermore, most of the participants have already had experience participating in EC ILCs; 57 laboratories had participated in the 2014 EC ILC and 26 in both the 2003 and 2014 EC ILCs.

The second part of the questionnaire contained questions related to the technical part of the measurements of radionuclides in air filter. Information on the air filter type used, the air sampling methods and the radioactivity measurement was gathered. As expected, air filters of various types and sizes are employed by the participating laboratories (*Annex 5*). The most popular air filter materials were glass fibre, nitrocellulose and polypropylene. According to size, they could be grouped in three categories. On the basis of the second part of the questionnaire, it could be concluded that all of the laboratories determine the radionuclides in air filters routinely, but there are no harmonised protocols. No correlation was found between the type and size of air filters, the sampling period, the total volume of air sampled per filter and the sampling frequency, except the fact that if the air filter size is larger, then the total volume of air sampled per filter is usually higher as well. The sampling frequency can be on a daily, weekly, monthly or annual basis or only occasionally.

Finally, in order to decide on the spiking of the individual filters, information on the minimum detectable activity (*MDA*), the typical activity level of ^{137}Cs , ^{134}Cs and ^{131}I measured per filter and any particular wish for a given spiking pattern or packaging were collected in the second part of the questionnaire. The ^{137}Cs activity per filter (or bunch of filters) measured routinely by the participating laboratories varied from 0.0001 to a few Bq with a mean of 9.62 Bq and a median of 0.10 Bq, but in most cases it is below their detection limit. Similarly, the ^{134}Cs activity per filter (or bunch of filters) measured routinely by the participating laboratories varied from 0.0001 to a few Bq with a mean of 12.30 Bq and a median of 0.03 Bq. For the ^{131}I activity, the participants reported activities from 0.004 to a few Bq with a mean of 50 Bq and a median of 0.04 Bq. As many of the values reported were elevated, it can be that mistakenly detection limits were reported instead. Additionally, besides the commonly measured radionuclides, several natural and artificial radionuclides are routinely determined by the laboratories, as indicated in *Table 4*.

Table 4. Radionuclides measured by the participating laboratories.

Nuclide	Number of laboratories reported measuring
^7Be	60
^{137}Cs	58
^{131}I	49
^{40}K	47
^{134}Cs	46
^{210}Pb	42
^{226}Ra	10
^{214}Pb , ^{214}Bi	8
^{212}Pb , ^{228}Ac	7
^{90}Sr , ^{60}Co	5
^{212}Bi , ^{238}U , ^{228}Ra	4
^{106}Ru , ^{125}Sb , ^{208}Tl , ^{243}Am , Gross alpha/beta	2
^{22}Na , ^{75}Se , ^{95}Nb , ^{95}Zr , ^{103}Pu , ^{141}Ce , ^{144}Ce , ^{192}Ir , ^{232}Th , ^{234}Th , ^{238}Pu , ^{239}Pu , ^{240}Pu , NORM	1

5. Questionnaire on analytical and measurement procedures

The descriptions of the analytical and measurement procedures applied by the participants were collected by means of a questionnaire together with the reporting of the results (*Annex 9*). This questionnaire was divided into three parts: (1) sample treatment, (2) equipment used and (3) measurement and data evaluation including uncertainty budget. The evaluation of the answers provided by the participants to the questionnaire is described in this chapter and in *Annex 10*.

Most of the laboratories (62 or 92.5% of the total) reported that they followed their routine procedures for the determination of activity in the spiked air filter. Three laboratories (Laboratories 19, 22 and 65) mentioned that they usually measure a stack of several (4-6) exposed filters in their routine procedure, one (Laboratory 57) mentioned that folded the spiked filter instead of cutting it and another laboratory (Laboratory 25) mentioned that they used software to correct for the difference in active area from their usual geometry.

It was advised to the participants to measure also the inner protective plastic bag, either together with the spiked air filter or separately, to ensure that no activity from the filter was left in the bag. That proved to be the case for ^{131}I and it is explained later. Measuring the plastic bag is not straightforward, though, especially those of the large filters. Usually, it is not easy to fit the air filters together with the plastic bag into the routinely used geometry or the geometry of the spiked air filter was not the same as their routine samples. Thirty three laboratories (49%) measured the protective

plastic bag together with the spiked air filter, 26 (39%) measured them separately and 8 (12%) did not measure the plastic bag at all.

The filter preparation for the measurement in the participating laboratories varied from no particular preparation at all, to pressing/compressing the spiked air filters, folding, packing together with blank air filters or the combination of the previous steps.

The majority of the laboratories used high-purity germanium (HPGe) detectors to measure the spiked air filters, of which 79% were using digital signal processing and 21% analogue electronics. Six participants (Laboratories 8, 13, 15, 22, 28 and 54) used self-developed software for the peak area determination and data evaluation, while the rest used commercially available software. For the determination of the counting peak efficiency of the detectors almost half of the laboratories (31 laboratories or 46%) used reference air filters made by spiking single or multiple-nuclide standard solutions onto blank filters. Software was used by 15 laboratories (22%) to determine the counting efficiency by computer simulation. Nineteen laboratories (28%) used a combination of a multigamma standard source and appropriate software. Two laboratories, probably erroneously, reported that no efficiency calibration was performed.

The laboratories measured the sources for a minimum of 1 hour and up to 233 hours (mean 53 h, median 47.9 h) as shown in *Table A-10 of Annex 10*. Respectively, the background was measured for a period from 15 to 333 hours. For the determination of the *MDA*, the laboratories used commercially available software and/or different calculation methods based on the Currie method, ISO 11929/2010, Risø method, DIN 25482 or other methods.

6. Evaluation and comparison of the results

6.1 Identification of outliers and normal distribution test

In order to evaluate the performance of the laboratories, their measurement results need to be compared to the individual reference activity value, i.e. the spiked activity on each filter. The individual reported activities were normalised to their assigned reference spiked activity by calculating the measured/reported-to-reference activity ratio. $u_c(k=1)$ was determined according to the following equation:

$$u_c(A/A_0) = \sqrt{\left[\frac{u_c(A_0)}{A_0}\right]^2 + \left[\frac{u_c(A)}{A}\right]^2} \quad (1)$$

where

- A is the activity value reported by the participating laboratory
- $u_c(A)$ is the standard uncertainty of a participant's result ($k=1$)
- A_0 is the assigned activity reference value
- $u_c(A_0)$ is the standard uncertainty of the assigned reference value ($k=1$)
- $u_c(A/A_0)$ is the combined standard uncertainty of the ratio A/A_0 ($k=1$)

Due to the relatively small uncertainty of the reference values, the uncertainty of the activity ratios measured versus spiked is dominated by the uncertainty of the measured values.

The limit of acceptable precision was set to $\pm 20\%$ deviation from the reference value. A more tolerant $\pm 33\%$ limit was also included for backwards compatibility with the previous ILCs of 2014 and 2003. The 20% value was chosen as it was assumed that

gamma-ray spectrometry – even with the low activity deposited by spiking on the filters using a technique which cannot provide as homogeneous distribution of the activity on the surface as the normal air sampling – can easily be performed within such uncertainty.

The presence of statistical outliers among the reported results was investigated using the Grubbs' test at a level of significance $\alpha=1\%$ and $\alpha=5\%$, as suggested in ISO/IEC 5725-2 (ISO, 1994). Statistical analysis of the results was carried out for the different working dilutions (E1 to E5) and for the ratio of the measured-to-reference radionuclide activity, separately. The reference value of the working dilutions is based on the standardised reference value of the mother solution and the gravimetrically determined dilution factors. The laboratory values used for the statistical analysis were calculated for each laboratory according to the following formula and grouped on the basis of the working dilution used (E1 to E5):

$$C = \frac{A}{m} \quad (2)$$

where

A is the activity value reported by the participating laboratory (Bq)
 m is the mass of the dilution spiked by JRC on the surface of the air filter (g)

Moreover, the distribution of the data was tested using the normal probability plot and the frequency histogram.

The z values are calculated according to the ISO 13528:2015(E), (ISO, 2015) in the following way:

$$z = \frac{\left(\frac{A}{A_0} - 1\right)}{\sigma_{pt}} \quad (3)$$

where

A is the activity value reported by the participating laboratory
 A_0 is the assigned activity reference value
 σ_{pt} is the standard deviation for proficiency assessment

6.2 Scores and evaluation criteria

6.2.1 Percentage difference

An alternative way of presentation albeit yielding no additional information versus the ratio, is the percentage difference (ISO, 2005a). The percentage difference from the reference activity value is calculated using the formula:

$$D_{\%} = \frac{A - A_0}{A_0} \cdot 100\% \quad (4)$$

where

A is the activity value reported by the participating laboratory
 A_0 is the assigned activity reference value

For the environmental radioactivity measurements the criterion of $\pm 20\%$ difference from the reference value is mostly used.

6.2.2 E_n number

To take the expanded uncertainty of the reported results and that of the reference values into account in the analysis, a performance test using E_n numbers was applied (ISO, 2005a). The calculation of the E_n numbers was carried out according to the following formula:

$$E_n = \frac{A - A_0}{\sqrt{U(A)^2 + U(A_0)^2}} \quad (5)$$

where

A	is the activity value reported by the participating laboratory
A_0	is the assigned activity reference value
$U(A)$	is the expanded uncertainty of a participant's result ($k=2$)
$U(A_0)$	is the expanded uncertainty of the assigned reference value ($k=2$)

When uncertainties are estimated according to the Guide to the Expression of Uncertainty Measurement (GUM) (ISO, 2008), a measurement result with its uncertainty interval giving a level of confidence of 95% should overlap with the reference value and its expanded uncertainty. Therefore, E_n numbers are interpreted as following:

- If $|E_n| \leq 1$, the laboratory values are compatible with the reference value;
- If $|E_n| > 1$, the laboratory values differ significantly from the reference values, the sources of deviation should be investigated and corrected, "warning signal";
- If $|E_n| > 1.5$, there is an urgent need to investigate and find the sources of the large deviation, "action signal".

6.2.3 Compatibility test

The reference values, the reported values, the ratios, the percentage differences and the E_n numbers are given for each laboratory and for all three radionuclides in the *Tables 5, 10, 14*.

For results to be compatible with the assigned reference values, they have to satisfy both the percentage difference, $D\%$ and E_n tests. This means that the percentage difference should be within the $\pm 20\%$ limit from the reference value and the absolute E_n should be smaller than or equal to unity. The compatibility for each laboratory and for the three radionuclides is given in the last three columns of *Table 21*.

6.2.4 PomPlot

In order to compare the results, a modern type of graph, the "PomPlot", which underlines the importance of the assigned uncertainties is applied.

The PomPlot, an intuitive graphical method, is used to produce an overview of the results (Pommé, 2016). It displays the relative deviations (D/MAD) of the individual results A from the reference value A_0 on the horizontal axis and relative uncertainties (u/MAD) on the vertical axis (*Figure 5*). For both axes, the variables are expressed as multiples of MAD , which is defined as the median of absolute deviation from the reference value

$$MAD = \text{Median}|D_i|, (i = 1, \dots, n) \quad (1)$$

where D_i is the difference between the reported and the reference activity:

$$D_i = \frac{A_i}{A_0} - 1 \quad (2)$$

The median absolute deviation MAD is used because of its robustness.

For every datum point the uncertainty is calculated as an independent sum of the reported combined uncertainties on A_i and A_0

$$u_i^2 = u_c^2(A_i) + u_c^2(A_0) \quad (3)$$

where $u_c(A_i) = U(A)_i/k$ and $u_c(A_0) = U(A_0)/k$ (4)




Fig. 5. Interpretation of the PomPlot (Pommé, 2016)

The ζ -scores, $|\zeta| = |D/u| = 1, 2$ and 3 , are represented by diagonal solid lines, creating the aspect of a pyramidal structure. The ζ -score is a measure for the deviation between laboratory result and reference value relative to the total uncertainty (ISO, 2005a). The points on the right-hand side of the graph correspond to results that are higher than the reference value whereas lower values are situated on the left. When the uncertainty is small, the corresponding point is situated high in the graph. The most accurate results should be situated close to the top of the pyramid. Points outside of the $\zeta = \pm 3$ lines are probably inconsistent with the reference value.

7. Results

All 67 participating laboratories registered to this ILC reported valid results for all three radionuclides. Two laboratories reported shortly after the submission was closed and their results were considered valid and were retained.

The Laboratory Code numbers in this report are those communicated to each participant individually and they are confidential. They have no correlation with the laboratory codes used during the exercise.

7.1 ^{137}Cs

7.1.1 Reported results

Table 5 presents the reference activities of each individual spiked air filter, the activity results as reported by the participating laboratories, the ratio of the reported to the reference activity, the percentage difference ($D\%$) of the reported to the reference activity and the E_n number, with their associated standard uncertainties ($k=1$) for the case of ^{137}Cs . The Laboratory Code numbers are the confidential code numbers communicated to the participants and have no correlation with the laboratory codes used during the exercise.

Table 5. ^{137}Cs reference (spiked) activity, activity result as reported by the participating laboratories, ratio of the reported to the reference activity, percentage difference ($D\%$) of the reported to the reference activity and E_n number, with their associated standard uncertainties ($k=1$).

Lab Code	Reference activity A_0 (Bq)	Uncertainty $u_c(A_0)$ (Bq) $k=1$	Reported activity A (Bq)	Reported uncertainty $u_c(A)$ (Bq)	Reported k factor	Uncertainty $u_c(A)$ (Bq) $k=1$	Relative unc. $u_c(A)/A$ (%) $k=1$	Ratio (A/A_0)	Standard uncertainty $u_c(A/A_0)$ $k=1$	Percentage Difference $D\%$	Standard uncertainty $u_c(A/A_0)$	E_n
1	0.398	0.003	0.34	0.09	2	0.045	13.2	0.85	0.11	-14.5	1.9	-0.6
2	0.266	0.002	0.247	0.019	1	0.019	7.7	0.93	0.07	-7.2	0.6	-0.5
3	0.450	0.004	0.46	0.05	2	0.025	5.4	1.02	0.06	2.3	0.1	0.2
4	0.364	0.003	0.31	0.031	1	0.031	10.0	0.85	0.09	-14.9	1.5	-0.9
5	0.379	0.003	0.42	0.05	2	0.025	6.0	1.11	0.07	10.9	0.7	0.8
6	0.510	0.004	0.521	0.05	1.65	0.03	5.8	1.02	0.06	2.2	0.1	0.2
7	0.512	0.004	0.447	0.04	1	0.04	8.9	0.87	0.08	-12.6	1.1	-0.8
8	4.287	0.036	5.3	0.6	2	0.3	5.7	1.24	0.07	23.6	1.4	1.7
9	0.365	0.003	0.384	0.048	2	0.024	6.3	1.05	0.07	5.2	0.3	0.4
10	0.197	0.002	0.165	0.017	2	0.0085	5.2	0.84	0.04	-16.4	0.9	-1.9
11	0.215	0.002	0.22	0.04	2	0.02	9.1	1.02	0.09	2.2	0.2	0.1
12	2.121	0.018	1.94	0.14	1	0.14	7.2	0.91	0.07	-8.5	0.6	-0.6
13	0.339	0.003	0.35	0.028	2	0.014	4.0	1.03	0.04	3.1	0.1	0.4
14	0.341	0.003	0.3349	0.0087	1	0.0087	2.6	0.98	0.03	-1.9	0.1	-0.3
15	0.280	0.002	0.35	0.03	1	0.03	8.6	1.25	0.11	24.8	2.1	1.2
16	0.286	0.002	0.354	0.0448	2	0.0224	6.3	1.24	0.08	23.7	1.5	1.51
17	0.252	0.002	0.25	0.02	1	0.02	8.0	0.99	0.08	-0.8	0.1	0.0

Lab Code	Reference activity A_0 (Bq)	Uncertainty $u_c(A_0)$ (Bq) $k=1$	Reported activity A (Bq)	Reported uncertainty $u_c(A)$ (Bq)	Reported k factor	Uncertainty $u_c(A)$ (Bq) $k=1$	Relative unc. $u_c(A)/A$ (%) $k=1$	Ratio (A/A_0)	Standard uncertainty $u_c(A/A_0)$ $k=1$	Percentage Difference $D\%$	Standard uncertainty $u_c(A/A_0)$	E_n
18	0.419	0.004	0.3505	0.0297	1	0.0297	8.5	0.84	0.07	-16.3	1.4	-1.1
19	0.493	0.004	0.464	0.027	1	0.027	5.8	0.94	0.06	-5.8	0.3	-0.5
20	0.294	0.003	0.33	0.03	2	0.015	4.5	1.12	0.05	12.1	0.6	1.2
21	0.511	0.004	0.459	0.026	1	0.026	5.7	0.90	0.05	-10.1	0.6	-0.98
22	0.347	0.003	0.372	0.031	1	0.031	8.3	1.07	0.09	7.2	0.6	0.4
23	0.574	0.005	0.6	0.04	2	0.02	3.3	1.05	0.04	4.5	0.2	0.6
24	0.441	0.004	0.42	0.033	2	0.017	3.9	0.95	0.04	-4.9	0.2	-0.6
25	0.210	0.002	0.194	0.04	2	0.02	10.3	0.92	0.10	-7.6	0.8	-0.4
26	0.388	0.003	0.97	0.19	2	0.095	9.8	2.50	0.25	149.7	14.7	3.1
27	4.995	0.042	4.41	0.1	2	0.05	1.1	0.88	0.01	-11.7	0.2	-4.5
28	0.372	0.003	0.534	0.139	2	0.070	13.0	1.43	0.19	43.4	5.7	1.2
29	0.390	0.003	0.3384	0.0291	1.65	0.0176	5.2	0.87	0.05	-13.2	0.7	-1.4
30	0.390	0.003	0.464	0.013	2	0.007	1.4	1.19	0.02	19.1	0.3	5.1
31	0.299	0.003	0.277	0.016	2	0.008	2.9	0.93	0.03	-7.5	0.2	-1.3
32	0.286	0.002	0.3	0.05	2	0.025	8.3	1.05	0.09	4.9	0.4	0.3
33	0.339	0.003	0.37	0.04	2	0.02	5.4	1.09	0.06	9.0	0.5	0.8
34	0.524	0.004	0.492	0.048	1	0.048	9.8	0.94	0.09	-6.2	0.6	-0.3
35	0.485	0.004	0.596	0.244	2	0.122	20.5	1.23	0.25	22.8	4.7	0.5
36	0.499	0.004	0.521	0.038	2	0.019	3.6	1.04	0.04	4.5	0.2	0.6

Lab Code	Reference activity A_0 (Bq)	Uncertainty $u_c(A_0)$ (Bq) $k=1$	Reported activity A (Bq)	Reported uncertainty $u_c(A)$ (Bq)	Reported k factor	Uncertainty $u_c(A)$ (Bq) $k=1$	Relative unc. $u_c(A)/A$ (%) $k=1$	Ratio (A/A_0)	Standard uncertainty $u_c(A/A_0)$ $k=1$	Percentage Difference $D\%$	Standard uncertainty $u_c(A/A_0)$	E_n
37	0.332	0.003	0.35	0.03	2	0.015	4.3	1.05	0.05	5.4	0.2	0.6
38	0.379	0.003	0.38	0.03	1	0.03	7.9	1.00	0.08	0.2	0.0	0.0
39	0.368	0.003	0.3	0.03	1	0.03	10.0	0.82	0.08	-18.4	1.8	-1.1
40	0.461	0.004	0.57	0.05	2	0.025	4.4	1.24	0.06	23.6	1.1	2.1
41	0.364	0.003	0.373	0.01	2	0.005	1.3	1.03	0.02	2.5	0.0	0.8
42	0.370	0.003	0.296	0.017	1	0.017	5.7	0.80	0.05	-19.9	1.2	-2.1
43	0.244	0.002	0.26	0.04	2	0.02	7.7	1.07	0.08	6.8	0.5	0.4
44	0.379	0.003	0.3	0.05	2	0.025	8.3	0.79	0.07	-20.8	1.7	-1.6
45	0.670	0.006	0.705	0.026	1	0.026	3.7	1.05	0.04	5.2	0.2	0.7
46	0.343	0.003	0.3	0.05	1	0.05	16.7	0.87	0.15	-12.6	2.1	-0.4
47	0.434	0.004	0.39	0.021	1	0.021	5.4	0.90	0.05	-10.1	0.5	-1.02
48	0.388	0.003	0.43	0.02	1	0.02	4.7	1.11	0.05	10.7	0.5	1.03
49	0.354	0.003	0.4474	0.028	2	0.014	3.1	1.26	0.04	26.4	0.9	3.3
50	0.290	0.002	0.29	0.01	1	0.01	3.4	1.00	0.04	0.1	0.0	0.0
51	0.343	0.003	0.333	0.01	1	0.01	3.0	0.97	0.03	-2.8	0.1	-0.5
52	0.467	0.004	0.431	0.054	2	0.027	6.3	0.92	0.06	-7.6	0.5	-0.7
53	0.531	0.005	0.594	0.055	2	0.028	4.6	1.12	0.05	11.9	0.6	1.1
54	1.024	0.009	0.592	0.015	1	0.015	2.5	0.58	0.02	-42.2	1.1	-12.5
55	0.308	0.003	0.25	0.02	2	0.01	4.0	0.81	0.03	-18.7	0.8	-2.8

Lab Code	Reference activity A_0 (Bq)	Uncertainty $u_c(A_0)$ (Bq) $k=1$	Reported activity A (Bq)	Reported uncertainty $u_c(A)$ (Bq)	Report -ed k factor	Uncertainty $u_c(A)$ (Bq) $k=1$	Relative unc. $u_c(A)/A$ (%) $k=1$	Ratio (A/A_0)	Standard uncertainty $u_c(A/A_0)$ $k=1$	Percentage Difference $D\%$	Standard uncertainty $u_c(A/A_0)$	E_n
56	0.222	0.002	0.235	0.018	1	0.018	7.7	1.06	0.08	5.7	0.4	0.4
57	0.776	0.007	0.74	0.056	2	0.028	3.8	0.95	0.04	-4.7	0.2	-0.6
58	0.343	0.003	0.34	0.01	1	0.01	2.9	0.99	0.03	-1.0	0.0	-0.2
59	0.414	0.004	0.398	0.022	1	0.022	5.5	0.96	0.05	-3.8	0.2	-0.3
60	0.375	0.003	0.33	0.02	1	0.02	6.1	0.88	0.05	-12.0	0.7	-1.1
61	0.238	0.002	0.2697	0.0394	1	0.0394	14.6	1.13	0.17	13.1	1.9	0.4
62	0.367	0.003	0.342	0.01	1	0.01	2.9	0.93	0.03	-6.7	0.2	-1.2
63	0.394	0.003	0.29988	0.02976	2	0.01488	5.0	0.76	0.04	-24.0	1.2	-3.1
64	0.230	0.002	0.196	0.017	1	0.017	8.7	0.85	0.07	-14.8	1.3	-0.99
65	0.415	0.004	0.3986	0.023	1	0.023	5.8	0.96	0.06	-4.0	0.2	-0.4
66	0.414	0.004	0.375	0.02	1	0.02	5.3	0.90	0.05	-9.5	0.5	-0.97
67	0.476	0.004	0.456	0.023	1	0.023	5.0	0.96	0.05	-4.2	0.2	-0.4

7.1.2 Evaluation

The presence of statistical outliers among the reported results was investigated using the Grubbs' test at a level of significance $\alpha=1\%$ and $\alpha=5\%$, as suggested in ISO/IEC 5725-2 (ISO, 1994). Statistical analysis of the results was carried out for the different working dilutions (E1 to E5) and for the ratio of the measured-to-reference ^{137}Cs activity separately. The reference value of the working dilutions is based on the standardised reference value of the mother solution and the gravimetrically determined dilution factors.

In the case of the ratio values according to the Grubbs' test at 1% three results were indicated as outliers: Laboratories 26, 28 and 54. The outlying values were not discarded, but were included in further evaluations, unless it is declared differently.

Table 6 presents the statistical analysis of the results reported by the participating laboratories per dilution used for spiking the air filters. Moreover, the distribution of the data was tested using the normal probability plot and the frequency histogram. According to both of these graphs presented in Figure 6, the ^{137}Cs data are distributed normally and unimodally. The mean values of the reported values for dilutions E1 and E5 are slightly higher than the reference values and slightly lower for the dilutions E2A, E3 and E4. The mean of the ratio is very close to unity.

Table 6. Statistical analysis of the laboratory reported results for ^{137}Cs , per dilution. In the last column, the ratios of the measured to the reference values are given.

		E1	E2A	E3	E4	E5	Ratio
All reported results	Number of laboratories	2	20	4	9	32	67
	Min (Bq g^{-1})	32.31	2.97	2.69	0.70	0.34	0.58
	Max (Bq g^{-1})	45.24	4.83	4.47	0.99	1.04	2.50
	Median (Bq g^{-1})	38.78	3.64	4.32	0.81	0.44	0.98
	Mean (Bq g^{-1})	38.78	3.72	3.95	0.81	0.46	1.01
	<i>Standard deviation</i> (Bq g^{-1})	9.15	0.51	0.84	0.09	0.12	0.23
Omitting outliers (5%)	Number of outliers	0	0	1	0	2	3
	Mean (Bq g^{-1})	38.78	3.72	4.37	0.81	0.46	0.99
	<i>Standard deviation</i> (Bq g^{-1})	9.15	0.51	0.11	0.09	0.12	0.12
Omitting outliers (1%)	Number of outliers	0	0	1	0	2	3
	Mean (Bq g^{-1})	38.78	3.72	4.37	0.81	0.46	0.99
	<i>Standard deviation</i> (Bq g^{-1})	9.15	0.51	0.11	0.09	0.12	0.12
	Ref value (Bq g^{-1})	36.6	3.91	4.66	0.878	0.416	1
	Expanded unc. (Bq g^{-1})	0.6	0.07	0.08	0.015	0.007	0.01
	Rel. exp. unc. (%)	1.70	1.70	1.70	1.70	1.70	0.01




Fig. 6. Normal probability plot and frequency histogram of the ^{137}Cs results after exclusion of the three outliers (see text). The red curve in the frequency histogram is the normal probability distribution.

In *Figure 7*, the ratio of the ^{137}Cs activity per spiked air filter as measured and reported by the participating laboratory over the individual spiked activity on the filter (JRC reference value) is plotted. The error bars show the expanded uncertainty ($k=2$). The blue dashed lines indicate the $\pm 20\%$ limit from the JRC reference value. For backwards compatibility with previous ILCs, the $\pm 33\%$ limit from the JRC reference value is also indicated by the red dashed lines.




Fig. 7. Ratio of ^{137}Cs activity per spiked air filter as measured by the participating laboratory to the individual spiked activity on the filter (JRC reference value). Blue dashed lines indicate the $\pm 20\%$ limit from the reference value and the red dashed lines the $\pm 33\%$ limit. The error bars show the expanded uncertainty ($k=2$).

In Figure 8, the percentage difference, $D\%$, of the results reported by the participants from the reference activity values are plotted in ascending order.




Fig. 8. Percentage difference of the ^{137}Cs results reported by the participating laboratories from the reference activity values, plotted in ascending order. Blue colour indicates the results within the $\pm 20\%$ range from the reference value and red indicates results outside that range.

The majority of the laboratories reported satisfactory results. From Table 5 and Figures 7 and 8 it can be seen that 11 laboratories of the 67 reported ^{137}Cs results discrepant by more than 20% from the reference value; this represents 16% of the participating laboratories. Only three (3) laboratories reported results which were outside the more tolerant limit of 33% from the reference value. This is just 6% of the participating laboratories.

To take the expanded uncertainty of the reported results and that of the reference values into account in the analysis, a performance test using E_n numbers was applied (ISO, 2005a). The E_n numbers sorted in ascending order are graphically presented in Figure 9.

Under the conditions of this test, 42 (63%) out of the 67 reported results for ^{137}Cs are compatible with the reference value while 25 are not. Among those 25, 12 (18%) laboratories reported results with $1 < |E_n| < 1.5$ and 13 (19%) laboratories reported incompatible results with $|E_n| > 1.5$. Comparing Figure 8 and Figure 9, it is obvious that results with significant deviations from the reference value are scoring badly with E_n numbers as well. However, that comparison also shows that some laboratories with an acceptably small deviation from the reference value are assigned with an incompatible E_n number, as they probably underestimated the uncertainty on their reported value.




Fig. 9. E_n numbers for the ^{137}Cs results, plotted in ascending order. Green color indicates compatible results, yellow indicates warning signal and red indicates action signal.

Another graphical way to compare the results, which underlines the importance of the assigned uncertainties, is the “PomPlot”. The theoretical description of the PomPlot can be found in *Section 6.2.4*. The PomPlot created on the basis of the reported results for ^{137}Cs is presented in *Figure 10*. The many points outside the $|\zeta|=3$, indicate that laboratories underestimated the uncertainties.




Fig. 10. PomPlot of the ^{137}Cs data. Green, blue and red solid lines indicate ζ -scores=1, 2 and 3, respectively.

7.1.3 Parameters possibly influencing the results

Based on the results reported by the participants and the answers provided to the questionnaire, *Table 7* contains the detailed evaluation of possible influencing parameters to the measurement performance and therefore the $D_{\%}$ and E_n number. There is no obvious correlation between laboratories with and without accreditation or certification and higher ratio of compatible results. However, laboratories measuring a large number (>100) of filters per year showed better performance. No strong correlation can be found between the air filter size, spiked activity level and sample preparation and the deviation of the submitted result from the reference value. In the case of laboratories which did not or could not follow their routine procedure, they all scored very well.

Table 7. Possible influencing parameters; statistics and categorisation of the reported results by the participating laboratories.

		Total Nr of Labs	$D_{\%}>20$ (Nr of labs)	Perce- tage (%)	$E_n>1$ (Nr of labs)	Perce- tage (%)
All participants		67	11	16.4	25	37.3
Country	EU	62	10	16.1	22	35.5
	Non-EU	5	1	20	3	60
Accreditation	Accredited (ISO 17025)	30	4	13.3	12	40
	Authorised	13	1	7.7	5	38.5
	Certified (ISO 9000)	2	0	0	0	0
	Certified (ISO 9000) and Accredited (ISO 17025)	10	3	30	3	30
^{137}Cs measurements per year	<25	9	2	22.2	5	55.6
	25-100	21	3	14.3	9	42.9
	>100	37	6	16.2	11	29.7
Air filter size	Small	29	5	17.2	11	37.9
	Medium	11	2	18.2	5	45.5
	Large	27	4	14.8	9	33.3
Spiked activity level (Bq)	0-0.3	15	2	13.3	5	33.3
	0.3-0.4	27	5	18.5	13	48.1
	0.4-0.5	13	2	15.4	3	23.1
	0.5-1	8	0	0	1	12.5
	>1	4	2	50	3	75

		Total Nr of Labs	$D_{\%}>20$ (Nr of labs)	Perce- tage (%)	$E_n>1$ (Nr of labs)	Perce- tage (%)
Routine conditions	Yes	62	11	17.7	25	40.3
	No	5	0	0	0	0
Measurement of plastic bag	Together with air filter	33	5	15.2	12	36.4
	Plastic bag measured separately	26	5	19.2	9	34.6
	Plastic bag not measured	8	1	12.5	4	50
Sample preparation	No sample preparation	30	4	13.3	11	36.7
	Pressing/ compressing	26	3	11.5	9	34.6
	Folding	23	4	17.4	8	34.8
	Packing with blank filters	1	0	0	0	0

7.1.4 Comparison to the 2003 and the 2014 EC ILC

Cesium-137 was a common measurand for all three ILCs conducted in 2003, 2014 and 2016 by the JRC. This makes it possible to make statistics and draw some conclusions. *Table 8* shows the statistics for the three ILCs involving the measurement of ^{137}Cs in air filters. One can observe a slight improvement over the years, especially for the number of results outside the limits for the ratio and equivalently the percentage difference; the percentage of values outside the $\pm 20\%$ range from the reference activity dropped from 38% in 2003, to 20% in 2014 and to 16% in 2016. However, the number of laboratories with $|E_n|>1$ went from 25% in 2003, to 32% in 2014 and to 37% in 2016. The estimation of the uncertainty is a critical point in the case of reporting results and this is reflected in the E_n scores.

More information can be found in *Tables 9a* and *9b*. In the first, the performance of laboratories which participated in two of the ILCs, either in 2003 and 2014 or in 2014 and 2016 is shown, In the second, the performance of laboratories which participated in all three ILCs, is presented.

Thirty laboratories participated in both the 2003 and 2014 air filter EC ILCs and, since one laboratory measured 2 air filters in 2003, reported 31 results in 2003 and 30 results in 2014. Nine laboratories improved their performance since they were outliers and/or in the incompatible E_n number groups in the 2003 exercise and reported compatible values in 2014. Two laboratories, which were in action level regarding E_n number provided values in warning level, show a slightly improvement. However, 7 laboratories which reported compatible values in 2003 are outliers and/or reported results giving incompatible E_n numbers in 2014; this could be an indication that the quality of the results provided by these laboratories is not stable over time, or alternatively, are overconfident on the accuracy of their results.

Fifty seven laboratories participated in the 2014 EC ILC and the 2016 ENV57/MetroERM ILC. The number of laboratories reporting results outside the limits is quasi-identical; 5% of the laboratories reported values outside the $\pm 33\%$ range for the ratio, 16% outside the $\pm 20\%$ range and 12% with $|E_n|>1$.

Table 8. Comparison of the spiked activity levels and the performance of the laboratories in the 2003 and 2014 EC ILC and the 2016 ENV57/MetroERM ILC.

	2003	2014	2016
Number of laboratories	48	76	67
Spiked ^{137}Cs activity levels in Bq			
Minimum	0.015	0.070	0.197
Maximum	0.564	2.34	5.00
Average	0.160	0.359	0.545
Median	0.123	0.195	0.379
Number of laboratory results outside the limits			
Ratio $> \pm 33\%$	6 (13%)	4 (5%)	3 (4.5%)
$D_{\%} > \pm 20\%$	18 (38%)	15 (20%)	11 (16%)
$1 < E_n < 1.5$	3 (6%)	11 (14%)	12 (18%)
$ E_n > 1.5$	9 (19%)	12 (18%)	13 (19%)

Table 9a. Comparison of the spiked activity levels and the performance of the laboratories in the 2003 and 2014 EC ILCs and the 2014 EC ILC and the 2016 ENV57/MetroERM ILC

	Laboratories participated in 2 ILC		Laboratories participated in 2 ILC	
	2003	2014	2014	2016
Number of laboratories	30		57	
Spiked ¹³⁷ Cs activity levels in Bq				
Minimum	0.029	0.083	0.073	0.197
Maximum	0.401	2.310	2.310	5.00
Average	0.162	0.324	0.342	0.516
Median	0.126	0.194	0.178	0.379
Number of laboratory results outside the limits				
Ratio>±33%	2 (7%)	2 (7%)	3 (5%)	3 (5%)
D ₀ >±20%	6 (20%)	4 (13%)	9 (16%)	9 (16%)
1< E _n <1.5	3 (10%)	6 (20%)	7 (12%)	7 (12%)
E _n >1.5	6 (20%)	2 (7%)	7 (12%)	10 (18%)

Table 9b. Comparison of the spiked activity levels and the performance of the laboratories in the 2003 and 2014 EC ILC and the 2016 ENV57/MetroERM ILC.

	Laboratories participated in 3 ILC		
	2003	2014	2016
Number of laboratories	26		
Spiked ¹³⁷ Cs activity levels in Bq			
Minimum	0.055	0.083	0.84
Maximum	0.404	2.31	2.50
Average	0.174	0.369	1.08
Median	0.136	0.200	1.02
Number of laboratory results outside the limits			
Ratio>±33%	2 (8%)	2 (8%)	2 (8%)
D ₀ >±20%	7 (27%)	5 (19%)	5 (19%)
1< E _n <1.5	1 (4%)	6 (23%)	4 (15%)
E _n >1.5	4 (15%)	2 (8%)	3 (12%)

Finally, 26 laboratories participated in all three ILCs. The overall performance is rather stable over the years; reported results outside the $\pm 33\%$ range for the ratio was 8% for all ILCs, results outside the $\pm 20\%$ range was 27% in 2003 and 19% for both 2014 and 2016. The results scoring $|E_n| > 1$ were 19% in 2003, 31% in 2014 and 27% in 2016.

It is worth mentioning that ^{137}Cs is a rather easy to measure radionuclide and the laboratories are scoring consistently well for that nuclide over the years.

7.2 ^{134}Cs

7.2.1 Reported results

Table 10 presents the reference activities of each individual spike air filter, the activity results as reported by the participating laboratories, the ratio of the reported to the reference activity, the percentage difference ($D_{\%}$) of the reported to the reference activity and the E_n number, with their associated standard uncertainties ($k=1$) for the case of ^{134}Cs . The Laboratory Code numbers are the confidential code numbers communicated to the participants and have no correlation with the laboratory codes used during the exercise.

Table 10. ^{134}Cs reference spiked activity, activity result as reported by the participating laboratories, ratio of the reported to the reference activity, percentage difference ($D\%$) of the reported to the reference activity and E_n number, with their associated standard uncertainties ($k=1$).

Lab Code	Reference activity A_0 (Bq)	Uncertainty $u_c(A_0)$ (Bq) $k=1$	Reported activity A (Bq)	Reported uncertainty $u_c(A)$ (Bq)	Reported k factor	Uncertainty $u_c(A)$ (Bq) $k=1$	Relative unc. $u_c(A)/A$ (%) $k=1$	Coinc. Summing Corr.	Ratio (A/A_0)	Standard uncertainty $u_c(A/A_0)$	Percentage Difference $D\%$	Standard uncertainty $u_c(A/A_0)$	E_n
1	0.681	0.007	0.6	0.1	2	0.05	8.3	yes	0.88	0.07	-12.0	1.0	-0.8
2	0.533	0.005	0.525	0.038	1	0.038	7.2	no	0.99	0.07	-1.4	0.1	-0.1
3	0.771	0.008	0.75	0.08	2	0.04	5.3	yes	0.97	0.05	-2.7	0.1	-0.3
4	0.790	0.008	0.75	0.022	1	0.022	2.9	no	0.95	0.03	-5.0	0.2	-0.8
5	0.821	0.008	0.89	0.07	2	0.04	3.9	yes	1.08	0.04	8.5	0.3	0.97
6	1.105	0.011	1.11	0.09	1.65	0.05	4.9	yes	1.00	0.05	0.5	0.0	0.0
7	0.877	0.009	0.73	0.044	1	0.044	6.0	yes	0.83	0.05	-16.7	1.0	-1.6
8	5.243	0.052	5.1	0.6	2	0.3	5.9	no	0.97	0.06	-2.7	0.2	-0.2
9	0.791	0.008	0.814	0.085	2	0.043	5.2	yes	1.03	0.05	2.9	0.2	0.3
10	0.395	0.004	0.344	0.032	2	0.016	4.7	yes	0.87	0.04	-12.9	0.6	-1.55
11	0.431	0.004	0.32	0.05	2	0.025	7.8	yes	0.74	0.06	-25.7	2.0	-2.2
12	3.578	0.036	3.8	0.46	1	0.46	12.1	yes	1.06	0.13	6.2	0.8	0.2
13	0.735	0.007	0.733	0.059	2	0.030	4.0	yes	1.00	0.04	-0.3	0.0	0.0
14	0.739	0.007	0.743	0.021	1	0.021	2.8	yes	1.00	0.03	0.5	0.0	0.1
15	0.608	0.006	0.68	0.05	1	0.05	7.4	no	1.12	0.08	11.9	0.9	0.7
16	0.620	0.006	0.664	0.0582	2	0.0291	4.4	yes	1.07	0.05	7.1	0.3	0.7
17	0.546	0.005	0.55	0.03	1	0.03	5.5	yes	1.01	0.06	0.7	0.0	0.1

Lab Code	Reference activity A_0 (Bq)	Uncertainty $u_c(A_0)$ (Bq) $k=1$	Reported activity A (Bq)	Reported uncertainty $u_c(A)$ (Bq)	Report -ed k factor	Uncertainty $u_c(A)$ (Bq) $k=1$	Relative unc. $u_c(A)/A$ (%) $k=1$	Coinc. Summing Corr.	Ratio (A/A_0)	Standard uncertainty $u_c(A/A_0)$	Percentage Difference $D\%$	Standard uncertainty $u_c(A/A_0)$	E_n
18	0.718	0.007	0.606	0.0303	1	0.0303	5.0	yes	0.84	0.04	-15.5	0.8	-1.8
19	0.831	0.008	0.754	0.042	1	0.042	5.6	yes	0.91	0.05	-9.3	0.5	-0.9
20	0.638	0.006	0.59	0.05	2	0.025	4.2	yes	0.93	0.04	-7.5	0.3	-0.9
21	0.875	0.009	0.77	0.048	1	0.048	6.2	yes	0.88	0.06	-12.0	0.8	-1.1
22	0.752	0.008	0.755	0.057	1	0.057	7.5	no	1.00	0.08	0.4	0.0	0.0
23	0.983	0.010	1	0.1	2	0.05	5.0	yes	1.02	0.05	1.7	0.1	0.2
24	0.757	0.008	0.647	0.034	2	0.017	2.6	yes	0.86	0.02	-14.5	0.4	-2.9
25	0.420	0.004	0.382	0.078	2	0.039	10.2	yes	0.91	0.09	-9.1	0.9	-0.5
26	0.842	0.008	1.53	0.3	2	0.15	9.8	yes	1.82	0.18	81.8	8.1	2.3
27	6.109	0.061	4.35	0.1	2	0.05	1.1	no	0.71	0.01	-28.8	0.4	-11.1
28	0.807	0.008	0.793	0.143	2	0.072	9.0	no	0.98	0.09	-1.8	0.2	-0.1
29	0.780	0.008	0.6731	0.0501	1.65	0.0304	4.5	yes	0.86	0.04	-13.7	0.6	-1.7
30	0.844	0.008	0.746	0.012	2	0.006	0.8	no	0.88	0.01	-11.7	0.1	-4.7
31	0.599	0.006	0.548	0.027	2	0.014	2.5	yes	0.91	0.02	-8.5	0.2	-1.7
32	0.620	0.006	0.6	0.1	2	0.05	8.3	yes	0.97	0.08	-3.2	0.3	-0.2
33	0.736	0.007	0.79	0.08	2	0.04	5.1	yes	1.07	0.06	7.4	0.4	0.7
34	0.898	0.009	0.74	0.073	1	0.073	9.9	yes	0.82	0.08	-17.6	1.7	-1.1
35	0.832	0.008	0.795	0.478	2	0.239	30.1	no	0.96	0.29	-4.4	1.3	-0.1
36	0.854	0.009	0.626	0.04	2	0.02	3.2	no	0.73	0.02	-26.7	0.9	-5.3

Lab Code	Reference activity A_0 (Bq)	Uncertainty $u_c(A_0)$ (Bq) $k=1$	Reported activity A (Bq)	Reported uncertainty $u_c(A)$ (Bq)	Reported k factor	Uncertainty $u_c(A)$ (Bq) $k=1$	Relative unc. $u_c(A)/A$ (%) $k=1$	Coinc. Summing Corr.	Ratio (A/A_0)	Standard uncertainty $u_c(A/A_0)$	Percentage Difference $D\%$	Standard uncertainty $u_c(A/A_0)$	E_n
37	0.720	0.007	0.75	0.06	2	0.03	4.0	yes	1.04	0.04	4.2	0.2	0.5
38	0.822	0.008	0.59	0.04	1	0.04	6.8	no	0.72	0.05	-28.2	1.9	-2.8
39	0.630	0.006	0.51	0.02	1	0.02	3.9	yes	0.81	0.03	-19.1	0.8	-2.9
40	0.790	0.008	0.75	0.05	2	0.025	3.3	no	0.95	0.03	-5.1	0.2	-0.8
41	0.788	0.008	0.694	0.015	2	0.008	1.1	no	0.88	0.01	-12.0	0.2	-4.3
42	0.634	0.006	0.405	0.025	1	0.025	6.2	yes	0.64	0.04	-36.1	2.3	-4.4
43	0.528	0.005	0.6	0.07	2	0.04	5.8	yes	1.14	0.07	13.7	0.8	1.02
44	0.758	0.008	0.49	0.06	2	0.03	6.1	yes	0.65	0.04	-35.4	2.2	-4.3
45	1.452	0.015	1.57	0.06	1	0.06	3.8	yes	1.08	0.04	8.1	0.3	0.95
46	0.744	0.007	0.64	0.1	1	0.1	15.6	yes	0.86	0.13	-14.0	2.2	-0.5
47	0.743	0.007	0.605	0.054	1	0.054	8.9	yes	0.81	0.07	-18.6	1.7	-1.3
48	0.841	0.008	0.9	0.03	1	0.03	3.3	yes	1.07	0.04	7.0	0.2	0.9
49	0.767	0.008	0.8311	0.0291	2	0.0146	1.8	no	1.08	0.02	8.4	0.2	1.9
50	0.628	0.006	0.48	0.022	1	0.022	4.6	no	0.76	0.04	-23.6	1.1	-3.2
51	0.742	0.007	0.598	0.033	1	0.033	5.5	no	0.81	0.05	-19.5	1.1	-2.1
52	0.800	0.008	0.787	0.094	2	0.047	6.0	yes	0.98	0.06	-1.6	0.1	-0.1
53	0.910	0.009	0.757	0.064	2	0.032	4.2	no	0.83	0.04	-16.8	0.7	-2.3
54	1.727	0.017	0.838	0.021	1	0.021	2.5	no	0.49	0.01	-51.5	1.4	-16.3
55	0.667	0.007	0.48	0.04	2	0.02	4.2	no	0.72	0.03	-28.0	1.2	-4.4

Lab Code	Reference activity A_0 (Bq)	Uncertainty $u_c(A_0)$ (Bq) $k=1$	Reported activity A (Bq)	Reported uncertainty $u_c(A)$ (Bq)	Report -ed k factor	Uncertainty $u_c(A)$ (Bq) $k=1$	Relative unc. $u_c(A)/A$ (%) $k=1$	Coinc. Summing Corr.	Ratio (A/A_0)	Standard uncertainty $u_c(A/A_0)$	Percentage Difference $D_{\%}$	Standard uncertainty $u_c(A/A_0)$	E_n
56	0.482	0.005	0.436	0.028	1	0.028	6.4	no	0.90	0.06	-9.5	0.6	-0.8
57	1.330	0.013	1.406	0.341	2	0.171	12.1	no	1.06	0.13	5.7	0.7	0.2
58	0.744	0.007	0.72	0.01	1	0.01	1.4	yes	0.97	0.02	-3.2	0.1	-0.96
59	0.896	0.009	0.83	0.037	1	0.037	4.5	yes	0.93	0.04	-7.4	0.3	-0.9
60	0.813	0.008	0.72	0.02	1	0.02	2.8	yes	0.89	0.03	-11.4	0.3	-2.1
61	0.477	0.005	0.5244	0.05	1	0.05	9.5	no	1.10	0.11	9.9	0.9	0.5
62	0.794	0.008	0.751	0.023	1	0.023	3.1	yes	0.95	0.03	-5.5	0.2	-0.9
63	0.676	0.007	0.49003	0.04499	2	0.022495	4.6	yes	0.72	0.03	-27.5	1.3	-4.0
64	0.460	0.005	0.39	0.03	1	0.03	7.7	yes	0.85	0.07	-15.2	1.2	-1.2
65	0.701	0.007	0.6265	0.034	1	0.034	5.4	yes	0.89	0.05	-10.6	0.6	-1.1
66	0.710	0.007	0.6	0.05	1	0.05	8.3	yes	0.84	0.07	-15.5	1.3	-1.1
67	0.816	0.008	0.738	0.059	1	0.059	8.0	yes	0.90	0.07	-9.5	0.8	-0.7

7.2.2 Evaluation

The presence of statistical outliers among the reported results was investigated using the Grubbs' test at a level of significance $\alpha=1\%$ and $\alpha=5\%$, as suggested in ISO/IEC 5725-2 (ISO, 1994). Statistical analysis of the results was carried out for the different working dilutions (E1 to E5), as explained in *Section 7.1.2*.

In the case of the ratio values according to the Grubbs' test one result was indicated as outlier: Laboratory 26. The outlying value was not discarded, but was included in further evaluations, unless it is declared differently.

Table 11 presents the statistical analysis of the results reported by the participating laboratories per dilution used for spiking the air filters. Moreover, the distribution of the data was tested using the normal probability plot and the frequency histogram. According to both of these graphs presented in *Figure 11*, the ^{134}Cs data are distributed normally and unimodally. The mean values of the reported values for all dilutions are lower than the reference values. The mean value of the ratio is lower than unity by approximately 10%.

Table 11. Statistical analysis of the laboratory reported results for ^{134}Cs , per dilution. In the last column, the ratios of the measured to the reference values are given.

		E1	E2A	E3	E4	E5	Ratio
All reported results	Number of laboratories	2	20	4	9	32	67
	Min (Bq g^{-1})	31.87	4.28	3.81	1.14	0.65	0.49
	Max (Bq g^{-1})	43.54	7.07	8.35	1.93	1.64	1.82
	Median (Bq g^{-1})	37.70	5.69	7.08	1.53	0.89	0.91
	Mean (Bq g^{-1})	37.70	5.81	6.58	1.54	0.89	0.93
	<i>Standard deviation</i> (Bq g^{-1})	8.25	0.69	1.94	0.23	0.17	0.17
Omitting outliers (5%)	Number of outliers	0	0	0	0	1	1
	Mean (Bq g^{-1})	37.70	5.81	6.58	1.54	0.87	0.91
	<i>Standard deviation</i> (Bq g^{-1})	8.25	0.69	1.94	0.23	0.10	0.13
Omitting outliers (1%)	Number of outliers	0	0	0	0	1	1
	Mean (Bq g^{-1})	37.70	5.81	6.58	1.54	0.87	0.91
	<i>Standard deviation</i> (Bq g^{-1})	8.25	0.69	1.94	0.23	0.10	0.13
	Ref value (Bq g^{-1})	44.8	6.69	7.86	1.757	0.902	1
	Expanded unc. (Bq g^{-1})	0.9	0.13	0.16	0.035	0.018	0.01
	Rel. exp. unc. (%)	2.00	2.00	2.00	2.00	2.00	0.01




Fig. 11. Normal probability plot and frequency histogram of the ^{134}Cs results after exclusion of the outlier (see text). The red curve in the frequency histogram is the normal probability distribution.

In *Figure 12*, the ratio of the ^{134}Cs activity per spiked air filter as measured and reported by the participating laboratory over the individual spiked activity on the filter (JRC reference value) is plotted. The error bars show the expanded uncertainty ($k=2$). The blue dashed lines indicate the $\pm 20\%$ limit from the JRC reference value and the red dashed lines the $\pm 33\%$ limit from the JRC reference value.




Fig. 12. Ratio of ^{134}Cs activity per filter as measured by the participating laboratory to the individual spiked activity on the filter (JRC reference value). Blue dashed lines indicate the $\pm 20\%$ limit from the JRC reference value and the red dashed lines the $\pm 33\%$ limit. The error bars show the expanded uncertainty ($k=2$).




Fig. 13. Percentage difference of the ^{134}Cs results reported by the participating laboratories from the reference activity values, plotted in ascending order. Blue color indicates the results within the range $\pm 20\%$ from the reference value and red indicates results outside that range.

The majority of the laboratories obtained satisfactory results. From *Table 10* and *Figures 12 and 13* it can be seen that 11 laboratories of the 67 (16.4%) reported ^{134}Cs results discrepant by more than $\pm 20\%$ from the reference value, which is the individual reference activity spiked on each filter. This represents 16% of the participating laboratories. Only 4 laboratories reported results which were outside the more tolerant limit of $\pm 33\%$ from the reference value. This is just 6.7% of the participating laboratories.




Fig. 14. E_n numbers for the ^{134}Cs results, plotted in ascending order. Green color indicates compatible results, yellow indicates warning signal and red indicates action signal.

To take the expanded uncertainty of the reported results and that of the reference values into account in the analysis, a performance test using E_n numbers was applied (ISO, 2005a). The E_n numbers sorted in ascending order are graphically presented in Figure 14.

Under the conditions of this test, 36 out of the 67 (53.7%) reported results for ^{134}Cs are compatible with the reference value, while 31 are not. Among those 31, 24 (35.8%) laboratories reported results with $1 < |E_n| < 1.5$ and 7 (10.4%) laboratories reported incompatible results with $|E_n| > 1.5$.

Another graphical way to compare the results, which underlines the importance of the assigned uncertainties, is the "PomPlot". For the theoretical description of the PomPlot, refer to Section 6.2.4. The PomPlot created on the basis of the reported results for ^{134}Cs is presented in Figure 15. The many points outside the $|\zeta| = 3$, indicate that laboratories underestimated the uncertainties.




Fig. 15. PomPlot of the ^{134}Cs data. Green, blue and red solid lines indicate ζ -scores=1, 2 and 3, respectively.

7.2.3 Parameters possibly influencing the results: Coincidence summing correction

Cesium-134 is relatively easy to measure as it has two relatively abundant gamma-ray lines at relatively high energies at 604.7 and 795.9 keV. However, it suffers from coincidence summing, which occurs for radionuclides emitting two or more photons in sequence within the resolving time of the spectrometer. This results in recording a sum pulse and the consequent loss of the event from the full energy peak of the individual photons. The probability for the summing effect increases with increasing total efficiency, that is with decreasing source-to-detector distance. It is independent of the count rate and thus the activity of the source.

A correction for the coincidence summing can be calculated either using explicit formulae as those given in the literature (Schima and Hoppes, 1983; Debertain and Helmer, 1988) and in Table 12, or appropriate computer codes or by consulting available tables (Debertain and Schötzig, 1990). In Table 12, the formulae from Schima and Hoppes (1983) for the reciprocal correction factor are given, i.e. the count rates

must be divided by the factor calculated using the formulae to obtain the count rate without coincidence summing. Numbers in parentheses must be replaced by the total efficiency at that energy and numbers in curly brackets must be replaced by the full energy peak efficiency at that energy. As indicative values, those for the two detectors used and described in *Section 3.1* are given as calculated for an active filter of 50 and 90 mm in diameter and measured by placing it on the detector endcap. The correction can be omitted only in the case where the sample is measured relative to a standard of the same nuclide and in the same geometry. The coincidence summing correction factors used and reported by the participants are given in *Table A-11 of Annex 11*.

Table 12. Formulae from Schima and Hoppes (1983) for the reciprocal correction factor and indicative values for Detectors A and B (*Section 3.1*) as calculated for an active filter of 50 and 90 mm in diameter and measured on the detector endcap (see also text).

Energy (keV)	Correction factor (reciprocal)	Detector A		Detector B	
		36% rel. eff.		92% rel. eff.	
	Active area on filter	ø50 mm	ø90 mm	ø50 mm	ø90 mm
563.2	1-0.007(KX)-0.993(604.7)-0.997(801.8)	1.370	1.221	1.663	1.439
569.3	1-0.007(KX)-0.993(604.7)-0.997(795.9)	1.366	1.218	1.658	1.436
604.7	1-0.004(KX)-0.074(563.2)-0.890(795.9)-0.032(1365)-0.075(801.8)-0.160(569.3)	1.201	1.124	1.342	1.237
795.9	1-0.006(KX)-0.993(604.7)-0.179(569.3)	1.213	1.131	1.361	1.248
801.8	1-0.008(KX)-0.814(604.7)-0.180(1168)-0.813(563.2)	1.334	1.203	1.573	1.387

From *Table 13* it can be concluded that the percentage of compatible and non-compatible results is practically independent of the application of the coincidence summing correction. Maybe this is due to the fact that a large number of laboratories used reference spiked filters for the efficiency calibration and possibly spiked with ^{134}Cs as well, so the correction became unnecessary.

Table 13. Distribution of the corrected and non-corrected for coincidence summing results of ^{134}Cs , as compatible and non-compatible according to the $D_{\%}$ and E_n criteria.

	Corrected for coincidence summing		Non-Corrected for coincidence summing	
	45 (67%)		22 (33%)	
	Compatible	Non-compatible	Compatible	Non-compatible
$D_{\%}$	40 (89%)	5 (11%)	16 (73%)	6 (27%)
E_n	25 (56%)	20 (44%)	11 (50%)	11 (50%)

7.3 ^{131}I

7.3.1 Reported results

Table 14 presents the reference activities of each individual spiked air filter, the activity results as reported by the participating laboratories, the ratio of the reported to the reference activity, the percentage difference ($D\%$) of the reported to the reference activity and the E_n number, with their associated standard uncertainties ($k=1$) for the case of ^{131}I . The Laboratory Code numbers are the confidential code numbers communicated to the participants and have no correlation with the laboratory codes used during the exercise.

Table 14. ^{131}I reference spiked activity, activity result as reported by the participating laboratories, ratio of the reported to the reference activity, percentage difference ($D\%$) of the reported to the reference activity and E_n number, with their associated standard uncertainties ($k=1$).

Lab Code	Reference activity A_0 (Bq)	Uncertainty $u_c(A_0)$ (Bq) $k=1$	Reported activity A (Bq)	Reported uncertainty $u_c(A)$ (Bq)	Report -ed k factor	Uncertainty $u_c(A)$ (Bq) $k=1$	Relative unc. $u_c(A)/A$ (%) $k=1$	Coincidence Summing Correction	Decay during the counting interval	Plastic bag measured	Ratio (A/A_0)	Standard uncertainty $u_c(A/A_0)$	Percentage Difference $D\%$	Standard uncertainty $u_c(A/A_0)$	E_n
1	0.822	0.008	0.7	0.1	2	0.05	7.1	yes	yes	Together	0.85	0.06	-14.9	1.1	-1.2
2	0.863	0.009	0.286	0.023	1	0.023	8.0	no	yes	Together	0.33	0.03	-66.9	5.4	-11.7
3	0.930	0.009	0.53	0.05	2	0.025	4.7	no	no	Together	0.57	0.03	-43.0	2.1	-7.5
4	1.266	0.013	0.0005	0.00033	1	0.00033	66.0	no	no	Together	0.0004	0.00	-100.0	66.0	-49.8
5	1.316	0.013	1.15	0.12	2	0.06	5.2	no	yes	Separately	0.87	0.05	-12.6	0.7	-1.4
6	1.772	0.018	1.51	0.15	1.65	0.09	6.0	yes	no	Not measured	0.85	0.05	-14.8	0.9	-1.4
7	1.058	0.011	0.17	0.012	1	0.012	7.1	yes	yes	Together	0.16	0.01	-83.9	6.0	-27.7
8	7.440	0.075	8	0.9	2	0.45	5.6	no	yes	Separately	1.08	0.06	7.5	0.4	0.6
9	1.269	0.013	1.159	0.122	2	0.061	5.3	yes	yes	Separately	0.91	0.05	-8.6	0.5	-0.9
10	0.640	0.006	0.13	0.02	2	0.01	7.7	yes	yes	Together	0.20	0.02	-79.7	6.2	-21.5
11	0.698	0.007	0.59	0.08	2	0.04	6.8	yes	yes	Together	0.85	0.06	-15.5	1.1	-1.3
12	5.032	0.050	1.462	0.088	1	0.088	6.0	yes	yes	Separately	0.29	0.02	-70.9	4.3	-17.6

Lab Code	Reference activity A_0 (Bq)	Uncertainty $u_c(A_0)$ (Bq) $k=1$	Reported activity A (Bq)	Reported uncertainty $u_c(A)$ (Bq)	Reported k factor	Uncertainty $u_c(A)$ (Bq) $k=1$	Relative unc. $u_c(A)/A$ (%) $k=1$	Coincidence Summing Correction	Decay during the counting interval	Plastic bag measured	Ratio (A/A_0)	Standard uncertainty $u_c(A/A_0)$	Percentage Difference $D\%$	Standard uncertainty $u_c(A/A_0)$	E_n
13	1.179	0.012	0.303	0.026	2	0.013	4.3	yes	yes	Separately	0.26	0.01	-74.3	3.3	-24.9
14	1.186	0.012	0.944	0.021	1	0.021	2.2	yes	no	Not measured	0.80	0.02	-20.4	0.5	-5.0
15	0.974	0.010	1.18	0.09	1	0.09	7.6	no	yes	Not measured	1.21	0.09	21.1	1.6	1.1
16	0.994	0.010	0.119	0.0265	2	0.0133	11.1	yes	yes	Separately	0.12	0.01	-88.0	9.8	-26.4
17	0.876	0.009	0.81	0.05	1	0.05	6.2	yes	yes	Separately	0.92	0.06	-7.5	0.5	-0.6
18	0.866	0.009	0.2652	0.02709	1	0.02709	10.2	yes	yes	Together	0.31	0.03	-69.4	7.1	-10.6
19	1.169	0.012	0.229	0.013	1	0.013	5.7	no	yes	Together	0.20	0.01	-80.4	4.6	-26.8
20	1.023	0.010	1.04	0.09	2	0.045	4.3	yes	no	Not measured	1.02	0.05	1.7	0.1	0.2
21	1.056	0.011	0.245	0.017	1	0.017	6.9	yes	yes	Together	0.23	0.02	-76.8	5.4	-20.2
22	1.206	0.012	1.13	0.082	1	0.082	7.3	no	no	Together	0.94	0.07	-6.3	0.5	-0.5
23	1.186	0.012	0.76	0.08	2	0.04	5.3	yes	no	Separately	0.64	0.03	-35.9	1.9	-5.1
24	0.913	0.009	0.057	0.014	2	0.007	12.3	yes	yes	Together	0.06	0.01	-93.8	11.6	-37.1
25	0.681	0.007	0.4	0.081	2	0.041	10.1	yes	yes	Together	0.59	0.06	-41.3	4.2	-3.4

Lab Code	Reference activity A_0 (Bq)	Uncertainty $u_c(A_0)$ (Bq) $k=1$	Reported activity A (Bq)	Reported uncertainty $u_c(A)$ (Bq)	Reported k factor	Uncertainty $u_c(A)$ (Bq) $k=1$	Relative unc. $u_c(A)/A$ (%) $k=1$	Coincidence Summing Correction	Decay during the counting interval	Plastic bag measured	Ratio (A/A_0)	Standard uncertainty $u_c(A/A_0)$	Percentage Difference $D\%$	Standard uncertainty $u_c(A/A_0)$	E_n
26	1.350	0.014	0.43	0.077	2	0.039	9.0	yes	yes	Separately	0.32	0.03	-68.1	6.1	-11.3
27	8.669	0.087	0.9	0.02	2	0.01	1.1	no	yes	Together	0.10	0.00	-89.6	1.3	-44.3
28	1.294	0.013	0.556	0.126	2	0.063	11.3	no	yes	Separately	0.43	0.05	-57.0	6.5	-5.8
29	1.264	0.013	0.4588	0.0567	1.65	0.0344	7.5	yes	yes	Separately	0.36	0.03	-63.7	4.8	-11.0
30	1.354	0.014	1.418	0.032	2	0.016	1.1	no	no	Not measured	1.05	0.02	4.7	0.1	1.52
31	0.971	0.010	0.197	0.012	2	0.006	3.0	no	yes	Separately	0.20	0.01	-79.7	2.6	-33.8
32	0.994	0.010	0.92	0.13	2	0.065	7.1	yes	yes	Separately	0.93	0.07	-7.4	0.5	-0.6
33	1.180	0.012	0.78	0.07	2	0.035	4.5	yes	yes	Together	0.66	0.03	-33.9	1.6	-5.4
34	1.084	0.011	0.272	0.028	1	0.028	10.3	no	no	Together	0.25	0.03	-74.9	7.7	-13.5
35	1.003	0.010	1.19	0.48	2	0.24	20.2	no	yes	Together	1.19	0.24	18.6	3.8	0.4
36	1.031	0.010	0.174	0.021	2	0.011	6.0	no	yes	Not measured	0.17	0.01	-83.1	5.1	-29.1
37	1.155	0.012	1.27	0.08	2	0.04	3.1	yes	yes	Separately	1.10	0.04	10.0	0.3	1.4
38	1.318	0.013	0.93	0.09	1	0.09	9.7	no	no	Separately	0.71	0.07	-29.4	2.9	-2.1
39	0.760	0.008	0.08	0.02	1	0.02	25.0	yes	yes	Together	0.11	0.03	-89.5	22.4	-15.9

Lab Code	Reference activity A_0 (Bq)	Uncertainty $u_c(A_0)$ (Bq) $k=1$	Reported activity A (Bq)	Reported uncertainty $u_c(A)$ (Bq)	Reported k factor	Uncertainty $u_c(A)$ (Bq) $k=1$	Relative unc. $u_c(A)/A$ (%) $k=1$	Coincidence Summing Correction	Decay during the counting interval	Plastic bag measured	Ratio (A/A_0)	Standard uncertainty $u_c(A/A_0)$	Percentage Difference $D\%$	Standard uncertainty $u_c(A/A_0)$	E_n
40	0.954	0.010	0.25	0.04	2	0.02	8.0	no	yes	Together	0.26	0.02	-73.8	5.9	-15.9
41	1.264	0.013	0.993	0.015	2	0.008	0.8	no	no	Not measured	0.79	0.01	-21.5	0.3	-9.2
42	0.764	0.008	0.088	0.007	1	0.007	8.0	yes	yes	Together	0.12	0.01	-88.5	7.1	-32.6
43	0.846	0.008	0.26	0.08	2	0.04	15.4	yes	yes	Separately	0.31	0.05	-69.3	10.7	-7.2
44	1.228	0.012	0.23	0.07	2	0.035	15.2	yes	yes	Separately	0.19	0.03	-81.3	12.4	-13.4
45	2.329	0.023	1.87	0.06	1	0.06	3.2	yes	yes	Separately	0.80	0.03	-19.7	0.7	-3.6
46	1.193	0.012	0.96	0.14	1	0.14	14.6	yes	yes	Separately	0.80	0.12	-19.5	2.9	-0.8
47	0.896	0.009	0.104	0.008	1	0.008	7.7	yes	yes	Together	0.12	0.01	-88.4	6.9	-32.9
48	1.349	0.014	1.41	0.05	1	0.05	3.5	yes	yes	Separately	1.04	0.04	4.5	0.2	0.6
49	1.230	0.012	0.6205	0.0323	2	0.0162	2.6	no	no	Together	0.50	0.01	-49.6	1.4	-15.0
50	1.007	0.010	0.81	0.044	1	0.044	5.4	no	no	Separately	0.80	0.04	-19.6	1.1	-2.2
51	1.191	0.012	0.993	0.035	1	0.035	3.5	no	yes	Together	0.83	0.03	-16.6	0.6	-2.7
52	0.965	0.010	0.072	0.013	2	0.007	9.0	yes	yes	Together	0.07	0.01	-92.5	8.4	-38.3
53	1.098	0.011	0.047	0.01	2	0.005	10.6	no	no	Not measured	0.04	0.00	-95.7	10.2	-43.4

Lab Code	Reference activity A_0 (Bq)	Uncertainty $u_c(A_0)$ (Bq) $k=1$	Reported activity A (Bq)	Reported uncertainty $u_c(A)$ (Bq)	Reported k factor	Uncertainty $u_c(A)$ (Bq) $k=1$	Relative unc. $u_c(A)/A$ (%) $k=1$	Coincidence Summing Correction	Decay during the counting interval	Plastic bag measured	Ratio (A/A_0)	Standard uncertainty $u_c(A/A_0)$	Percentage Difference $D\%$	Standard uncertainty $u_c(A/A_0)$	E_n
54	2.429	0.024	0.344	0.02	1	0.02	5.8	no	yes	Together	0.14	0.01	-85.8	5.1	-33.1
55	1.069	0.011	0.76	0.06	2	0.03	3.9	no	no	Together	0.71	0.03	-28.9	1.2	-4.9
56	0.773	0.008	1.25	0.083	1	0.083	6.6	no	yes	Separately	1.62	0.11	61.8	4.1	2.9
57	1.605	0.016	0.228	0	2	0	0.0	no	no	Together	0.14	0.00	-85.8	0.9	-42.7
58	1.193	0.012	0.93	0.02	1	0.02	2.2	yes	yes	Separately	0.78	0.02	-22.1	0.5	-5.6
59	1.437	0.014	0.295	0.02	1	0.02	6.8	yes	no	Separately	0.21	0.01	-79.5	5.4	-23.2
60	1.303	0.013	1.1	0.03	1	0.03	2.7	yes	no	Together	0.84	0.02	-15.6	0.5	-3.1
61	0.773	0.008	0.6817	0.04	1	0.04	5.9	no	yes	Separately	0.88	0.05	-11.8	0.7	-1.1
62	1.274	0.013	0.391	0.016	1	0.016	4.1	yes	yes	Separately	0.31	0.01	-69.3	2.9	-21.6
63	0.815	0.008	0.088108	0.050394	2	0.02519 ₇	28.6	yes	no	Together	0.11	0.03	-89.2	25.5	-13.7
64	0.745	0.007	0.14	0.011	1	0.011	7.9	yes	yes	Together	0.19	0.01	-81.2	6.4	-22.8
65	0.985	0.010	0.1707	0.0104	1	0.0104	6.1	yes	yes	Together	0.17	0.01	-82.7	5.1	-28.4
66	0.857	0.009	0.07	0.007	1	0.007	10.0	yes	yes	Together	0.08	0.01	-91.8	9.2	-35.5
67	0.984	0.010	0.246	0.025	1	0.025	10.2	yes	yes	Together	0.25	0.03	-75.0	7.7	-13.7

7.3.2 Evaluation

The presence of statistical outliers among the reported results was investigated using the Grubbs' test at a level of significance $\alpha=1\%$ and $\alpha=5\%$, as suggested in ISO/IEC 5725-2 (ISO, 1994). Statistical analysis of the results was carried out for the different working dilutions (E1 to E5) as explained in Section 7.1.2.

In the case of the ratio values according to the Grubbs' test two results were indicated as outliers: Laboratories 1 and 35. The outlying values were not discarded, but were included in further evaluations, unless it is declared differently.

Table 15 presents the statistical analysis of the results reported by the participating laboratories per dilution used for spiking the air filters. Moreover, the distribution of the data was tested using the normal probability plot and the frequency histogram. According to both of these graphs presented in Figure 16, the ^{131}I data are not distributed normally and unimodally. The mean of the reported values for all dilutions is lower than the reference values and the mean of the ratio is only 62 to 66% of the unity.

Table 15. Statistical analysis of the laboratory reported results for ^{131}I , per dilution. In the last column, the ratios of the measured to the reference values are given.

		E1	E2A	E3	E4	E5	Ratio
All reported results	Number of laboratories	2	20	4	9	32	67
	Min (Bq g^{-1})	6.59	0.35	1.57	0.53	0.00	0.00
	Max (Bq g^{-1})	68.29	9.58	3.21	2.51	2.34	2.94
	Median (Bq g^{-1})	37.44	1.33	2.04	0.94	1.16	0.66
	Mean (Bq g^{-1})	37.44	2.31	2.21	1.20	1.06	0.66
	<i>Standard deviation</i> (Bq g^{-1})	43.63	2.42	0.71	0.80	0.50	0.44
Omitting outliers (5%)	Number of outliers	0	2	0	0	0	2
	Mean (Bq g^{-1})	37.44	1.66	2.43	1.20	1.06	0.62
	<i>Standard deviation</i> (Bq g^{-1})	43.63	1.33	0.69	0.80	0.50	0.32
Omitting outliers (1%)	Number of outliers	0	4	0	0	0	4
	Mean (Bq g^{-1})	37.44	1.25	2.43	1.20	1.06	0.62
	<i>Standard deviation</i> (Bq g^{-1})	43.63	0.66	0.69	0.80	0.50	0.32
	Ref value (Bq g^{-1})	63.5	8.07	11.05	2.846	1.446	1
	Expanded unc. (Bq g^{-1})	0.3	0.03	0.05	0.012	0.006	0.01
	Rel. exp. unc. (%)	0.43	0.43	0.43	0.43	0.43	0.01




Fig. 16. Normal probability plot and frequency histogram of the ^{131}I results. The red curve in the frequency histogram is the normal probability distribution.

In Figure 17, the ratio of the ^{131}I activity per spiked air filter as measured and reported by the participating laboratory over the individual spiked activity on the filter (JRC reference value) is plotted. The error bars show the expanded uncertainty ($k=2$). The blue dashed lines indicate the $\pm 20\%$ limit from the JRC reference value and the red dashed lines the $\pm 33\%$ limit from the JRC reference value.




Fig. 17. Ratio of ^{131}I activity per spiked air filter as measured by the participating laboratory to the individual spiked activity on the filter (JRC reference value). Blue dashed lines indicate the $\pm 20\%$ limit from the JRC reference value and the red dashed lines the $\pm 33\%$ limit. The error bars show the expanded uncertainty ($k=2$).




Fig. 18. Percentage difference of the ^{131}I results reported by the participating laboratories from the reference activity values, plotted in ascending order. Blue color indicates the results within the range $\pm 20\%$ from the reference value and red indicates results outside that range.

The majority of the laboratories reported results below the reference activity value. From Table 14 and Figures 17 and 18 it can be seen that 47 of the 67 laboratories reported ^{131}I results discrepant by more than $\pm 20\%$ from the reference value, which is the individual reference activity spiked on each filter. This represents 70% of the participating laboratories. Forty one laboratories reported results which were outside the more tolerant limit of $\pm 33\%$ from the reference value. This is just 61% of the participating laboratories.




Fig. 19. E_n numbers for the ^{131}I results plotted in ascending order. Green color indicates compatible results, yellow indicates warning signal and red indicates action signal.

To take the expanded uncertainty of the reported results and that of the reference values into account in the analysis, a performance test using E_n numbers was applied (ISO, 2005a). The E_n numbers sorted in ascending order are graphically presented in Figure 19.




Fig. 20a. PomPlot of the ^{131}I data. Green, blue and red solid lines indicate ζ -scores=1, 2 and 3, respectively.




Fig. 20b. PomPlot of the ^{131}I data within the $\pm 20\%$ range from their respective reference values. Green, blue and red solid lines indicate ζ -scores=1, 2 and 3, respectively.

Under the conditions of this test, 9 (13%) out of the 67 reported results for ^{131}I are compatible with the reference value, while 58 are not. Among those 58, 7 (10%) laboratories reported incompatible results with $1 < |E_n| < 1.5$ and 51 (76%) reported incompatible results with $|E_n| > 1.5$.

The PomPlot (see theoretical description in *Section 6.2.4*) created on the basis of the reported results for ^{131}I is presented in *Figure 20a*. Because of most results being far from their respective reference values, the MAD value is abnormally high and this results in the peculiar PomPlot of *Figure 20a*. If only results within the $\pm 20\%$ range from their respective reference values are taken into account, the calculated PomPlot is that depicted in *Figure 20b*. The points outside the $|\zeta|=3$, indicate those laboratories that probably underestimated the uncertainties on their reported results for ^{131}I .

7.3.3 Parameters possibly influencing the results

Although measuring ^{131}I does not present any problem, as its main gamma-ray lines are emitted at relatively high energies, at 284.3, 364.5 and 637.0 keV with emission probabilities of 6.14, 81.2 and 7.12%, respectively, there are some issues to be brought to the attention of the analyst.

7.3.3.1 Decay during measurement

Because of the relatively short half-life of ^{131}I (8.0233 (19) d), besides the usual decay correction applied for reporting the results on the reference date of 1 March 2016 0:00 UTC, a second correction maybe necessary for longer acquisition periods to compensate for the decay during the measurement period (Debertin and Helmer, 1988). As it is shown in *Table 16*, this correction is negligible for an acquisition time of 1 hour, it becomes 1% for a 6-hour acquisition and close to 9% for a 2-day acquisition. An alternative correction is to take in the decay correction calculations as start time of the measurement the mid-point of the measurement time.

Table 16. Correction factors for the decay during measurement of ^{131}I .

Counting interval	Correction factor
1 h	1.002
2 h	1.004
6 h	1.011
12 h	1.022
24 h	1.044
2 d	1.089
3 d	1.135
5 d	1.231
7 d	1.333

Table 17 shows the distribution of the corrected and non-corrected for decay during measurement results of ^{131}I , as reported by the participants, as compatible and non-compatible according to the $D_{\%}$ and E_n criteria. Nineteen laboratories did not correct for decay during measurement, but the compatible and non-compatible distribution of the

reported results does not differ from the distribution of those reported by laboratories which corrected for decay during measurement.

Table 17. Distribution of the corrected and non-corrected for decay during measurement results of ^{131}I , as reported by the participants, as compatible and non-compatible according to the $D_{\%}$ and E_n criteria.

	Corrected for decay during measurement		Non-Corrected for decay during measurement	
	48 (72%)		19 (28%)	
	Compatible	Non-compatible	Compatible	Non-compatible
$D_{\%}$	14 (29%)	34 (71%)	6 (32%)	13 (68%)
E_n	7 (15%)	41 (85%)	2 (10%)	17 (90%)

7.3.3.2 Coincidence summing correction

The three principal gamma-ray lines of ^{131}I at 284.3, 364.5 and 637.0 keV suffer from coincidence summing (for explanation refer to *Section 7.2.3*). However, the correction factors are close to unity, with the exception of the 284.3 keV line in combination with high efficiency detectors. The formulae from Schima and Hoppes (1983) for the reciprocal correction factor and indicative values for Detectors A and B (*Section 3.1*) as calculated for an active filter of 50 and 90 mm in diameter and measured on the detector endcap are given in *Table 18*.

Table 18. Formulae from Schima and Hoppes (1983) for the reciprocal correction factor and indicative values for Detectors A and B (*Section 3.1*) as calculated for an active filter of 50 and 90 mm in diameter and measured on the detector endcap (see also *Section 7.2.3*).

Energy (keV)	Correction factor (reciprocal)	Detector A		Detector B	
		36% rel. eff.		92% rel. eff.	
	Active area on filter	ø50 mm	ø90 mm	ø50 mm	ø90 mm
284.3	$1 - 0.467(KX) - 0.389(80.2)$	1.044	1.019	1.276	1.169
364.5	$1 + 0.029\{80.2\}\{284.3\}/\{364.5\}$	0.997	0.999	0.989	0.995
637	$1 + 0.007\{364.5\}\{272.0\}/\{637.0\}$	0.999	0.999	0.998	0.999

7.3.3.3 Radioactivity transferred to the protective plastic bag

During the exercise it became obvious that there was transfer of a variable quantity of the ^{131}I activity from the spiked air filter to the containing (inner) protective plastic bag. More affected were the small filters being directly in contact with the plastic bag. The larger filters were folded with the activity inwards and so protected. This activity transfer concerns only ^{131}I ; the ^{137}Cs and ^{134}Cs were not affected.

At the JRC we measured originally the quality control spiked air filters in their plastic bags. To check for the activity transfer, we measured two spiked air filters (IRMM-03 and IRMM-06) alone and their corresponding plastic bags separately. The findings are

given in *Table 19*; a fraction of 3 to 6% of the ^{131}I activity had been transferred to the plastic bag.

Table 19. Count rate results for the 364 keV line of ^{131}I from measuring filters IRMM-03 and IRMM-06 and their plastic bags. The percent ratio of the count rate obtained from the plastic bag to that from the filter is given in the last column.

Item	Detector	Count rate (364 keV)	Ratio %
Filter_IRMM-06 in original bag	A	0.0712	
Filter_IRMM-06 Plastic bag only		0.0023	3.2
Filter_IRMM-06 in new bag		0.0700	
SUM		0.0723	
Filter_IRMM-03 in original bag	B	0.0538	
Filter_IRMM-03 Plastic bag only		0.0046	5.6
Filter_IRMM-03 in new bag		0.0480	
SUM		0.0526	
Filter_IRMM-06 in original bag	B	0.1711	
Filter_IRMM-06 Plastic bag only		0.0058	3.5
Filter_IRMM-06 in new bag		0.1671	
SUM		0.1729	

Although it was originally recommended to the participants to measure the plastic bag as well, they were warned in the course of the exercise (see communication in *Annex 2E*) to measure both the filter and the plastic bag (either together or separately) and add the measured activities. It is evident that the counting efficiency of the plastic bag is different than that of the filter, but even with an estimation of the counting efficiency, adding the transferred activity would be an improvement on the ^{131}I result. *Table 20* gives the distribution of the results for ^{131}I corrected and non-corrected for the transferred activity to the protective plastic bag, as reported by the participants, as compatible and non-compatible according to the $D_{\%}$ and E_n criteria. Eight laboratories reported that they did not measure the protective plastic bags, but there is no significant difference in the distribution of the reported results as compatible or non-compatible compared to the laboratories that they carried out the appropriate corrections.

Table 20. Distribution of the results for ^{131}I corrected and non-corrected for the transferred activity to the protective plastic bag, as reported by the participants, as compatible and non-compatible according to the $D_{\%}$ and E_n criteria.

	Plastic bag measured (together with filter or separately)		Plastic bag not measured	
	59 (88%)		8 (12%)	
	Compatible	Non- compatible	Compatible	Non- compatible
$D_{\%}$	17 (29%)	42 (71%)	3 (37.5%)	5 (62.5%)
E_n	8 (14%)	51 (86%)	1 (12.5%)	7 (87.5%)

7.3.4 Conclusions

By measuring the quality control spiked air filters and point sources it became obvious that although the results of the activity were as expected for ^{137}Cs and ^{134}Cs , they were systematically low and with significant spread for ^{131}I . For logistics reasons, the quality control sources were measured at the JRC after dispatching the spiked air filters to the participants. It is worth noticing that both filters and point sources exhibit equally lower than expected activities for ^{131}I . The first occurrence was observed with point source PS014. Point sources and liquid scintillation sources made before that gave excellent results with deviations from the expected values of less than 1%.

The cause of this decrease of the ^{131}I activity is not clear. All point sources and spiked air filters were prepared gravimetrically in the same way and by the same operator under controlled conditions of temperature, pressure and humidity. The drying of the sources was done overnight and not in a forced way (e.g. heating). Volatility of iodine could be the culprit; a solution of $25 \text{ mg g}^{-1} \text{ KI}$ and $25 \text{ mg g}^{-1} \text{ Na}_2\text{S}_2\text{O}_3$ was used to keep the iodine in solution, but it might be that the salt concentration was not adequate.

In an attempt to explain the discrepancies, the ratio of the measured to the reference activity values of ^{131}I was plotted in a number of ways, of which two are shown in *Figures 21 and 22*. In *Figure 21* the ratio was plotted per day of production of the spiked air filters and with ascending of spiked activity (per day). Notice that smaller filters were spiked first and then the larger ones. Dilution E5 shows better performance, but being the most dilute was used for the larger filters. No effect can be observed in terms of activity levels spiked or mass of the spiked solution.




Fig. 21. Ratio of ^{131}I activity per spiked air filter as measured by the participating laboratory over the individual spiked activity on the filter (JRC reference value) plotted per day of preparation and with ascending spiked activity (per day). Blue dashed lines indicate the $\pm 20\%$ limit from the reference value and red dashed lines the $\pm 33\%$ limit. The error bars show the expanded uncertainty ($k=2$). The laboratory numbers are arbitrary.

In *Figure 22* the ratio was plotted per dilution used for spiking, with indication of the filter size in color. It can be observed that larger filters behaved better in retaining ^{131}I spiked on them. This can be due to the material of the filter itself; filters with fiberglass composition behaved better than those from cellulose. This is not absolute though and does not explain why the point sources prepared by depositing aliquots of the radioactive solution onto a plastic foil, exhibit the same behaviour.

Because the mechanism of depositing the activity by spiking is different than that by collecting air particulates, we estimate that such effect should not be observed in air filters loaded by exposing them to an aerosol.




Fig. 22. Ratio of ^{131}I activity per spiked air filter as measured by the participating laboratory over the individual spiked activity on the filter (JRC reference value) plotted per dilution used for spiking and with indication (in color) of the filter size. Blue dashed lines indicate the $\pm 20\%$ limit from the reference value and red dashed lines the $\pm 33\%$ limit. The error bars show the expanded uncertainty ($k=2$). The laboratory numbers are arbitrary.

The spiked air filters could also lose some ^{131}I activity by being exposed to heat. Some participants mentioned in private communication (but not reported in the questionnaire) heating the filters prior to measuring in order to dry them, as routine in their procedure.

One way to compensate for the loss of ^{131}I and reflect the spread of the effective spiking results in the calculation of the E_n number could be to artificially increase the uncertainty on the reference value to 22.5% (ne half of the mean discrepancy of the quality control sources measured activity from their corresponding expected activity values).

Figure 23 shows the E_n numbers for the ^{131}I results plotted in ascending order, when the uncertainty on the reference value is increased to 22.5%. In such case, 30 (45%) laboratories score $|E_n| \leq 1$, 6 (9%) laboratories score $1 < |E_n| < 1.5$ and 31 (46%) score $|E_n| > 1.5$. The laboratories in the three E_n categories can be identified from *Figure 23*.




Fig. 23. E_n numbers for the ^{131}I reported results plotted in ascending order, when the uncertainty on the reference value is increased to 22.5%. Green color indicates compatible results, yellow indicates warning signal and red indicates action signal.

8. Conclusions

The 2016 ENV57/MetroERM Interlaboratory Comparison was organized and conducted by the JRC in the frame of the ENV57/MetroERM "Metrology for radiological early warning networks in Europe", a European Metrology Research Program (EMRP). All 67 participating laboratories reported valid activity results for all of the three radionuclides of interest, namely ^{137}Cs , ^{134}Cs and ^{131}I . Figure 24 is a color coded cumulative plot of the ratio of the measured to the reference activity values to judge the performance of the laboratories for all three radionuclides simultaneously.

Table 21 lists the percentage difference, $D_{\%}$, and the E_n number for each of the three radionuclides, ^{137}Cs , ^{134}Cs and ^{131}I . Values of $D_{\%}$ outside the $\pm 20\%$ limit from the reference value are in yellow. Absolute values of E_n between 1 and 1.5 are also in yellow and those above 1.5 are in red.

The compatibility criterion is satisfied when both the percent difference, $D_{\%}$, being within the $\pm 20\%$ limit from the reference value and the absolute E_n being smaller than unity. The compatibility for each laboratory and for the three radionuclides is given in the last three columns of Table 21.

In Table 22 the number of results (and the percentage over all results in parentheses) satisfying the $< \pm 20\%$ and $< \pm 33\%$ discrepancy from the reference values is given. The performance of the laboratories is similar for measuring ^{137}Cs and ^{134}Cs with 84% of the laboratories reporting values within the $\pm 20\%$ limit from the reference values for both nuclides and 63% and 54%, respectively, scoring $|E_n| \leq 1$. Equally, 61% and 54% of the laboratories reported compatible results, satisfying both criteria. However, only 40% of the laboratories reported compatible results for both ^{137}Cs and ^{134}Cs (Table 23). This is strange as one would expect that laboratories measuring correctly one of the two nuclides should be able to measure correctly the other as well.




Fig. 24. Ratio of ^{137}Cs (in blue), ^{134}Cs (in red) and ^{131}I (in green) activity per spiked air filter as measured by the participating laboratory to the individual spiked activity on the filter (JRC reference value). Blue dashed lines indicate the $\pm 20\%$ limit from the reference value and the red dashed lines the $\pm 33\%$ limit. The error bars show the expanded uncertainty ($k=2$).

For ^{131}I only 30% of the results were within the $\pm 20\%$ limit from the reference values and 13% obtained $|E_n| \leq 1$. At least, this poor performance is due partly to the quality of the spiked air filters, in what concerns ^{131}I (refer to *Section 7.3.3*). It was the first attempt to include ^{131}I in an ILC organised by JRC; it is a radionuclide which could be released in case of a nuclear incident or accident and therefore of interest to the laboratories monitoring the environment. Because of its properties and its rather short half-life it is not very often included as measurand in ILCs or PTs.

Nevertheless, 30% of the participating laboratories reported ^{131}I activities within the $\pm 20\%$ limit from the reference values and 39% the more tolerant $\pm 33\%$ range. In those circumstances, it is not surprising that only 13% of the laboratories reported compatible results for ^{131}I , 13% for both ^{134}Cs and ^{131}I and only 7.5%, i.e. 5 laboratories, for ^{137}Cs and ^{131}I or for all three radionuclides simultaneously.

The lower success rate for the E_n number, compared to the percentage difference, $D_\%$, reflects the fact that some laboratories underestimated the uncertainties on their results and so they were penalized in the E_n number test. The latter takes into account the uncertainty and although the results pass the $D_\%$ test, they fail the E_n test. In *Table A-12* of *Annex 12* the detailed uncertainty budgets, as reported by the participating laboratories, are given for each of the three radionuclides. Some laboratories did not report the individual components of their uncertainty budgets. In column 13 the combined standard uncertainty as reported by the laboratories is given, in column 15 the calculated from the components combined standard uncertainty is given and in column 16 the combined standard uncertainty as deduced from the reported results. Some discrepancies were observed and the concerned laboratories should improve the way they calculate their uncertainty budgets by following the "GUM approach" and establish a complete evaluation of uncertainty sources. A workshop was organised by JRC on 7 and 8 April, 2016 (*Annex 2H*) to bring together the participants of this ILC to discuss all aspects of the ILC and provide ideas and suggestions to improve the performance of the laboratories.

Table 21. Performance of the laboratories: percentage difference, $D_{\%}$, and E_n number for each of the three radionuclides, ^{137}Cs , ^{134}Cs and ^{131}I . Values of $D_{\%}$ outside the $\pm 20\%$ limit from the reference value are in yellow. Absolute values of E_n between 1 and 1.5 are also in yellow and those above 1.5 are in red.

	^{137}Cs		^{134}Cs		^{131}I		Compatibility		
Lab Code	$D_{\%}$	E_n	$D_{\%}$	E_n	$D_{\%}$	E_n	^{137}Cs	^{134}Cs	^{131}I
1	-14.5	-0.6	-12.0	-0.8	-14.9	-1.2	yes	yes	no
2	-7.2	-0.5	-1.4	-0.1	-66.9	-11.7	yes	yes	no
3	2.3	0.2	-2.7	-0.3	-43.0	-7.5	yes	yes	no
4	-14.9	-0.9	-5.0	-0.8	-100.0	-49.8	yes	yes	no
5	10.9	0.8	8.5	0.96	-12.6	-1.4	yes	yes	no
6	2.2	0.2	0.5	0.0	-14.8	-1.4	yes	yes	no
7	-12.6	-0.8	-16.7	-1.6	-83.9	-27.7	yes	no	no
8	23.6	1.7	-2.7	-0.2	7.5	0.6	no	yes	yes
9	5.2	0.4	2.9	0.3	-8.6	-0.9	yes	yes	yes
10	-16.4	-1.9	-12.9	-1.55	-79.7	-21.5	no	no	no
11	2.2	0.1	-25.7	-2.2	-15.5	-1.3	yes	no	no
12	-8.5	-0.6	6.2	0.2	-70.9	-17.6	yes	yes	no
13	3.1	0.4	-0.3	0.0	-74.3	-24.9	yes	yes	no
14	-1.9	-0.3	0.5	0.1	-20.4	-5.0	yes	yes	no
15	24.8	1.2	11.9	0.7	21.1	1.1	no	yes	no
16	23.7	1.51	7.1	0.7	-88.0	-26.4	no	yes	no
17	-0.8	0.0	0.7	0.1	-7.5	-0.6	yes	yes	yes
18	-16.3	-1.1	-15.5	-1.8	-69.4	-10.6	no	no	no
19	-5.8	-0.5	-9.3	-0.9	-80.4	-26.8	yes	yes	no
20	12.1	1.2	-7.5	-0.9	1.7	0.2	no	yes	yes
21	-10.1	-0.98	-12.0	-1.1	-76.8	-20.2	yes	no	no
22	7.2	0.4	0.4	0.0	-6.3	-0.5	yes	yes	yes
23	4.5	0.6	1.7	0.2	-35.9	-5.1	yes	yes	no
24	-4.9	-0.6	-14.5	-2.9	-93.8	-37.1	yes	no	no
25	-7.6	-0.4	-9.1	-0.5	-41.3	-3.4	yes	yes	no
26	149.7	3.1	81.8	2.3	-68.1	-11.3	no	no	no
27	-11.7	-4.5	-28.8	-11.1	-89.6	-44.3	no	no	no
28	43.4	1.2	-1.8	-0.1	-57.0	-5.8	no	yes	no
29	-13.2	-1.4	-13.7	-1.7	-63.7	-11.0	no	no	no
30	19.1	5.1	-11.7	-4.7	4.7	1.52	no	no	no
31	-7.5	-1.3	-8.5	-1.7	-79.7	-33.8	no	no	no
32	4.9	0.3	-3.2	-0.2	-7.4	-0.6	yes	yes	yes
33	9.0	0.8	7.4	0.7	-33.9	-5.4	yes	yes	no

	¹³⁷ Cs		¹³⁴ Cs		¹³¹ I		Compatibility		
Lab Code	D%	E _n	D%	E _n	D%	E _n	¹³⁷ Cs	¹³⁴ Cs	¹³¹ I
34	-6.2	-0.3	-17.6	-1.1	-74.9	-13.5	yes	no	no
35	22.8	0.5	-4.4	-0.1	18.6	0.4	no	yes	yes
36	4.5	0.6	-26.7	-5.3	-83.1	-29.1	yes	no	no
37	5.4	0.6	4.2	0.5	10.0	1.4	yes	yes	no
38	0.2	0.0	-28.2	-2.8	-29.4	-2.1	yes	no	no
39	-18.4	-1.1	-19.1	-2.9	-89.5	-15.9	no	no	no
40	23.6	2.1	-5.1	-0.8	-73.8	-15.9	no	yes	no
41	2.5	0.8	-12.0	-4.3	-21.5	-9.2	yes	no	no
42	-19.9	-2.1	-36.1	-4.4	-88.5	-32.6	no	no	no
43	6.8	0.4	13.7	1.02	-69.3	-7.2	yes	no	no
44	-20.8	-1.6	-35.4	-4.3	-81.3	-13.4	no	no	no
45	5.2	0.7	8.1	0.95	-19.7	-3.6	yes	yes	no
46	-12.6	-0.4	-14.0	-0.5	-19.5	-0.8	yes	yes	yes
47	-10.1	-1.02	-18.6	-1.3	-88.4	-32.9	no	no	no
48	10.7	1.03	7.0	0.9	4.5	0.6	no	yes	yes
49	26.4	3.3	8.4	1.9	-49.6	-15.0	no	no	no
50	0.1	0.0	-23.6	-3.2	-19.6	-2.2	yes	no	no
51	-2.8	-0.5	-19.5	-2.1	-16.6	-2.7	yes	no	no
52	-7.6	-0.7	-1.6	-0.1	-92.5	-38.3	yes	yes	no
53	11.9	1.1	-16.8	-2.3	-95.7	-43.4	no	no	no
54	-42.2	-12.5	-51.5	-16.3	-85.8	-33.1	no	no	no
55	-18.7	-2.8	-28.0	-4.4	-28.9	-4.9	no	no	no
56	5.7	0.4	-9.5	-0.8	61.8	2.9	yes	yes	no
57	-4.7	-0.6	5.7	0.2	-85.8	-42.7	yes	yes	no
58	-1.0	-0.2	-3.2	-0.96	-22.1	-5.6	yes	yes	no
59	-3.8	-0.3	-7.4	-0.9	-79.5	-23.2	yes	yes	no
60	-12.0	-1.1	-11.4	-2.1	-15.6	-3.1	no	no	no
61	13.1	0.4	9.9	0.5	-11.8	-1.1	yes	yes	no
62	-6.7	-1.2	-5.5	-0.9	-69.3	-21.6	no	yes	no
63	-24.0	-3.1	-27.5	-4.0	-89.2	-13.7	no	no	no
64	-14.8	-0.99	-15.2	-1.2	-81.2	-22.8	yes	no	no
65	-4.0	-0.4	-10.6	-1.1	-82.7	-28.4	yes	no	no
66	-9.5	-0.97	-15.5	-1.1	-91.8	-35.5	yes	no	no
67	-4.2	-0.4	-9.5	-0.7	-75.0	-13.7	yes	yes	no

From the comparison of the statistics of the reported results for ^{137}Cs , it is concluded that the performance of the laboratories analysing this nuclide is rather stable. The results for ^{134}Cs were equally satisfactory, albeit not necessarily from the same laboratories reporting compatible results for ^{137}Cs . The observed negative bias on the reported results compared to the reference values, might be partly attributed to a non-adequate summing correction applied by some laboratories. However, the statistical results show that the coincidence summing correction was taken into account rather successfully, either by correcting or by using suitable sources for the counting efficiency calibration, as there is no significant difference in the distribution of compatible and non-compatible results (Table 13).

In the case of ^{131}I , the transfer of activity to the protective plastic bag was a minor problem and most of the participants measured the plastic bag and added the activity to the activity measured on the spiked air filter.

Table 22. Cumulative table showing the performance of the laboratories for the three radionuclides, ^{137}Cs , ^{134}Cs and ^{131}I . The numbers give the number of laboratories with the percent fraction in parentheses.

	Number of laboratories		
Nuclides	^{137}Cs	^{134}Cs	^{131}I
$ D\% \leq 20$	56 (83.6%)	56 (83.6%)	20 (29.9%)
$ D\% > 20$	11 (16.4%)	11 (16.4%)	47 (70.1%)
$ D\% \leq 33\%$	64 (95.5%)	63 (94.0%)	26 (38.8%)
$ D\% > 33\%$	3 (4.5%)	4 (6.0%)	41 (61.2%)
$ E_n \leq 1$	42 (62.7%)	36 (53.7%)	9 (13.4%)
$1 < E_n < 1.5$	12 (17.9%)	7 (10.4%)	7 (10.4%)
$ E_n > 1.5$	13 (19.4%)	24 (35.8%)	51 (76.1%)

The integrity of the spiked ^{131}I activity was a major problem and because of that we cannot arrive at a clear conclusion on the performance of the participating laboratories for this radionuclide. We can however conclude that:

- For laboratories which reported compatible results, it means that they had received an air filter with the integral of the spiked ^{131}I activity and that they had measured ^{131}I correctly.
- For laboratories which did not report compatible results, the reasons can be multiple:
 - Air filter maintaining only part of the spiked ^{131}I activity;
 - Measurement of ^{131}I not performed correctly;
 - Both of the above.

This exercise highlights the need for additional studies of the radioiodine behaviour on the air filters.

A number of laboratories (8 or 12%) did not use the recommended values for the half-lives and the gamma-ray emission probabilities. However, given the values used and the measurement conditions, we estimate that the discrepancies from using the

recommended values would be completely negligible for ^{134}Cs (<0.0001%) and very small (<0.4%) in the case of ^{131}I .

Table 23. Cumulative table showing the number of laboratories (percent fraction in parentheses) satisfying both the $|D\%|\leq 20$ and $|E_n|\leq 1$ criteria for the three radionuclides, ^{137}Cs , ^{134}Cs and ^{131}I , for any pair of them and for all three simultaneously.

	Number of laboratories		
Nuclides	^{137}Cs	^{134}Cs	^{131}I
Single	41 (61.2%)	36 (53.7%)	9 (13.4%)
Double	27 (40.3%)		
Double		9 (13.4%)	
Double	->		5 (7.5%)
Triple	5 (7.5%)		

It is worth mentioning that one participant (Laboratory 26) informed us on 11 October 2016 that they had by mistake reported the results in a cyclic permutation. This means that the activity result for ^{137}Cs was reported for ^{131}I , the one for ^{134}Cs as that for ^{137}Cs and the one for ^{131}I as that for ^{134}Cs . Table 24 lists the correct order of results, the original (calculated with the originally reported results) and the new $D\%$ and E_n numbers and indeed, it is a set of very good results, compatible for all three nuclides.

Table 24. Correct order of results for Laboratory 26 and the original and the new $D\%$ and E_n numbers.

Nuclide	Reported activity A (Bq)	Reported uncertainty $u_c(A)$ (Bq)	Reported k factor	Original $D\%$	New $D\%$	Original E_n	New E_n
^{137}Cs	0.43	0.077	2	150	10.7	3.1	0.54
^{134}Cs	0.97	0.19	2	82	15.2	2.3	0.67
^{131}I	1.53	0.30	2	-68	13.3	-1.5	0.26

Unfortunately, we could not accept modification of the results reported after the deadline and the originally reported results were kept for Laboratory 26. An important phase of an analysis campaign is reporting of the results. The most common mistake is to use the wrong units or misplace the decimal point, but other mistakes are not excluded. To err is human and the usefulness of such exercise is to discover and correct any mistake.

We are aware of the complexity of the questionnaire that has to be filled by the participants, but it contains information which can help us to understand how environmental laboratories perform their tasks of air monitoring for radioactivity and to deliver reference samples suitable for such measurements and eventually solve problems that some laboratories could encounter.

The answers given to the accompanying questionnaires revealed a diversity of sampling devices and sampling procedures, leaving much room for variation in the measurement

geometry and sample preparation of air filters. If just one type of air filter with the same activity would have been sent to the participating laboratories, a comparison of results would not have reflected the routine measurement conditions. The radioactivity loaded air filters were prepared on air filters provided by the participants by individually spiking gravimetrically a radioactive solution containing the nuclides of interest.

The fact that some problems are always encountered, which can be detected and corrected confirms that there is a permanent need for such comparisons to reaffirm the performance of the laboratories. The "Metrology for radiological early warning networks in Europe" (ENV57-MetroERM) project, coordinated in the frame of the European Metrology Research Programme and funded by the European Association of National Metrology Institutes (EURAMET) and the EU in 2014, aims as well to the harmonisation of the data provided by the radiological early warning networks in Europe (MetroERM, 2016).

The JRC will conduct more ILCs in the future and has developed to that purpose a special automatic dispenser capable of accurately delivering aliquots of a solution as small as 20 μL in any given pattern, based on a X-Y coordinate system. This dispenser can be used in the next ILC to be organized on radioactivity measurement in spiked air filters and it will improve on the homogeneity of the final product.

References

2000/473/Euratom, 2000. Commission recommendation of 8 June 2000 on the application of Article 36 of the Euratom Treaty concerning the monitoring of the levels of radioactivity in the environment for the purpose of assessing the exposure of the population as a whole, OJ L 191:37-46.

87/600/Euratom, 1987. Council Decision of 14 December 1987 on Community arrangements for the early exchange of information in the event of a radiological emergency, OJ L 371: 76-78.

Altitzoglou, T., Máté, B., 2016. Evaluation of the 2014 EC measurement comparison on simulated airborne particulates: ^{137}Cs in air filters; EUR 27969 EN; doi:10.2787/26918

Campion, P.J., 1975. Procedures for accurately diluting and dispensing radioactive solutions. Monographie BIPM-1, Bureau International des Poids et Mesures, Sèvres, France.

DDEP, 2016. Laboratoire National Henri Becquerel, 2016. Decay Data Evaluation Program, Recommended data. http://www.nucleide.org/DDEP_WG/DDEPdata.htm

Debertin, K., Helmer, R.G., 1988. Gamma- and X-ray spectrometry with semiconductor detectors, North-Holland, Amsterdam, 1988.

Debertin, K., Schötzig, U., 1990. Bedeutung von Summationskorrekturen bei der Gammastrahlen-Spektrometrie mit Germaniumdetektoren. PTB-Bericht PTB-Ra-24, 1990.

Euratom, 2012. Consolidated version of the Treaty establishing the European Atomic Energy Community, 2012. OJ C 327/01:1-107.

EURDEP, 2016. What is EURDEP? The official webpage of EURDEP, 2016. <https://remon.jrc.ec.europa.eu/About>

Grau Malonda, A., Garcia-Toraño, E., 1982. Evaluation of counting efficiency in liquid scintillation counting of pure β -ray emitters. Int. J. Appl. Radiat. Isot. 33, 249-253.

Günther, E., 2002. Determination of the activity of radionuclide sources emitting beta and gamma radiation using the CIEMAT/NIST method. In: Proceedings of International Conference Advances in Liquid Scintillation Spectrometry LSC- 2001, 2001.05.07-11, Karlsruhe, Germany, 57-63.

ISO, 1994. ISO/IEC 5725-2:1994, Accuracy (trueness and precision) of measurement methods and results – Part 2: Basic method for the determination of repeatability and reproducibility of a standard measurement method. ISO – International Organisation for Standardization, Geneva, Switzerland.

ISO, 2005a. ISO 13528:2005(E), Statistical methods for use in proficiency testing by interlaboratory comparisons. ISO – International Organisation for Standardization, Geneva, Switzerland.

ISO, 2008. ISO/IEC Guide 98-3:2008, Uncertainty of measurement - Part 3: Guide to the expression of uncertainty in measurement (GUM:1995). ISO – International Organisation for Standardization, Geneva, Switzerland.

Lépy, M.-C., Altitzoglou, T., Anagnostakis, M.J., Arnold, D., Capogni, M., Ceccatelli, A., De Felice, P., Dersch, R., Dryak, P., Fazio, A., Ferreux, L., Guardati, M., Han, J.B., Hurtado, S., Karfopoulos, K.L., Klemola, S., Kovar, P., Lee, K.B., Ocone, R., Ott, O., Sima, O., Sudar, S., Svec, A., Chau Van Tao, Tran Thien Thanh, Vidmar, T. (2010). Intercomparison of methods for coincidence summing corrections in gamma-ray spectrometry, *Applied Radiation and Isotopes* 68 (2010) 1407–1412.

Makepeace, J., Altitzoglou, T., Cassette, P., Dryák, P., Günther, E., Verzezen, F., Broda, R., Simpson, B., Unterweger, M., 1998. International Comparison of Measurements of the Specific Activity of Tritiated Water, *Appl. Radiat. Isot.*, 49, 1411-1416.

MetroERM, 2016. The official webpage of the MetroERM project, 2016. <http://earlywarning-emrp.eu/>

Ratel, G., Michotte, C., Hino, Y., Wätjen, U., 2005a. Activity measurements of the radionuclide ^{137}Cs for the NMJ, Japan and the IRMM, Geel in the ongoing comparison BIPM.RI(II)-K1.Cs-137 and update of the KCRV, *Metrologia*, 42, Tech. Suppl. 06009.

Ratel, G., Michotte, C., Wätjen, U., Janßen, H., Szücs, L., Coursol, N., Hino, Y., 2005b. Activity measurements of the radionuclide ^{134}Cs for the IRMM, Geel, PTB, Germany, OMH, Hungary, LNE-LNHB, France and NMJ, Japan in the ongoing comparison BIPM.RI(II)-K1.Cs-134. Final Update Report for Cs-134 (2) 2005/09/15. BIPM, Sèvres, France.

Ratel, G., Michotte, C., Leena, J., Iglíki, A., Sahagia, M., Hino, Y., 2007. Update of the BIPM comparison BIPM.RI(II)-K1.Cs-134 of activity measurements of the radionuclide ^{134}Cs to include the 2005 results of the BARC (India) and the CNEA (Argentina), the 2006 result of the IFIN-HH (Romania) and the link for the 2005 regional comparison APMP.RI(II)-K2.Cs-134 to include the VNIIM and the INER. Final Update Report for Cs-134 (3) 2007/09/24. BIPM, Sèvres, France.

Schima, F.J., Hoppes, D.D., 1983. Tables for cascade-summing corrections in gamma-ray spectrometry, *Int. J. Appl. Radiat. Isot.*, 34:8, pp.1109-1114.

Sibbens, G., Altitzoglou, T., 2007. Preparation of radioactive sources for radionuclide metrology. *Metrologia* 44, S71–S78.

Solé, V. A., 1990. A Monte Carlo program to calculate said angle and transmission corrections in counting systems. IRMM Internal Report GE/R/RN/12/90.

Pommé, S., 2016. When the model doesn't cover reality: examples from radionuclide metrology. *Metrologia* 53, S55-S64.

Spernal, A., Denecke, B., 1964. Präzise Absolutmessung der Aktivität von Tritium – III, *Int. J. Appl. Radiat. Isot.*, 15, 241-254.

Wätjen, U., Szánto, Zs., Altitzoglou, T., Sibbens, G., Keightley, J., Van Ammel, R., Hult, M., De Cort, M., 2007. Evaluation of EC measurement comparison on simulated airborne particulates – ^{137}Cs in air filters. Report EUR 22926 EN, ISBN 978-92-79-06962-8.

Wätjen, 2008. Wätjen, U., 2008. European Measurement Comparisons of Environmental Radioactivity. AIP Conference Proceedings, 1036, 85-99.

List of abbreviations and definitions

BIPM	Bureau International des Poids et Mesures
BEGe	Broad Energy Germanium detector
CIEMAT	Centro de Investigaciones Energéticas, Medioambientales y Tecnológicas
CCRI(II)	Comité Consultatif des Rayonnements Ionisants, Section 2
DDEP	Decay Data Evaluation Program
EC	European Commission
EMRP	European Metrology Research Programme
EU MS	European Union Member States
EURATOM	European Atomic Energy Community
EURAMET	European Association of National Metrology Institutes
EURDEP	European Union Radiological Data Exchange Platform
GUM	Guide to the Expression of Uncertainty in Measurement
HPGe	High-Purity Germanium detector
ICS-REM	International Comparison Scheme for Radioactivity Environmental Monitoring
EC ILC	Interlaboratory comparison organised by JRC-IRMM
ISO	International Organization for Standardization
JRC	Directorate General Joint Research Centre
JRC-IRMM	JRC Institute for Reference Materials and Measurements
JRC-ITU	JRC Institute for Transuranium Elements
KCRV	Key Comparison Reference Value
LSC	liquid scintillation counter, liquid scintillation counting
MetroERM	Metrology for radiological early warning networks in Europe, ENV57
NIST	National Institute of Standards and Technology
NMI	National Metrology Institute
SIR	Système International de Référence, International Reference System for radionuclides
UTC	Coordinated Universal Time
<i>A</i>	activity measured by participating laboratory
<i>A₀</i>	activity reference value, spiked activity
<i>D_%</i>	percentage difference
<i>E_n</i>	performance statistic <i>E_n</i> number
<i>k</i>	coverage factor according to GUM
<i>MAD</i>	median absolute deviation
<i>MDA</i>	Minimum Detectable Activity
<i>stdev</i>	standard deviation, standard uncertainty in counting alone

u_c	combined standard uncertainty according to GUM
U	expanded uncertainty according to GUM
$U(A)$	expanded uncertainty of laboratory result ($k=2$)
$U(A_0)$	expanded uncertainty of reference value ($k=2$)
$u(A)$	standard uncertainty of laboratory result ($k=1$)
$u(A_0)$	standard uncertainty of reference value ($k=1$)
$u_c(A/A_0)$	combined standard uncertainty of ratio A/A_0 ($k=1$)

List of figures

Fig. 1. Key comparisons of CCRI(II) and traceability of the reference values for samples provided by JRC for the interlaboratory comparisons amongst monitoring laboratories (KCRV = Key Comparison Reference Value).

Fig. 2. Typical gamma-ray spectrum of a mixture of the three radionuclides of interest, ^{137}Cs , ^{134}Cs and ^{131}I . The most abundant peaks are indicated.

Fig. 3. Preparation scheme of the ^{137}Cs , ^{134}Cs and ^{131}I dilutions.

Fig. 4. Different cases of spiked air filters, which have been modelled for Monte Carlo simulations to calculate counting efficiencies with some typical detectors.

a) $\varnothing 70$ -mm filter with 9 active spots

b) $\varnothing 110$ -mm filter with 16 active spots and

c) $\varnothing 130$ -mm filter with 16 active spots and (including the red dots) 24 active spots.

Fig. 5. Interpretation of the PomPlot (Pommé, 2016).

Fig. 6. Normal probability plot and frequency histogram of the ^{137}Cs results after exclusion of the three outliers (see text). The red curve in the frequency histogram is the normal probability distribution.

Fig. 7. Ratio of ^{137}Cs activity per spiked air filter as measured by the participating laboratory to the individual spiked activity on the filter (JRC reference value). Blue dashed lines indicate the $\pm 20\%$ limit from the reference value and the red dashed lines the $\pm 33\%$ limit. The error bars show the expanded uncertainty ($k=2$).

Fig. 8. Percentage difference of the ^{137}Cs results reported by the participating laboratories from the reference activity values, plotted in ascending order. Blue colour indicates the results within the $\pm 20\%$ range from the reference value and red indicates results outside that range.

Fig. 9. E_n numbers for the ^{137}Cs results, plotted in ascending order. Green color indicates compatible results, yellow indicates warning signal and red indicates action signal.

Fig. 10. PomPlot of the ^{137}Cs data. Green, blue and red solid lines indicate ζ -scores=1, 2 and 3, respectively.

Fig. 11. Normal probability plot and frequency histogram of the ^{134}Cs results after exclusion of the outlier (see text). The red curve in the frequency histogram is the normal probability distribution.

Fig. 12. Ratio of ^{134}Cs activity per filter as measured by the participating laboratory to the individual spiked activity on the filter (JRC reference value). Blue dashed lines indicate the $\pm 20\%$ limit from the JRC reference value and the red dashed lines the $\pm 33\%$ limit. The error bars show the expanded uncertainty ($k=2$).

Fig. 13. Percentage difference of the ^{134}Cs results reported by the participating laboratories from the reference activity values, plotted in ascending order. Blue color indicates the results within the range $\pm 20\%$ from the reference value and red indicates results outside that range.

Fig. 14. E_n numbers for the ^{134}Cs results, plotted in ascending order. Green color indicates compatible results, yellow indicates warning signal and red indicates action signal.

Fig. 15. PomPlot of the ^{134}Cs data. Green, blue and red solid lines indicate ζ -scores=1, 2 and 3, respectively.

Fig. 16. Normal probability plot and frequency histogram of the ^{131}I results. The red curve in the frequency histogram is the normal probability distribution.

Fig. 17. Ratio of ^{131}I activity per spiked air filter as measured by the participating laboratory to the individual spiked activity on the filter (JRC reference value). Blue dashed lines indicate the $\pm 20\%$ limit from the JRC reference value and the red dashed lines the $\pm 33\%$ limit. The error bars show the expanded uncertainty ($k=2$).

Fig. 18. Percentage difference of the ^{131}I results reported by the participating laboratories from the reference activity values, plotted in ascending order. Blue color indicates the results within the range $\pm 20\%$ from the reference value and red indicates results outside that range.

Fig. 19. E_n numbers for the ^{131}I results plotted in ascending order. Green color indicates compatible results, yellow indicates warning signal and red indicates action signal.

Fig. 20a. PomPlot of the ^{131}I data. Green, blue and red solid lines indicate ζ -scores=1, 2 and 3, respectively.

Fig. 20b. PomPlot of the ^{131}I data within the $\pm 20\%$ range from their respective reference values. Green, blue and red solid lines indicate ζ -scores=1, 2 and 3, respectively.

Fig. 21. Ratio of ^{131}I activity per spiked air filter as measured by the participating laboratory over the individual spiked activity on the filter (JRC reference value) plotted per day of preparation and with ascending spiked activity (per day). Blue dashed lines indicate the $\pm 20\%$ limit from the reference value and the red dashed lines the $\pm 33\%$ limit. The error bars show the expanded uncertainty ($k=2$). The laboratory numbers are arbitrary.

Fig. 22. Ratio of ^{131}I activity per spiked air filter as measured by the participating laboratory over the individual spiked activity on the filter (JRC reference value) plotted per dilution used for spiking and with indication (in color) of the filter size. Blue dashed lines indicate the $\pm 20\%$ limit from the reference value and red dashed lines the $\pm 33\%$ limit. The error bars show the expanded uncertainty ($k=2$). The laboratory numbers are arbitrary.

Fig. 23. E_n numbers for the ^{131}I reported results plotted in ascending order, when the uncertainty on the reference value is increased to 22.5%. Green color indicates compatible results, yellow indicates warning signal and red indicates action signal.

Fig. 24. Ratio of ^{137}Cs (in blue), ^{134}Cs (in red) and ^{131}I (in green) activity per spiked air filter as measured by the participating laboratory to the individual spiked activity on the filter (JRC reference value). Blue dashed lines indicate the $\pm 20\%$ limit from the reference value and the red dashed lines the $\pm 33\%$ limit. The error bars show the expanded uncertainty ($k=2$).

Fig. A-1. Ratio of measured-to-reference activity and expanded uncertainty ($k=2$) of the quality control **filters** prepared from the ^{137}Cs , ^{134}Cs and ^{131}I dilutions used for spiking the air filters. The blue diamond symbol is used for results obtained with Detector A and the red diamond symbol for those obtained with Detector B. The solid blue line is the Reference value (unity), the blue dashed lines designate the $\pm 20\%$ range and the red dashed lines the $\pm 33\%$ range. **Top:** ^{137}Cs , **Middle:** ^{134}Cs , **Bottom:** ^{131}I .

Fig. A-2. Ratio of measured-to-reference activity and expanded uncertainty ($k=2$) of the quality control **point sources** prepared from the ^{137}Cs , ^{134}Cs and ^{131}I dilutions used for spiking the air filters. The blue diamond symbol is used for results obtained with Detector A and the red diamond symbol for those obtained with Detector B. The solid blue line is the Reference value (unity), the blue dashed lines designate the $\pm 20\%$ range and the red dashed lines the $\pm 33\%$ range. **Top:** ^{137}Cs , **Middle:** ^{134}Cs , **Bottom:** ^{131}I .

List of tables

Table 1. Recommended nuclear data for the three radionuclides of interest (combined standard uncertainties in parenthesis ($k=1$)) (DDEP, 2016).

Table 2. Dilutions and dilution factors with their combined standard uncertainties u_c (in parentheses).

Table 3a. Deviation of the counting efficiency of a spiked filter from that of a filter with homogeneous distribution of the activity as calculated using the Monte Carlo code GEOLEP (see text). Calculations are performed for two different detectors (A and B), two different sizes of the homogeneous active area ($\varnothing 50$ and 60 mm) and three different diameters of the 9 active spots ($\varnothing 3$, 7 and 10 mm).

Table 3b. Deviation of the counting efficiency of a spiked filter from that of a filter with homogeneous distribution of the activity as calculated using the Monte Carlo code GEOLEP (see text). Calculations are performed for two different detectors (A and B), two different sizes of the homogeneous active area ($\varnothing 90$ and 110 mm) and three different diameters of the 16 active spots ($\varnothing 3$, 7 and 10 mm). A third detector (C) was introduced as well and the counting efficiency for filters with homogeneous active area of $\varnothing 110$ mm and $\varnothing 100$ mm was simulated for 16 and 24 active spots (see *Figure 4c*).

Table 4. Radionuclides measured by the participating laboratories.

Table 5. ^{137}Cs reference (spiked) activity, activity result as reported by the participating laboratories, ratio of the reported to the reference activity, percentage difference ($D\%$) of the reported to the reference activity and E_n number, with their associated standard uncertainties ($k=1$).

Table 6. Statistical analysis of the laboratory reported results for ^{137}Cs , per dilution. In the last column, the ratios of the measured to the reference values are given.

Table 7. Possible influencing parameters; statistics and categorisation of the reported results by the participating laboratories.

Table 8. Comparison of the spiked activity levels and the performance of the laboratories in the 2003 and 2014 EC ILCs and the 2016 ENV57/MetroERM ILC.

Table 9a. Comparison of the spiked activity levels and the performance of the laboratories in the 2003 and 2014 EC ILCs and the 2014 EC ILC and the 2016 ENV57/MetroERM ILC.

Table 9b. Comparison of the spiked activity levels and the performance of the laboratories in the 2003 and 2014 EC ILCs and the 2016 ENV57/MetroERM ILC.

Table 10. ^{134}Cs reference spiked activity, activity result as reported by the participating laboratories, ratio of the reported to the reference activity, percentage difference ($D\%$) of the reported to the reference activity and E_n number, with their associated standard uncertainties ($k=1$).

Table 11. Statistical analysis of the laboratory reported results for ^{134}Cs , per dilution. In the last column, the ratios of the measured to the reference values are given.

Table 12. Formulae from Schima and Hoppes (1983) for the reciprocal correction factor and indicative values for Detectors A and B (*Section 3.1*) as calculated for an active filter of 50 and 90 mm in diameter and measured on the detector endcap (see also text).

Table 13. Distribution of the corrected and non-corrected for coincidence summing results of ^{134}Cs , as compatible and non-compatible according to the $D\%$ and E_n criteria.

Table 14. ^{131}I reference spiked activity, activity result as reported by the participating laboratories, ratio of the reported to the reference activity, percentage difference ($D\%$)

of the reported to the reference activity and E_n number, with their associated standard uncertainties ($k=1$).

Table 15. Statistical analysis of the laboratory reported results for ^{131}I , per dilution. In the last column, the ratios of the measured to the reference values are given.

Table 16. Correction factors for the decay during measurement of ^{131}I .

Table 17. Distribution of the corrected and non-corrected for decay during measurement results of ^{131}I , as compatible and non-compatible according to the $D_\%$ and E_n criteria.

Table 18. Formulae from Schima and Hoppes (1983) for the reciprocal correction factor and indicative values for Detectors A and B (*Section 3.1*) as calculated for an active filter of 50 and 90 mm in diameter and measured on the detector endcap (see also *Section 7.2.3*).

Table 19. Count rate results for the 364 keV line of ^{131}I from measuring filters IRMM-03 and IRMM-06 and their plastic bags. The percent ratio of the count rate obtained from the plastic bag to that from the filter is given in the last column.

Table 20. Distribution of the results for ^{131}I corrected and non-corrected for the transferred activity to the protective plastic bag, as compatible and non-compatible according to the $D_\%$ and E_n criteria.

Table 21. Performance of the laboratories: percentage difference, $D_\%$, and E_n number for each of the three radionuclides, ^{137}Cs , ^{134}Cs and ^{131}I . Values of $D_\%$ outside the $\pm 20\%$ limit from the reference value are in yellow. Absolute values of E_n between 1 and 1.5 are also in yellow and those above 1.5 are in red.

Table 22. Cumulative table showing the performance of the laboratories for the three radionuclides, ^{137}Cs , ^{134}Cs and ^{131}I . The numbers give the number of laboratories with the percent fraction in parentheses.

Table 23. Cumulative table showing the number of laboratories (percent fraction in parentheses) satisfying both the $|D_\%| \leq 20$ and $|E_n| \leq 1$ criteria for the three radionuclides, ^{137}Cs , ^{134}Cs and ^{131}I , for any pair of them and for all three simultaneously.

Table 24. Correct order of results for Laboratory 26 and the original and the new $D_\%$ and E_n numbers.

Table A-1. General information on the 2016 MetroERM ILC participating laboratories.

Table A-2. Air filter size categories.

Table A-3. Information on the air sampling methods.

Table A-4. Information on the technical part of the measurements.

Table A-5. Type, material size and form of air filters used by the participants and as reported by the participants.

Table A-6. Sequence of spiked filters and quality control sources prepared for the 2016 ENV57/MetroERM ILC exercise.

Table A-7a. Gamma-ray spectrometric results of the quality control filters; measured values and the corresponding uncertainties.

Table A-7b. Gamma-ray spectrometric results of the quality control filters; their relative deviation from the reference values and their ratio to the reference values.

Table A-8a. Gamma-ray spectrometric results of the quality control point sources; measured values and the corresponding uncertainties.

Table A-8b. Gamma-ray spectrometric results of the quality control point sources; their relative deviation from the reference values and their ratio to the reference values.

Table A-9. Treatment of spiked air filters before measurement by the participants and equipment used for the measurement of the spiked air filters.

Table A-10. Measurement parameters; acquisition time, count rates, MDA.

Table A-11. Coincidence summing correction factors as applied and reported by the participants.

Table A-12. Detailed uncertainty budget table showing the uncertainty components as reported by the participating laboratories, for all three radionuclides. In column 13 the combined relative uncertainty as reported by the participants is given. For comparison, in column 15 the combined relative uncertainty calculated as square root of the quadratic sum of the uncertainty components given in the table is given. Finally, in column 16 the combined relative uncertainty as calculated from the uncertainty accompanying the results is given.

ANNEXES

Annex 1: List of participating laboratories per country (in alphabetic order)

AUSTRIA

Mr. Johannes Klimstein
Austrian Agency for Health and Food Safety (AGES)
Department for Radon and Radio
Wieningerstrasse 8
4020 Linz

Mr. Florian Smecka
AGES Vienna
Radiation Protection & Radioch
Spargelfeldstraße 191
1220 Vienna

BELGIUM

Dr. Michel Bruggeman
SCK-CEN
LRM
Boeretang 200
2400 Mol

Mr. Tony Dieudonne
IRE ELIT
BUS
Avenue de l'Esperance 1
6220 Fleurus

BULGARIA

Mrs. Bistra Hristova
Regional Health Inspectorate - Varna
Public Health Department
3, Bregalnitsa str
9000 Varna
BULGARIA

Mrs. Cvetelina Kamenova
Executive Environment Agency
Regional Laboratory Vratza
Exarh Josif 81
3000 Vratza
BULGARIA

Dr. Rositza Kamenova-Totzeva
National Center of Radiobiology and Radiation Protection
Public Exposure Monitoring Lab
Georgi Sofiiski Blvd. No 3
1606 Sofia
BULGARIA

Mrs. Lena Krumova
Executive Environment Agency
Regional Laboratory - Montana

4 "Julius Irasek" Str.
3400 Montana
BULGARIA

Mrs. Kalinka Stoyanova
Executive Environment Agency
Lab for radiation measurements
blvd. Tzar Boris III 136
1618 Sofia
BULGARIA

CROATIA

Dr. Tomislav Bituh
Institute for Medical Research and Occupational Health
Radiation Protection Unit
Ksaverska Cesta 2
P.O. Box 291
10000 Zagreb
CROATIA

CYPRUS

Mrs. Anastasia Caballero
State General Laboratory of Cyprus
Kimonos Street 44
1451 Nicosia
CYPRUS

CZECH REPUBLIC

Mr. Zdenek Borecky
Statni ustav radiacni ochrany v.v.i.
Branch Hradec Kralove
Pileticka 57/15A
500 03 Hradec Kralove
CZECH REPUBLIC

Mrs. Helena Malá
NRPI (SÚRO)
Monitoring dep.
Bartošková 28
140 00 Prague
CZECH REPUBLIC

Mr. Jiri Rada
Statni ustav radiacni ochrany, v.v.i.
Syllabova 21
703 00 Ostrava
CZECH REPUBLIC

DENMARK

Dr. Sven Nielsen
Technical University of Denmark
Center for Nuclear Technologie
Frederiksborgvej 399
4000 Roskilde
DENMARK

Mr. Asser Nyander Poulsen
The Danish Health Authority
Radiation Protection
Knapholm 7
2730 Herlev
DENMARK

ESTONIA

Ms. Eia Jakobson
Environmental Board Republic of Estonia
Radiation Safety Department
Kopli 76
10416 Tallinn
ESTONIA

FINLAND

Dr. Jani Turunen
STUK - Radiation and Nuclear Safety Authority
VALO
Laippatie 4
FI-00880 Helsinki
FINLAND

FRANCE

Dr. Rodolfo Gurriaran
IRSN
STEME/LMRE
Batiment 501
Campus universitaire d'Orsay
Bois des Rames
91400 Orsay
FRANCE

Mr. Romain VIDAL
Institut de Radioprotection et de sureté Nucléaire (IRSN)
PRP-ENV / STEME/LMN
31 Rue de l'écuse
78116 Le Vésinet
FRANCE

GERMANY

Mrs. Jacqueline Bieringer
Bundesamt für Strahlenschutz (Federal Office for Radiation Protection)
SW 2.5
Rosastrasse 9
79098 Freiburg
GERMANY

GREECE

Prof. Alexandros Clouvas
Aristotle University of Thessaloniki
Egnatia Street
Polytechnical School
Building Delta - Fifth floor
54124 Thessaloniki
GREECE

Dr. Maria Nikolaki
Greek Atomic Energy Commission
Environmental Radioactivity
Patriatchou Grigoriou & Neapoleos
P.O. Box 60092
15310 Agia Paraskevi Attiki
GREECE

HUNGARY

Ms. Júlia Kövöndiné Kónyi
National Public Health Center
National Directorate for Radiobiology and Radiohygiene
Environmental Radiohygiene
Anna street nr. 5
1221 Budapest
HUNGARY

IRELAND

Mr. Stephen Somerville
Environmental Protection Agency (ORP)
Radiation Monitoring
3 Clonskeagh Square
Clonskeagh Road
Dublin 14
D14 Dublin
IRELAND

ITALY

Mr. Pietro Badalamenti
ARPA Lombardia
Office for Radiation Protect.
via Juvara 22
20129 Milano
ITALY

Dr. Giovanna Belmusto
ARPACAL
Department of Reggio Calabria
via troncovito snc
89135 Reggio Calabria
ITALY

Dr. Claudio Cristofaro
ARPA Molise - Centro di Riferimento Regionale per il Controllo della Radioattività
Ambientale
Campobasso Department
Contrada Selva Piana snc
86100 Campobasso
ITALY

Dr. Giorgio Evangelisti
ARPA Lazio
Sede di Viterbo
Via Montezebio 17
01100 Viterbo
ITALY

Dr. Massimo Faure Ragani
ARPA Valle di Aosta
Environmental radioactivity
Loc Grande Charrière 44
11020 Saint Christophe
ITALY

Dr. Mauro Magnoni
ARPA Piemonte
Dipartimento Radiazioni
Via Jervis, 30
10015 Ivrea
ITALY

Dr. Stefano Pegoretti
Agenzia Provinciale Protezione Ambiente
Settore Laboratorio
via Lidorno, 1
38123 Trento
ITALY

Dr. Laura Porzio
ARPA Piemonte
SS Siti Nucleari
Via Trino 89
13100 Vercelli
ITALY

Dr. Giuseppe Roselli
ARPA Puglia
Dipartimento Bari - Polo R. I.
Via Oberdan 18/E
70126 Bari
ITALY

Dr. Paola Sabatini
ARPA Umbria
UOLM- Serv. Rad. Ionizzanti
Via Pievaiola 207 B-3
06132 Perugia
ITALY

Dr. Roberto Sogni
ARPA Emilia Romagna
CTR Radioattività Ambientale
via XXI Aprile, 48
29121 Piacenza
ITALY

LITHUANIA

Dr. Arunas Gudelis
Center for Physical Sciences and Technology (FTMC)
Metrology Department
Savanoriu Ave. 231
LT-02300 Vilnius
LITHUANIA

Dr. Beata Silobritiene
Environmental Protection Agency
Environmental Research
A.Juozapaviciaus str. 9
09311 Vilnius
LITHUANIA

LUXEMBOURG

Dr. Marielle Lecomte
Ministry of Health
Radioprotection
Villa Louvigny
Allée Marconi
2120 Luxembourg
LUXEMBOURG

MALTA

Ms. Doris Gambin
Public Health Lab.
Evans building
Lower Merchants street
VLT 1179 Valletta
MALTA

NETHERLANDS

Dr. Ronald Overwater
RIVM
VLH
Antonie van Leeuwenhoeklaan 9
3721MA Bilthoven
NETHERLANDS

NORWAY

Mr. Alexander Mauring
Norwegian Radiation Protection Authority
Monitoring and Research
Grini Næringspark 13
1361 Østerås
NORWAY

Mr. Bredo Møller
Norwegian Radiation Protection Authority
Svanhovd Emergency Prep. Unit
Svanhovd
9925 Svanvik
NORWAY

POLAND

Mr. Michal Bonczyk
Central Mining Institute (GIG)
Environmenal Radioactivity
Plac Gwarkow 1
40-166Katowice
POLAND

Mr. Jerzy Chytla
Voivodship Sanitary and Epidemiological Station in Rzeszow (WSSE RZESZOW)
Radiation Laboratory (LPP)
Jezierskiego 39
38-500Sanok
POLAND

Mr. Krzysztof Isajenko
Central Laboratory for Radiological Protection
Dosimetry Department
Konwaliowa St. 7
PL-03-194 Warszawa
POLAND

Prof. Jerzy W. Mietelski
Institute of Nuclear Physics (IFJ PAN)
Radzikowskiego 152
31-342Krakow
POLAND

Dr. Jakub Ośko
National Centre for Nuclear Research
LPD
A. Sołtana 7
05-400Otwock
POLAND

PORTUGAL

Dr. Maria José Madruga
Instituto Superior Técnico/laboratório de Proteção e Segurança Radiológica
Estrada Nacional 10, km 139,7
2695-066 Bobadela LRS
PORTUGAL

ROMANIA

Ms. Luminita Cojocaru
Environmental Protection Agency Dolj
Radioactivity Laboratory Cv.
150 Calea Bucuresti street,
Craiova, Dolj, Romania
200344 Craiova
ROMANIA

Mrs. Nelida Florea
Environmental Protection Agency
Radioactivity Laboratory
Str. Splaiul Muresului FN
310132 Arad
ROMANIA

Mr. Vlad Florin
Environmental Protection Agency
Radioactivity
Iza 1A
430073 Baia Mare
ROMANIA

Mrs. Claudia Puscasu
Environmental Protection Agency Constanta
Radioactivity Laboratory Cta
300 Mamaia B-dul,
Room nr.19
(C.M.R." Dobrogea" Building)
900581 Constanta
ROMANIA

Mrs. Elena Simion
National Environmental Protection Agency
Nat. Ref. Radioactivity Lab.
294, Splaiul Independentei
Sector 6
060031 Bucharest
ROMANIA

Mr. Claudia Tăbăcaru
Environmental Protection Agency
Radioactivity Laboratory
Th. Vascauteanu 10 bis
700464 Iași
ROMANIA

SLOVENIA

Mr. Peter Jovanovič
Zvd Zavod za Varstvo Pri Delu d. O. O.
Chengdujska Cesta 25
1260 Ljubljana Polje
SLOVENIA

Mr. Branko Vodenik
Jozef Stefan Institute
F2
Jamova cesta 39
1000 Ljubljana
SLOVENIA

SPAIN

Dr. Javier Guillén
University of Extremadura
LARUEX, Dpt. applied Physics
Faculty of Veterinary
University of Extremadura
Avda. Universidad, s/n
10003 Cáceres
SPAIN

Mrs. Isabel Serrano
Universitat Politècnica de Catalunya
INTE-ETSEIB
C / Diagonal, 647
08028 Barcelona
SPAIN

Dr. Jose Antonio Suarez Navarro
CIEMAT
Environmental Radioactivity
Avenida Complutense 40
28040 Madrid
SPAIN

SWEDEN

Sofia Eriksson
Swedish Radiation Safety Authority
Solna Strandväg 122
17154 Solna
SWEDEN

Dr. Johan Kastlander
Swedish Defence Research Agency
Gullfossgratan 6 (Kista)
164 90 Stockholm
SWEDEN

SWITZERLAND

Mr. Giovanni Ferreri
Federal Office of Public Health
Environmental Radioactivity
Schwarzenburgstrasse 157
3003 Bern
SWITZERLAND

TURKEY

Mr. Memduh Fatih ÇINAR
Turkish Atomic Energy Authority
Sarayköy Nuclear Research and Training Center
Rad. & Ana. Meas. Division
Saray Mah. Atom Cad. No:27
06893 Kazan/Ankara
TURKEY

Dr. Nurdan Güngör
Turkish Atomic Energy Authority
Cekmece Nuclear Research and Training Center
Radioactivity Measurement Unit
Yarımburgaz Mah. Nukleer Arast
Merkezi Yolu No 10
34303 Istanbul
TURKEY

UNITED KINGDOM

Mr. Alasdair Morgan
Cavendish Nuclear Ltd
Radiometric Spectrometry
Greeson Court
Westlakes Science Park
CA24 3HZ Moor Row
UNITED KINGDOM

Annex 2: Communication to the participants

2A. Expression of interest

Ref. Ares(2015)5386740 - 26/11/2015

From: ALTITZOGLOU Timotheos (JRC-GEEL)
Sent: 23 November 2015 12:25
To: 'Claudia.Landstetter@ages.at'; 'wolfgang.ringer@ages.at';
'benoit.deconinck@ire-elit.eu'; 'metrolou@ire-elit.eu';
'lsnevers@sccken.be'; 'm.shishenkov@yahoo.com';
'shishenkov@dea.government.bg'; 'v.badulin@ncrrp.org';
'sdimadov.bg'; 'rivmadadov'; 'ortice@zti-varna.com';
'r.torzeva@ncrrp.org'; 'cvetelina.1980@abv.bg'; 'lab_vrhm-real.net';
'jibieringer@bfs.de'; 'axel.dalheimer@dwd.de'; 'eva.sindejkova@suib.cz';
'zdenek.borckysuro.cz'; 'jiri.radasuro.cz'; 'apo@sis.dk';
'spnifdu.dk'; 'etia.jakobson@ksskonnamet.ee'; 'gtenorio@us.es';
'antonis.maltzos@ecae.gr'; 'kioanni@cc.uoi.gr'; 'catalina.gasco@ciemat.es';
'ymiralte@ciemat.es'; 'isabel.serrano@upc.edu'; 'finerall@ull.es';
'marc.giez@ciemat.es'; 'th.tunelini@knytooski.hu';
'leandro.nagao@suprabiente.it'; 'il.porzio@arpa.piemonte.it';
'p.sabatini@arpa.umbria.it'; 'mauro.magnoni@arpa.piemonte.it';
'macapatti@arpa.sardegna.it'; 'rusconi@arpa.lombardia.it';
'm.faureragani@arpa.vda.it'; 'r.trozzo@arpacal.it';
'carmela.fortunato@arpab.it'; 'cristofaro.arpam@libero.it';
'sogni@arpa.emr.it'; 'stefano.pegoretti@provincia.tn.it';
'l.vitucci@arpa.puglia.it'; 'risco.rc@arpacal.it'; 'silvia.paci@arpalazio.it';
'beatrice.gilbratien@aaam.lt'; 'gudai@skt1.mil.it';
'marilje.lecomte@ns.etat.lu'; 'andris.abramenkovs@lvgmc.lv';
'public.health.lab@gov.mt'; 'pieter.kwakman@rivm.nl';
'dyrekto@for.waw.pl'; 'isaajenko@clor.waw.pl';
'Tomasz.Pijaszczynski@ncbj.gov.pl'; 'madruga@ctn.ist.utl.pt';
'elena.simonianpm.ro'; 'lmlanpm.ro'; 'ana.gheras@lanpm.ro';
'acratova@yahoo.com'; 'stracta@eml.ro'; 'stracta@clicknet.ro';
'statie.ra@apmm.anpm.ro'; 'johan.kastlander@ofp.se';
'sofia.eriksson@ssm.se'; 'mike.froggatt@cavendishnuclear.com';
'michael.davidson@phe.gov.uk';
cc: 'stefan.neumaier@ptb.de'; 'harald.dombrowski@ptb.de';
'Valentin.BLIDEANU@cea.fr'; 'DE CORT Marc (JRC-ISPRA)';
'Stratos.valaki@jrc.ec.europa.eu'; 'MONDELAERS WITTY (JRC-GEEL)'; HULT
WIKEL (JRC-GEEL) MALO PECEVA (JRC-GEEL); DEHOUCK Pieter (JRC-GEEL);
ALTITZOGLOU Timotheos (JRC-GEEL)
Subject: METROERM Interlaboratory Comparison: Expression of interest

Dear colleague,

The JRC is partner of the European Metrology Research Programme (EMRP) EN57
Metrology for
Radiological Early Warning Networks in Europe' project
[http://earlywarning-emrp.eu/].

The JRC-IRMM under Task 3.2 'Traceability management of airborne radioactivity'
[http://earlywarning-emrp.eu/?page_id=1106] will organise and conduct an
interlaboratory comparison
(ILC) on the measurement of
ILC-REM exercise in 2014.
Such exercises contribute to the harmonisation of the radiological early warning
networks in Europe.

The particularity of the ILC is that the spiked filters will be made to the
specifications of each participant
on blank filters supplied by them. This time, the filters will be spiked with
137Cs, 134Cs and 131I (if the latter
is available timely), all three gamma-ray emitters.

The spiked air filters are expected to be dispatched to the participating
laboratories by end of January
2016. The participating laboratories will be requested to send their results to
JRC-IRMM before the end
of February 2016.

The ILC will be conducted according to ISO 17043. The reported results will be
treated confidentially and
they will be presented without association with the name of the corresponding
lab (but they will be
available to the concerned lab).

You have participated at the 2014 ILC-REM exercise and you might be interested
in the new Metrology for
Radiological Early Warning Networks in Europe' project.
Replying to this e-mail as soon
as possible and no later than 4 December 2015. In case of too many interested
laboratories and because
of our limited human resources, the selection will be based on a combination of
geographical and 'first-
come, first-served' criteria.
Please, include in your e-mail the name of the responsible/contact person,
the address, telephone
number and e-mail for the communication.

Looking forward to hearing from you.

Best regards,

Timotheos Altitzoglou

PS1 If you have received this e-mail more than once, we apologise.
PS2 In case you know other laboratories not in the mailing list that
might be interested, please
feel free to forward this e-mail to them.

The views expressed are purely those of the writer and may not any circumstances
be regarded as stating
an official position of the European Commission.

Timotheos ALTITZOGLOU

Project Leader

European Commission
Directorate-General Joint Research Centre
Institute for Reference Materials and Measurements
Geel, Belgium
B-2440 Geel, Belgium
8-2440 Geel, Belgium

JRC-IRMM-REM-COMPARISONS@ec.europa.eu
timotheos.altitzoglou@ec.europa.eu
http://ec.europa.eu/dgs/jrc
http://irmm.jrc.ec.europa.eu

From: ALTZITZOGLOU Timotheos (JRC-GEEL)
Sent: 23 November 2015 12:39
To: 'ibieringer@bfs.de'; 'mbruggem@sccken.be';
'nina.cernohlawe@bml.fwg.at'; 'dabrow@paa.gov.pl';
'hasan.dikmen@taek.gov.tr'; 'sybille.estier@bag.admin.ch';
'Laurent.ferreux@cea.fr'; 'catalina.gasco@ciemat.es';
'ana.gherasimian@pmr.ro'; 'Peter.Jovanovic@czvz.si'; 'pkovar@cmi.cz';
'Alexander.Maurin@empa.ch'; 'maria.mikolajewicz@gr';
's.somervill@epa.ie'; 'Halim.taskin@taek.gov.tr'; 'jani.turunen@stuk.fi';
'branko.vodeni@ijs.si'; 'herbert.wershofen@ptb.de';
'tviomar@sccken.be'; 'jcamp@sccken.be'; 'jsuran@cmi.cz';
'milan.bunata@envinet.cz'; 'jan.kubancak@envinet.cz';
'ustoeher@bfs.de'; 'mzaehring@bfs.de'; 'mbteher@bfs.de';
'dirk.arnold@ptb.de'; 'daniel.zapata@ptb.de'; 'patrick.kessler@ptb.de';
'arturo.vargas@upc.edu.es'; 'josep.sempau@upc.edu'; 'jeo@ciemat.es'; 'jean-
marc.bordy@cea.fr'; 'valerie.tourenco@cea.fr';
'christophe.debay@irsn.fr'; 'erwan.manach@irsn.fr';
'adeline.bardet@irsn.fr'; 'clouvas@eng.auth.gr';
'pierino.defelice@enea.it'; 'mariapia.toni@enea.it';
'maurizio.bovi@enea.it'; 'steven.judge@npl.co.uk';
'steven.bell@npl.co.uk'; 'laurence.brice@npl.co.uk';
'sergey@frim.rdg.ac.uk'; 'denis.chirio@ijs.si'; 'stefan.neumaier@ptb.de';
'namikkamal.sahin@taek.gov.tr'; 'yucel@taek.gov.tr';
Cc: 'stefan.neumaier@ptb.de'; 'harald.dombrowski@ptb.de';
'valentin.BLIDEANU@cea.fr'; 'DE CORT Marc (JRC-ISPRa);
'stratos.valakis@jrc.ec.europa.eu'; 'MONDELAERS Willy (JRC-GEEL); HULT
Mikael (JRC-GEEL); 'MALO Petya (JRC-GEEL); DEHOUCQ Pieter (JRC-GEEL);
ALTZITZOGLOU Timotheos (JRC-GEEL)
Subject: MetroERM Interlaboratory Comparison: Expression of interest

Dear colleague,

The JRC is partner of the European Metrology Research Programme (EMRP) EN57
'Metrology for
Radiological Early Warning Networks in Europe' project
[<http://earlywarning-emrp.eu/>].

The JRC-IRMM under Task 3.2 'Traceability management of airborne radioactivity'
[http://earlywarning-emrp.eu/?page_id=1106] will organise and conduct an
interlaboratory comparison
(ILC) on the measurement of radionuclides in air filters, similar to the recent
ILC-REM exercise in 2014.
Such exercises contribute to the harmonisation of the radiological early warning
networks in Europe.

The particularity of the ILC is that the spiked filters will be made to the
specifications of each participant
on blank filters supplied by them. This time, the filters will be spiked with
137Cs, 134Cs and 131I (if the latter
is available timely), all three gamma-ray emitters.

The spiked air filters are expected to be dispatched to the participating
laboratories by end of January
2016. Participants' laboratories will be requested to send their results to
JRC-IRMM before the end
of February 2016.

The ILC will be conducted according to ISO 17043. The reported results will be
treated confidentially and
they will be presented without association with the name of the corresponding
lab (but they will be
available to the concerned lab).

You have initially expressed your interest in participating at the MetroERM ILC
by responding to the
corresponding questionnaire or you are a MetroERM partner. It would be
appreciated if you could

confirm your interest in participating at the MetroERM ILC by replying to this
e-mail as soon as possible
and no later than 4 December 2015. In case of too many interested laboratories
and because of our
limited human resources, the selection will be based on a combination of
geographical and 'first-come,
first-served' criteria.
Please, include in your e-mail the name of the responsible/contact person,
the telephone number and e-mail for the communication.

Looking forward to hearing from you.

Best regards,

Timotheos Altitzoglou
PS1 If you have received this e-mail more than once, we apologise.
PS2 If you know of other laboratories not in the mailing list that
might be interested, please
feel free to forward this e-mail to them.

The views expressed are purely those of the writer and may not any circumstances
be regarded as stating
an official position of the European Commission.

Timotheos ALTZITZOGLOU
Project Leader

European Commission
Directorate-General
Institute for Reference Materials and Measurements
Unit 04, Standards for Nuclear Safety, Security & Safeguards
Geel, Belgium
B-2440 Geel
+32 14 571266
JRC-IRMM-REM-COMPARISONS@ec.europa.eu
timotheos.altitzoglou@ec.europa.eu
<http://ec.europa.eu/dgs/jrc>
<http://irmm.jrc.ec.europa.eu>

2B. Invitation e-mail and accompanying letter

To:

skalinn@gmail.com; florian.smecka@ages.at; fath.cinar@taek.gov.tr;
lena_iv.1960@abv.bg; racraiova@yahoo.com;
giorgio.evangelisti@arpalazio.it; p.sabatini@arpa.umbria.it;
m.faureragan@arpa.vda.it; cristofaro.arpanm@ilberio.it;
rusconi@arpalombardia.it; bus@ire-elt.eu;
beata.slibritiene@aaa.am.it; Alasdair.morgan@cavendishnuclear.com;
elena.simion@anpm.ro; isajenko@clor.waw.pl; Peter.Jovanovic@zvd.si;
arunas.gudelis@fmcc.lt; Jani.Turunen (jani.turunen@stuk.fi);
maria.nikolaki@eeae.gr; r.totzeva@ncrrp.org; Branko Vodenik
(branko.vodenik@ijs.si); pieter.kwakman@rvm.nl; konyil@osski.hu;
cipuscasu@yahoo.com; denis.cindro@ijs.si; spni@dtu.dk;
zdenek.borecky@suro.cz; leen.verheyen@sccken.be;
Tomasz.Pliszczynski@ncbj.gov.pl; maduga@ctn.technico.ulisboa.pt;
marie.lecomte@ms.etat.lu; S.Somerville@epa.ie;
Florin.VLAD@apmm.anpm.ro; rodolfo.gurriaran@irsu.fr;
laura.porzio@arpa.piemonte.it; claudia.tabacaru@apmis.anpm.ro;
tbituh@imi.hr; acaballero@agl.moh.gov.cy;
mauro.magnoni@arpa.piemonte.it; rsogni@arpa.emr.it;
stefano.pegoretti@provincia.tn.it; johan.kastlander@foi.se;
Sofia.Eriksson@ssm.se; Alexander.Mauring
(alexander.mauring@nrpa.no); Bredo.Moller@nrpa.no;
arno.achatz@ages.at; jerzy.chytila@wsse.rzeszow.pl; fguelen@unex.es;
jbierring@bfs.de; APO@sis.dk; reszkam@op.pl;
jacek.kapala@umb.edu.pl; cvetelina.1980@abv.bg; romain.vidal@irsu.fr;
Michal.Bonczyk (mbonczyk@gig.eu); isabel.serrano@upc.edu;
clouvas@eng.auth.gr; Eia.Jakobson@keskkonnaamet.ee;
petr.rulik@suro.cz; jerzy.mietelski@ifj.edu.pl; Halim.Taskin@taek.gov.tr;
Jiri.Rada@sub.cz; nelida.florea@apmar.anpm.ro; fisico.rc@arpacal.it;
office@rzi-varna.com; giovanni.ferreri@bag.admin.ch;
mbruggem@sccken.be; nina.cernohlawek@bmlfuw.gv.at;
dabrow@paa.gov.pl; hasan.dikmen@taek.gov.tr; Laurent.ferreux@cea.fr;
catalina.gasco@ciemat.es; Petr.Kovar (pkovar@cmi.cz);
andris.abramenkovs@lvgtmc.lv; past@lvgtmc.lv; mary-
doris.gambin@gov.mt; public.health.lab@gov.mt;
faig.mutallimov@gmail.com; milli.monitoring@mail.ru
stefan.neumaier@ptb.de; Valentin.BUDEANU@cea.fr; DE CORT Marc
(JRC-ISPRA); DE LA CALLE GUNTINAS Maria Beatriz (JRC-GEEL); MALO
Petra (JRC-GEEL); MONDELAERS Willy (JRC-GEEL); HULT Mikael (JRC-
GEEL); DEHOUCQ Pieter (JRC-GEEL); ALTITZOGLOU Timotheos (JRC-GEEL)
Invitation to the MetroERM/ENV57 interlaboratory comparison on Cs-137,
Cs-134 and I-131 measurement in air filters
MetroERM-ILC_Invitation_Letter.pdf

Cc:

Subject:

Attachments:

Dear colleague,

Your laboratory has expressed interest in participating in the MetroERM/ENV57 interlaboratory comparison on Cs-137, Cs-134 and I-131 measurement in air filters organised by JRC-IRMM.

Please see the attached letter and follow the instructions.

In order to provide you with air filter spiked with the appropriate activity level, similar to what you measure routinely, and allow you to measure following your usual procedure and measurement geometry, we kindly ask you to reply to a questionnaire on sampling method, air filter used and measurement conditions and send us **two (2) new** (unused, blank) **air filters**. Both the questionnaire and the air filters must reach JRC-IRMM before **8 January 2016**.

The questionnaire can be found at:

<https://ec.europa.eu/eusurvey/runner/ILCMetroERM2016>

The password is: MetroERM2016 (case sensitive)

If you have further questions, please contact us. We will be happy to help you.

Email: JRC-IRMM-REM-COMPARISONS@ec.europa.eu

Looking forward to your participation in this comparison.

Yours sincerely,

Timotheos ALTITZOGLOU
Project Leader

Petya Malo
Logistics Assistant



European Commission

Directorate-General Joint Research Centre
Institute for Reference Materials and Measurements
Unit D4. Standards for Nuclear Safety, Security & Safeguards
Retiesweg 111

B-2440 Geel, Belgium

Tel.: +32 (0)15 275 2525

JRC-IRMM-REM-COMPARISONS@ec.europa.eu

<https://ec.europa.eu/jrc/en/irmm>

"The views expressed are purely those of the writer and may not in any circumstances be regarded as stating an official position of the European Commission."



Geel, 11 December 2015
JRC.D.47A/PM/ARES

Subject: Invitation to the MetroERM/ENV57 interlaboratory comparison on Cs-137, Cs-134 and I-131 measurement in air filters

Dear colleague,

The JRC is partner of the European Metrology Research Programme (EMRP) ENV57 'Metrology for Radiological Early Warning Networks in Europe' project (<http://earlywarning-emrp.eu/>).

The JRC-IRMM under Task 3.2 'Traceability management of airborne radioactivity' (http://earlywarning-emrp.eu/?page_id=1106) and to fulfil deliverables D3.2.4 to D3.2.7 will organise and conduct an interlaboratory comparison (ILC) on the measurement of ¹³⁷Cs, ¹³⁴Cs and ¹³¹I in air filters. Such exercises contribute to the harmonisation of the radiological early warning networks in Europe.

Your laboratory has expressed interest in participating in the above mentioned interlaboratory comparison organised by JRC-IRMM. In order to provide you with air filter spiked with the appropriate activity level, similar to what you measure routinely, and allow you to measure following your usual procedure and measurement geometry, we kindly ask you to reply to a questionnaire on sampling method, air filter used and measurement conditions. The questionnaire can be found at:

<https://ec.europa.eu/eusurvey/runner/ILCMetroERM2016>

The password is: MetroERM2016 (case sensitive)

You are also requested to send us **two (2) new** (unused, blank) **air filters** to the attention of:

Mrs. Petya MALO
European Commission
Directorate-General Joint Research Centre
Institute for Reference Materials and Measurements
Standards for Nuclear Safety, Security & Safeguards
Retieseweg 111
B-2440 Geel, Belgium

Please, reply to the questionnaire and send the blank air filters as soon as possible, but no later than **8 January 2016**.

Retieseweg, B-2440 Geel - Belgium. Telephone: +32 14 571 211 Fax: +32 14 594 273
Telephone: direct line +32 14 571 286, +32 14 571 286
E-mail: JRC-IRMM-REM-COMPARISONS@ec.europa.eu
Internet: <https://ec.europa.eu/jrc/institutes/irmm/>

Once we receive your answers to the questionnaire and the blank filters, we will give you instructions on how to register online for the interlaboratory comparison. The spiked samples with further instructions will be sent to the address specified in the registration form. The spiked filters are expected to be dispatched to the participating laboratories by mid February 2016. The participating laboratories will be requested to send their results to JRC IRMM via a web-based questionnaire in early March 2016. The report of the exercise is foreseen to be available in June 2016.

The reported results will be treated confidentially without disclosing the name of the corresponding lab. It is worth mentioning that this comparison is **not** in the series of interlaboratory comparisons organised for REM.

This interlaboratory comparison exercise is free of charge for the participating laboratories.

Should you have any question, please feel free to contact us at:

Email: JRC-IRMM-REM-COMPARISONS@ec.europa.eu

Looking forward to your participation in this comparison.


Yours sincerely,


Timotheos Alizizoglou

Project Leader

cc Messrs. Stefan Numeier (PTB, DE), Valentin Blideanu (CEA, FR)
Mr. Marc de Cort (JRC ITU)
Messrs. Willy Mondelaers, Pieter Dehouck, Mikael Hult, Ms. Beatriz de la Calle, Petya Malo (JRC IRMM)

2C. Registration information (e-mail)

<p>From: JRC-IRMM-REM-COMPARISONS To: tbtuh@imi.hr; Branko Vodenik (branko.vodenik@ijs.si); clouvas@eng.auth.gr; spni@dfu.dk; giovanni.ferreri@bag.admin.ch; nelida.florea@apmar.anpm.ro; cipuscasi@yahoo.com; radioactivitate@apnis.anpm.ro; nurdan.gungor@taek.gov.tr; racraiova@yahoo.com; rogn@arpa.emr.it; cristofaro.arpan@libero.it; elena.simion@anpm.ro; skalima@gmail.com; rodolfo.gurriaran@irs.fr; fisco.rc@arpacal.it; s.somerville@epa.ie; m.faureragan@arpa.vda.it; madrugac@ctn.tecnico.ulisboa.pt; lena_iv.1960@abv.bg; cvetelina.1980@abv.bg; eia.jakobson@keskkonnamet.ee; zdenek.borecky@suro.cz; florin.vlad@apmm.anpm.ro; florian.smecka@ages.at; Alexander.Mauring (alexander.mauring@nrpa.no); p.sabatini@arpa.umbria.it; romain.vidal@irs.fr; helena.mala@suro.cz; i.porzio@arpa.piemonte.it; fath.cinar@taek.gov.tr; Michal Bonczyk (mbonczyk@gig.eu); ronald.overwater@rivm.nl; strahlenschutz.linz@ages.at; guillen@unex.es; leen.verheyen@sccken.be; isabel.serrano@upc.edu; jerzy.chytla@wsse.rzeszow.pl; stefano.pegoretti@provincia.tn.it; johan.kastlander@roi.se; jiri.rada@suro.cz; isajenko@clor.waw.pl; ja.suarez@ciemmat.es; g.roselli@arpa.puglia.it; arunas.gudelis@ftmc.lt; tony.dieudonne@ire.eu; acaballero@sgl.moh.gov.cy; jacek.kapala@umb.edu.pl; jerzy.mietelski@fj.edu.pl; gorgio.evangelisti@arpalazio.it; bisi.hristova@mail.bg; m.magnoni@arpa.piemonte.it; r.totzeva@ncrrp.org; alasdair.morgan@cavendishnuclear.com; Jani Turunen (jani.turunen@stuk.fi); beata.silobritiene@aaa.am.lt; R.RUSCONI@ARPALOMBARDIA.IT; j.bieringer@bfs.de; Peter.Jovanovic@zvd.si; Sofia.Eriksson@ssm.se; APO@sis.dk; jakub.osko@ncbj.gov.pl; mary.doris.gambin@gov.mt; Bredo.Moller@nrpa.no; konyi@oski.hu; maria.nikolaki@eeae.gr; marielle.lecomte@ms.etat.lu</p> <p>Cc: MONDELAERS Willy (JRC-GEEL); HULT Mikael (JRC-GEEL); DEHOUCQ Pieter (JRC-GEEL); DE CORT Marc (JRC-SPRA); VALAKIS Stratos (JRC-SPRA); Stefan.Neumaier@ptb.de; Valentin.BUDEANU@cea.fr</p> <p>Subject: EMRP MetroRM/ENV57 interlaboratory comparison on Cs-137, Cs-134 and I-131 measurement in air filters - REGISTRATION</p> <p>Attachments: Instructions_for_registration_MetroRM_ILC_2016.docx</p> <p>Dear colleagues,</p> <p>Thank you for sending blank air filters to us. We are now in the process of spiking the filters with ¹³⁷Cs, ¹³⁴Cs and ¹³¹I and we expect to have them ready for dispatching to the participating laboratories by February 22, 2016.</p> <p>The reporting of laboratory results will be done via Internet. Therefore, we kindly ask you to register your laboratory via the following WEB link and send us the signed registration form: https://web.jrc.ec.europa.eu/jrcRegistrationWeb/registration/registration.do?seComparison=1561</p>	<p>The deadline for the registration is 26 February 2016</p> <p>Please read carefully the attached instructions for registration. Note that the registration is only definitive if we receive a signed registration form from you. You can send the signed form by e-mail, by fax or by post.</p> <p>Should you have any question, please feel free to contact us at: Email: JRC-IRMM-REM-COMPARISONS@ec.europa.eu</p> <p>Looking forward to hearing from you, we remain,</p> <p>Yours sincerely,</p> <p>Ms Petya MALO Administrative Assistant</p>  <p>European Commission Directorate-General Joint Research Centre (JRC) Institute for Reference Materials and Measurements (IRMM) Unit D.4. Standards for Nuclear Safety, Security and Safeguards Retieseweg 111 B-2440 Geel, Belgium phone: +32 14 571 286</p> <p>petya.malo@ec.europa.eu https://ec.europa.eu/jrc/en/institutes/irmm</p> <p><i>The views expressed are purely those of the writer and may not in any circumstances be regarded as stating an official position of the European Commission</i></p>
--	---

2D. Spiked air filter dispatch (e-mail) and accompanying letter

Ref. Ares(2016)968046 - 25/02/2016

Geel, 23 February 2016

Subject: MetroERM/ENV57 interlaboratory comparison on Cs-137, Cs-134 and I-131 measurement in air filters – spiked air filter dispatch

Dear colleague,

Today, the parcel containing the spiked air filter was dispatched to your laboratory by DHL courier from our site (IRMM, Geel, Belgium).

Please confirm reception of the sample by e-mail to JRC-IRMM-REM-COMPARISONS@ec.europa.eu. Check the spiked air filter and in case of any damage report to the above e-mail address.

You can track your parcel using the following URL

<http://www.dhl.com/en/express/tracking.html> and code «DHL_tracking_no».

The parcel contains:

- 1) Test item: spiked air filter
- 2) Accompanying letter (with instructions for your measurements and calculations)

The reporting of the results is done via the login page using the following URL:

<https://web.jrc.ec.europa.eu/ilcReportingWeb>

To report your results you need a password key which is unique to this interlaboratory comparison and your laboratory. You will receive your password key after the registration is closed (deadline 29 February 2016).

The deadline for reporting results and completing the questionnaire is **Tuesday, 29 March 2016**.

Should you have any question, please do not hesitate to contact us.

We wish you success with your measurements.

Kind regards,

Timotheos ALTITZOGLOU
Project Leader

Petya Malo
Logistics Assistant



European Commission
Directorate-General Joint Research Centre
Institute for Reference Materials and Measurements
Unit D4. Standards for Nuclear Safety, Security & Safeguards
Ref:esewig 111
B-2440 Geel, Belgium
+32 14 571266
PRC-IRMM-REF-COMPARISONS@ec.europa.eu
<https://ec.europa.eu/jrc>
<https://ec.europa.eu/jrc/institutes/jrcmm>

"The views expressed are purely those of the writer and may not in any circumstances be regarded as stating an official position of the European Commission."



EUROPEAN COMMISSION
DIRECTORATE-GENERAL
JOINT RESEARCH CENTRE
Directorate D - Institute for Reference Materials and Measurements
Standards for Nuclear Safety, Security and Safeguards

Geel, 22 February 2016
JRC.D.4/TA/pm/ARES(2016)

«Title» «First_Name» «Surname»
«Company»
«Department»
«Street_Name»
«Zip_Code» «City»
«Country»

Participation in the MetroERM/ENV57 interlaboratory comparison on the 'Measurement of ^{137}Cs , ^{134}Cs and ^{131}I in air filters'

Dear «Title» «Surname»,

Thank you for participating in the EC interlaboratory comparison on ^{137}Cs , ^{134}Cs and ^{131}I measurement in air filters.

Enclosed, you will find one of your blank filters spiked with ^{137}Cs , ^{134}Cs and ^{131}I activity either at a level you normally measure or higher than the threshold you indicated in the questionnaire.

Please confirm the receipt of the sample by e-mail to JRC-IRMM-REM-COMPARISONS@ec.europa.eu. Please check the spiked air filter and in case of any damage report to the above e-mail address.

This parcel contains:

- Test item: spiked air filter
- This accompanying letter.

Please keep this letter; you will need it to report your results.

On the filter you can distinguish the active spots in blue colour. This should make further preparation of the filter for measuring easier. Large filters had to be folded for shipment; the activity is on the inner side.

For the measurement and analysis, you should follow the procedure you routinely use.

Rellesweg 111, B-2440 Geel - Belgium. Telephone: (32-14)571-211.
Office: 010 00087. Telephone: direct line (32-14)571-266.
E-mail: JRC-IRMM-REM-COMPARISONS@ec.europa.eu

In case you do not measure the filter in the inner plastic bag, it is advisable to measure also the inner bag to make sure that no activity from the filter is remaining in the plastic bag.

The results must be reported for the **Reference Date of 1 March 2016 0:00 UTC**.

For your calculations, we recommend to use the data provided by the Decay Data Evaluation Project (DDEP) at http://www.nucleide.org/DDEP_WG/DDEPdata.htm. For your convenience, the most useful data are listed in the table below.

Nuclide	Half-life (days)	Energy (keV)	Photons per 100 disintegrations
^{137}Cs	10975 (29)	661.7	84.99 (20)
^{134}Cs	754.0 (5)	563.2	8.342 (15)
		569.3	15.368 (21)
		604.7	97.63 (8)
		795.9	85.47 (9)
^{131}I	8.0233 (19)	802.0	8.694 (16)
		284.3	6.14 (6)
		364.5	81.2 (5)
		637.0	7.12 (7)

The numbers in parentheses are the uncertainties on the last digit of the corresponding values.

In the case of ^{134}Cs and ^{131}I you may consider applying corrections for the true coincidence summing effects, if that is necessary. As usually the measurements are relatively long compared to the half-life of ^{131}I , a correction for the decay during the counting interval might also be necessary.

Finally, we recommend you to measure the spiked air filters as soon as possible after reception, as the activity of the relatively short-lived ^{131}I will be significantly reduced by time.

Reporting of the results

The reporting of the results is done via the login page using the following URL: <https://web.jrc.ec.europa.eu/jrcReportingWeb>

To report your results you need a password key which is unique to this interlaboratory comparison and your laboratory. Your password key will be communicated to you after the deadline for the registration (26 February 2016).

Please note that only **submitted** results will be taken into account, therefore, do not only **Save** your results but also click on the **Submit** button. Once you have submitted your results and questionnaire, please remember to send us a signed copy by e-mail to IRMM-REM-COMPARISONS@ec.europa.eu.

As you have been informed, the description of your analytical and measurement procedures will be collected via questionnaire using the same URL link as for reporting the results. We kindly ask you to answer all relevant questions regarding the procedures you employed for the measurement of the filter sample. Disregard questions which are

not relevant to the methods used in your laboratory. The uncertainty of the result must be reported in the same units as the activity concentration (i.e. Bq).

Please, notice that during the reporting of your results the *Cancel* button serves as an exit or return button.

The deadline for reporting results and completing the questionnaire is **Tuesday, 29 March 2016**.

Should you have any question, please do not hesitate to contact us.

We wish you success with your measurements.

Kind regards,

Timotheos ALTITZOGLOU
Project Leader

2E. Information to participants during the measurement exercise

From:
Sent:
To:

IRC-IRMM-REM-COMPARISONS

IRCI-IRMM-REM-COMP
15 March 2016 12:04

JRC-IRMM-REM-COMPARISONS

15 March 2016 22:04

florian.smecka@ages.at'; strahlenschutz.linz@ages.at';
henk.verheyen@skcken.be'; tony.dedonne@irelle.eu;
skckskalima@gmail.com'; lena_iv1960@adv.bg'; cirelina.1980@abv.bg';
bisl.hristova@mail.bg'; rtotzeva@ncrrp.org'; tbiuh@imi.hr;
accaballer@sgf.moh.gov.tr'; zdenek.borecky@suro.cz;
helena.mala@suro.cz'; jiri.rada@suro.cz'; spni@dtu.dk'; apo@sik.dk;
jane.jakobson@keskonnamet.se'; jani.turunen@stuk.fi;
rodolfo.gurriaran@esl.es'; romain.via@irsin.fr'; bjerneger@bfs.de;
clouvas@eng.auth.gr'; maria.nikolaik@eeae.gr'; konyi@oskii.hu';
s.somerville@epa.ie'; soggi@arpa.emi.it'; cristofaro@arpa@libero.it';
f.pisano@arpa.it'; f.pisano@arpa.piemonte.it';
p.sabatini@arpa.umbria.it'; l.porzio@arpa.piemonte.it';
stefano.pegoretti@provincia.tn.it'; g.roselli@arpa.puglia.it';
giorgio.evangelisti@arpalazio.it'; m.magnoni@arpa.piemonte.it';
r.usconi@arpalombardia.it'; a.unas.gedulis@frnc.lt';
maria.slabovirtova@aaai.it'; marriele.leclerc@ms.etat.lu'; hary-
alexander.gambin@gov.mt'; tonald.orewiler@ivm.nl';
Alexander.Mauring@nrpa.no'; breddo.moller@nrpa.no';
mbonczyk@sig.eu'; jerry.chylita@wawse.rzeszow.pl';
mbonczyk@lor.waw.pl'; jacek.kapala@umb.edu.pl';
jerry.j.mietelski@fj.edu.pl'; jakub.oskoc@ncbj.gov.pl';
madrugue@ctn.tecnico.ulisboa.pt'; helida.floreza@apmar.anpm.ro';
pascuas@yehoo.com'; radioactivitate@apmis.anpm.ro';
racraiova@yahoo.com'; elena.simion@anpm.ro';
florin.vlad@apmm.anpm.ro'; branko.vodenik@ijs.si';
peter.jovanovic@svd.si'; igulen@unex.es'; isabel.serrano@upc.edu';
jua.guerrero@ciemat.es'; pohan.kastlander@foi.se'; sofia.eriksson@smn.se';
giuseppe.ferrari@bag.adm.int'; nurdan.gunter@taek.gov.tr';
ninhai.kaya@taek.gov.tr'; alasdair.morgan@cawendishnuclear.com';
michel.bruggeman@skcken.be'; p.badalament@arpalombardia.it';
hasan.dikmen@taek.gov.tr'; ALTIZITOGUL Timotheos (JRC-GEEL); MALO
steyfa (JRC-GEEL); HULT Mikael (JRC-GEEL); DE CORT Marc (JRC-SPRA);
petra.finn-neumaier@ptb.de

Cc:

Subject:

Dear colleague.

You are probably in the course of measuring and evaluating the results of the spiked air filter for the MetroERM ILC.

As found by other participants and confirmed by our own measurements, part of the I-131 activity can be found in the inner plastic bag protecting your spiked filter during transport.

measured some of our control spiked filters (ϕ 70 mm) and found that 3–5 % of the activity was left in the plastic bag. For larger filters the fraction can be even 10% of the spiked activity, as reported by participants.

This makes our recommendation to measure the inner plastic bag more important.

If you measure the filter in its plastic bag, you need to do nothing else.

In case you measure the filter alone, it is strongly advisable to measure also the inner plastic bag and add the results to the value measured for the filter itself. Please, give the individual values in the comments at the end of the questionnaire as well.

My measurements showed no indication for Cs-137 and Cs-134. Both Cs-137 and Cs-134 do not leave traces in the plastic bag.

Good luck with your measurements.

Best regards,

Timos

Ms Petya MALO
Administrative Assistant



European Commission

Directorate-General Joint Research Centre (JRC)
Institute for Reference Materials and Measurements (IRMM)
Unit D.4, Standards for Nuclear Safety, Security and Safeguards

Retieseweg 111

8-2440 Geel, Belgium
phone: +32 14 571 286

petya.malo@ec.europa.eu

<https://ec.europa.eu/jrc>

<https://ec.europa.eu/jrc/en/institutes/irmm>

The views expressed are purely those of the writer and may not in any circumstances be regarded as stating an official position of the European Commission

2F. Reporting results

Ref. Ares(2016)1168938 - 08/03/2016

From: JRC-IRMM-REM-COMPARISONS
Sent: 07 March 2016 17:26
To: ALTITZOGLOU Timotheos (JRC-GEEL)
Cc: MALO Petya (JRC-GEEL)
Subject: MetroERM/ENV57 ILC on Cs-137, Cs-134 and I-131 measurement in air filters – submission of results
Attachments: MetroERM_2016_ILC_instructions_submission_results.pdf

Dear colleague,

The online system is now open to submit your results of the measurement of Cs-137, Cs-134 and I-131 in air filters for the ENV57/MetroERM ILC.

To access the online system, go to the login page using the URL <https://web.jrc.ec.europa.eu/ilcReportingWeb> and fill in your password key.

For your laboratory, the password key is **XXXXXXXXXXXX**.

Once logged in, you can report your results (ignore the column on the technique) and fill in the corresponding questionnaire (see attached file).

Please do not report your results through Excel.

You can save your results and answers to the questionnaire but they will be ready for submission only when you save + validate.

After you submit your results and questionnaire, please print them both by clicking at the appropriate link, sign them and send them by e-mail (JRC-IRMM-REM-COMPARISONS@ec.europa.eu), by fax (+32 14 584 273) or by post, to finalise your results submission.

We remind you that:

- Both the activity results and their uncertainties should be reported in the unit **becquerel (Bq)** (not mBq or %).
- The reference date is **01 March 2016, 00:00 UTC**.

Finally, the deadline for submitting your results and the questionnaire is **Tuesday, 29 March 2016**.

Should you face any problem with reporting, please feel free to contact us at JRC-IRMM-REM-COMPARISONS@ec.europa.eu

Best regards,

Timotheos ALTITZOGLOU
Project Leader

Petya Malo
Logistics Assistant



European Commission
Directorate-General Joint Research Centre
Institutional Relations and Communications
Unit D4: Standards for Nuclear Safety, Security & Safeguards
Ruesseweg 111
B-2440 Geel, Belgium
+32 14 571251
JRC-IRMM-REM-COMPARISONS@ec.europa.eu
<https://ec.europa.eu/jrc/en/institution>

"The views expressed are purely those of the writer and may not in any circumstances be regarded as stating an official position of the European Commission."

2G. Communication of preliminary results to the participants

 Ref. Ares(2016)5457470 - 20/09/2016

Geel, 20 September 2016

Subject: Preliminary results of the 2016 EMRP ENV57/MetroERM interlaboratory comparison (ILC) and new Laboratory Code

Dear «Title» «Surname»,

Thank you for your participation at the 2016 EMRP ENV57/MetroERM interlaboratory comparison (ILC) for the measurement of ^{137}Cs , ^{134}Cs and ^{131}I in air filters. Currently, we are working on the preparation of the final report. However, for your information we are sending you a preliminary evaluation of the results of this comparison in the attached PDF.

Anonymity being a requirement, each laboratory was assigned a new code number (different than the original one you received with the spiked filter), which is kept **confidential**.

The new code number for your laboratory is «New_Lab_Code».

In this preliminary report, the results are presented for all three radionuclides measured, i.e. ^{137}Cs , ^{134}Cs and ^{131}I , in the form of the ratio of the laboratory reported value to the reference value and the E_n number. More details you can find in the report itself. You can identify your laboratory in the graphs and find out how your reported results compare to the reference values.

The final report of this comparison exercise is foreseen to be available later in 2016. If you have any further questions with respect to this comparison, please feel free to contact me at JRC-IRMM-REM-COMPARISONS@ec.europa.eu.

Best regards,

Timos Altitzoglou

Timotheos ALTITZOGLOU
Project Leader

Petya Malo
Logistics Assistant




European Commission
Joint Research Centre (JRC)
Standards for Nuclear Safety, Security & Safeguards

Retieseweg 111
B-2440 Geel, Belgium
+32 14 571266
JRC-IRMM-REM-COMPARISONS@ec.europa.eu

EU Science Hub: <https://ec.europa.eu/jrc>

Connect with us on:





Contact: T. Altritzoglou
Timotheos.altritzoglou@ec.europa.eu
 Tel. +32 14 571266

Geel, September 19, 2016

Subject: Report of preliminary results of the 2016 ENV57/MetroERM Interlaboratory Comparison

In December 2015, the European Commission Directorate-General Joint Research Centre, Institute for Reference Materials and Measurements (JRC-IRMM) invited laboratories routinely measuring radioactivity in exposed air filters to participate at the 2016 EMRP ENV57/MetroERM interlaboratory comparison (ILC). The ILC was organized as part of the JRC contribution to the European Metrology Research Project (EMRP) ENV57/MetroERM "Metrology for radiological early warning networks in Europe". In fact, this work covers Deliverables D3.2.1 to D3.2.7 of Task 3.2 "Traceability management for airborne radioactivity".

The aim of this ILC (or Proficiency Test (PT)) exercise was to quantify airborne radioactivity field station measurement performance. The participating laboratories provided blank air filters of the type they routinely use, which the JRC spiked gravimetrically with known amount of radioactivity individually for each laboratory and returned to them on February 22 and 23, 2016. For this exercise, the filters were spiked with ¹³⁷Cs, ¹³⁴Cs and ¹³¹I with activities in the range of several tenths of Bq to a few Bq, close to those the laboratories routinely measure.

The ILC was accompanied by two relevant questionnaires to collect information about sample collection, measurement and analysis, as they are routinely carried out by the participating laboratories. The deadline to report the measurement results and submit the answers to the second questionnaire was set to March 29, 2016.

In total, 67 laboratories from 29 countries monitoring radioactivity in the environment participated in this ILC and the preliminary results are presented in this note in two ways:

1. As the ratio, R, of the reported by the participating laboratory activity value to the reference value, according to the formula $R = \frac{A}{A_0}$

where

A is the activity value reported by the participating laboratory
 A₀ is the assigned activity reference value

The ratios are graphically presented in Figure 1 for ¹³⁷Cs, Figure 3 for ¹³⁴Cs and Figure 5 for ¹³¹I.

Reading the ratio from the provided graphs and since each participant knows his/her reported results, he/she can easily calculate to a good accuracy the activity reference values, A₀, for his/her particular sample.

2. As the E_n number, according to the formula:

$$E_n = \frac{A - A_0}{\sqrt{U(A)^2 + U(A_0)^2}}$$

where

A is the activity value reported by the participating laboratory
 A₀ is the assigned activity reference value
 U(A) is the expanded uncertainty of a participant's result (k=2)
 U(A₀) is the expanded uncertainty of the assigned reference value (k=2)

The E_n number takes the expanded uncertainty of the reported results and that of the reference values into account in the analysis. The E_n numbers are interpreted as follows:

- If |E_n| ≤ 1, the laboratory values are compatible with the reference value;
- If |E_n| > 1, the laboratory values differ significantly from the reference values, the sources of deviation should be investigated and corrected, "warning signal";
- If |E_n| > 1.5, there is an urgent need to investigate and find the sources of the large deviation, "action signal".

The E_n numbers sorted in ascending order are graphically presented in Figure 2 for ¹³⁷Cs, Figure 4 for ¹³⁴Cs and Figure 6 for ¹³¹I.

To preserve anonymity, each participating laboratory has been given a unique Laboratory Code number (which you can find in the accompanying e-mail); using the Laboratory Code you can consult the graphs below to obtain your particular scores.

In the following pages you can find the graphs of the ratios and the E_n numbers, together with some comments, for all three radionuclides ¹³⁷Cs, ¹³⁴Cs and ¹³¹I, that were the subject of this ILC. A full report is in preparation, including details for the filter spiking and the analysis of the results reported by the participants.

The Reference Date for all results is: **1-Mar-2016 0:00 UTC**

Thank you once more for participating at this 2016 EMRP ENV57/MetroERM interlaboratory comparison.

Yours sincerely,

Timos Altritzoglou

2016 ENV57 / MetroERM Interlaboratory Comparison
Report of preliminary results

Nuclide: ¹³⁷Cs




Fig. 1. Ratio of ¹³⁷Cs activity per filter as reported by the participating laboratory over the spiked activity on the filter (JRC reference value) sorted in ascending order of the Laboratory Code. Red dashed lines indicate the 33% limit and blue dashed lines the 20% limit from the JRC reference value.




Fig. 2. E_n numbers for ¹³⁷Cs plotted in ascending order. Green color indicates compatible results, yellow indicates warning signal and red indicates action signal.

The majority of the laboratories reported reliable measurement results; only 3 out of 67 participants reported values with a percentage difference larger than $\pm 33\%$ of the reference value. Even with a stricter criterion, only 11 laboratories reported values with deviation larger than 20% from the reference value. Furthermore, 42 laboratories passed successfully the criterion of the compatibility test based on E_n numbers, 12 laboratories reported incompatible results with $|E_n| > 1$ and 13 with $|E_n| > 1.5$. The performance of the laboratories is very similar compared to the 2014 and 2003 EC interlaboratory comparisons on the measurement of ¹³⁷Cs in air filters.

Nuclide: ¹³⁴Cs




Fig. 3. Ratio of ¹³⁴Cs activity per filter as reported by the participating laboratory over the spiked activity on the filter (JRC reference value) sorted in ascending order of the Laboratory Code. Red dashed lines indicate the 33% limit and blue dashed lines the 20% limit from the JRC reference value.




Fig. 4. E_n numbers for ¹³⁴Cs plotted in ascending order. Green color indicates compatible results, yellow indicates warning signal and red indicates action signal.

Most of the laboratories reported reliable measurement results; only 4 out of 67 participants reported values with a percentage difference larger than $\pm 33\%$ of the reference value. Even with a stricter criterion, only 11 laboratories reported values with deviation larger than 20% from the reference value. For the E_n number compatibility test, 36 laboratories fulfilled the criterion, 7 laboratories reported incompatible results with $|E_n| > 1$ and 24 with $|E_n| > 1.5$.

Nuclide: ^{131}I




Fig. 5. Ratio of ^{131}I activity per filter as reported by the participating laboratory over the spiked activity on the filter (JRC reference value) sorted in ascending order of the Laboratory Code. Red dashed lines indicate the 33% limit from the JRC reference value.




Fig. 6. E_n numbers for ^{131}I plotted in ascending order. Green color indicates compatible results, yellow indicates warning signal and red indicates action signal.

In the case of ^{131}I , 32 of the 67 participating laboratories reported measurement results within the 33% limit from the reference value and 35 reported values with a percentage difference larger than $\pm 33\%$ from the reference value. Furthermore, only 11 laboratories passed successfully the criterion of the compatibility test based on E_n numbers; 9 laboratories reported incompatible results with $|E_n| > 1$ and 47 with $|E_n| > 1.5$. The majority of the laboratories reported lower activity values than the nominal reference activity value.




Analyzing the quality control air filters and point sources that were prepared together with the spiked air filters sent to the participating laboratories, it was found that the amount of activity measured on the air filters was lower than expected. As iodine is volatile, and despite the precautions taken, one can think that a variable amount of activity was lost from each spiked air filter by evaporation during the preparation phase. In addition, as initially indicated by some participants, some activity was also transferred to the plastic bag containing the air filter. Although the relative amount of that activity was not high, measuring the air filter without its plastic bag resulted in lower activity values.




Because of these two observations, the results from the measurement of ^{131}I in the air filters are non-conclusive. Indeed, those laboratories that reported correct values for ^{131}I , it is safe to say that they received air filters with the integral spiked activity and that they performed correctly the measurements. For those laboratories that reported significantly lower activity values for ^{131}I than expected, it could mean that

1. the laboratory received a spiked air filter with lower ^{131}I activity than the nominal, but performed the measurement correctly, or
2. the laboratory received a spiked air filter with the nominal ^{131}I activity, but did not perform correctly the measurement (including any eventual sample preparation), or
3. the laboratory received a spiked air filter with lower ^{131}I activity than the nominal, and did not perform correctly the measurement (including any eventual sample preparation).

We further investigate the case and more details will be given in the final report on the 2016 ENV57/MetroERM Interlaboratory Comparison.

2H. Invitation to the Workshop

 <p>EUROPEAN COMMISSION DIRECTORATE-GENERAL JOINT RESEARCH CENTRE Directorate D - Institute for Reference Materials and Measurements Standards for Nuclear Safety, Security and Safeguards</p>	<p>Ref. Ares(2016)583899 - 02/02/2016</p> <p>Geel, 2 February 2016 JRC.D.4/mh/MVDL/ARES(2016)</p>	<p>Subject: Invitation to ICS-REM and MetroERM air filter workshop on 7 April 2016</p>	<p>The JRC-IRMM is pleased to announce that the ICS-REM and MetroERM workshop on the measurement of radionuclides on air filters to be held on 7 April 2016 at its premises. The workshop will be organised in connection to the EURATOM Article 35/36 experts' meeting which will take place on 5 and 6 April, also in Geel.</p> <p>The most recent laboratory intercomparison (measurement of ^{137}Cs on air filters) performed under the scope of EURATOM Treaty Article 35 will be presented. Also the intercomparison performed under the EMRP-project (European Metrological Research Programme) MetroERM¹ involving the measurement of ^{131}I, ^{134}Cs and ^{137}Cs on air filters will be described.</p> <p>The workshop will include (i) background information, (ii) description of the sample preparation, (iii) presentation of the results, (iv) discussions on specific pit-falls and (v) presentations by some participants.</p> <p>The workshop will start at 09:00 and finish at 17:00 h. The workshop is free of charge.</p> <p>For registration, please contact our secretary: mira.van-de-lucht@ec.europa.eu.</p>	 Willy MONDELAERS Head of Unit SN3S	 Timothee ALTITZOGLOU Radionuclide Sector
			<p>¹ Metrology for early warning radiological networks in Europe; project number ENV57</p>		
			<p>Ref: Ares(2016)583899 - 02/02/2016</p>		

 <p>EUROPEAN COMMISSION DIRECTORATE-GENERAL JOINT RESEARCH CENTRE Directorate D - Institute for Reference Materials and Measurements Standards for Nuclear Safety, Security and Safeguards</p>	<p>Ref. Ares(2016)583899 - 02/02/2016</p> <p>Geel, 2 February 2016 JRC.D.4/mh/MVDL/ARES(2016)</p>	<p>Subject: Invitation to training course on basics of gamma-ray spectrometry - with special focus on analysis of air filters on 8 April 2016</p>	<p>The JRC-IRMM is pleased to announce that a training course on basics of gamma-ray spectrometry - with special focus on analysis of air filters will be given on April 8, 2016 at its premises. The training will be organised in connection to the EURATOM Article 35/36 experts' meeting which will take place on 5 and 6 April, and a workshop on intercomparisons of radioactivity on air filters on 7 April, also in Geel.</p> <p>The course will include lectures on:</p> <ul style="list-style-type: none"> • Basics of gamma-ray spectrometry • Low-level measurements • Some pit-falls in gamma-ray spectrometry • Specific issues related to gamma-ray spectrometry of air-filters • Sampling • Dispersion of radioactivity in air <p>The course will start at 09:00 and finish at 17:00 h. The course is free of charge.</p> <p>For registration, please contact our secretary: mira.van-de-lucht@ec.europa.eu.</p>	 Willy MONDELAERS Head of Unit SN3S	 Mikael HULT Head of Radionuclide Sector
			<p>Ref: Ares(2016)583899 - 02/02/2016</p>		

To:

Johannes.klimstein@ages.at; strahlenschutz.linz@ages.at;
florian.smecka@ages.at; faig.mutallimov@gmail.com;
milli.monitoring@mail.ru; liesel.sneyers@skcen.be; lsneyers@skcen.be;
tony.diudonne@ire-elt.eu; kdim@abv.bg; lab_vr@m-real.net;
cvetelina.1980@abv.bg; r.totzeva@ncrrp.org; bisi.hristova@mail.bg;
office@rzi-varna.com; d.deljkic@zjzfbih.ba; czcn1@zjzfbih.ba;
giovanni.ferreri@bag.admin.ch; stefan.roellin@babs.admin.ch;
acaballero@sgl.moh.gov.cy; zdenek.borecky@suro.cz; jiri.rada@suro.cz;
eva.sindelkova@subj.cz; petr.rulik@suro.cz; jbliringer@bfs.de;
Axel.Dalheimer@dwd.de; spni@dtu.dk; apo@sis.dk;
eia.jakobson@keskonnaamet.ee; stoulos@auth.gr;
constantinos.potiriadis@eeae.gr; kioanmid@cc.uoi.gr; f.legarda@ehu.es;
catalina.gasco@ciemat.es; alejandrosalasgarcia@gmail.com;
isabel.serrano@upc.edu; gtenorio@us.es; fimerall@ull.es;
rodolfo.gurriaran@irsu.fr; tbtuh@imi.hr; sebestyent@nebih.gov.hu;
kony@oski.hu; radiol@oski.hu; s.somerville@epa.ie;
fisco.rc@arpacal.it; carmela fortunato@arpab.it;
cristofaro.arpam@libero.it; lvitucci@arpa.puglia.it;
m.faureragani@arpa.vda.it; g.durante@arpacal.it;
r.rusconi@arpalombardia.it; p.sabatini@arpa.umbria.it;
l.porzio@arpa.piemonte.it; rogni@arpa.emr.it;
mauro.bonomi@provincia.tn.it; labor.apa@provincia.tn.it;
giorgio.evangelisti@arpalazio.it; leandro.magro@sprambiente.it;
m.magnoni@arpa.piemonte.it; mcappai@arpa.sardegna.it;
b.silobritiene@aaa.am.it; gudelisi@ktl.mil.it; marie.lecomte@ms.etat.lu;
andris.abramenkova@lvgnrc.lv; pasts@lvgnrc.lv; mary-
doris.gambin@gov.mt; gert-jan.knetsch@rvm.nl; bredo.moller@nrpa.no;
Alexander.Mauring@nrpa.no; isajenko@clor.waw.pl;
t.pilczynski@ncbj.gov.pl; madruga@ctn.ist.utl.pt; stracta@my-email.ro;
clpuscasu@yahoo.com; elena.simion@anpm.ro; racraiova@yahoo.com;
Florin.VLAD@apmm.anpm.ro; nelda.florea@apmar.anpm.ro;
radioactivitate@apmis.anpm.ro; catharina.soderstrom@foi.se;
sofia.eriksson@sm.se; jinkolic@vinca.rs; fath.cinar@taek.gov.tr;
keith.bulloch@phe.gov.uk; mike.froggatt@cavendishnuclear.com;
R.RUSCONI@ARPALOMBARDIA.IT; m.magnoni@arpa.piemonte.it;
tony.diudonne@ire.eu; bus@ire-elt.eu; beata.silobritiene@aaa.am.it;
giorgio.evangelisti@arpalazio.it; p.sabatini@arpa.umbria.it;
alasdair.morgan@cavendishnuclear.com; elena.simion@anpm.ro;
Isajenko@clor.waw.pl; m.faureragani@arpa.vda.it;
peter.jovanovic@zvd.si; arunas.gudelis@ftmc.it; jani.turunen@stuf.fi;
maria.nikolaki@eeae.gr; r.totzeva@ncrrp.org; branko.vodenik@ijs.si;
fath.cinar@taek.gov.tr; ronald.overwater@rvm.nl;
pieter.kwakman@rvm.nl; rsogni@arpa.emr.it; kony@oski.hu;
johan.kastlander@foi.se; clpuscasu@yahoo.com; denis.cindro@ijs.si;
cristofaro.arpam@libero.it; spni@dtu.dk; zdenek.borecky@suro.cz;
leen.verheyen@skcen.be; stefano.pegoretti@provincia.tn.it;
jakub.osko@ncbj.gov.pl; Tomasz.Pilczynski@ncbj.gov.pl;
racraiova@yahoo.com; sofia.eriksson@sm.se;

madruga@ctn.tecnico.ulisboa.pt; marie.lecomte@ms.etat.lu;
Bredo.Moller@nrpa.no; s.somerville@epa.ie;
Alexander.Mauring@nrpa.no; florin.vlad@apmm.anpm.ro;
skalinn@gmail.com; strahlenschutz.linz@ages.at; arno.achat@ages.at;
lena_iv.1960@abv.bg; rodolfo.gurriaran@irsu.fr;
l.porzio@arpa.piemonte.it; laura.porzio@arpa.piemonte.it;
radioactivitate@apmis.anpm.ro; claudia.tabacaru@apmis.anpm.ro;
tbtuh@imi.hr; acaballero@sgl.moh.gov.cy; florian.smecka@ages.at;
jbliringer@bfs.de; fguillen@unex.es; apo@sis.dk;
jerzy.chyha@wsse.rzeszow.pl; jerzy.mietelski@fj.edu.pl;
nurdan.gunor@taek.gov.tr; Halim.Taskin@taek.gov.tr;
eia.jakobson@keskonnaamet.ee; mbontczyk@ig.eu;
helena.mala@suro.cz; petr.rulik@suro.cz; cvetelina.1980@abv.bg;
romain.vida@irsu.fr; isabel.serrano@upc.edu; clouvas@eng.auth.gr;
jacek.kapala@umb.edu.pl; nelda.florea@apmar.anpm.ro;
jiri.rada@suro.cz; fisco.rc@arpacal.it; bisi.hristova@mail.bg;
giovanni.ferreri@bag.admin.ch; office@rzi-varna.com;
mbuggem@skcen.be; hasan.dikmen@taek.gov.tr; ja.suarez@ciemat.es;
catalina.gasco@ciemat.es; mary-doris.gambin@gov.mt;
public.health.lab@gov.mt; g.roselli@arpa.puglia.it
MONDELAERS Willy (JRC-GEL); DE CORT Marc (JRC-ISPR); HULT Mikael
(JRC-GEL); stefan.neumaier@ptb.de
Contribution to the Workshop on the 2014 EC interlaboratory comparison
and the MetroRM/ENV57 interlaboratory comparison on radioactivity
measurement in air filters

Cc:

Subject:

Dear colleague,

You have already been informed about the Workshop on the 2014 EC interlaboratory comparison and the MetroRM/ENV57 interlaboratory comparison on radioactivity measurement in air filters, to be held at EC-JRC-IRMM in Geel on April 7, 2016.

During the Workshop, we will present you with the details of these two ILCs and we will show and discuss the results.

As we see this Workshop as a forum to discuss the measurement of radioactivity in air monitoring filters

and to exchange ideas and experiences, we invite you to actively participate by presenting

- the method you use to prepare the filters for measurement

- the way you perform the measurements in your laboratory

- which geometry you use for the measurements and how you calibrate your detectors

- how you perform the analysis and the calculations

and by sharing your experiences and discussing any problem encountered.

We encourage you to send an e-mail to linotheos.alizoglou@ec.europa.eu with a few words on what you would like to present and a tentative title. Depending on the number of presentations, they can be from 5 to 15 minutes each. The sooner we get your expression of interest, the sooner we will be able to finalise the Workshop agenda.

Looking forward to hearing from you,

Best regards,

Timos Altizoglou
Project Leader

The views expressed are purely those of the writer and may not in any circumstances be regarded as stating an official position of the European Commission.

Dr. TIMOS ALTITZOGLOU



European Commission
Directorate-General Joint Research Centre
Institute for Reference Materials and Measurements
Unit D.4. Standards for Nuclear Safety, Security & Safeguards
Retieseweg 111
B-2440 Geel, Belgium
+32 14 571266
timotheos.altizoglou@ec.europa.eu
<http://ec.europa.eu/dg/jrc>
<http://irmm.jrc.ec.europa.eu>



OUR MISSION

As the Commission's in-house science service, the Joint Research Centre's mission is to provide EU policies with independent, evidence-based scientific and technical support throughout the whole policy cycle.

Working in close cooperation with policy Directorates-General, the JRC addresses key societal challenges while stimulating innovation through developing new methods, tools and standards, and sharing its know-how with the Member States, the scientific community and international partners.

Joint Research Centre

The European Commission's
in-house science service

European Commission
Joint Research Centre (JRC)
Institute for Reference Materials and Measurements
Retieseweg 111
2440 Geel, Belgium

Contact: Timotheos Altitzoglou

Tel.: +32 (0)14 571 266

E-mail: timotheos.altitzoglou@ec.europa.eu
mira.van-de-lucht@ec.europa.eu (secretary)

Serving society
Stimulating innovation
Supporting legislation

<https://ec.europa.eu/jrc/>

Measurement on simulated airborne particulates – ^{137}Cs in air filters

Workshop for the participants of the 2014 EC and the 2016 EVN57 – MetroERM measurement comparisons

Programme

Thursday 7 April 2016

Joint Research Centre
Institute for Reference Materials
and Measurements

© European Union, 2016



Measurement on simulated airborne particulates – ^{137}Cs in air filters JRC-IRMM – 7 April 2016

08:15 *Pick up at the hotel, Geel*

08:45 *Arrival at the JRC-IRMM entrance*

09:00 Welcome - Introduction – Overview
Timos Altitzoglou (JRC)

09:15 About EURDEP / REM
Marc De Cort (JRC)

09:45 Scope and organization of the
measurement comparisons
Borbala Maté & Timos Altitzoglou
(JRC)

10:05 Radionuclide standardization and
filter spiking
Timos Altitzoglou (JRC)

10:30 *Coffee break*

11:00 Filter measurement and Quality
Control
Timos Altitzoglou (JRC)

11:30 Results-Performance-Conclusions:
REM ILC
Borbala Maté (JRC)

11:50 Results-Performance-Conclusions:
MetroERM ILC
Timos Altitzoglou (JRC)

12:15 *Discussion*

12:45 *Lunch*

14:00 Radioactivity Air Monitoring
Stefan Neumaier (PTB)

14:30 Contribution by participants –
Short presentations by:
Sven Poul Nielsen (DTU, DK),
Bredo Møller (NRPA, NO),
Tereza Jezková (SÚRO, CZ),
Asser Nyander Poulsen (SIS, DK),
Paola Sabatini (ARPA Umbria, IT),
Lidia Silva (C2TN, PT),
Giovanni Ferreri (FOPH, CH),
Michèle Pallmer (Division de la
Radioprotection, LU)

15:45 *Coffee break*

16:00 *Laboratory visit*

17:00 *End*

17:15 *Departure to the hotel, Geel*



OUR MISSION

As the Commission's in-house science service, the Joint Research Centre's mission is to provide EU policies with independent, evidence-based scientific and technical support throughout the whole policy cycle.

Working in close cooperation with policy Directorates-General, the JRC addresses key societal challenges while stimulating innovation through developing new methods, tools and standards, and sharing its know-how with the Member States, the scientific community and international partners.

Joint Research Centre

The European Commission's
in-house science service

European Commission
Joint Research Centre (JRC)
Institute for Reference Materials and Measurements
Retieseweg 111
2440 Geel, Belgium

Contact: Mikael Hult

Tel: +32 (0)14 571 269

E-mail: mikael.hult@ec.europa.eu
mira.van-de-lucht@ec.europa.eu (secretary)

Serving society
Stimulating innovation
Supporting legislation

<https://ec.europa.eu/jrc/>

Basics of gamma-ray spectrometry and analysis of air filters

Training - Programme

Friday 8 April 2016

Joint Research Centre
Institute for Reference Materials
and Measurements

© European Union, 2016

Joint
Research
Centre

Basics of gamma-ray spectrometry and analysis of air filters

JRC-IRMM – 8 April 2016

08:15	Pick up at the hotel, Geel	14:45	Aerosol physics <i>Christina Isaxon, Lund University</i>
08:45	Arrival at the JRC-IRMM	15:30	Coffee break
09:00	Welcome and introduction to the SN35-unit <i>Wim Mondelaers</i>	16:00	Dispersion of radioactivity in the atmosphere <i>Johan Camps, SCK-CEN</i>
09:15	The Radionuclide Metrology Sector <i>Mikael Hult, JRC</i>	16:45	Discussion and evaluation <i>Mikael Hult, JRC</i>
09:30	Legislation and BSS: Marc de Cort, JRC	17:00	End
09:45	The basic equation for gamma-ray spectrometry <i>Mikael Hult, JRC</i>	17:15	Departure to the hotel, Geel
10:15	Coffee break		
10:45	Full Energy Peak Efficiency <i>Sean Collins, NPL</i>		
11:15	Coincidence summing <i>Tim Vidmar, SCK-CEN</i>		
11:45	Instrumentation and Low-level gamma- ray spectrometry <i>Mikael Hult, JRC</i>		
12:15	Lunch		
13:45	Gamma-ray spectrometry of air-filters <i>Michel Bruggeman, SCK-CEN</i>		
14:15	Quality System and LIMS <i>Michel Bruggeman, SCK-CEN</i>		

Annex 3: First questionnaire sent to the participants to collect information on air filters and sampling

Questionnaire for participants in ILC MetroERM 2016

Fields marked with * are mandatory.

1 Organisation and contact details

1.1 Person details

1.1.1 Title

1.1.2 First name

1.1.3 Family name

1.1.4 E-mail address

1.1.5 Gender

1.1.6 Telephone (with country code)

1.1.7 Fax (with country code)

1.2 Organisation details

1.2.1 Organisation

1.2.2 Organisation department

1.2.3 Organisation address

1.2.4 Zip code

1.2.5 City

1.2.6 Country

1.2.7 Mailing address (to which the filter will be sent back):

1.2.8 Please provide the complete mailing address here:

2 Laboratory

2.1 What is the type of your laboratory (more than one choice is possible)?

2.2 Please specify here:

★ 2.3 Is your laboratory certified, accredited or authorised (more than one choice is possible)?

- ☐ Certified (ISO 9000)
☐ Accredited according to ISO 17025
☐ Accredited according to other ISO(s)
☐ Authorised

2.4 Please provide for which methods and for which radionuclides you are accredited:

3 Measurement

- Please respect the indicated units when providing the values.
- Use a point as a decimal mark.

3.1 Sampling: provide the following parameters for the air sampling method

3.1.1 Supplier and model of air sampler:

3.1.2 Total air volume sampled: m3/filter

★ 3.1.3 Sampling period: hours

★ 3.1.4 Sampling frequency: ☐ daily ☐ weekly ☐ n

3.1.5 Please define here:

3.1.6 How many measurements does your laboratory perform per year?

- ☐ < 25
- ☐ 25-100
- ☐ > 100

3

3.2 Air filters: define the following information on the air filter material used

★ 3.2.1 Supplier of the air filter

★ 3.2.2 Type and material of the air filter

★ 3.2.3 Shape of the air filter:

- ☐ circular
- ☐ square
- ☐ rectangular

★ 3.2.4 Size of the air filter (in cm)

3.2.5 Blank weight (areal density): mg/cm²

3.2.6 Typical aerosol deposit collected: mg/cm²

3.2.7 Typical activity level measured for:

	Bq/filter
Cs-137	
Cs-134	
I-131	

★ 3.2.8 Would you like to get the filters spiked with coloured solution in order to optically identify the active spots?

- ☐ yes ☐ no

★ 3.2.9 Specify the requirements for the spiked filters (more than one choice is possible):

- ☐ Pattern ☐ Margin ☐ Other ☐ No special requirements

4

★ 3.2.10 Specify any required pattern for spiking the activity on the air filter:

★ 3.2.11 Specify any required margin to be left blank while spiking the air filter:

★ 3.2.12 Specify any other requirements:

3.3 Sample preparation and measurements

★ 3.3.1 What type of detector is used for the measurements?

- ☐ Ge(Li) detector
☐ HPGe detector
☐ BEGe detector
☐ Well type detector
☐ NaI(Tl) detector
☐ Other

*** 3.3.2 Please specify here:**

★ 3.3.3 Describe briefly the measurement methods, including sample preparation if any:

★ 3.3.4 What is the typical detection limit of your routine method for: (insert "no" if you do not measure this nuclide)

	Bq
Cs-137	
Cs-134	
I-131	

4 Additional information

4.1 Which radionuclides do you determine routinely in air filters?

- ☐ Be-7 ☐ K-40 ☐ Cs-134 ☐ Cs-137 ☐ I-131 ☐ Pb-210 ☐ Sr-90
- ☐ other

5

★ 4.2 Please specify here:

4.3 Comments on sampler, filter, sampling, measurements, etc.:

4.4 Comments on the questionnaire:

6

Annex 4: Responses of the participants to the questionnaire for collecting information on air filters and sampling

To determine the % ratios in *Table A-1* and the subsequent tables, the 67 participants were taken as 100%.

Table A-1. General information on the 2016 MetroERM ILC participating laboratories.

Question	Number of answers	Ratio (%)
<i>"What is the type of your laboratory (more than one choice is possible)?"</i>		
Research and development	23	34
Radioactivity in the environment	63	94
Monitoring of nuclear facilities	15	22
Fissile material control or safeguards	2	3
Governmental laboratory	32	48
University laboratory	6	9
Other	2	3
<i>"Is your laboratory certified, accredited or authorised (more than one choice is possible)?"</i>		
Certified (ISO 9000) and Accredited (ISO 17025)	10	15
Certified (ISO 9000)	2	3
Accredited according to ISO 17025	30	45
Authorised	13	19

Table A-2. Air filter size categories.

Size group	Size of air filter	Number of air filters	Ratio (%)
small	Ø 4.7-13 cm	29	44
intermediate	18x20-23.5x28 cm	11	17
large	24.6x41-60x70 cm	26	39

Table A-3. Information on the air sampling methods.

Question	Answer		
	Range	Average	Median
Small size group			
Total volume of air sampled/filter (m ³)	0.1-14000	1340	127.5
Sampling period (h)	3.5-300	75	24
Sampling frequency	Mostly on daily basis		
Intermediate size group			
Total volume of air sampled/filter (m ³)	18-30000	13029	11000
Sampling period (h)	2-720	206	168
Sampling frequency	Mostly on a weekly basis		
Large size group			
Total volume of air sampled/filter (m ³)	600-300000	105733	100000
Sampling period (h)	84-720	193	168
Sampling frequency	Mostly on a weekly basis		

Table A-4. Information on the technical part of the measurements.

Question	Number of answers	Ratio (%)
<i>"How many measurements (¹³⁷Cs in air filters) does your laboratory perform per year?"</i>		
< 25	9	13
25-100	21	31
> 100	37	55
<i>"Would you like to get the filters spiked with coloured solution in order to optically identify the active spots?"</i>		
Yes	60	90
No	7	10

<i>"Specify the requirements for the spiked filters:"</i>		
No special requirement	39	58
Pattern - requests for homogeneous distribution	15	22
Margin - active area	15	22
Other - air filter's side to spike, special folding, sealed bag/holder requirements	3	4.5
<i>"What type of detector is used routinely for the determination of ¹³⁷Cs in air filters?"</i>		
Ge(Li) detector	2	3
HPGe detector	58	87
BEGe detector	13	19
Well type detector	1	1.5
NaI(Tl) detector	0	0
Other	0	0
<i>"Do you determine any additional radionuclide(s) besides ¹³⁷Cs in the air filters?"</i>		
Yes	63	94
No	4	6

Annex 5: Type, material, size and form of the air filters used by the participants

Table A-5. Type, material size and form of air filters used by the participants and as reported by the participants.

Lab Code	Type and material of the air filter	Supplier	Filter size (cm)
1	cotton filters (high-quality cotton linters	Whatman	ø 4.7
2	HB5773	Hollingsworth & Vose	11x16
3	The Filters are made of refined pulp and linters with over 98% alpha-cellulose content	Sartorius	ø 5.5
4	PVC		60x60
5	Polypropylen filter type G-3	PTI Physik-Technik-Innovation	44x44
6	Polypropylen-Filter Typ G-3	PTI Physik-Technik-Innovation	60x60
7	glass fibre	F&J Specialty Products, Inc	ø 4.7
8	Model : No.880052.01-4	micropo	ø 10.2
9	Petrianov Filter FPP-15-1,5	PTI Physik-Technik-Innovation	44x44
10	Glass fiber. Model no: FP102M2	F&J Specialty Products, Inc	ø 10.2
11	Cellulose	Bernard Dumas, France	ø 11.5
12	Glass Microfibre Filters GF/A NO. 1820-866	Whatman	20x25
13	GF/A glass microfibre filter	Whatman	45x56
14	Petrianov filters type FPP-15-1.5	PTI Physik-Technik-Innovation	44x44
15	Petrianov FPP 15-1.5, PVC	Russian made	44x44
16	Glass microfiber filters, Grade GF/A	Whatman	30x23
17	Polymer	VF, a.s., Cerna Hora, Czech Republic	30x23
18	Glass microfibre, type GF/A 55 mm ø	Whatman	ø 5.5
19	glass fibre	F&J Specialty Products, Inc	ø 4.7
20	Petrianova FPP	CLOR	40x40
21	FP-47m , glass fiber filter	F&J Specialty Products, Inc	ø 4.7
22	Polypropylene	PTI Physik-Technik-Innovation	45x56
23	Part no. FP2063-47, Hydrophobic borosilicate glass microfibre with acrylic resin binder	Hi-Q Environmental Products, USA	ø 4.7

Lab Code	Type and material of the air filter	Supplier	Filter size (cm)
24	AP4004705 - glass fiber filters - with a 0.7 µm pore size	Millipore	ø 4.7
25	Polypropylene	3M	ø 12.7
26	Cellulose	PTI Physik-Technik-Innovation	44x44
27	Model NO. FP – 47	F&J Specialty Products, Inc	ø 4.7
28	Whatman GF/A Glassfibre	Whatman	45x56
29	Glass Microfibre Filters GF/A	Whatman	20x25
30	Petrianov fabric FPP-15-1.5 (Postchlorinated Polyvinylchloride PCV Fiber)		44x44
31	Whatman Glass Microfiber Filter GF/A	Whatman	45x56
32	Polypropylene (3M)		40x40
33	Chlorinated vinyl polychloride filter	Sorbent, Russian Federation	44x44
34	Glass fiber paper MODEL NO. FP-47M	F&J Specialty Products, Inc	ø 4.7
35	Filter Paper (Whatman)	Whatman	ø 4.7
36	Glass Fibre (Circular Plastic Holder)		ø 4.7
37	FPM 1545	AS ESFIL TEHNO, Estonia	45x56
38	ФПП-15-1,5 (FPP-15-1,5), a layer of ultra-thin fibers with an average size of 1.5 microns, deposited on the gauze	JSC Sorbent	60x70
39	Glass Fibre GF/A (Whatman)	Whatman	30x23
40	glass fiber filters (steel impactors)		ø 4.7
41	Fiberglass	TECNASA S.A.	44x44
42	FP-47m glass fiber filter	F&J Specialty Products, Inc	ø 4.7
43	Glass fibre sheets	Munktell	20x25
44	GF/A Cat. No 1820-866	Whatman	20x25
45	FPP-15	Esfil Tehno AS Sillamae, Estonia	87x74
46	Filter, polypropylene; type G3, PTI	PTI Physik-Technik-Innovation	45x56
47	Glass Fibre Filter	GELMAN SCIENCES	ø 4.7
48	Petryanov FPP_15_1.5	PTI Physik-Technik-Innovation	45x56
49	Petryanov's filters FPP-15-1.5		40x40
50	3M polypropylene	3m	45x56
51	Petrianov filter type FPP-15-1.5	(Russia)	44x44

Lab Code	Type and material of the air filter	Supplier	Filter size (cm)
52	GF6 Glass Fiber Filter	Whatman	ø 4.7
53	0.8 microns, nitrocellulose	Millipore	ø 4.7
54	MODEL FP-4.0 M	F&J Specialty Products	ø 10.2
55	PETRIANOV FILTER, FPP-15-1.5	Central Laboratory for Radiological Protection	40x40
56	type FPM 1530	ESFIL TEHNO AS, Estonia	30x23
57	quartz microfibre filters	FILTER -LAB	ø 10.2
58	HB5773, Glasfiber	Hollingworth	45x56
59	Whatman Glass Microfiber Filters GF/A	GE Healthcare Life Sciences	45x56
60	Pads A500 GS, glass fibre	Camfil	ø 7.7
61	EPM 2000 Glass Microfibre Filter	Whatman International Ltd, Maidstone, England.	20x25
62	Flat filter paper, SIO2	TRM Filter d.o.o.	30x30
63	Glass fibre papers	Schleicher & Schuell	ø 4.7
64	F&J model no. FP-4.0M, borosilicate microfiber	F&J Specialty Products, Inc	ø 10.2
65	glas fibre	F&J Specialty Products, Inc	ø 4.7
66	Glass Fiber Filter	GELMAN SCIENCES	ø 4.7
67	Glass Fiber Filter	F&J Specialty Products, Inc	ø 4.7

Annex 6: Sequence of spiked filters and quality control sources prepared for the 1026 ENV57/MetroERM ILC exercise

Table A-6. Sequence of spiked filters and quality control sources prepared for the 2016 ENV57/MetroERM ILC exercise.







Lab Code	Sample ID	Source Type	Solution
	IRMM-01	Filter	E3
65	Filter_006	Filter	E3
19	Filter_023	Filter	E3
54	Filter_057	Filter	E3
12	Filter_053	Filter	E3
	PS_020	Point Source	E3
	IRMM-02	Filter	E4
	PS_021	Point Source	E4
53	Filter_041	Filter	E2A
52	Filter_019	Filter	E2A
42	Filter_051	Filter	E2A
18	Filter_055	Filter	E2A
	IRMM-03	Filter	E2A
3	Filter_047	Filter	E2A
35	Filter_043	Filter	E2A
36	Filter_045	Filter	E2A
47	Filter_013	Filter	E2A
	IRMM-04	Filter	E2A
67	Filter_010	Filter	E2A
24	Filter_007	Filter	E2A
23	Filter_002	Filter	E2A
57	Filter_028	Filter	E2A

Lab Code	Sample ID	Source Type	Solution
66	Filter_011	Filter	E2A
21	Filter_029	Filter	E2A
40	Filter_067	Filter	E2A
7	Filter_026	Filter	E2A
34	Filter_020	Filter	E2A
	IRMM-05	Filter	E2A
	PS_022	Filter	E2A
	IRMM-06	Filter	E1
8	Filter_059	Filter	E1
27	Filter_064	Filter	E1
	PS_023	Point Source	E1
25	Filter_012	Filter	E4
44	Filter_024	Filter	E4
10	Filter_027	Filter	E4
29	Filter_015	Filter	E4
11	Filter_005	Filter	E4
61	Filter_054	Filter	E4
	IRMM-07	Filter	E4
	PS_024	Point Source	E4
2	Filter_058	Filter	E4
64	Filter_065	Filter	E4
49	Filter_004	Filter	E5

Lab Code	Sample ID	Source Type	Solution
62	Filter_066	Filter	E5
16	Filter_034	Filter	E5
	IRMM-08	Filter	E5
59	Filter_032	Filter	E5
9	Filter_008	Filter	E5
	PS_025	Point Source	E5
20	Filter_022	Filter	E5
32	Filter_001	Filter	E5
60	Filter_046	Filter	E5
56	Filter_039	Filter	E5
48	Filter_025	Filter	E5
15	Filter_052	Filter	E5
51	Filter_016	Filter	E5
46	Filter_017	Filter	E5
38	Filter_040	Filter	E5
	IRMM-09	Filter	E5
31	Filter_063	Filter	E4
63	Filter_056	Filter	E2A
37	Filter_033	Filter	E5
13	Filter_037	Filter	E5
55	Filter_049	Filter	E5
28	Filter_061	Filter	E5
	IRMM-10	Filter	E5
5	Filter_042	Filter	E5
43	Filter_031	Filter	E5
4	Filter_060	Filter	E5
	IRMM-11	Filter	E5

Lab Code	Sample ID	Source Type	Solution
30	Filter_048	Filter	E5
26	Filter_062	Filter	E5
45	Filter_021	Filter	E5
	IRMM-12	Filter	E2A
1	Filter_038	Filter	E2A
39	Filter_003	Filter	E2A
50	Filter_018	Filter	E5
22	Filter_036	Filter	E5
58	Filter_014	Filter	E5
	IRMM-14	Filter	E5
	PS-026	Filter	E5
41	Filter_050	Filter	E5
17	Filter_030	Filter	E5
14	Filter_035	Filter	E5
	IRMM-15	Filter	E5
33	Filter_044	Filter	E5
6	Filter_009	Filter	E5

Annex 7: Examples of spiked filters

		
a)	b)	c)
		
d)	e)	
		
f)		

Annex 8: Quality control filters and sources; measurement results and their deviation from the reference activity values

Table A-7a. Gamma-ray spectrometric results of the quality control filters; measured values and the corresponding uncertainties.

Sample_ ID (IRMM- ##)	DILUTION	DETECTOR	¹³⁷ Cs		¹³⁴ Cs		¹³¹ I	
			Measured Activity (Bq)	Uncertainty (Bq) (k=1)	Measured Activity (Bq)	Uncertainty (Bq) (k=1)	Measured Activity (Bq)	Uncertainty (Bq) (k=1)
01	E3	B	0.540	0.032	0.800	0.036	0.626	0.024
02	E4	B	0.089	0.013	0.200	0.026	0.164	0.012
03	E2A	B	0.546	0.038	0.887	0.046	0.442	0.018
04	E2A	B	0.484	0.027	0.733	0.031	0.414	0.015
04	E2A	B	0.482	0.028	0.736	0.031	0.427	0.017
05	E2A	B	0.483	0.034	0.655	0.037	0.369	0.016
06	E1	B	1.935	0.092	2.138	0.080	1.408	0.050
07	E4	A	0.082	0.011	0.145	0.016	0.103	0.024
08	E5	B	0.101	0.011	0.168	0.018	0.121	0.007
09	E5	B	0.141	0.010	0.264	0.019	0.254	0.011
10	E5	B	0.087	0.007	0.176	0.013	0.210	0.009
11	E5	A	0.115	0.012	0.223	0.014	0.200	0.009
12	E2A	B	0.380	0.019	0.574	0.022	0.747	0.038
14	E5	B	0.089	0.007	0.185	0.014	0.175	0.011
15	E5	A	0.081	0.008	0.186	0.012	0.126	0.008

Table A-7b. Gamma-ray spectrometric results of the quality control filters; their relative deviation from the reference values and their ratio to the reference values.

Sample_ID (IRMM-##)	¹³⁷ Cs			¹³⁴ Cs			¹³¹ I		
	Deviation from Ref. Activity (%)	Activity RATIO (Measured to Reference)	Uncertainty (k=2)	Deviation from Ref. Activity (%)	Activity RATIO (Measured to Reference)	Uncertainty (k=2)	Deviation from Ref. Activity (%)	Activity RATIO (Measured to Reference)	Uncertainty (k=2)
01	6.1	1.06	0.13	-4.4	0.96	0.09	-47	0.53	0.04
02	-6.0	0.94	0.28	8.4	1.08	0.28	-45	0.55	0.08
03	-1.1	0.99	0.14	-3.7	0.96	0.10	-60	0.40	0.03
04	-6.1	0.94	0.10	-14.7	0.85	0.07	-60	0.40	0.03
04	-6.5	0.94	0.11	-14.4	0.86	0.07	-59	0.41	0.03
05	4.3	1.04	0.15	-15.2	0.85	0.10	-60	0.40	0.03
06	3.8	1.04	0.10	-3.7	0.96	0.07	-55	0.45	0.03
07	-5.5	0.94	0.26	-14.7	0.85	0.19	-62	0.38	0.17
08	25.2	1.25	0.27	-0.8	0.99	0.22	-55	0.45	0.05
09	8.6	1.09	0.15	-3.4	0.97	0.14	-42	0.58	0.05
10	12.2	1.12	0.19	8.3	1.08	0.16	-20	0.80	0.07
11	10.4	1.10	0.24	1.0	1.01	0.13	-44	0.56	0.05
12	-16.5	0.84	0.08	-24.5	0.75	0.06	-18	0.82	0.08
14	5.2	1.05	0.16	3.7	1.04	0.16	-39	0.61	0.07
15	-0.3	1.00	0.21	8.8	1.09	0.14	-54	0.46	0.06

Table A-8a. Gamma-ray spectrometric results of the quality control point sources; measured values and the corresponding uncertainties.

Sample_ ID (PS_##)	DILUTION	DETECTOR	¹³⁷ Cs		¹³⁴ Cs		¹³¹ I	
			Measured Activity (Bq)	Uncertainty (Bq) (k=1)	Measured Activity (Bq)	Uncertainty (Bq) (k=1)	Measured Activity (Bq)	Uncertainty (Bq) (k=1)
15	D1	A	30.9	1.1	37.0	1.3	24.2	0.9
15	D1	B	31.0	1.4	33.5	1.2	24.1	0.9
16	D3	B	1.40	0.07	2.13	0.08	2.30	0.09
17	D4	A	0.23	0.02	0.45	0.03	0.53	0.03
18	D5	B	0.36	0.03	0.69	0.04	0.94	0.05
19	D2A	A	1.93	0.08	3.42	0.13	2.23	0.09
20	E3	B	0.64	0.04	0.95	0.04	0.77	0.03
21	E4	A	0.11	0.01	0.20	0.02	0.19	0.01
22	E2A	A	0.65	0.03	1.05	0.05	0.59	0.03
23	E1	B	1.41	0.07	1.58	0.06	0.83	0.04
24	E4	B	0.18	0.01	0.14	0.02	0.12	0.01
25	E5	B	0.058	0.006	0.12	0.02	0.12	0.01
26	E5	A	0.063	0.010	0.13	0.01	0.15	0.01

Table A-8b. Gamma-ray spectrometric results of the quality control point sources; their relative deviation from the reference values and their ratio to the reference values.

Sample _ID (PS_# #)	¹³⁷ Cs			¹³⁴ Cs			¹³¹ I		
	Deviation from Ref. Activity (%)	Activity RATIO (Measured to Reference)	Uncer- tainty (k=2)	Deviation from Ref. Activity (%)	Activity RATIO (Measured to Reference)	Uncer- tainty (k=2)	Deviation from Ref. Activity (%)	Activity RATIO (Measured to Reference)	Uncer- tainty (k=2)
15	0.33	1.00	0.07	1.1	1.01	0.07	-53	0.47	0.03
15	0.7	1.01	0.09	-8.4	0.92	0.07	-54	0.46	0.03
16	1.45	1.01	0.10	-6.1	0.94	0.07	-28	0.72	0.05
17	5.8	1.06	0.18	6.9	1.07	0.13	-22	0.78	0.08
18	4.4	1.04	0.15	-4.5	0.95	0.10	-19	0.81	0.08
19	-3.5	0.97	0.08	2.4	1.02	0.08	-45	0.55	0.04
20	-5.6	0.94	0.10	-14.6	0.85	0.07	-51	0.49	0.04
21	4.4	1.04	0.22	0.8	1.01	0.15	-41	0.59	0.08
22	-5.25	0.95	0.09	-8.9	0.91	0.08	-57	0.43	0.04
23	-1.5	0.98	0.10	-7.3	0.93	0.07	-66	0.34	0.03
24	2.3	1.02	0.19	-34	0.66	0.17	-64	0.36	0.06
25	16.0	1.16	0.24	16.3	1.16	0.34	-28	0.72	0.11
26	13.8	1.14	0.35	11.6	1.12	0.22	-21	0.79	0.10




Fig. A-1. Ratio of measured-to-reference activity and expanded uncertainty ($k=2$) of the quality control **filters** prepared from the ^{137}Cs , ^{134}Cs and ^{131}I dilutions used for spiking the air filters. The blue diamond symbol is used for results obtained with Detector A and the red diamond symbol for those obtained with Detector B. The solid blue line is the Reference value (unity), the blue dashed lines designate the $\pm 20\%$ range and the red dashed lines the $\pm 33\%$ range.

Top: ^{137}Cs , **Middle:** ^{134}Cs , **Bottom:** ^{131}I




Fig. A-2. Ratio of measured-to-reference activity and expanded uncertainty ($k=2$) of the quality control **point sources** prepared from the ^{137}Cs , ^{134}Cs and ^{131}I dilutions used for spiking the air filters. The blue diamond symbol is used for results obtained with Detector A and the red diamond symbol for those obtained with Detector B. The solid blue line is the Reference value (unity), the blue dashed lines designate the $\pm 20\%$ range and the red dashed lines the $\pm 33\%$ range.

Top: ^{137}Cs , **Middle:** ^{134}Cs , **Bottom:** ^{131}I

Annex 9: Reporting form and second questionnaire for submitting the measurement results

Milc questionnaire

Comparison for Cs-137, Cs-134 and I-131 measurement in air filters

As it was previously announced, the reporting of your measurement method and results will be done via an online questionnaire. We kindly ask you to answer all relevant questions regarding the procedures you used for the measurement of the filter sample. Questions marked with an asterisk (*) are mandatory to answer. Please note that only upon reception of the signed reporting form your submission is final! Be aware that the deadline for the submission of the results and the questionnaire is Tuesday, 29 March 2016. Thank you for your cooperation!

Submission Form

1. Sample treatment

1.1. Was the comparison sample treated according to the same analytical procedure as routinely used in your laboratory for the same type of samples? *

☐ a) Yes
☐ b) No

1.1.1. If not, please specify the differences here: *

1.2. Did you measure the sample together with the inner plastic foil or did you measure the foil separately? *

☐ a) Together
☐ b) The plastic foil was measured separately
☐ c) The plastic foil was NOT measured

1.2.1. If separately, please provide the measured activity of the plastic foil (in Bq) with its uncertainty: *

- Page 1 of 10 -

1.3. Did you apply any preconcentration or chemical treatment? *

☐ a) Yes
☐ b) No

1.3.1. If yes, please describe the method (chemicals and procedures used, chemical recovery obtained, method of chemical recovery determination, etc.): *

1.4. Describe the method used for the sample preparation (more than one choice is possible): *

☐ a) Packing together with blank air filters
☐ b) Folding
☐ c) Pressing/compressing
☐ d) No sample preparation
☐ e) Other

1.4.1. If other, please specify here: *

2. Equipment

2.1. Which type of detector was used for the measurements (more than one choice is possible)? *

☐ a) Ge(Li) detector
☐ b) HPGe detector
☐ c) BEGe detector
☐ d) Well type detector
☐ e) NaI(Tl) detector
☐ f) Other

- Page 2 of 10 -

124

2.1.1. If other, please specify here: *

2.1.2. Provide the supplier, the model of the detector and its relative efficiency:

2.1.3. Describe how the efficiency calibration of the detector is done (to determine the counting efficiency for the air filter measurement) and provide the source of nuclear data used: *

2.2. What type of electronics and data acquisition was used? *

☐ a) Analog signal processing (spectroscopy amplifier, etc)

☐ b) Digital signal processing

☐ c) Other

2.2.1. If other, please specify here: *

2.3. Specify the software for peak area determination and data evaluation used (more than one choice is possible): *

☐ a) Genie 2000 family

☐ b) GammaVision

☐ c) InterWinner

☐ d) Gamma-W

☐ e) EMCA+

☐ f) Gamma-track

☐ g) SAMPO

☐ h) HYPERMET

- Page 3 of 10 -

☐ i) Other

2.3.1. If other, please specify here: *

3. Measurement and data evaluation

3.1. Indicate the period during which you performed the measurements. *

See table **Measurements period** at bottom

3.2. Fill in the following table: *

See table **Acquisition time and measurement cycles** at bottom

3.3. Provide the limit of detection (value in Bq): *

See table **Detection limit** at bottom

3.4. Provide information on the calculation method of the limit of detection (software and/or method used): *

☐ a) Genie 2000

☐ b) GammaVision

☐ c) Curie method

☐ d) ISO 11929

☐ e) DIN 25482

☐ f) Risø method

☐ g) Other method and/or software

- Page 4 of 10 -

125

3.4.1. If other method and/or software, please specify here: *

3.5. Did you use the recommended half-lives and gamma-ray emission intensities? *

☐ a) Yes

☐ b) No

3.5.1. If not, please indicate the values you used: *

See table **Half-lives and gamma-ray emission intensities** at bottom

3.5.2. Please indicate the reference for the values you used:

3.6. Did you apply corrections for the coincidence summing? *

☐ a) Yes

☐ b) No

3.6.1. If yes, please give the correction factors you used in the table below: *

See table **Coincidence summing correction factors** at bottom

3.7. Did you apply a correction for the decay of I-131 during the counting interval? *

☐ a) Yes

☐ b) No

- Page 5 of 10 -

3.7.1. If yes, please insert the correction factor below: *

4. Uncertainty budget *

See table **Uncertainty budget** at bottom

4.1. If you indicated "Other" in the table above, please specify the component(s):

5. Additional information

5.1. Difficulties encountered

5.2. Comments on this interlaboratory comparison exercise:

5.3. Comments on this questionnaire:

- Page 6 of 10 -

Acquisition time and measurement cycles

Questions/Response table	Acquisition time (sec)	Measurement cycles	Count rate for Cs-137 peak (counts/sec)	Count rate for the 605-KeV Cs-134 peak (counts/sec)	Count rate for the 364-KeV I-131 peak (counts/sec)
Sample					
Background/blank					

Coincidence summing correction factors

Questions/Response table	Coincidence summing correction factor
Cs-134 (604.7 KeV)	
Cs-134 (795.9 KeV)	
I-131 (364.5 KeV)	
I-131 (637.0 KeV)	

Detection limit

- Page 7 of 10 -

Questions/Response table	Detection limit (Bq)
Cs-137	
Cs-134	
I-131	

Half-lives and gamma-ray emission intensities

Questions/Response table	Half-life used (days)	Gamma-ray emission intensity used (in %)
Cs-137 (661.7 KeV)		
Cs-134 (604.7 KeV)		
Cs-134 (795.9 KeV)		
I-131 (364.5 KeV)		
I-131 (637.0 KeV)		

Measurements period

Questions/Response table	Start date	End date
Measurements period		

- Page 8 of 10 -

Uncertainty budget

The combined relative standard uncertainty (i.e. the quadratic sum of the uncertainty components) should be the same as the standard uncertainty reported with your results.

<i>Questions/Response table</i>	<i>Relative uncertainty component (%) - Cs-137</i>	<i>Relative uncertainty component (%) - Cs-134</i>	<i>Relative uncertainty component (%) - I-131</i>
<i>Counting statistics</i>			
<i>Blank/background measurement</i>			
<i>Efficiency calibration (including the activity of the calibration source)</i>			
<i>Chemical yield (if applicable)</i>			
<i>Self-absorption in the source</i>			
<i>Activity of the calibration source</i>			
<i>Decay correction</i>			
<i>Other 1</i>			
<i>Other 2</i>			
<i>Other 3</i>			
<i>Combined relative standard uncertainty (quadratic sum of the components)</i>			

Annex 10: Responses to the second questionnaire accompanying the reporting form

To determine the ratios, the 67 participants were taken as 100%.

Table A-9. Treatment of spiked air filters before measurement by the participants and equipment used for the measurement of the spiked air filters.

Question	Number of answers	Ratio (%)
<i>"Were the comparison sample treated according to the same analytical procedure as routinely used in your laboratory for the same type of samples?"</i>		
Yes	62	92.5
No	5	7.5
<i>"Did you measure the sample together with the plastic foil or did you measure the foil separately?"</i>		
Together	33	49
Separately	26	39
The plastic foil was not measured	8	12
<i>"Did you apply any pre-concentration or chemical treatment?"</i>		
Yes	0	0
No	67	100
<i>"Which type of detector was used for the determination of ^{137}Cs in the air filter?"</i>		
Ge(Li) detector	2	3
HPGe detector	58	87
BEGe detector	13	19
Well type detector	1	1.5
NaI(Tl) detector	0	0
Other	0	0
<i>"What type of electronics and data acquisition was used?"</i>		
a) Analog signal processing (spectroscopy amplifier, etc.)	19	21
b) Digital signal processing	72	79
c) Other	0	0
Use of recommended half-lives		
Yes	59	88

No	8	12
Coincidence summing correction applied		
Yes	48	72
No	19	28
Decay during measurement correction applied (for ^{131}I)		
Yes	48	72
No	19	28

Table A-10. Measurement parameters; acquisition time, count rates, MDA.

	Minimum	Maximum	Mean	Median
Measurement time of sample (h)	1	233	53.0	47.9
Count rate for ^{137}Cs peak in sample (cps)	0.0011	0.153	0.0176	0.0132
Count rate for the 605 keV ^{134}Cs peak (cps)	0.0015	0.174	0.0267	0.0264
Count rate for the 364 keV ^{131}I peak (cps)	0	0.314	0.0347	0.0209
Measurement time of background (h)	15	333	80.8	63.0
Count rate for ^{137}Cs peak in background (cps)	0	0.124	0.0032	-
Count rate for the 605 keV ^{134}Cs peak in background (cps)	0	0.161	0.0038	-
Count rate for the 364 keV ^{131}I peak in background (cps)	0	0.039	0.0022	-
MDA for ^{137}Cs (Bq)	0.003	0.12	0.039	0.029
MDA for ^{134}Cs (Bq)	0.002	0.53	0.046	0.025
MDA for ^{131}I (Bq)	0.0002	0.50	0.051	0.025

Annex 11: Coincidence summing correction factors as applied by the participants

Table A-11. Coincidence summing correction factors as applied and reported by the participants.

Lab Code	¹³⁴ Cs		¹³¹ I	
	604.7 keV	795.9 keV	364.5 keV	637.0 keV
1	n.a.- software determined	n.a.- software determined	n.a.- software determined	n.a.- software determined
2				
3	1.02	1.018		
4				
5	1.15	1.15		
6	0.867	0.864	1.005	1
7	0.823	0.817	1.01	1
8				
9	0.801	0.799	1	not used
10	0.76532	0.76795	1	not used
11	1.17	1.168	0.997	0.999
12	0.788	0.787	free	free
13	1.22	1.23	1	1
14	0.76497	0.76133	1.0092	1.0017
15				
16	by software	by software	by software	by software
17	1.137	1.138		
18	0.899	0.898	1.004	
19	0.818	0.815		
20	1.25	1.25	1	1
21	0.795	0.789	1.01	free
22				
23	0.87	0.87	1.01	1
24	1.269	1.267	0.995	0.998
25	0.78363	0.78101	1.0097	1
26	0.851	0.85	1.003	free
27				
28				
29	0.939	0.928	1.009	
30				
31	1.118	1.117		
32	1.277		1	
33	0.87175	0.87174	1	
34	0.0825	0.821		
35				
36				
37	1.14	1.14	0.998	0.999
38				
39	0.947	0.946	1.002	

Lab Code	¹³⁴ Cs		¹³¹ I	
	604.7 keV	795.9 keV	364.5 keV	637.0 keV
40				
41				
42	0.828	0.825	1	free
43	1.07	1.14	0.97	1.08
44	0.79499	0.79405	1.0092	
45	NA	0.93	0.97	NA
46	0.901	0.9	1.003	1
47	1.18	1.2	1	
48	0.87	0.869	1.004	1
49				
50				
51				
52	0.712	0.692	1	
53				
54				
55				
56				
57				
58	1.24	1.24	1	1
59	0.849	0,848	1	
60	1.299	1.303	0.99	1
61				
62	1.279	1.279	0.988	0.995
63	1.35	1.35	1	1
64	1.252	1.249	0.995	
65	0.863	0,859		
66	1.18	1.2	1	
67	1.197	1.199	0.991	

Note: The coincidence summing correction factors in the table above are given as reported by the participants. As there is net summing in for the ¹³⁴Cs peaks and summing out for the ¹³¹I, the correction factors should be >1 for the two ¹³⁴Cs peaks and <1 for the two ¹³¹I peaks. It is possible that some participants reported the reciprocal of the correction factor.

Annex 12: Uncertainty budget as reported by the participating laboratories

Table A-12. Detailed uncertainty budget table showing the uncertainty components as reported by the participating laboratories, for all three radionuclides. In column 13 the combined relative uncertainty as reported by the participants is given. For comparison, in column 15 the combined relative uncertainty calculated as square root of the quadratic sum of the uncertainty components given in the table is given. Finally, in column 16 the combined relative uncertainty as calculated from the uncertainty accompanying the results is given.

1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17
Lab Code	Nuclide	Counting statistics	Blank/background measurement	Efficiency calibration (including the activity of the calibration source)	Chemical yield (if applicable)	Self-absorption in the source	Activity of the calibration source	Decay correction	Other 1	Other 2	Other 3	Combined relative standard uncertainty	Reported k factor	Calculated from components as quadratic sum	Calculated from the reported results	Explanation of "other" components, notes
1	Cs-137	26.4	n.d.	1.20	n.d.	n.d.	4	n.d.	2.5	8		28	2	28.0	26.5	O1 (reproducibility, stability, weight of multigamma source)+ O2 (other software calculation, not explicit)
1	Cs-134	17.1	n.d.	1.20	n.d.	n.d.	4	n.d.	2.5	8		19.5	2	19.5	16.7	
1	I-131	16.8	n.d.	1.20	n.d.	n.d.	4	n.d.	2.5	8		19.2	2	19.2	14.3	
2	Cs-137											see 4,1	1	0.0	7.7	We do not, as of yet, report the individual contribution of factors to the uncertainty budget. Instead, we have a generously set value for random uncertainties, which includes factors such as sample weight. This value is 5 %.
2	Cs-134											see 4,1	1	0.0	7.2	
2	I-131											see 4,1	1	0.0	8.0	

1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17
Lab Code	Nuclide	Counting statistics	Blank/background measurement	Efficiency calibration (including the activity of the calibration source)	Chemical yield (if applicable)	Self-absorption in the source	Activity of the calibration source	Decay correction	Other 1	Other 2	Other 3	Combined relative standard uncertainty	Reported k factor	Calculated from components as quadratic sum	Calculated from the reported results	Explanation of "other" components, notes
																The uncertainties relating to the bac
3	Cs-137	5.053											2	5.1	10.9	
3	Cs-134	2.893											2	2.9	10.7	
3	I-131	4.052											2	4.1	9.4	
4	Cs-137	100											1	100.0	10.0	
4	Cs-134	100											1	100.0	2.9	
4	I-131	100											1	100.0	66.0	
5	Cs-137	4		1.1	0.96							5.95	2	4.3	11.9	
5	Cs-134	3.2		1								3.95	2	3.4	7.9	
5	I-131	2.1		4.3								5.2	2	4.8	10.4	
6	Cs-137	7.3		2.35			1.2		3	3.4		9	1.65	9.0	9.6	sample preparation
6	Cs-134	6.1		2.35			1.2		3	3.4		8	1.65	8.0	8.1	
6	I-131	8.5		2.35			1.2		3	3.4		10	1.65	10.0	9.9	
7	Cs-137	2.1	7	5		1		1	0.24	1	0	8.91	1	9.0	8.9	1-inhomogeneity, 2-random error, 3-true coincidence correction factor
7	Cs-134	1.3		5		2		1	0.24	1	2	6.06	1	6.1	6.0	
7	I-131	2.7		5		2		2	0.24	2	2	6.95	1	7.0	7.1	
8	Cs-137												2	0.0	11.3	
8	Cs-134												2	0.0	11.8	
8	I-131												2	0.0	11.3	
9	Cs-137	6	2	5		5	3	1	5	2		11.4	2	11.4	12.5	Sample preparation; peak area determination
9	Cs-134	4	2	5		5	3	1	5	2		10.4	2	10.4	10.4	
9	I-131	4	2	5		5	3	1	5	2		10.1	2	10.4	10.5	

1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17
Lab Code	Nuclide	Counting statistics	Blank/background measurement	Efficiency calibration (including the activity of the calibration source)	Chemical yield (if applicable)	Self-absorption in the source	Activity of the calibration source	Decay correction	Other 1	Other 2	Other 3	Combined relative standard uncertainty	Reported k factor	Calculated from components as quadratic sum	Calculated from the reported results	Explanation of "other" components, notes
10	Cs-137	2.5	0	3.18	0	0	0.95	0.00021	0.0000092	1.27	3	4.9	2	5.3	10.3	(1) Correction factor during counting period, (2) Repeatability (Equipment) (3) Reproducibility
10	Cs-134	1.7	0	3.24	0	0	0.95	0.0015	0.00065	1.27	3	4.6	2	5.0	9.3	
10	I-131	5.6	0	3.61	0	0	1.68	0.021	0.0086	1.27	3	7.3	2	7.6	15.4	
11	Cs-137	7.16	2	3		1	2	0.27	3.7	2	0.2	9.3	2	9.3	18.2	positioning of the sample; summing correction; combination (diameter; sample density; sample composition)
11	Cs-134	2.1	0	4		1	2	0.024	4.5	2	0.2	7	2	7.0	15.6	
11	I-131												2	0.0	13.6	
12	Cs-137	1.87		1.25					0.27	0.02	6.70	7.20	1	7.1	7.2	Other1=unc. on decay constant; Other2=unc. on emission yield; Other3=repeatability(+summing correction for Cs-134)
12	Cs-134	1.61		1.94					0.02	0.23	6.70	7.20	1	7.2	12.1	
12	I-131												1	0.0	6.0	
13	Cs-137												2	0.0	8.0	
13	Cs-134												2	0.0	8.0	
13	I-131												2	0.0	8.6	
14	Cs-137	2.53	0	0.89	0	0	0	0	0	0	0	2.6	1	2.7	2.6	Other 1: Gamma emission probability. Other 2: Coincidence summing. Other 3:
14	Cs-134	1.37	0	0.9	0	0	0	0	0.2	1	2.1	2.83	1	2.9	2.8	

1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17
Lab Code	Nuclide	Counting statistics	Blank/background measurement	Efficiency calibration (including the activity of the calibration source)	Chemical yield (if applicable)	Self-absorption in the source	Activity of the calibration source	Decay correction	Other 1	Other 2	Other 3	Combined relative standard uncertainty	Reported k factor	Calculated from components as quadratic sum	Calculated from the reported results	Explanation of "other" components, notes
14	I-131	0.88	0	1.43	0	0	0	0	0.25	1	1	2.22	1	2.2	2.2	Type A uncertainty evaluation
15	Cs-137	3.78	19.4										1	19.8	8.6	
15	Cs-134	2.45											1	2.5	7.4	
15	I-131	1.31											1	1.3	7.6	
16	Cs-137	5.17	-	3.65	-	-	-	0.099	-	-	-	6.33	2	6.3	12.7	
16	Cs-134	2.42	-	3.65	-	-	-	0.242	-	-	-	4.39	2	4.4	8.8	
16	I-131	6.88	-	8.75	-	-	-	0.037	-	-	-	11.13	2	11.1	22.3	
17	Cs-137	3.7	0	1.2		0.2			5			6.3	1	6.3	8.0	1 - unhomogeneity. 2 - summing correction
17	Cs-134	1.4	0	1.2		0.2			5	0.3		5.2	1	5.3	5.5	
17	I-131	1	0	1.2		0.2		2	5			5.6	1	5.6	6.2	
18	Cs-137			7.3					x			8.5	1	7.3	8.5	Uncertainty is combined by Genie2000, budget not specified.
18	Cs-134			7.3					x			5	1	7.3	5.0	
18	I-131			9.3					x			10.2	1	9.3	10.2	
19	Cs-137	2.6		4				1	1	3		5.81	1	5.8	5.8	1-Self absorption in the sample; 2-Radioactivity distribution on the filter
19	Cs-134	1.8		4				1.5	1	3		5.61	1	5.6	5.6	
19	I-131	2.6		4				2	1	3		6.06	1	6.1	5.7	
20	Cs-137	80	5	5	0	5	5					0.02	2	80.6	9.1	
20	Cs-134	80	5	5	0	5	5					0.05	2	80.6	8.5	
20	I-131	70	5	5		5	5	10				0.09	2	71.4	8.7	

1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17
Lab Code	Nuclide	Counting statistics	Blank/background measurement	Efficiency calibration (including the activity of the calibration source)	Chemical yield (if applicable)	Self-absorption in the source	Activity of the calibration source	Decay correction	Other 1	Other 2	Other 3	Combined relative standard uncertainty	Reported k factor	Calculated from components as quadratic sum	Calculated from the reported results	Explanation of "other" components, notes
21	Cs-137	2.12	0	5	0	1	0	1	0.24	1	0	5.71	1	5.7	5.7	Other 1=INHOMOGENEITY; Other 2=RANDAOM ERROR; Other 3=TRUE COINCIDENCE CORRECTION FACTOR
21	Cs-134	1.97	0	5	0	2	0	1	0.24	1	2	6.24	1	6.2	6.2	
21	I-131	2.79	0	5	0	2	0	2	0.24	2	2	6.99	1	7.0	6.9	
22	Cs-137								7				1	7.0	8.3	Generic 7% uncertainty added to uncertainties from weighted averages across 3 detectors
22	Cs-134								7				1	7.0	7.5	
22	I-131								7				1	7.0	7.3	
23	Cs-137												2	0.0	6.7	
23	Cs-134												2	0.0	10.0	
23	I-131												2	0.0	10.5	
24	Cs-137											7.9	2	0.0	7.9	
24	Cs-134											5.2	2	0.0	5.3	
24	I-131											23.8	2	0.0	24.6	
25	Cs-137	5.36	0	20	0	0	3	0	0.5	0	0	20.81	2	20.9	20.6	emission intensity
25	Cs-134	3.77	0	20	0	0	3	0	2	0	0	20.45	2	20.7	20.4	
25	I-131	2.64	0	20	0	0	3	0	1.25	0	0	20.29	2	20.4	20.3	
26	Cs-137	6		5			5		2.89			14	2	9.7	19.6	combined operational uncertainty
26	Cs-134	6		5			5		2.89			14	2	9.7	19.6	
26	I-131	6		5			5		2.89			14	2	9.7	17.9	
27	Cs-137											1 171	2	0.0	2.3	

1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17
Lab Code	Nuclide	Counting statistics	Blank/background measurement	Efficiency calibration (including the activity of the calibration source)	Chemical yield (if applicable)	Self-absorption in the source	Activity of the calibration source	Decay correction	Other 1	Other 2	Other 3	Combined relative standard uncertainty	Reported k factor	Calculated from components as quadratic sum	Calculated from the reported results	Explanation of "other" components, notes
27	Cs-134											1 153	2	0.0	2.3	
27	I-131											3 781	2	0.0	2.2	
28	Cs-137	10		5					2				2	11.4	26.0	Correction for density and fill height
28	Cs-134	10		5					2				2	11.4	18.0	
28	I-131	10		5					2				2	11.4	22.6	
29	Cs-137	6.81		2.32			1.2		3	3.4		8.59	1.65	8.5	8.6	preparation and handling
29	Cs-134	5.32		2.32			1.2		3	3.4		7.46	1.65	7.4	7.4	
29	I-131	30.98		2.79			2.2		3	3.4		31.51	1.65	31.4	12.4	
30	Cs-137	2.8		2.02									2	3.5	2.8	
30	Cs-134	1.61		2.01									2	2.6	1.6	
30	I-131	2.28		3.03									2	3.8	2.3	
31	Cs-137	3	2.5	4			2.5					5.9	2	5.6	5.8	
31	Cs-134	1.9	2.5	4				1.5				5	2	5.1	4.9	
31	I-131	2.5	2.5	4				1.8				6.2	2	5.3	6.1	
32	Cs-137	1.20 %	0	1.5		3			3	3	3		2	6.2	16.7	coincidence calculation / efficiency transfer gespecor / systematic bias due to compression
32	Cs-134	0.8	0	1.4		3			3	3	3		2	6.2	16.7	
32	I-131	0.5	0	1.4		3			3	3	3		2	6.2	14.1	
33	Cs-137	3.65		3.301	0.24				0.13	1.6	0.2	5.2	2	5.2	10.8	Geometry Repeatability, Volume& system stability, Tcc error
33	Cs-134	2.19		4.32	0.35				0.13	1.6	0.2	5.1	2	5.1	10.1	
33	I-131	2.22		3.53	0.74				0.13	1.6	0.2	4.5	2	4.5	9.0	

1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17
Lab Code	Nuclide	Counting statistics	Blank/background measurement	Efficiency calibration (including the activity of the calibration source)	Chemical yield (if applicable)	Self-absorption in the source	Activity of the calibration source	Decay correction	Other 1	Other 2	Other 3	Combined relative standard uncertainty	Reported k factor	Calculated from components as quadratic sum	Calculated from the reported results	Explanation of "other" components, notes
34	Cs-137	4		6			1.2	0.2	5	4		9.72	1	9.6	9.8	the inhomogeneity of the sample and approximating a circular shape for the active spots
34	Cs-134	4.2		6			1.2	0.39	5	4		9.81	1	9.7	9.9	
34	I-131	4.4		7			1.2	1	5	4		10.76	1	10.5	10.3	
35	Cs-137					0			8	19		20	2	20.6	40.9	Other 1: Uncertainty budget without intra-reproducibility; Other 2: Intra-reproducibility for this geometry
35	Cs-134					0			6	19		20	2	19.9	60.1	
35	I-131					0			7	19		20	2	20.2	40.3	
36	Cs-137	8.23	0	3.16	0		3	0.5	5	0	0	10.58	2	10.1	7.3	Systematic errors
36	Cs-134	4.21	0	3.16	0		3	0.1	5	0	0	7.85	2	7.3	6.4	
36	I-131	21.27	0	3.16	0		3	0.1	5	0	0	22.28	2	22.1	12.1	
37	Cs-137	4.42	0	6					0			7.5	2	7.5	8.6	Coincidence summing correction uncertainty
37	Cs-134	3.92	0	6					2			7.4	2	7.4	8.0	
37	I-131	1.79	0	6					0			6.3	2	6.3	6.3	
38	Cs-137	3.43	6.61	3.54			1.51		5			7.84	1	9.6	7.9	Other 1: the sample homogeneity, Other 2: the activity left on a groundwork
38	Cs-134	2.33	6.14	3.54			1.51		5			6.87	1	9.0	6.8	
38	I-131	1.56	4.63	3.54			1.51		5	7		9.59	1	10.5	9.7	
39	Cs-137								x				1	0.0	10.0	uncertainty budget put in edit and

1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17
Lab Code	Nuclide	Counting statistics	Blank/background measurement	Efficiency calibration (including the activity of the calibration source)	Chemical yield (if applicable)	Self-absorption in the source	Activity of the calibration source	Decay correction	Other 1	Other 2	Other 3	Combined relative standard uncertainty	Reported k factor	Calculated from components as quadratic sum	Calculated from the reported results	Explanation of "other" components, notes
39	Cs-134								x				1	0.0	3.9	calculated by Genie analysis software
39	I-131								x				1	0.0	25.0	
40	Cs-137								9				2	9.0	8.8	The calculations were performed using Gamma Vision Software
40	Cs-134								7				2	7.0	6.7	
40	I-131								15				2	15.0	16.0	
41	Cs-137					1						2.68	2	1.0	2.7	
41	Cs-134					2						2.16	2	2.0	2.2	
41	I-131					2						1.51	2	2.0	1.5	
42	Cs-137	2.08	0	5	0	-	0	1	0.24	1	0	5.69	1	5.6	5.7	Other 1=INHOMOGENEITY; Other 2=RANDOM ERROR; Other 3=TRUE COINCIDENCE CORRECTION FACTOR
42	Cs-134	1.54	0	5	0	-	0	1	0.24	1	2	6.12	1	5.8	6.2	
42	I-131	5.21	0	5	0	-	0	2	0.24	2	2	8.26	1	8.0	8.0	
43	Cs-137	11.45	-	2.08	-	1	1.6	-	5	-	-	12.77	2	12.8	15.4	Error of repeatability
43	Cs-134	5.85	-	2.11	-	1	-	-	5	-	-	7.98	2	8.0	11.7	
43	I-131	32.97	-	1.66	-	1	-	-	5	-	-	33.39	2	33.4	30.8	
44	Cs-137	6.4	113	1.75		NA	1		3.3			7.5	2	113.2	16.7	repeatability
44	Cs-134	4.2	69.6	1.75		NA	1		4.1			6.2	2	69.9	12.2	
44	I-131	12.4	54	1.75		NA	1		6.7			14.2	2	55.8	30.4	

1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17
Lab Code	Nuclide	Counting statistics	Blank/background measurement	Efficiency calibration (including the activity of the calibration source)	Chemical yield (if applicable)	Self-absorption in the source	Activity of the calibration source	Decay correction	Other 1	Other 2	Other 3	Combined relative standard uncertainty	Reported k factor	Calculated from components as quadratic sum	Calculated from the reported results	Explanation of "other" components, notes
45	Cs-137	3.29	NA	1.3	NA						1	3.68	1	3.7	3.7	1. coincidence summing; 2. efficiency curve interpolation; 3. sample fitting to the standard geometry
45	Cs-134	2.57	NA	1.5	NA				1.8	1.2	1	3.82	1	3.8	3.8	
45	I-131	1.4	NA	1.5	NA			0.03	1.5	1.2	1	2.99	1	3.0	3.2	
46	Cs-137	4.41		10					10.2			14.8	1	14.9	16.7	Other 1: Folding and positioning of the filter on the detector
46	Cs-134	2.53		10					10.2			14.5	1	14.5	15.6	
46	I-131	3.88		10					10.2			14.9	1	14.8	14.6	
47	Cs-137	1.7		0.5		-			5			5.3	1	5.3	5.4	geometry factor, coincidence correction
47	Cs-134	1.6		0.5		-			5	7.2		8.9	1	8.9	8.9	
47	I-131	5.5		2.1		-			5			7.7	1	7.7	7.7	
48	Cs-137	2.827	-	1.17	-		1.17	1.67E-05	1.71	0.23 5	1.2	3.89	1	3.9	4.7	systematic, yield, monte carlo corrections
48	Cs-134	1.645	-	2.12	-		1.17	0.00061	1.71	0.08 2	1.2	3.6	1	3.6	3.3	
48	I-131	0.983 5	-	1.6846	-		1.17	0.0021	1.71	0.61 6	1.2	3.15	1	3.1	3.5	
49	Cs-137	6.02				not applicable						6.02	2	6.0	6.3	
49	Cs-134	3.8										3.8	2	3.8	3.5	
49	I-131	5.2										5.2	2	5.2	5.2	
50	Cs-137	2.4	in countin	2.4	not applica	0	includ ed in	included in result				3.4	1	3.4	3.4	

1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17
Lab Code	Nuclide	Counting statistics	Blank/background measurement	Efficiency calibration (including the activity of the calibration source)	Chemical yield (if applicable)	Self-absorption in the source	Activity of the calibration source	Decay correction	Other 1	Other 2	Other 3	Combined relative standard uncertainty	Reported k factor	Calculated from components as quadratic sum	Calculated from the reported results	Explanation of "other" components, notes
			g statistics		ble		efficiency									
50	Cs-134					0							1	0.0	4.6	
50	I-131					0							1	0.0	5.4	
51	Cs-137	2.103	4.677	1.3264	0		1.43	0.26	0	0	0	5.487	1	5.5	3.0	
51	Cs-134	1.494	0	1.5172	0		1.43	0.07	0	0	0	2.565	1	2.6	5.5	
51	I-131	1.156	0	2.7404	0		2.04	0.02	0	0	0	3.607	1	3.6	3.5	
52	Cs-137	2.35		5.49			2.27		2	0.58		6.33	2	6.7	12.5	Other 1: Additional random uncertainty; Other 1: Additional systematic uncertainty
52	Cs-134	7.17		5.49			2.27		2	0.58		9.27	2	9.5	11.9	
52	I-131												2	0.0	18.1	
53	Cs-137	91.44		8.66									2	91.8	9.3	
53	Cs-134	91.44		8.66									2	91.8	8.5	
53	I-131	91.44		8.66									2	91.8	21.3	
54	Cs-137	4.5	2.3	2.8								5.8	1	5.8	2.5	
54	Cs-134	4.5	2.3	2.8								5.8	1	5.8	2.5	
54	I-131	4.5	2.3	2.8								5.8	1	5.8	5.8	
55	Cs-137	2.9	0.5	3.5		0.1	3.4	0.2				5.7	2	5.7	8.0	
55	Cs-134	1.7	0	3.5		0.1	3.4	0.2				5.2	2	5.2	8.3	
55	I-131	1		4.7		0.5	3.4	0.1				5.9	2	5.9	7.9	
56	Cs-137	2	13.2	2			1.4		0.15			13.6	1	13.6	7.7	Gamma-ray emission intensity used
56	Cs-134	1.2	0	10					0.1			10	1	10.1	6.4	
56	I-131	0.94	0	10					1			10	1	10.1	6.6	

1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17
Lab Code	Nuclide	Counting statistics	Blank/background measurement	Efficiency calibration (including the activity of the calibration source)	Chemical yield (if applicable)	Self-absorption in the source	Activity of the calibration source	Decay correction	Other 1	Other 2	Other 3	Combined relative standard uncertainty	Reported k factor	Calculated from components as quadratic sum	Calculated from the reported results	Explanation of "other" components, notes
57	Cs-137	6.99		3					0.24			7.62	2	7.6	7.6	Gamma photons emission probability
57	Cs-134	24.12		3					0.08			24.3	2	24.3	24.3	
57	I-131												2	0.0	0.0	
58	Cs-137	2.7	0	1.4					0.5				1	3.1	2.9	Relative positioning difference between calibration source and sample, decay correction uncert. added by Genie2000, eff. calib. uncertainty 1.4% due to differences in calibration source and sample, in addition to the uncertainty in the efficiency curve ca
58	Cs-134	1.4	0	1.4					0.5				1	2.0	1.4	
58	I-131	1.2	0	1.4				0	0.5				1	1.9	2.2	
59	Cs-137	4.1		3.7								5.52	1	5.5	5.5	
59	Cs-134	2		3.9								4.4	1	4.4	4.5	
59	I-131	4.3		4.24								6.04	1	6.0	6.8	
60	Cs-137											5.3	1	0.0	6.1	
60	Cs-134											2.6	1	0.0	2.8	
60	I-131											2.7	1	0.0	2.7	
61	Cs-137											14.61	1	0.0	14.6	
61	Cs-134											9.53	1	0.0	9.5	
61	I-131											5.86	1	0.0	5.9	

1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17
Lab Code	Nuclide	Counting statistics	Blank/background measurement	Efficiency calibration (including the activity of the calibration source)	Chemical yield (if applicable)	Self-absorption in the source	Activity of the calibration source	Decay correction	Other 1	Other 2	Other 3	Combined relative standard uncertainty	Reported k factor	Calculated from components as quadratic sum	Calculated from the reported results	Explanation of "other" components, notes
62	Cs-137	2.92	22	5.6			2.5		1			6.4	1	23.0	2.9	Other 1 = Inhomogeneity. Self absorption uncertainty contribution is included in Efficiency calibration uncertainty. Sample was measured on several spectrometers, the uncertainty budget is given for one measurement only.
62	Cs-134	1.36		2.86			2.5		1			4	1	4.2	3.1	
62	I-131	2.4		5.35			2.5		1			6.4	1	6.5	4.1	
63	Cs-137	4.174		1.192			3		2.3			9.92	2	5.8	9.9	other 1 is budget about inhomogeneity
63	Cs-134	3.195		1.192			3		2.3			9.18	2	5.1	9.2	
63	I-131	26		1.192			3		2.3			28.6	2	26.3	57.2	
64	Cs-137	6	3.7	4.5								8.4	1	8.4	8.7	
64	Cs-134	5	2	5.5								7.7	1	7.7	7.7	
64	I-131	5.4	3.7	4.5								7.9	1	7.9	7.9	
65	Cs-137	3.2		4				1	2.5			5.78	1	5.8	5.8	distribution of blank filters spiked, is random
65	Cs-134	2.4		4				1.5	2.5			5.5	1	5.5	5.4	
65	I-131	3.4		4				2	2.5			6.14	1	6.1	6.1	
66	Cs-137	1.9	0	0.5			0.8	0.2	4.6			5	1	5.1	5.3	geometry factor, coincidence

1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17
Lab Code	Nuclide	Counting statistics	Blank/background measurement	Efficiency calibration (including the activity of the calibration source)	Chemical yield (if applicable)	Self-absorption in the source	Activity of the calibration source	Decay correction	Other 1	Other 2	Other 3	Combined relative standard uncertainty	Reported k factor	Calculated from components as quadratic sum	Calculated from the reported results	Explanation of "other" components, notes
66	Cs-134	1.2	0	0.6					4.5	6.5		8	1	8.0	8.3	correction
66	I-131	8.9	0	0.7					4.5			10	1	10.0	10.0	
67	Cs-137	1.95		0.5			0.8	0.24	4.5			5	1	5.0	5.0	geometry factor, coincidence correction, Homogeneity
67	Cs-134	1.5		0.7				0.1	4.5	6.4		8	1	8.0	8.0	
67	I-131	6.1		2.7				0.62	4.5		6.3	10	1	10.2	10.2	

***Europe Direct is a service to help you find answers
to your questions about the European Union.***

Freephone number (*):

00 800 6 7 8 9 10 11

(*) The information given is free, as are most calls (though some operators, phone boxes or hotels may charge you).

More information on the European Union is available on the internet (<http://europa.eu>).

HOW TO OBTAIN EU PUBLICATIONS

Free publications:

- one copy:
via EU Bookshop (<http://bookshop.europa.eu>);
- more than one copy or posters/maps:
from the European Union's representations (http://ec.europa.eu/represent_en.htm);
from the delegations in non-EU countries (http://eeas.europa.eu/delegations/index_en.htm);
by contacting the Europe Direct service (http://europa.eu/europedirect/index_en.htm) or
calling 00 800 6 7 8 9 10 11 (freephone number from anywhere in the EU) (*).

(*) The information given is free, as are most calls (though some operators, phone boxes or hotels may charge you).

Priced publications:

- via EU Bookshop (<http://bookshop.europa.eu>).

JRC Mission

As the science and knowledge service of the European Commission the Joint Research Centre's mission is to support EU policies with independent evidence throughout the whole policy cycle.



EU Science Hub

ec.europa.eu/jrc



@EU_ScienceHub



EU Science Hub - Joint Research Centre



Joint Research Centre



EU Science Hub



Publications Office

doi:10.2760/030337

ISBN 978-92-79-65276-9