

# JRC REFERENCE MATERIALS REPORT

## CERTIFICATION REPORT

### Preparation and certification of $^{243}\text{Am}$ spike reference material: IRMM-0243

*Certified reference material for  
the amount content of  $^{243}\text{Am}$   
and  $n(^{241}\text{Am})/n(^{243}\text{Am})$  isotope  
amount ratio*

Jakopič, R., Fankhauser, A., Aregbe, Y., Crozet,  
M., Maillard, C., Richter, S., Rivier, C., Roudil, D.,  
Altitzoglou, T., Pommé, S., Marouli, N., Tzika, F.

2017



This publication is a Reference Materials report by the Joint Research Centre (JRC), the European Commission's science and knowledge service. It aims to provide evidence-based scientific support to the European policymaking process. The scientific output expressed does not imply a policy position of the European Commission. Neither the European Commission nor any person acting on behalf of the Commission is responsible for the use that might be made of this publication.

**Contact information**

Name: Rožle Jakopič

Address: Joint Research Centre, Retieseweg 111, 2440 Geel, Belgium

Email: rozle.jakopic@ec.europa.eu

Tel.: + 32 (0) 14 571 617

**JRC Science Hub**

<https://ec.europa.eu/jrc>

JRC107840

EUR 28748 EN

PDF ISBN 978-92-79-72281-3 ISSN 1831-9424 Doi:10.2760/40283

Luxembourg: Publications Office of the European Union, 2017

© European Atomic Energy Community, 2017

Reuse is authorised provided the source is acknowledged. The reuse policy of European Commission documents is regulated by Decision 2011/833/EU (OJ L 330, 14.12.2011, p. 39).

For any use or reproduction of photos or other material that is not under the EU copyright, permission must be sought directly from the copyright holders.

How to cite this report: Author(s), *Title*, EUR (where available), Publisher, Publisher City, Year of Publication, ISBN (where available), doi (where available), PUBSY No.

All images © European Atomic Energy Community, 2017

## Table of contents

<b>1. Introduction</b>	<b>8</b>
1.1 Background	8
1.2 Choice of the material	8
1.3 Design of the project	8
<b>2. Participants</b>	<b>11</b>
<b>3. Material processing and process control</b>	<b>11</b>
3.1 Origin of the starting material	11
3.2 Processing	11
3.3 Process control	12
<b>4. Homogeneity</b>	<b>12</b>
4.1 Between-unit homogeneity for the amount content of $^{243}\text{Am}$	12
4.2 Homogeneity of the Am isotope ratios	15
4.3 Within-unit homogeneity and minimum sample intake	15
<b>5. Stability</b>	<b>15</b>
5.1 Short-term stability	16
5.2 Long-term stability	16
<b>6. Characterisation</b>	<b>16</b>
6.1 $^{241}\text{Am}$ in-house spike	17
6.2 Chemical purification and isotope measurements	17
6.4 The amount content of $^{243}\text{Am}$ and its uncertainty	18
6.5 The amount ratios of $n(^{241}\text{Am})/n(^{243}\text{Am})$ and $n(^{242\text{m}}\text{Am})/n(^{243}\text{Am})$ and their uncertainties	20
6.6 Verification measurements	21
6.7 Inter-laboratory comparison	21
<b>7. Value Assignment</b>	<b>22</b>
7.1 Certified values and their uncertainties	22
7.2 Indicative values and their uncertainties	24
<b>8. Metrological traceability and commutability</b>	<b>25</b>
8.1 Metrological traceability	25
8.2 Commutability	25
<b>9. Instructions for use</b>	<b>26</b>
9.1 Safety information	26
9.2 Storage conditions	26
9.3 Preparation and use of the material	26
9.4 Minimum sample intake	26
9.5 Use of the certified value	26
<b>10. Conclusions</b>	<b>27</b>
<b>References</b>	<b>28</b>
<b>List of abbreviations and definitions</b>	<b>30</b>
<b>List of figures</b>	<b>32</b>

<b>List of tables .....</b>	<b>33</b>
<b>Annexes .....</b>	<b>34</b>

## Acknowledgement

The authors would like to thank C. Hennessy, F. Kehoe and S. Werelds from the JRC-Geel for assistance in the preparation of the reference material, Monika Sturm from the IAEA for support in separation chemistry and americium mass spectrometry and the Laboratory for Target Preparation at JRC-Geel for providing the  $^{241}\text{Pu}$  material.

The authors would also like to thank R. Van Ammel (JRC-Geel) and the experts of the Certification Advisory Panel, S. D. Balsley (International Atomic Energy Agency, Vienna, Austria), P. Mason (NBL Program Office, US Department of Energy, Argonne, U.S.A.) and Richard Essex (National Institute of Standards and Technology, Gaithersburg, U.S.A.) for reviewing the certification report.

### Authors

Jakopič, R.<sup>1</sup>, Fankhauser, A.<sup>1</sup>, Aregbe, Y.<sup>1</sup>, Crozet, M.<sup>2</sup>, Maillard, C.<sup>2</sup>, Richter, S.<sup>1</sup>, Rivier, C.<sup>2</sup>, Roudil, D.<sup>2</sup>, Altitzoglou, T.<sup>1</sup>, Pommé, S.<sup>1</sup>, Marouli, M.<sup>1</sup>, Tzika, F.<sup>1</sup>

<sup>1</sup>) European Commission, Joint Research Centre, Directorate G - Nuclear Safety and Security, Geel, Belgium

<sup>2</sup>) Commissariat à l'Énergie Atomique et aux Énergies Alternatives (CEA), Direction de l'Énergie Nucléaire (DEN), Marcoule, France

## Abstract

This report describes the preparation and certification of IRMM-0243, a  $^{243}\text{Am}$  spike reference material. It is certified for the amount content of  $^{243}\text{Am}$  and the  $n(^{241}\text{Am})/n(^{243}\text{Am})$  isotope amount ratio. Furthermore, the material is certified for the amount contents of  $^{241}\text{Am}$  and total Am, the mass fractions of  $^{243}\text{Am}$ ,  $^{241}\text{Am}$  and total Am, the isotope amount and mass fractions (e.g. isotopic composition) and the molar mass of Am. The material was produced in compliance with ISO/IEC 17034:2016 [1] and certified in accordance with ISO Guide 35:2006 [2].

The material was prepared by dilution of an americium starting solution in nitric acid and dispensing of the solution into glass ampoules. In total 587 units were produced.

Between-unit homogeneity was quantified and stability during dispatch and storage were assessed in accordance with ISO Guide 35:2006 [2].

The characterisation of the amount content of  $^{243}\text{Am}$  was performed by Isotope Dilution Mass Spectrometry (IDMS) using a  $^{241}\text{Am}$  spike, produced from highly enriched  $^{241}\text{Pu}$  material. The  $n(^{241}\text{Am})/n(^{243}\text{Am})$  isotope amount ratio was measured by Thermal Ionisation Mass Spectrometry (TIMS).

The certified values were verified by alpha particle spectrometry, alpha particle counting at a defined solid angle (DSA) and high-resolution gamma-ray spectrometry as independent verification methods.

The uncertainties of the certified values were estimated in compliance with the Guide to the Expression of Uncertainty in Measurement (GUM) [3] and include uncertainties related to possible inhomogeneity, instability and characterisation.

The indicative value for the  $n(^{242\text{m}}\text{Am})/n(^{243}\text{Am})$  isotope amount ratio and its uncertainty were derived from the inter-laboratory comparison (ILC) organised by CEA/CETAMA using IRMM-0243 as the test sample.

The main purpose of this material is for use as a spike isotopic reference material for the quantification of americium by IDMS in unknown samples. A unit of IRMM-0243 consists of a glass ampoule with a screw cap containing about 3.5 mL nitric acid solution ( $c = 1$  mol/L) with an americium mass fraction of about 1.5  $\mu\text{g/g}$  solution.

The material is a true solution; therefore there is no recommended minimum sample intake<sup>1</sup> to be taken into account.

The following values were assigned:

---

<sup>1</sup> Size of aliquot for analysis

IRMM-0243	AMERICIUM IN NITRIC ACID SOLUTION		
		Certified value <sup>1)</sup>	Uncertainty <sup>2)</sup>
Amount contents [nmol/g (solution)]	<sup>241</sup> Am <sup>3)</sup>	0.7754	0.0015
	<sup>243</sup> Am	5.696	0.011
	Am <sup>3)</sup>	6.472	0.012
Isotope amount ratios [mol/mol]	$n(^{241}\text{Am})/n(^{243}\text{Am})$	0.136138	0.000054
Mass fractions <sup>3)</sup> [μg/g (solution)]	<sup>241</sup> Am	0.18692	0.00036
	<sup>243</sup> Am	1.3845	0.0026
	Am	1.5716	0.0030
Isotope amount fractions (·100) <sup>3)</sup> [mol/mol]	$n(^{241}\text{Am})/n(\text{Am})$	11.9810	0.0042
	$n(^{243}\text{Am})/n(\text{Am})$	88.0069	0.0042
Isotope mass fractions (·100) <sup>3)</sup> [g/g]	$m(^{241}\text{Am})/m(\text{Am})$	11.8940	0.0042
	$m(^{243}\text{Am})/m(\text{Am})$	88.0940	0.0042
Molar mass <sup>3)</sup> [g/mol]	$M(\text{Am})$	242.821094	0.000085
		Indicative value <sup>4)</sup>	Uncertainty <sup>2)</sup>
Isotope amount ratios [mol/mol]	$n(^{242\text{m}}\text{Am})/n(^{243}\text{Am})$	0.0001373	0.0000024
Isotope amount fractions (·100) [mol/mol]	$n(^{242\text{m}}\text{Am})/n(\text{Am})$	0.01208	0.00021
Isotope mass fractions (·100) [g/g]	$m(^{242\text{m}}\text{Am})/m(\text{Am})$	0.01205	0.00021

<sup>1)</sup> The certified values are traceable to the International System of units (SI) via the values on the respective certificates of IRMM-049d and IRMM-290b-A3 and via the half-life of <sup>241</sup>Pu. The reference date for the certified values is January 1, 2017.

<sup>2)</sup> The uncertainty is the expanded uncertainty with a coverage factor  $k = 2$  corresponding to a level of confidence of about 95 % estimated in accordance with ISO/IEC Guide 98-3, Guide to the Expression of Uncertainty in Measurement (GUM:1995), ISO, 2008.

<sup>3)</sup> The derived certified values are calculated from the certified amount content of <sup>243</sup>Am, certified  $n(^{241}\text{Am})/n(^{243}\text{Am})$  isotope amount ratio and the atomic masses according to Wang et al. (The AME 2012 atomic mass evaluation (II). Tables, Graphs and References, Chinese Physics C, Vol. 36, No. 12, 1603-2014, 2012).

<sup>4)</sup> The indicative value for the  $n(^{242\text{m}}\text{Am})/n(^{243}\text{Am})$  isotope amount ratio was derived from the <sup>243</sup>Am inter-laboratory comparison (ILC) organised by CEA/CETAMA using IRMM-0243 as the test sample. It represents the DerSimonian-Laird weighted mean of the results from nine participating laboratories. The indicative values for the  $m(^{242\text{m}}\text{Am})/m(\text{Am})$  mass and  $n(^{242\text{m}}\text{Am})/n(\text{Am})$  amount fractions were calculated using the indicative value of  $n(^{242\text{m}}\text{Am})/n(^{243}\text{Am})$  and the certified values. The reference date for the indicative values is January 1, 2017.

The half-lives of radionuclides were obtained from DDEP-BIPM (Table of radionuclides) and R. Wellum et al. (A new evaluation of the half-life of <sup>241</sup>Pu, J. Anal. At. Spectrom., 24, 801-807, 2009).

# 1. Introduction

## 1.1 Background

In nuclear safeguards and security, accurate isotopic measurements are required in order to draw correct conclusions. The accuracy, reliability and traceability of such measurements depend heavily on suitable isotopic reference materials. There is a wide range of uranium and plutonium Certified Reference Materials (CRMs) for quality control, method validation and instrument calibration in mass spectrometry. On the contrary, the availability of americium reference materials is limited. Currently, there is no  $^{243}\text{Am}$  spike CRM available, although it is indispensable for accurate mass spectrometry measurements of  $^{241}\text{Am}$  in nuclear materials. Such material can be used in nuclear forensics to determine the 'model age' of a (seized) plutonium material, i.e. the time elapsed since its last chemical purification (via  $^{241}\text{Am}/^{241}\text{Pu}$  chronometer), in radiation waste management for characterisation of irradiated nuclear fuel/waste for long-term disposal at geological repositories and in nuclear data applications for determination of radionuclide half-lives [4].

The provision of nuclear reference materials is regularly addressed among reference materials providers and users, e.g. in the frame of the Working Group on Techniques and Standards for Destructive Analysis (WGDA) of the European Safeguards and Research Association (ESARDA), the International Atomic Energy Agency (IAEA), and the Nuclear Forensics International Technical Working Group (ITWG). The need for an americium spike CRM was expressed at the 2014 IAEA Technical Meeting on Reference Materials for Destructive Analysis in the Nuclear Fuel Cycle as well as at the 2016 Nuclear Security Summit: Certified Reference Material Fact Sheet [5].

To fill this gap, a novel  $^{243}\text{Am}$  spike (IRMM-0243) has been prepared and certified for the amount content and isotopic composition in collaboration with the French Nuclear and Alternative Energies Commission (CEA), Nuclear Energy Division – CEA, Marcoule.

## 1.2 Choice of the material

The IRMM-0243 reference material was prepared from an americium starting solution, containing  $^{243}\text{Am}$  and  $^{241}\text{Am}$  isotopes with relative isotope mass fractions  $m(^{243}\text{Am})/m(\text{Am})$  and  $m(^{241}\text{Am})/m(\text{Am})$  of about 88 % and 12 %, respectively. The solution was diluted in nitric acid to achieve an Am mass fraction of  $1.5 \mu\text{g/g}$  and dispensed into individual glass ampoules. This concentration of americium is considered suitable for various mass spectrometry measurements (e.g. TIMS, ICP-MS).

## 1.3 Design of the project

IRMM-0243 was characterised for the amount contents of  $^{243}\text{Am}$ ,  $^{241}\text{Am}$  and total Am, the mass fractions of  $^{243}\text{Am}$ ,  $^{241}\text{Am}$  and total Am, the isotope amount ratios of  $n(^{241}\text{Am})/n(^{243}\text{Am})$  and  $n(^{242\text{m}}\text{Am})/n(^{243}\text{Am})$ , the isotope amount and mass fractions (e.g. isotopic composition) and the molar mass of Am.

Characterisation measurements, homogeneity and stability studies were performed on one set of randomly selected ampoules.

The characterisation of the amount content of  $^{243}\text{Am}$  was performed by Isotope Dilution Mass Spectrometry (IDMS) [6]. In IDMS, the amount of an element in the sample is determined on the basis of additions of known amounts of the same element (called spike) whose isotopic composition differs from that of the unknown sample. By measuring the change in isotopic composition of the blend (sample-spike mixture) by mass spectrometry, the unknown amount of the element in the sample can be calculated. In absence of a suitable  $^{241}\text{Am}$  spike CRM for IDMS measurements, an alternative approach was applied. The  $^{241}\text{Am}$  spike was produced from highly enriched  $^{241}\text{Pu}$  ( $m(^{241}\text{Pu})/m(\text{Pu}) = 0.993$ ) material available at JRC-Geel. After the initial


purification of the  $^{241}\text{Pu}$  material, the ingrown  $^{241}\text{Am}$ , produced by beta-decay of  $^{241}\text{Pu}$ , was used as spike for the measurement of the amount content of  $^{243}\text{Am}$  by IDMS (see Section 6.1). The isotope amount ratios of  $n(^{241}\text{Am})/n(^{243}\text{Am})$  and  $n(^{242\text{m}}\text{Am})/n(^{243}\text{Am})$  were measured by Thermal Ionisation Mass Spectrometry (TIMS). All isotope ratio measurements were performed using the total evaporation method on a multi-collector Triton TIMS.

Independent verification measurements were performed by alpha particle spectrometry, alpha particle counting at a defined solid angle (DSA) and high-resolution gamma-ray spectrometry.

Prior to release of this  $^{243}\text{Am}$  spike CRM, that same material was used in an inter-laboratory comparison (ILC) organised by CEA/CETAMA. The purpose of this exercise was to assess the measurement performance of participating laboratories. It also served as an external verification of the certified values.

Figure 1 shows all the steps in the preparation and certification of the IRMM-0243 material in a flow chart.

**Figure 1** Flow chart for the preparation and certification of IRMM-0243



## 2. Participants

The americium starting solution, the purification and the impurity analysis were provided by the CEA/L2AT (ATalante Analysis Laboratory, Marcoule, France).

Value assignment, homogeneity and stability studies, including data evaluation and the overall project management have been performed at the European Commission`s Joint Research Centre, Directorate G - Nuclear Safety and Security, G.2 - Standards for Nuclear Safety, Security and Safeguards, Geel, Belgium.

## 3. Material processing and process control

### 3.1 Origin of the starting material

About 4 mg of americium starting material was made available by the CEA/L2AT (ATalante Analysis Laboratory, Marcoule, France).

### 3.2 Processing

The processing of IRMM-0243 material included the purification of the americium starting solution, dilution, and subsequent dispensing into glass ampoules.

#### Purification of Am solution

The americium starting solution was purified by extraction chromatography to remove the impurities (Ca, Al, Fe, Co, Ni, etc.). The americium solution was loaded on a pre-conditioned TRU-Spec column (Triskem International, Bruz, France) and the impurities washed through the column with 20 mL of nitric acid solution ( $c = 1$  mol/L). Americium (and the lanthanides) was then eluted with 5 mL dilute nitric acid solution ( $c = 0.01$  mol/L).


#### Cleaning

The glass ampoules, the flask for the dilution and the dispenser were cleaned prior to use with ultra-pure Milli Q water (18.2 M $\Omega$  cm at 25 °C, Millipore, Belgium) and in 10 % (m/m) nitric acid solution (p.a., Merck, Darmstadt, Germany) for 24 hours. The flask and the ampoules were dried in an oven at 60 °C.

#### Dilution and dispensing

The purified americium solution (about 1 g/L) was diluted with 2400 mL nitric acid solution ( $c = 1$  mol/L, p.a., Merck, Darmstadt, Germany) in a 3 L glass flask to obtain an americium mass fraction of 1.5  $\mu$ g/g. The solution was allowed to homogenise for a week with occasional stirring by hand. Aliquots of 3.5 mL of the mother solution were dispensed into glass ampoules with screw-caps. The ampoules were closed immediately after the dispensing to minimise the evaporation of the solution. The dispensing of all 587 units was completed within a week. The dilution of the americium starting solution and the dispensing into ampoules were carried out in a dedicated glove box. Finally, the ampoules were taken out of the glove box, checked individually for surface contamination, sealed in PVC bags and labelled. The major processing steps are depicted in Figure 2.

**Figure 2** Processing steps of the americium reference material: a) dispensing of a solution, b) closing of an ampoule with a screw-cap, c) contamination check and PVC sealing and d) a unit of IRMM-0243



### 3.3 Process control

The results of the impurity analysis after purification of the americium source solution are summarised in Annex A.

## 4. Homogeneity

A key requirement for any reference material is the equivalence between the various units. In this respect, it is relevant whether the variation between units is significant compared to the uncertainty of the certified value. In contrast to that it is not relevant if this variation between units is significant compared to the analytical variation. Consequently, ISO 17034:2016 [1] requires RM producers to quantify the between-unit variation. This aspect is covered in between-unit homogeneity studies.

The within-unit inhomogeneity does not influence the uncertainty of the certified value when the minimum sample intake is respected, but determines the minimum size of an aliquot that is representative for the whole unit.

### 4.1 Between-unit homogeneity for the amount content of $^{243}\text{Am}$

The between-unit homogeneity was evaluated to ensure that the certified values of the CRM are valid for all units of the material, within the stated uncertainty.

The number of selected units (18 units in total) is greater than the cubic root of the total number of the produced units (587). The eighteen units were selected using a random stratified sampling scheme covering the whole batch for the between-unit homogeneity test.

For this project, the homogeneity, stability and characterisation studies were performed on the same units. The data obtained from the between-unit homogeneity assessment were also used to assign the certified values of the material (see Section 6 and 7).

A chemical purification was carried out prior to isotope ratio measurements on each unit. The measurements were carried on sets of two units over the period from January to July 2016. This allowed enough time for the weighing, chemical preparation and isotope measurements on each set. It also gave rise to sets with slightly increased amounts of ingrown <sup>241</sup>Am over time. Details are described in section 6 (Characterisation) of this report.

Using this approach, the measurements were performed under intermediate precision conditions rather than repeatability conditions within short intervals of time. Consequently, significant day-to-day effects occurred (confirmed by analysis of variance, ANOVA) that could mask the between-unit variation. The correction of the data for day-to-day variations was deemed not necessary since the uncertainty from the homogeneity study (see Table 2) was three times smaller compared to the uncertainty from the characterisation (see Section 6) and, therefore contributed only to a small extent to the overall uncertainty of this reference material.

The respective samples of each unit of IRMM-0243 were measured in a randomised manner to be able to separate a potential analytical drift from a trend in the filling sequence. The measurement results of the unit No. 240 and 437 had to be excluded from the evaluation due to technical reasons, such as e.g. loss of sample prior to measurement, high background from the filament due to unusual high filament temperatures or very low signal intensity.

Regression analyses were performed to evaluate potential trends in the analytical sequence as well as trends in the filling sequence. A trend in the analytical sequence was observed at the 95 % confidence level. The analytical trend arose from the data points at the beginning of the analytical sequence (see Annex B). Since no technical reason was found to exclude these data points, they were all retained for statistical evaluation. Correction of this trend is therefore expected to improve the sensitivity of the subsequent statistical analysis through a reduction in analytical variation without masking potential between-unit heterogeneities. As the analytical sequence and the unit numbers were not correlated, the analytical trend was corrected as shown below (Equation 1):

$$x_{i-corr} = x_i - b \cdot i \quad \text{Equation 1}$$

*i*                    position of the result in the analytical sequence  
*b*                    slope of the linear regression for the analytical run

No trend in the filling sequence was visible. The trend corrected data were tested for consistency using Grubbs outlier test on a confidence level of 99 % on the individual results and on the unit means. No outlying individual result or outlying unit means were detected.

The results of the homogeneity study are shown in Annex B

Quantification of between-unit inhomogeneity was accomplished by analysis of variance (ANOVA), which can separate the between-unit standard deviation (*s<sub>bb</sub>*) from the within-unit standard deviation (*s<sub>wb</sub>*). The latter is equivalent to the method intermediate precision if the individual samples are representative for the whole unit.

Evaluation by ANOVA requires unit means that follow at least a unimodal distribution and results for each unit that follow unimodal distributions with approximately the same standard deviations (homoscedasticity). Distribution of the unit means was visually tested using histograms and normal probability plots. Minor deviations from unimodality of the individual values do not significantly affect the estimate of between-ampoule standard deviations. Results of the statistical evaluation are given in Table 1.

**Table 1** Results of the statistical evaluation of the homogeneity study for the amount content of  $^{243}\text{Am}$

	Trends (before correction) <sup>1)</sup>		Outliers <sup>2)</sup>		Distribution	
	Analytical sequence	Filling sequence	Individual results	Unit means	Individual results	Unit means
Amount content of $^{243}\text{Am}$	yes	no	none	none	unimodal	unimodal

<sup>1)</sup> at 95 % confidence level

<sup>2)</sup> at 99 % confidence level

It should be noted that  $s_{bb,rel}$  (between-unit relative standard deviation) and  $s_{wb,rel}$  (within-unit relative standard deviation) are estimates of the true standard deviations and therefore subject to random fluctuations. Therefore, the mean square between groups ( $MS_{between}$ ) can be smaller than the mean squares within groups ( $MS_{within}$ ), resulting in negative arguments under the square root used for the estimation of the between-unit variation, whereas the true variation cannot be lower than zero. In this case, the maximum inhomogeneity,  $u_{bb}^*$ , that could be hidden by method repeatability (in our case intermediate precision) was calculated as described by Linsinger *et al.* [7].  $u_{bb}^*$  is comparable to the limit of detection of an analytical method, yielding the maximum inhomogeneity that might be undetected by the given study setup (alpha risk).

Relative within-unit standard deviation of method intermediate precision ( $s_{wb,rel}$ ), relative between-unit standard deviation ( $s_{bb,rel}$ ) and relative maximum inhomogeneity ( $u_{bb,rel}^*$ ) were calculated as:

$$s_{wb,rel} = \frac{\sqrt{MS_{within}}}{\bar{y}} \quad \text{Equation 2}$$

$$s_{bb,rel} = \frac{\sqrt{\frac{MS_{between} - MS_{within}}{N}}}{\bar{y}} \quad \text{Equation 3}$$

$$u_{bb,rel}^* = \frac{\sqrt{\frac{MS_{within}}{N}} \sqrt[4]{\frac{2}{v_{MS_{within}}}}}{\bar{y}} \quad \text{Equation 4}$$

- $MS_{within}$  mean square within-unit from an ANOVA
- $MS_{between}$  mean squares between-unit from an ANOVA
- $\bar{y}$  mean of all results of the homogeneity study
- $N$  mean number of replicates per unit
- $v_{MS_{within}}$  degrees of freedom of  $MS_{within}$

The uncertainty contribution for homogeneity was determined under intermediate precision conditions as described earlier in this section. The results of the evaluation of the between-unit variation are summarised in Table 2.

**Table 2** Results of the homogeneity study for the amount content of  $^{243}\text{Am}$

	$s_{wb,rel}$ [%]	$s_{bb,rel}$ <sup>1)</sup> [%]	$u_{bb,rel}^*$ [%]
Amount content of $^{243}\text{Am}$	0.038	0.029	0.0081

<sup>1)</sup> Dataset without trend correction

Therefore, the between-unit standard deviation ( $s_{bb}$ ) can be used as estimate of  $u_{bb}$ . As  $u_{bb}^*$  sets the limits of the study to detect inhomogeneity, the larger value of  $s_{bb}$  and  $u_{bb}^*$  is adopted as uncertainty contribution to account for potential inhomogeneity.

The correction of the data for the analytical trend as described earlier in this section (Equation 1) was applied solely for the purpose of the homogeneity evaluation. It was not applied for the value assignment (Sections 6 and 7) since its influence on the mean value for the amount content of  $^{243}\text{Am}$  was negligible and is included in the uncertainty contribution ( $s_{bb,rel}$ ).

## 4.2 Homogeneity of the Am isotope ratios

The IRMM-0243 reference material was prepared by dilution of an americium starting solution in nitric acid and dispensing into ampoules. Any differences in the isotope amount ratios could only stem from a contamination with americium of a different isotopic composition during processing. A dedicated glove box was used for the dilution and dispensing, in which no other americium source was kept. Therefore any contamination originating from another americium source can be excluded. This was confirmed by the characterisation study on randomly selected units where no heterogeneity in the isotope ratios was observed.

From these studies and considerations it is concluded that the homogeneity contribution to the uncertainty of the isotope ratios is negligible and therefore, no additional uncertainty component ( $u_{bb,rel} = 0$ ) was applied.

## 4.3 Within-unit homogeneity and minimum sample intake

The within-unit homogeneity is closely related to the minimum sample intake. The minimum sample intake is the minimum amount of sample that is representative for the whole unit and thus should be used in an analysis. Using sample sizes equal to or above the minimum sample intake guarantee the certified value within its stated uncertainty.

Quantification of within-unit inhomogeneity was not necessary for IRMM-0243 because the material is a true solution and as such can be regarded as completely homogeneous. Therefore, no minimum sample intake needs to be taken into account for analysis.

## 5. Stability

Contamination, evaporation of solvent and adsorption on glass walls are considered as the most relevant factors that could affect the stability of both the isotope ratios and the amount content of americium in IRMM-0243. Contamination prior to use (opening of the ampoules) is excluded since the IRMM-0243 reference material is supplied in glass ampoules with a screw-cap and sealed in PVC bags. Therefore only the influence of evaporation (temperature) and adsorption needed to be investigated.

Stability testing is necessary to establish conditions for storage (long-term stability) as well as conditions for dispatch to the customers (short-term stability). During transport, especially in summer time, temperatures up to 60 °C could be reached and stability under these conditions must be demonstrated. It should be noted that the term stability in this context does not refer to radioactive decay. It is self-evident that the radionuclides are decaying according to their half-lives, a process which is quantitatively predictable using the decay data [8, 9].

### **5.1 Short-term stability**

The shipment of nuclear materials follows the legal requirements related to radioprotection measures for transport of radioactive materials. The packaging of IRMM-0243 consists of two parts: an inner and an outer container according to regulations and respective procedures [10]. Units of IRMM-0243 are double sealed in PVC bags and put into a Type A inner container for radioactive materials. The inner container is then placed in a larger outer container, sealed and finally shipped to customers. From the package material specification and the fact that the transport does not take longer than one week, the ampoules of IRMM-0243 packed as described above are not exposed to temperatures outside the range of 4 to 60 °C.

No measurable evaporation or adsorption on the walls that would affect the concentration of americium in the solution is expected to occur during shipment. The isotope ratios are not affected by evaporation or adsorption. Taking this into account, the IRMM-0243 is considered stable regarding its isotopic composition and the amount content during dispatch and can be shipped to customers under normal temperature conditions. No additional uncertainty component ( $u_{\text{sts, rel}} = 0$ ) was applied.

### **5.2 Long-term stability**

No dedicated long-term stability study has been conducted for IRMM-0243. Experience acquired over the years with various nuclear CRMs at JRC-Geel has demonstrated that nitric acid solutions containing radioactive materials at various concentration levels can be stored at room temperature for long period with no measurable change in isotopic composition or in amount contents. In addition the measurements for the characterisation and homogeneity study (carried out between January and July 2016) were also used to demonstrate the stability of IRMM-0243 during the certification campaign. No significant change in the amount content of  $^{243}\text{Am}$  was observed in the selected units of IRMM-0243 (Annex C). Taking these considerations into account, no additional uncertainty component ( $u_{\text{lts, rel}} = 0$ ) was applied.

Nevertheless, the IRMM-0243 will be subjected to post-certification monitoring to control its stability. Two units of IRMM-0243 will be analysed every year to verify the certified values. The validity of the material certificate is 3 years and may be extended after further stability test are carried out.

## **6. Characterisation**

The material characterisation is the process of determining the property values of a reference material. The material characterisation was based on a primary method of measurement [6, 11], confirmed by an independent method.

*A primary method is "a method having the highest metrological qualities, whose operation(s) can be completely described and understood and for which a complete uncertainty statement can be written in terms of SI units. A primary ratio method measures the value of a ratio of an unknown to a standard of the same quantity; its operation must be completely described by a measurement equation [2, 11]".*

The material characterisation for the amount content  $^{243}\text{Am}$  was based on IDMS using  $^{241}\text{Am}$  spike produced in-house (see Section 6.1) and for the  $n(^{241}\text{Am})/n(^{243}\text{Am})$  and  $n(^{242\text{m}}\text{Am})/n(^{243}\text{Am})$  isotope amount ratios on TIMS. Independent verification measurements were carried out by alpha particle spectrometry, alpha particle counting at DSA and high-resolution gamma-ray spectrometry at the JRC.G.2 Laboratory for Radionuclide Metrology.

The measurements for the characterisation and homogeneity assessment were carried out together on the same selected ampoules (see Section 4).

## 6.1 $^{241}\text{Am}$ in-house spike

A  $^{241}\text{Am}$  spike material is required in order to determine the amount content of  $^{243}\text{Am}$  in selected ampoules of IRMM-0243 by IDMS. For that reason a  $^{241}\text{Am}$  spike was produced from highly enriched  $^{241}\text{Pu}$  material (mass fraction  $m(^{241}\text{Pu})/m(\text{Pu}) = 0.993$ ) available at JRC-Geel. The ingrown  $^{241}\text{Am}$ , which is produced by beta-decay of  $^{241}\text{Pu}$  after purification of the material, is used as a spike for the measurements of the  $^{243}\text{Am}$ .

The purification of about 2 mg  $^{241}\text{Pu}$  material was accomplished by anion exchange separations (Bio Rad AG<sup>®</sup> 1-X4 resin, 100-200 mesh, Hercules, USA) [12]. The purification procedure was performed three times in order to remove all the decay products that had been growing in since the production of the material in 1991. The final purification was carried out on June 10, 2014 at 15:10 CET, which marked the start of the  $^{241}\text{Am}$  in-growth (time zero). The completeness of the purification of  $^{241}\text{Pu}$  from  $^{241}\text{Am}$  was confirmed by high-resolution gamma-ray spectrometry.

The purified plutonium material was diluted with 100 mL nitric acid solution ( $c = 1$  mol/L) and characterised for the amount content of  $^{241}\text{Pu}$  and isotopic composition by IDMS (using a  $^{242}\text{Pu}$  spike CRM, IRMM-049d) and TIMS, respectively.

The value for the amount content of  $^{241}\text{Pu}$  was calculated from the mean of the 10 blends measured by IDMS. Each blend was measured in replicates on the Triton TIMS (ThermoFischer Scientific, Bremen, Germany). The mass fractionation correction was based on the measurement of the IRMM-290b-A3 Pu isotopic standard. The results of the characterisation measurement of  $^{241}\text{Pu}$  solution are shown in Table 3.

**Table 3** The amount content of  $^{241}\text{Pu}$  and the isotope amount fractions in the  $^{241}\text{Pu}$  solution after purification. The reference date of the values is June 10, 2014 15:10 CET (time zero)

		Value	Uncertainty <sup>1)</sup>
Amount content [nmol/g (solution)]	$^{241}\text{Pu}$	81.394	0.053
Isotope amount fractions ( $\cdot 100$ ) [mol/mol]	$n(^{238}\text{Pu})/n(\text{Pu})$	0.00097	0.00034
	$n(^{239}\text{Pu})/n(\text{Pu})$	0.00053	0.00018
	$n(^{240}\text{Pu})/n(\text{Pu})$	0.25741	0.00029
	$n(^{241}\text{Pu})/n(\text{Pu})$	99.29993	0.00052
	$n(^{242}\text{Pu})/n(\text{Pu})$	0.44116	0.00025

<sup>1)</sup> Expanded ( $k = 2$ ) uncertainty

## 6.2 Chemical purification and isotope measurements

Spiking was carried out about one and a half years after the purification of the  $^{241}\text{Pu}$  material. During this period a sufficient amount of  $^{241}\text{Am}$  spike had been produced for IDMS analysis. From each of the eighteen selected units, an aliquot (2.5-2.7 g) was taken and mixed with an aliquot (about 1.0 g) of the  $^{241}\text{Pu}$  spike solution with the ingrown  $^{241}\text{Am}$  for determination of the amount content of  $^{243}\text{Am}$  by IDMS, and another aliquot (un-spiked, about 0.7 g) for determination of the isotopic composition (e.g.

isotope amount ratios) by TIMS. Spiking was done by accurate weighing using the substitution method. The mass of the  $^{241}\text{Pu}$  spike solution was monitored to account for evaporation losses before each weighing.

In substitution weighing, the mass of a sample is determined through a series of mass determinations of an unknown (U) and a reference weight (S). The so called "SUUS" method was applied. The uncertainty contributions in substitution weighing are the uncertainties associated with the calibrated reference weights, air buoyancy correction and the variability of the balance used in "SUUS" method.

Prior to the isotope measurements, two purification steps were carried out to remove the  $^{241}\text{Pu}$  from the in-grown  $^{241}\text{Am}$  in the blends by means of UTEVA-spec and DGA extraction resins (Triskem International, Bruz, France), respectively [13]. The un-spiked samples underwent the same purification protocol as the blends.

Isotope ratio measurements were performed on the multi-collector Triton Thermal Ionisation Mass Spectrometer (ThermoFisher Scientific, Bremen, Germany) using the total evaporation (TE) method.

The measurements were carried out in a similar manner as routinely performed for uranium and plutonium samples of similar size. In the TE method, the evaporation filament is heated up to maintain a steady intensity and measured until the whole sample is consumed. In this way, the fractionation effects in the ion source are minimized [14, 15, 16, 17]. Degassed zone refined rhenium was used as ionization and evaporation filaments (double filament configuration). About 20 ng of americium nitrate solution was deposited on an evaporation filament, evaporated and mounted on a sample turret. All ion currents were measured simultaneously on Faraday cups.

For uranium and plutonium a mass fractionation correction by measurement of a CRM on the same magazine is recommended following ASTM C1672-17 [18]. According to ASTM C1672-17, the bias for uncorrected  $n(^{235}\text{U})/n(^{238}\text{U})$  and  $n(^{242}\text{Pu})/n(^{239}\text{Pu})$  ratios is below 0.05 % within 2 standard deviations ( $2\sigma$ ) for ratios spanning 3 mass units. In the absence of a suitable Am isotopic standard, it was not possible to perform a correction for mass fractionation during isotope ratio measurements for this project. Due to similar chemical behaviour and similar ionization energies, it can be assumed that americium behaves similar to uranium and plutonium during the TE measurement and that the bias statements for uranium and plutonium can equally be applied to americium. For the  $n(^{243}\text{Am})/n(^{241}\text{Am})$  ratio this maximum expected bias was applied as an additional uncertainty component and set to 0.033 % ( $2\sigma$ , for ratios spanning 2 mass units) and for the  $n(^{242\text{m}}\text{Am})/n(^{243}\text{Am})$  ratio to 0.017 % ( $2\sigma$ , for ratios spanning 1 mass unit), marked as "K-TE" in Figure 3.

In addition, an uncertainty of 0.020 % ( $2\sigma$ , for ratios spanning 3 mass units) derived from isotope measurements of the U and Pu quality control samples (gravimetrically prepared CRMs IRMM-074/2,3 for U and IRMM-290-G3 for Pu) was added as a conservative contribution to the uncertainty budget for Am, marked as "K-QC" in Figure 3.

## 6.4 The amount content of $^{243}\text{Am}$ and its uncertainty

The amount content of  $^{243}\text{Am}$  ( $c_{243\text{Am}}$ ) in the selected units was calculated using the equations below:

$$c_{243\text{Am}} = \frac{R_y - R_b}{R_b - R_x} \cdot R_x \cdot \frac{m_y}{m_x} \cdot c_{241\text{Am}} \quad \text{Equation 5}$$

$$c_{241\text{Am}} = c_{241\text{Pu}}^0 \frac{\lambda_{241\text{Pu}}}{\lambda_{241\text{Am}} - \lambda_{241\text{Pu}}} (e^{-\lambda_{241\text{Pu}} \Delta t} - e^{-\lambda_{241\text{Am}} \Delta t}) + c_{241\text{Am}}^0 e^{-\lambda_{241\text{Am}} \Delta t} \quad \text{Equation 6}$$

$R_x$	isotope amount ratio $n(^{243}\text{Am})/n(^{241}\text{Am})$ in the un-spiked sample [mol/mol]
$R_y$	isotope amount ratio $n(^{243}\text{Am})/n(^{241}\text{Am})$ in the spike [mol/mol]
$R_b$	isotope amount ratio $n(^{243}\text{Am})/n(^{241}\text{Am})$ in the blend [mol/mol]
$m_x$	mass of the sample [g]
$m_y$	mass of the spike [g]
$C_{241\text{Am}}$	amount content of the in-grown $^{241}\text{Am}$ spike [mol/g]
$C_{241\text{Pu}}^0$	amount content of $^{241}\text{Pu}$ spike solution at the time of purification [mol/g]
$C_{241\text{Am}}^0$	(residual) amount content of $^{241}\text{Am}$ in $^{241}\text{Pu}$ spike solution after purification [mol/g]
$\Delta t$	time for the ingrowth of $^{241}\text{Am}$ [a]
$\lambda_{241\text{Am}}$	decay constants of $^{241}\text{Am}$ [ $\text{a}^{-1}$ ]
$\lambda_{241\text{Pu}}$	decay constants $^{241}\text{Pu}$ [ $\text{a}^{-1}$ ]

The half-lives and associated uncertainties ( $k = 1$ ) used for the calculations of the respective decay constants ( $\lambda = \ln 2/T_{1/2}$ ) were  $14.325 \pm 0.012$  a [8] for  $^{241}\text{Pu}$  and  $432.6 \pm 0.6$  a [9] for  $^{241}\text{Am}$ . The residual amount of  $^{241}\text{Am}$  in the  $^{241}\text{Pu}$  spike solution after the purification (June 10, 2014 15:10 CET) was found negligible ( $C_{241\text{Am}}^0 = 0$ , Equation 6). This was confirmed by high-resolution gamma-ray spectrometry (see Section 6.1)

The amount content of  $^{243}\text{Am}$  and its associated expanded uncertainty was calculated according to the Guide to the Expression of Uncertainty in Measurement (GUM) [3] using the GUM Workbench Software. Detailed calculations and associated uncertainty budgets are given in Annex D.


The amount content of  $^{243}\text{Am}$  and its standard uncertainty from characterisation are given in Table 4.

**Table 4** The amount content of  $^{243}\text{Am}$  and its standard uncertainty from characterisation in IRMM-0243

	Mean [nmol/g (solution)]	$u_{\text{char}}$ [nmol/g (solution)]	$u_{\text{char, rel}}$ [%]
Amount content of $^{243}\text{Am}$	5.6960	0.0052	0.091

The uncertainty from characterisation of the amount content of  $^{243}\text{Am}$  is dominated by the uncertainties associated with the half-life of  $^{241}\text{Pu}$  and the amount content of the  $^{241}\text{Pu}$  spike solution. Relative contributions of various uncertainty components from characterisation are shown in Figure 3.

**Figure 3** Uncertainty contributions from characterisation of the amount content of  $^{243}\text{Am}$



### 6.5 The amount ratios of $n(^{241}\text{Am})/n(^{243}\text{Am})$ and $n(^{242\text{m}}\text{Am})/n(^{243}\text{Am})$ and their uncertainties

The Am isotope amount ratios and their standard uncertainties from characterisation are given in Table 5.

**Table 5** The  $n(^{241}\text{Am})/n(^{243}\text{Am})$  and  $n(^{242\text{m}}\text{Am})/n(^{243}\text{Am})$  isotope amount ratios and their standard uncertainties from characterisation in IRMM-0243

	Mean [mol/mol]	$U_{\text{char}}$ [mol/mol]	$U_{\text{char, rel}}$ [%]
$n(^{241}\text{Am})/n(^{243}\text{Am})$	0.136138	0.000027	0.020
$n(^{242\text{m}}\text{Am})/n(^{243}\text{Am})$ <sup>1)</sup>	0.00015063	0.00000077	0.51

<sup>1)</sup> See section 6.7

The uncertainty from characterisation of the  $n(^{241}\text{Am})/n(^{243}\text{Am})$  and  $n(^{242\text{m}}\text{Am})/n(^{243}\text{Am})$  isotope amount ratios consists of the standard deviation of the replicate measurements, the variability between selected ampoules, an uncertainty component [18] from the potential bias during the isotope ratio measurement (see Section 6.2) and an uncertainty component from the quality control measurements. In the case of  $n(^{242\text{m}}\text{Am})/n(^{243}\text{Am})$  amount ratio, an additional source of uncertainty was added to account for the  $^{242}\text{Cm}$  impurity in the IRMM-0243 material (see Annex A).

Results from the characterisation measurements for the amount content of  $^{243}\text{Am}$  and  $n(^{241}\text{Am})/n(^{243}\text{Am})$  and  $n(^{242\text{m}}\text{Am})/n(^{243}\text{Am})$  amount ratios are shown in Annex E.

## 6.6 Verification measurements

Independent verification measurement for the amount content of  $^{243}\text{Am}$  and  $^{241}\text{Am}$  and for the  $n(^{241}\text{Am})/n(^{243}\text{Am})$  amount ratios were performed applying radiometric methods at the JRC.G.2 Laboratory for Radionuclide Metrology.

High-resolution alpha particle spectrometry was performed to determine the  $A(^{243}\text{Am})/A(^{241}\text{Am})$  activity ratio of the Am reference solution [19, 20, 21]. The alpha source was prepared from the Am CRM solution by electrodeposition on a polished stainless steel disk with an active diameter of 18.6 mm. The measurement was performed using a passivated ion-implanted planar silicon detector (PIPS<sup>®</sup>, 150 mm<sup>2</sup> active area, Mirion Technologies (MGPI) SA, France). The total activity per unit mass was determined by means of alpha-particle counting at a defined solid angle (DSA) [19, 22]. Gravimetrically quantified drops of americium solution were deposited on 34 mm glass plates and covered with 20 µg/cm<sup>2</sup> VYNS foils to prevent material loss.

Two gamma measurement campaigns were carried out, one in the underground laboratory "HADES" [23] and one above ground in the radionuclide metrology laboratory at JRC-Geel. Point-like sources for gamma measurements were prepared gravimetrically by drop deposition on laminated plastic foils with a diameter of 34 mm. They were measured above ground using two coaxial HPGe gamma-ray spectrometers, one of 35 % and the other of 90 % relative efficiency (Mirion Technologies (MGPI) SA, France). One gamma source was selected for ultra-low-level gamma-ray spectrometry measurements (ULGS) in HADES laboratory using the Ge-8 detector. The Ge-8 is a HPGe detector of type Broad Energy Germanium Detector (BEGe) with a relative efficiency of 20 % (Canberra, U.S.A). The activities of  $^{241}\text{Am}$  and  $^{243}\text{Am}$  were calculated based on the main gamma-ray peaks of the nuclides, i.e. the 59.54 keV line for  $^{241}\text{Am}$  and the 74.66 keV line for  $^{243}\text{Am}$ , respectively [23, 24, 25].

The measurement results from the three independent methods confirmed the certified values of the  $^{243}\text{Am}$  and  $^{241}\text{Am}$  amount contents and  $n(^{241}\text{Am})/n(^{243}\text{Am})$  amount ratio within measurement uncertainties. Measurement uncertainties were estimated according to GUM [3] and mean values calculated by means of the power-moderated mean formalism [26]. The results of the verification measurements are shown in Annex F.

## 6.7 Inter-laboratory comparison

The IRMM-0243 certification project provided a unique opportunity for the measurement community to assess the current state of practice for americium measurements. To this end, the  $^{243}\text{Am}$  inter-laboratory comparison (ILC) was organised by CEA/CETAMA. This ILC also constitutes an efficient external control of the reference values certified by the JRC-Geel and consolidates the characterisation measurement data.

An ampoule of IRMM-0243 material with undisclosed values was distributed to the participants. The laboratories were asked to measure the amount content of  $^{241}\text{Am}$ ,  $^{243}\text{Am}$  and total Am and the  $n(^{241}\text{Am})/n(^{243}\text{Am})$  and  $n(^{242\text{m}}\text{Am})/n(^{243}\text{Am})$  isotope amount ratios using their analytical procedures. A detailed ILC report will be published in due time.

In this certification report, we compared the certified values ( $X_{\text{CRM}}$ ) with the weighted mean ( $X_{\text{ILC}}$ ) of the participants using the compatibility equation [27] below:

$$\text{compatibility} = \frac{|X_{\text{ILC}} - X_{\text{CRM}}|}{\sqrt{u_{\text{ILC}}^2 + u_{\text{CRM}}^2}} \quad \text{Equation 7}$$

$X_{\text{ILC}}$	weighted mean of the participant results (ILC)
$X_{\text{CRM}}$	certified value
$u_{\text{ILC}}$	standard uncertainty of the weighted mean (ILC)

$u_{CRM}$  standard uncertainty of the certified value

The results of the compatibility evaluation are shown in Table 6.

**Table 6** Results of the compatibility evaluation

Measurand	Compatibility
Amount content of $^{243}\text{Am}$	0.3
Amount content of $^{241}\text{Am}$	0.1
Amount content of Am	0.3
$n(^{241}\text{Am})/n(^{243}\text{Am})$	0.8
$n(^{242m}\text{Am})/n(^{243}\text{Am})$	9.4

It can be concluded from Table 6 that there is a good agreement between the certified values and the weighted mean values of the participants' results (compatibility  $\leq 2$  at a 95 % CI) [27] for all the measurands, except for the  $n(^{242m}\text{Am})/n(^{243}\text{Am})$  isotope amount ratio. For the  $n(^{242m}\text{Am})/n(^{243}\text{Am})$  isotope amount ratio, a lower value ( $0.0001373 \pm 0.0000024$ ) was obtained from the participants (ILC) compared to the characterised value by TIMS ( $0.0001506 \pm 0.0000015$ , see Section 6.5). Additional measurements of the procedural blank at JRC-Geel showed several hundred cps at masses 238, 239 and 242. This may explain the higher value of the  $n(^{242m}\text{Am})/n(^{243}\text{Am})$  isotope amount ratio. It is very likely that these interferences come from reagents and resins used in the chemical purification since they were not observed in the ion source (TIMS) or on the rhenium filaments. The  $n(^{242m}\text{Am})/n(^{243}\text{Am})$  isotope amount ratio could not be corrected quantitatively for this effect and, therefore, only an indicative value based on the ILC results has been assigned to the  $n(^{242m}\text{Am})/n(^{243}\text{Am})$  isotope amount ratio (see Section 7.2).

The results of the CEA/CETAMA  $^{243}\text{Am}$  ILC provided a valuable external verification for IRMM-0243 and underpinned confidence in the certified values.

## 7. Value Assignment

Certified values were assigned.

Certified values are values that fulfil the highest standards of accuracy. Certified values for IRMM-0243 were assigned on the basis of IDMS and TIMS and verified by independent measurement methods. Full uncertainty budgets in accordance with the GUM [3] were established.

### 7.1 Certified values and their uncertainties

The certified value for each parameter was assigned as the unweighted mean of the nine individual values from characterisation (Annex E).

The uncertainty of the assigned values consists of uncertainties related to characterisation,  $u_{char}$  (Section 6), potential between-ampoule inhomogeneity,  $u_{bb}$  (Section 4), and potential degradation during transport,  $u_{sts}$  and long-term storage,  $u_{lts}$  (Section 5). As described in Section 5 no additional uncertainty from stability (short- and long-term) was applied. These different contributions were combined to estimate the expanded, uncertainty of the certified value ( $U_{CRM}$ ) with a coverage factor  $k$  as:

$$U_{\text{CRM}} = k \cdot \sqrt{u_{\text{char}}^2 + u_{\text{bb}}^2} \quad \text{Equation 8}$$

- $u_{\text{char}}$  was estimated as described in Section 6
- $u_{\text{bb}}$  was estimated as described in Section 4.

Because of sufficient degrees of freedom of the different uncertainty contributions, a coverage factor  $k$  of 2, corresponding to a level of confidence of about 95 %, was applied to obtain the expanded uncertainties for all certified properties. The certified values and their uncertainties are summarised in Table 7 and Table 8.

**Table 7** Certified amount content of  $^{243}\text{Am}$  and its uncertainty in IRMM-0243

IRMM-0243	Certified value <sup>1)</sup> [nmol/g (solution)]	$u_{\text{char, rel}}$ [%]	$u_{\text{bb, rel}}$ [%]	$U_{\text{CRM, rel}}$ <sup>2)</sup> [%]	$U_{\text{CRM}}$ <sup>2)</sup> [nmol/g (solution)]
Amount content of $^{243}\text{Am}$	5.696	0.091	0.029	0.19	0.011

<sup>1)</sup> The reference date for the certified value is January 1, 2017.

<sup>2)</sup> Expanded ( $k = 2$ ) uncertainty

**Table 8** Certified isotope amount ratio of  $n(^{241}\text{Am})/n(^{243}\text{Am})$  and its uncertainty in IRMM-0243

IRMM-0243	Certified value <sup>1)</sup> [mol/mol]	$u_{\text{char, rel}}$ [%]	$U_{\text{CRM, rel}}$ [%]	$U_{\text{CRM}}$ <sup>2)</sup> [mol/mol]
$n(^{241}\text{Am})/n(^{243}\text{Am})$	0.136138	0.020	0.040	0.000054

<sup>1)</sup> The reference date for the certified values is January 1, 2017.

<sup>2)</sup> Expanded ( $k = 2$ ) uncertainty

The values for the amount contents of  $^{241}\text{Am}$  and total Am, the mass fractions of  $^{241}\text{Am}$ ,  $^{243}\text{Am}$  and total Am, the isotope amount and mass fractions (isotopic composition) and the molar mass of Am in IRMM-0243 were calculated by using the certified values of the amount content of  $^{243}\text{Am}$ , the  $n(^{241}\text{Am})/n(^{243}\text{Am})$  isotope amount ratio and the isotope atomic masses (Table 10). These derived certified values were calculated from the certified values from the characterisation study applying a well understood mathematical model. They are summarised in Table 9.

**Table 9** Derived certified amount contents, mass fractions, isotope mass and amount fractions, molar mass and their uncertainties

IRMM-0243		Certified value <sup>1) 3)</sup>	Uncertainty <sup>2)</sup>
Amount contents [nmol/g (solution)]	<sup>241</sup> Am	0.7754	0.0015
	Am	6.472	0.012
Mass fractions [μg/g (solution)]	<sup>241</sup> Am	0.18692	0.00036
	<sup>243</sup> Am	1.3845	0.0026
	Am	1.5716	0.0030
Am isotope amount fractions (·100) [mol/mol]	$n(^{241}\text{Am})/n(\text{Am})$	11.9810	0.0042
	$n(^{243}\text{Am})/n(\text{Am})$	88.0069	0.0042
Am isotope mass fractions (·100) [g/g]	$m(^{241}\text{Am})/m(\text{Am})$	11.8940	0.0042
	$m(^{243}\text{Am})/m(\text{Am})$	88.0940	0.0042
Molar mass of Am [g/mol]	$M(\text{Am})$	242.821094	0.000085

<sup>1)</sup> The reference date for the certified values is January 1, 2017.

<sup>2)</sup> Expanded ( $k = 2$ ) uncertainty

<sup>3)</sup> The derived certified values

**Table 10** Atomic masses of Am isotopes used for calculation of the derived values [28]

Isotope	Atomic mass [g/mol]	Uncertainty ( $k = 1$ ) [g/mol]
<sup>241</sup> Am	241.0568293	0.0000019
<sup>242m</sup> Am	242.0595494	0.0000019
<sup>243</sup> Am	243.0613813	0.0000024

## 7.2 Indicative values and their uncertainties

The indicative value for the  $n(^{242\text{m}}\text{Am})/n(^{243}\text{Am})$  isotope amount ratio and its uncertainty was assigned based on the <sup>243</sup>Am ILC organised by CEA/CETAMA using IRMM-0243 as the test sample. The assigned value is the DerSimonian-Laird weighted mean [29] of the results from nine participating laboratories (Annex G). The indicative values for the  $m(^{242\text{m}}\text{Am})/m(\text{Am})$  mass and  $n(^{242\text{m}}\text{Am})/n(\text{Am})$  amount fractions were calculated using the indicative value of  $n(^{242\text{m}}\text{Am})/n(^{243}\text{Am})$  isotope amount ratio and the certified values. The indicative values and their uncertainties are summarised in Table 11.

**Table 11** Indicative values and their uncertainties in IRMM-0243

IRMM-0243		Indicative value <sup>1)</sup>	Uncertainty <sup>2)</sup>
Isotope amount ratios [mol/mol]	$n(^{242m}\text{Am})/n(^{243}\text{Am})$	0.0001373	0.0000024
Isotope amount fractions ( $\cdot 100$ ) [mol/mol]	$n(^{242m}\text{Am})/n(\text{Am})$ <sup>3)</sup>	0.01208	0.00021
Isotope mass fractions ( $\cdot 100$ ) [g/g]	$m(^{242m}\text{Am})/m(\text{Am})$ <sup>3)</sup>	0.01205	0.00021

<sup>1)</sup> The reference date for the indicative values is January 1, 2017.

<sup>2)</sup> Expanded ( $k = 2$ ) uncertainty

<sup>3)</sup> The derived indicative values

## 8. Metrological traceability and commutability

### 8.1 Metrological traceability

#### Identity

The measurands are structurally defined and independent of the measurement method.

#### Quantity value

The certified values are traceable to the SI via the values on the respective certificates of IRMM-049d and IRMM-0290b-A3 and via the half-life of  $^{241}\text{Pu}$  [8]. The masses obtained by substitution weighing are traceable to the SI through the use of calibrated balances and reference weights.

### 8.2 Commutability

Many measurement procedures include one or more steps, which are selecting specific analytes (or specific groups of analytes) from the sample for the subsequent steps of the whole measurement process. Often the complete identity of these 'intermediate analytes' is not fully known or taken into account. Therefore, it is difficult to mimic all the analytically relevant properties of real samples within a CRM. The degree of equivalence in the analytical behaviour of real samples and a CRM with respect to various measurement procedures (methods) is summarised in a concept called 'commutability of a reference material'. There are various definitions expressing this concept. For instance, the CLSI Guideline C-53A [30] recommends the use of the following definition for the term *commutability*:

*"The equivalence of the mathematical relationships among the results of different measurement procedures for an RM and for representative samples of the type intended to be measured."*

The commutability of a CRM defines its fitness for use and, thus, is a crucial characteristic in case of the application of different measurement methods. When commutability of a CRM is not established in such cases, the results from routinely used methods cannot be legitimately compared with the certified value to determine whether a bias does not exist in calibration, nor can the CRM be used as a calibrant.

This reference material is a tailor-made  $^{243}\text{Am}$  spike CRM to be applied for quantification of americium content in various nuclear samples by IDMS. The analytical behaviour of the IRMM-0243 will be the same as for a routine sample.

## 9. Instructions for use

### 9.1 Safety information

The IRMM-0243 contains radioactive material. The ampoules should be handled with great care and by trained and experienced personnel in a laboratory suitably equipped for the safe handling of radioactive materials.

### 9.2 Storage conditions

The vials should be stored at  $+ 18\text{ °C} \pm 5\text{ °C}$ .

Please note that the European Commission cannot be held responsible for changes that happen during storage of the material at the customer's premises, especially for opened ampoules.

### 9.3 Preparation and use of the material

This material is ready for use after ampoule opening. Users are cautioned that once the ampoule is opened, the amount content of americium may be affected (e.g. by evaporation)

### 9.4 Minimum sample intake

The IRMM-0243 material is a true solution and, therefore no minimum sample intake needs to be taken into account for analysis.

### 9.5 Use of the certified value

The main purpose of the material is for use as a spike isotopic reference material for determination of the americium content in an unknown sample by isotope dilution through a measurement of the isotope amount ratio  $R_b = n(^{241}\text{Am})/n(^{243}\text{Am})$  in a blend.

The amount content ( $C_x$ ) of americium can be calculated with the aid of the following equation, which enables an easy quantification of the uncertainty sources in the procedure:

$$C_x = C_y \cdot \frac{m_y}{m_x} \cdot \frac{R_y - R_b}{R_b - R_x} \cdot \frac{\Sigma(R_i)_x}{\Sigma(R_i)_y} \quad \text{Equation 8}$$

$C_y$	element amount content of the spike material
$m_x$	mass of unknown sample used in the measurement
$m_y$	mass of the spike solution in the measurement
$R_x$	amount ratio $n(^{241}\text{Am})/n(^{243}\text{Am})$ in the unknown sample
$R_y$	amount ratio $n(^{241}\text{Am})/n(^{243}\text{Am})$ in the spike material
$R_b$	amount ratio $n(^{241}\text{Am})/n(^{243}\text{Am})$ in the blend
$\Sigma(R_i)_x$	sum of all isotope amount ratios in the unknown sample
$\Sigma(R_i)_y$	sum of all isotope amount ratios in the spike material

## 10. Conclusions

A novel  $^{243}\text{Am}$  spike reference material IRMM-0243 was prepared and certified for the amount content of  $^{243}\text{Am}$  and the  $n(^{241}\text{Am})/n(^{243}\text{Am})$  isotope amount ratio. The material was produced in compliance with international guidelines. Certified values for the amount content of  $^{243}\text{Am}$  and the  $n(^{241}\text{Am})/n(^{243}\text{Am})$  isotope amount ratios were established by IDMS and TIMS and confirmed by independent verification measurements. The indicative value for the  $n(^{242\text{m}}\text{Am})/n(^{243}\text{Am})$  isotope amount ratio and its uncertainty were derived from the  $^{243}\text{Am}$  inter-laboratory comparison (ILC) organised by CEA/CETAMA using IRMM-0243 as the test sample. The material is supplied in a glass ampule with a screw-cap containing about 3.5 mL of dilute nitric acid solution with an americium mass fraction of about 1.5  $\mu\text{g/g}$ . This material is for use as a spike isotopic reference material for determination of americium content by IDMS in unknown samples.

## References

- 1 ISO 17034, General requirements for the competence of reference materials producers, International Organization for Standardization, Geneva, Switzerland, 2016
- 2 ISO Guide 35, Reference materials – General and statistical principles for certification, International Organization for Standardization, Geneva, Switzerland, 2006
- 3 ISO/IEC Guide 98-3, Guide to the Expression of Uncertainty in Measurement (GUM, 1995), International Organization for Standardization, Geneva, Switzerland, 2008
- 4 S.K. Aggarwal, A review on the mass spectrometric studies of americium: Present status and future perspective, *Mass Spectrom. Review*, 2016, DOI: 10.1002/mas.21506
- 5 2016 Nuclear Security Summit: Certified Reference Material Fact Sheet: [http://www.belfercenter.org/sites/default/files/legacy/files/nuclearmatters/files/2016\\_nss\\_certified\\_reference\\_material\\_factsheet.pdf?m=1461101302](http://www.belfercenter.org/sites/default/files/legacy/files/nuclearmatters/files/2016_nss_certified_reference_material_factsheet.pdf?m=1461101302)
- 6 J. Vogl, W. Pritzkow, Isotope dilution mass spectrometry – a primary method of measurement and its role for RM certification, *J. Metrol. Soc. India*, 25, 135-164, 2010
- 7 T.P.J. Linsinger, J. Pauwels, A.M.H. van der Veen, H. Schimmel, A. Lamberty, Homogeneity and stability of reference materials, *Accred. Qual. Assur.*, 6, 20-25, 2001
- 8 R. Wellum, A. Verbruggen, R. Kessel, A new evaluation of the half-life of  $^{241}\text{Pu}$ , *J. Anal. At. Spectrom.*, 24, 801-807, 2009
- 9 DDEP-BIPM (2004-2013). Monographie BIPM-5. Table of Radionuclides. Retrieved from [http://www.nucleide.org/DDEP\\_WG/DDEPdata.htm](http://www.nucleide.org/DDEP_WG/DDEPdata.htm) "last checked on July 19, 2017"
- 10 IAEA Regulations for the Safe Transport of Radioactive Material, 1996 Edition, "Requirements", IAEA Safety Standards Series No. TS-R-1 (ST-1, Revised), IAEA, Vienna, 2000
- 11 M. J. Milton and T. J. Quinn, Primary methods for the measurement of amount of substance, *Metrologia*, 38, 289-296, 2001
- 12 R. Jakopic, A. Verbruggen, R. Eykens, F. Kehoe, F. Kuhn, H. Kushigeta, U. Jacobsson, J. Bauwens, S. Richter, R. Wellum, Y. Aregbe, An inter-calibration campaign using various selected Pu spike isotopic reference materials, *J. Radioanal. Nucl. Chem.*, 286, 449-454, 2010
- 13 Z.H. Kazi, J.R. Cornett, X. Zhao, L. Kieser, Americium and plutonium separation by extraction chromatography for determination by accelerator mass spectrometry, *Anal. Chim. Acta*, 829, 75-80, 2014
- 14 E.L. Callis, R.M. Abernathy, High precision isotopic analysis of uranium and plutonium by total sample volatilization and signal integration, *Int. J. Mass Spectrom. Ion Process*, 103, 322-327, 1991
- 14 S. Richter, S.A. Goldberg, Improved techniques for high accuracy isotope ratio measurement of nuclear materials using thermal ionisation mass spectrometry, *Int. J. Mass Spectrom.*, 229, 181-197, 2003
- 16 S. Bürger, S.D. Balsley, S.D. Baumann, S.F. Boulyga, J.A. Cunningham, S. Kappel, A. Kopft, J. Poths, Uranium and Plutonium analysis of nuclear material samples by

- multi-collector thermal ionisation mass spectrometry: quality control, measurement uncertainty, and metrological traceability, *Int. J. Mass Spectrom.*, 311, 40-50, 2012
- 17 S. Boulyga, S. Konnegger-Kappel, S. Richter, L. Sangley, Mass spectrometry analysis for nuclear safeguards, *J. Anal. At. Spectros.*, 30, 1469-1489, 2015
  - 18 Standard test Method for determination of uranium or plutonium isotopic composition or concentration by the total evaporation Method using a thermal ionization mass spectrometer, C1672-17, ASTM International
  - 19 S. Pommé, G. Sibbens, Alpha-particle counting and Spectrometry in a primary Standardisation laboratory, *Acta Chim. Slov.*, 52, 111-119, 2008
  - 20 S. Pommé, B.C. Marroyo, Improved peak shape fitting in alpha spectra, *App. Rad. Isot.*, 96, 148-153, 2015
  - 21 S. Pommé, Typical uncertainties in alpha-particle spectrometry, *Metrologia*, 52, S146-S155, 2015
  - 22 S. Pommé, The uncertainty of counting at a defined solid angle, *Metrologia*, 52, S73-S85, 2015
  - 23 M. Hult, W. Preusse, J. Gasparro, M. Kohler, Underground gamma-ray Spectrometry, *Acta Chim. Slov.*, 53, 1-7, 2006
  - 24 M. Hult, Low-level gamma-ray spectrometry using Ge-detectors, *Metrologia*, S87-S94, Erratum at *Metrologia*, 44/5, 425, 2007
  - 25 I. Kawrakow, D.W.O. Rogers, PIRS-701: The EGSnrc Code System: Monte Carlo simulation of Electron and Photon Transport, Ionizing Radiation Standards (NRC, Ottawa, Ontario), 2003
  - 26 S. Pommé, J. Keightley, Determination of a reference value and its uncertainty through a power-moderated mean, *Metrologia*, 52, 200-212, 2015
  - 27 JCGM 200:2012, International vocabulary of metrology – Basic and general concepts and associated terms (VIM), 2012
  - 28 M. Wang, G. Audi, A. H. Wapstra, F. G. Kondev, M. MacCormick, X. Xu, B. Pfeiffer, The AME2012 atomic mass evaluation (II). Tables, Graphs and References, *Chinese Physics C*, Vol. 36, No.12, 1603-2014, 2012
  - 29 C. Rivier, M. Désenfant, M. Crozet, C. Rigaux, D. Roudil, B. Tufféry, A. Ruas, Use of an excess variance approach for the certification of reference materials by interlaboratory comparison, *Accred. Qual. Assur.*, 19, 269-274, 2014
  - 30 H. Vesper, H. Emons, M. Gnezda, C. P. Jain, W. G. Miller, R. Rej, G. Schumann, J. Tate, L. Thienpont, J. E. Vaks, Characterization and Qualification of Commutable Reference Materials for Laboratory Medicine; Approved Guideline, CLSI document C53-A, Clinical and Laboratory Standards Institute, Wayne, PA, USA, 2010

## List of abbreviations and definitions

ANOVA	Analysis of variance
$c$	amount of substance concentration
CEA	Le Commissariat a l`energie atomique et aux energies alternatives, la Direction de l`energie nucléaire (Alternative Energies and Atomic Energy Commission)
CET	Central European Time
CETAMA	Commission d'Etablissement des Methodes d'Analyse
CI	Confidence interval
CLSI	Clinical and Laboratory Standards Institute
cps	Counts per second
CRM	Certified reference material
DSA	Defined solid angle
EC	European Commission
ESARDA	European Safeguards Research and Development Association
GUM	Guide to the Expression of Uncertainty in Measurement
IAEA	International Atomic Energy Agency
ICP-MS	Inductively coupled plasma mass spectrometry
IDMS	Isotope dilution mass spectrometry
ISO	International Organization for Standardization
ITWG	The Nuclear Forensics International Technical Working Group
JRC	Joint Research Centre of the European Commission
$k$	Coverage factor
$\lambda$	Decay constant
$m$	mass
$M$	Molar mass
$MS_{\text{between}}$	Mean of squares between-unit from an ANOVA
$MS_{\text{within}}$	Mean of squares within-unit from an ANOVA
$n$	amount of substance
p.a.	pro analysis
PIPS	passivated implanted planar silicon
PVC	Polyvinyl chloride
$R_b$	Isotope amount ratio in the blend
$R_x$	Isotope amount ratio in the un-spiked sample
$R_y$	Isotope amount ratio in the spike
rel	Index denoting relative figures (uncertainties etc.)
RM	Reference material

$s$	Standard deviation
$s_{bb}$	Between-unit standard deviation; an additional index "rel" is added when appropriate
SI	International System of Units
$s_{wb}$	Within-unit standard deviation
$T_{1/2}$	Half life
TE	Total evaporation
TIMS	Thermal Ionisation Mass Spectrometry
$u$	Standard uncertainty
$U$	Expanded uncertainty
$u_{bb}^*$	Standard uncertainty related to a maximum between-unit inhomogeneity that could be hidden by method repeatability; an additional index "rel" is added as appropriate
$u_{bb}$	Standard uncertainty related to a possible between-unit inhomogeneity; an additional index "rel" is added as appropriate
$u_{char}$	Standard uncertainty of the material characterisation; an additional index "rel" is added as appropriate
$u_{CRM}$	Combined standard uncertainty of the certified value; an additional index "rel" is added as appropriate
$U_{CRM}$	Expanded uncertainty of the certified value; an additional index "rel" is added as appropriate
$u_{lts}$	Standard uncertainty of the long-term stability; an additional index "rel" is added as appropriate
$u_{sts}$	Standard uncertainty of the short-term stability; an additional index "rel" is added as appropriate
ULGS	Ultra-low-level gamma-ray spectrometry
VVNS	polyvinylchloride- polyvinyl acetate copolymer
$\bar{y}$	Arithmetic mean
$\nu_{MS_{within}}$	Degrees of freedom of $MS_{within}$
WGDA	Working Group on Techniques and Standards for Destructive Analysis

## List of figures

Figure 1 Flow chart for the preparation and certification of IRMM-0243 .....	10
Figure 2 Processing steps of the americium reference material: a) dispensing of a solution, b) closing of an ampoule with a screw-cap, c) contamination check and PVC sealing and d) a unit of IRMM-0243 .....	12
Figure 3 Uncertainty contributions from characterisation of the amount content of <sup>243</sup> Am .....	20

## List of tables

Table 1 Results of the statistical evaluation of the homogeneity study for the amount content of $^{243}\text{Am}$ .....	14
Table 2 Results of the homogeneity study for the amount content of $^{243}\text{Am}$ .....	15
Table 3 The amount content of $^{241}\text{Pu}$ and the isotope amount fractions in the $^{241}\text{Pu}$ solution after purification. The reference date of the values is June 10, 2014 15:10 CET (time zero) .....	17
Table 4 The amount content of $^{243}\text{Am}$ and its standard uncertainty from characterisation in IRMM-0243 .....	19
Table 5 The $n(^{241}\text{Am})/n(^{243}\text{Am})$ and $n(^{242\text{m}}\text{Am})/n(^{243}\text{Am})$ isotope amount ratios and their standard uncertainties from characterisation in IRMM-0243 .....	20
Table 6 Results of the compatibility evaluation .....	22
Table 7 Certified amount content of $^{243}\text{Am}$ and its uncertainty in IRMM-0243 .....	23
Table 8 Certified isotope amount ratio of $n(^{241}\text{Am})/n(^{243}\text{Am})$ and its uncertainty in IRMM-0243 .....	23
Table 9 Derived certified amount contents, mass fractions, isotope mass and amount fractions, molar mass and their uncertainties .....	24
Table 10 Atomic masses of Am isotopes used for calculation of the derived values .....	24
Table 11 Indicative values and their uncertainties in IRMM-0243 .....	25

## Annexes

Annex A Results of the impurity analysis of the americium original starting solution ....	35
Annex B Results of the homogeneity measurements for IRMM-0243 .....	36
Annex C Results of the stability measurements for the IRMM-0243.....	39
Annex D Uncertainty budget for the calculation of the amount content of $^{243}\text{Am}$ .....	40
Annex E Results of the characterisation for IRMM-0243.....	69
Annex F Results of the radiometric verification measurements for IRMM-0243 .....	71
Annex G Results for the $n(^{242\text{m}}\text{Am})/n(^{243}\text{Am})$ amount ratio from the $^{243}\text{Am}$ ILC .....	72

**Annex A** Results of the impurity analysis of the americium original starting solution

**Table A1** Mass concentration of elemental impurities in the americium solution after purification measured by ICP-AES (date of analysis: April 24, 2015). The mass concentrations of the major impurities before purification are shown in parentheses.

Element	Mass concentration [mg/L]	Element	Mass concentration [mg/L]
Ag	< 10.0	Na	< 10.0
Al	< 10.0 (560)	Nd	< 10.0
Ba	< 10.0	Ni	< 10.0 (995)
Be	< 10.0	P	< 10.0
Ca	< 10.0 (250)	Pb	< 10.0
Cd	< 10.0	Pr	< 10.0
Ce	< 10.0	S	< 10.0 (30)
Co	< 10.0 (1750)	Sb	< 10.0
Cr	< 10.0	Si	< 10.0
Cu	< 10.0	Sm	< 10.0
Eu	< 10.0	Sn	< 10.0
Fe	< 10.0 (3665)	Sr	< 10.0
Gd	< 10.0	Ti	< 10.0
K	< 10.0	U	< 10.0
La	< 10.0	W	< 10.0
Mg	< 10.0	Y	< 10.0
Mn	< 10.0	Zn	< 10.0
Mo	< 10.0 (61)	Zr	< 10.0

**Table A2** Activity concentration of americium and curium isotopes in the americium starting solution after purification measured by alpha spectrometry (date of analysis: April 28, 2015)


Isotope	Activity concentration [Bq/L]	Relative Uncertainty [ $k = 2$ ]
<sup>241</sup> Am	$1.3 \cdot 10^{10}$	$\pm 10\%$
<sup>243</sup> Am	$5.9 \cdot 10^9$	$\pm 10\%$
<sup>242</sup> Cm	$3.1 \cdot 10^7$	$\pm 20\%$
<sup>244</sup> Cm	$7.3 \cdot 10^6$	$\pm 30\%$

**Annex B** Results of the homogeneity measurements for IRMM-0243


**Table B1** Results for the amount content of  $^{243}\text{Am}$  as a function of the analytical sequence. Data are normalised to January 1, 2017.

Unit	Sequence	$^{243}\text{Am}$ amount content [nmol/g]	Unit	Sequence	$^{243}\text{Am}$ amount content [nmol/g]
165	1	5.6900	564	36	5.6995
465	2	5.7028	134	37	5.6999
165	3	5.6940	345	38	5.6991
465	4	5.6931	134	39	5.6995
165	5	5.6939	345	40	5.6991
465	6	5.6903	134	41	5.6959
165	7	5.6912	345	42	5.6980
465	8	5.6929	134	43	5.6978
83	9	5.6946	345	44	5.6977
83	10	5.6939	194	45	5.6953
83	11	5.6957	194	46	5.6975
83	12	5.6906	194	47	5.6953
373	13	5.6965	194	48	5.6959
517	14	5.6950	194	49	5.6973
373	15	5.6938	284	50	5.6958
517	16	5.6963	508	51	5.6934
373	17	5.6978	284	52	5.6970
517	18	5.6938	508	53	5.6945
373	19	5.6952	284	54	5.6964
517	20	5.6980	508	55	5.6943
11	21	5.6965	284	56	5.6941
309	22	5.6960	508	57	5.6955
11	23	5.6968	284	58	5.6958
309	24	5.6970	508	59	5.6950
11	25	5.6977	222	60	5.7000
309	26	5.7029	497	61	5.6978
11	27	5.6966	222	62	5.6973
309	28	5.6952	497	63	5.6995
54	29	5.6986	222	64	5.7004
564	30	5.6953	497	65	5.7002
54	31	5.6980	222	66	5.7003
564	32	5.6976	497	67	5.6970
54	33	5.6981	222	68	5.6967
564	34	5.7005	497	69	5.6921
54	35	5.6972			


**Figure B2** The amount content of  $^{243}\text{Am}$  from homogeneity study. Replicates are shown as a function of the analytical sequence.



**Figure B3** The amount content of  $^{243}\text{Am}$  after trend-correction from homogeneity study. Replicates are shown as a function of the analytical sequence.




**Figure B4** Mean amount contents of  $^{243}\text{Am}$  from homogeneity study as a function of the units (filling sequence). The unit means are plotted with 95 % CI of the means.



**Annex C** Results of the stability measurements for the IRMM-0243

**Figure C1** The amount content of  $^{243}\text{Am}$  from stability study (trend-corrected) as a function of the time.



**Annex D** Uncertainty budget for the calculation of the amount content of <sup>243</sup>Am.

Am-243 amount content with Am-241 spike by IDMS		
<p><b>Am-243 amount content with Am-241 spike by IDMS</b></p> <p>Author: Jakopic</p> <p>adapted simplified equation spike <sup>241</sup>Am, sample <sup>243</sup>Am, ratio <sup>243</sup>Am/<sup>241</sup>Am for IDMS equation</p> <p>IDMS blends and isotopic abundance (IA) aliquots from the same selected ampoule measured on the same sample turret</p> <p><sup>241</sup>Am spike amount content calculated from the ingrowth from <sup>241</sup>Pu purified solution</p> <p>The completeness of separation (removal) of <sup>241</sup>Am from <sup>241</sup>Pu in the spike solution was confirmed by high resolution gamma-ray spectrometry. The residual amount of <sup>241</sup>Am in the <sup>241</sup>Pu spike solution after purification (June 10, 2014) was found negligible.</p> <p>weighing certificate E3871 set A: ampoules IDMS/IA: 165, 465 set B: ampoules IDMS/IA: 240, 83 set C: ampoules IDMS/IA: 373, 517 set D: ampoules IDMS/IA: 11, 309 evaporation factors (due to evaporation of the <sup>241</sup>Pu purified solution): K<sub>1</sub>, K<sub>2</sub>, K<sub>3</sub> (ampoules 165, 465, 240, 83, 373, 517, 11 and 309).</p> <p>weighing certificate E3873 set E: ampoules IDMS/IA: 54, 564 set F: ampoules IDMS/IA: 134, 345 set G: ampoules IDMS/IA: 194, 437 set H: ampoules IDMS/IA: 284, 508 set I: ampoules IDMS/IA: 222, 497 evaporation factors: K<sub>1</sub>, K<sub>2</sub>, K<sub>3</sub>, K<sub>4</sub> (ampoules 54, 564, 134, 345, 194, 437). Evaporation factors: K<sub>1</sub>, K<sub>2</sub>, K<sub>3</sub>, K<sub>4</sub>, K<sub>5</sub> (ampoules 284, 508, 222, 497)</p> <p>Reference date for certification: 1 January 2017 (individual results from the respective IDMS sets A-I are normalised to 1 January 2017)</p> <p>Evaporation of the Pu-241 spike solution (under weight control) for the ingrowth of Am-241 is taken into account</p> <p>Due to technical reasons ampoules 240 and 437 were discarded.</p> <p>Additional uncertainty contribution was added for <sup>241</sup>Am/<sup>243</sup>Am ratio of the blend and unspiked aliquots in IDMS calculation (0.05%, k = 2 spanning 3 amu, according to ASTM standard) as there was no Am isotopic standard available to correct for mass fractionation during TIMS measurement</p> <p>It was assumed that the Am behaves similar to U and Pu during the TE TIMS measurement. Therefore, an additional uncertainty of 0.02 % (k = 2) for 3 mass units was added (from U and Pu QC charts) to account for the lack of an Am isotopic standard.</p> <p>The impurity analysis of the Am mother solution after purification by TRU (carried out by CEA) showed the presence of <sup>242</sup>Cm and <sup>244</sup>Cm isotopes. These two isotopes were not removed by purification on UTEVA and DGA prior to TIMS measurement. Therefore the ratio <sup>242m</sup>Am/<sup>243</sup>Am measured by TIMS includes isobaric <sup>242</sup>Cm. Alpha spectrometry performed on the aliquot of IRMM-0243 solutions confirmed this. The level of <sup>242</sup>Cm impurity in the Am fraction was about 0.2 % (ICP-AES by CEA and alpha spectrometry at JRC Geel gave the same value) and was added as an additional uncertainty component to the overall uncertainty of the <sup>242m</sup>Am/<sup>243</sup>Am ratio.</p> <p>Uncertainty on the time difference is set to 12 h!</p> <p>uncertainty of the time for <sup>241</sup>Am ingrowth = 0.65 days (90% *12 h + 10%*48 h (average TIME between UTEVA 1 and UTEVA 2). It is assumed 90 % of Pu and Am is separated with UTEVA 1 and 10 % with UTEVA 2. This is a conservative estimation, in fact according to the literature more than 99 % separation is achieved with UTEVA 1!</p> <p>Uncertainty from the homogeneity assessment (softCRM, ANOVA) was added to the <sup>243</sup>Am amount content.</p> <p>The mean values are calculated as the average of 9 ampoules: 165, 83, 373, 11, 564, 345, 194, 284, 497 (from each set A-I one unit)</p> <p><b>Model Equation:</b></p> <p>{-----simplified IDMS equation-----}</p> <p><math>f_{c241Am}(m_x, m_y, R_D)</math></p>		
Date: 10/13/2017	File: Am-243 amount content with Am-241 spike by IDMS_edited.SMU	Page 1 of 29

Am-243 amount content with Am-241 spike by IDMS

$C_{241Am}, R_y, R_x = C_{241Am} * m_y / m_x * (R_y - R_b * K_{ASTM} * K_{QC}) / (R_b * K_{ASTM} * K_{QC} - R_x * K_{ASTM} * K_{QC}) * (R_x * K_{ASTM} * K_{QC});$   
 {-----set A-----}  
 $C_{243Am165} = f_{C_{241Am}}(m_{x1}, m_{y1}, R_{b1}, C_{241Am1}, R_{y1}, R_{x1}) * e^{(-\lambda_{243Am} \cdot \Delta T_A)} * 1 / (K_1 * K_2 * K_3);$   
 $C_{243Am456} = f_{C_{241Am}}(m_{x2}, m_{y2}, R_{b2}, C_{241Am1}, R_{y2}, R_{x2}) * e^{(-\lambda_{243Am} \cdot \Delta T_A)} * 1 / (K_1 * K_2 * K_3);$   
 $C_{241Am1} = \lambda_{241Pu} / (\lambda_{241Am} - \lambda_{241Pu}) * C_{241Pu} * (e^{(-\lambda_{241Pu} \cdot \Delta t_A)} - e^{(-\lambda_{241Am} \cdot \Delta t_A)}) * e^{(-\lambda_{241Am} \cdot \Delta T_{ATIMS});}$   
 {-----set B-----}  
 $C_{243Am83} = f_{C_{241Am}}(m_{x4}, m_{y4}, R_{b4}, C_{241Am2}, R_{y4}, R_{x4}) * e^{(-\lambda_{243Am} \cdot \Delta T_B)} * 1 / (K_1 * K_2 * K_3);$   
 $C_{241Am2} = \lambda_{241Pu} / (\lambda_{241Am} - \lambda_{241Pu}) * C_{241Pu} * (e^{(-\lambda_{241Pu} \cdot \Delta t_B)} - e^{(-\lambda_{241Am} \cdot \Delta t_B)}) * e^{(-\lambda_{241Am} \cdot \Delta T_{BTIMS});}$   
 {-----set C-----}  
 $C_{243Am373} = f_{C_{241Am}}(m_{x5}, m_{y5}, R_{b5}, C_{241Am3}, R_{y5}, R_{x5}) * e^{(-\lambda_{243Am} \cdot \Delta T_C)} * 1 / (K_1 * K_2 * K_3);$   
 $C_{243Am517} = f_{C_{241Am}}(m_{x6}, m_{y6}, R_{b6}, C_{241Am3}, R_{y6}, R_{x6}) * e^{(-\lambda_{243Am} \cdot \Delta T_C)} * 1 / (K_1 * K_2 * K_3);$   
 $C_{241Am3} = \lambda_{241Pu} / (\lambda_{241Am} - \lambda_{241Pu}) * C_{241Pu} * (e^{(-\lambda_{241Pu} \cdot \Delta t_C)} - e^{(-\lambda_{241Am} \cdot \Delta t_C)}) * e^{(-\lambda_{241Am} \cdot \Delta T_{CTIMS});}$   
 {-----set D-----}  
 $C_{243Am11} = f_{C_{241Am}}(m_{x7}, m_{y7}, R_{b7}, C_{241Am4}, R_{y7}, R_{x7}) * e^{(-\lambda_{243Am} \cdot \Delta T_D)} * 1 / (K_1 * K_2 * K_3);$   
 $C_{243Am309} = f_{C_{241Am}}(m_{x8}, m_{y8}, R_{b8}, C_{241Am4}, R_{y8}, R_{x8}) * e^{(-\lambda_{243Am} \cdot \Delta T_D)} * 1 / (K_1 * K_2 * K_3);$   
 $C_{241Am4} = \lambda_{241Pu} / (\lambda_{241Am} - \lambda_{241Pu}) * C_{241Pu} * (e^{(-\lambda_{241Pu} \cdot \Delta t_D)} - e^{(-\lambda_{241Am} \cdot \Delta t_D)}) * e^{(-\lambda_{241Am} \cdot \Delta T_{DTIMS});}$   
 {-----set E-----}  
 $C_{243Am54} = f_{C_{241Am}}(m_{x9}, m_{y9}, R_{b9}, C_{241Am5}, R_{y9}, R_{x9}) * e^{(-\lambda_{243Am} \cdot \Delta T_E)} * 1 / (K_1 * K_2 * K_3 * K_4);$   
 $C_{243Am564} = f_{C_{241Am}}(m_{x10}, m_{y10}, R_{b10}, C_{241Am5}, R_{y10}, R_{x10}) * e^{(-\lambda_{243Am} \cdot \Delta T_E)} * 1 / (K_1 * K_2 * K_3 * K_4);$   
 $C_{241Am5} = (\lambda_{241Pu} / (\lambda_{241Am} - \lambda_{241Pu})) * C_{241Pu} * (e^{(-\lambda_{241Pu} \cdot \Delta t_E)} - e^{(-\lambda_{241Am} \cdot \Delta t_E)}) * e^{(-\lambda_{241Am} \cdot \Delta T_{ETIMS});}$   
 {-----set F-----}  
 $C_{243Am134} = f_{C_{241Am}}(m_{x11}, m_{y11}, R_{b11}, C_{241Am6}, R_{y11}, R_{x11}) * e^{(-\lambda_{243Am} \cdot \Delta T_F)} * 1 / (K_1 * K_2 * K_3 * K_4);$   
 $C_{243Am345} = f_{C_{241Am}}(m_{x12}, m_{y12}, R_{b12}, C_{241Am6}, R_{y12}, R_{x12}) * e^{(-\lambda_{243Am} \cdot \Delta T_F)} * 1 / (K_1 * K_2 * K_3 * K_4);$   
 $C_{241Am6} = \lambda_{241Pu} / (\lambda_{241Am} - \lambda_{241Pu}) * C_{241Pu} * (e^{(-\lambda_{241Pu} \cdot \Delta t_F)} - e^{(-\lambda_{241Am} \cdot \Delta t_F)}) * e^{(-\lambda_{241Am} \cdot \Delta T_{FTIMS});}$   
 {-----set G-----}  
 $C_{243Am194} = f_{C_{241Am}}(m_{x13}, m_{y13}, R_{b13}, C_{241Am7}, R_{y13}, R_{x13}) * e^{(-\lambda_{243Am} \cdot \Delta T_G)} * 1 / (K_1 * K_2 * K_3 * K_4);$   
 $C_{241Am7} = \lambda_{241Pu} / (\lambda_{241Am} - \lambda_{241Pu}) * C_{241Pu} * (e^{(-\lambda_{241Pu} \cdot \Delta t_G)} - e^{(-\lambda_{241Am} \cdot \Delta t_G)}) * e^{(-\lambda_{241Am} \cdot \Delta T_{GTIMS});}$   
 {-----set H-----}  
 $C_{243Am284} = f_{C_{241Am}}(m_{x15}, m_{y15}, R_{b15}, C_{241Am8}, R_{y15}, R_{x15}) * e^{(-\lambda_{243Am} \cdot \Delta T_H)} * 1 / (K_1 * K_2 * K_3 * K_4 * K_5);$   
 $C_{243Am508} = f_{C_{241Am}}(m_{x16}, m_{y16}, R_{b16}, C_{241Am8}, R_{y16}, R_{x16}) * e^{(-\lambda_{243Am} \cdot \Delta T_H)} * 1 / (K_1 * K_2 * K_3 * K_4 * K_5);$   
 $C_{241Am8} = \lambda_{241Pu} / (\lambda_{241Am} - \lambda_{241Pu}) * C_{241Pu} * (e^{(-\lambda_{241Pu} \cdot \Delta t_H)} - e^{(-\lambda_{241Am} \cdot \Delta t_H)}) * e^{(-\lambda_{241Am} \cdot \Delta T_{HTIMS});}$   
 {-----set I-----}  
 $C_{243Am222} = f_{C_{241Am}}(m_{x17}, m_{y17}, R_{b17}, C_{241Am9}, R_{y17}, R_{x17}) * e^{(-\lambda_{243Am} \cdot \Delta T_I)} * 1 / (K_1 * K_2 * K_3 * K_4 * K_5);$   
 $C_{243Am497} = f_{C_{241Am}}(m_{x18}, m_{y18}, R_{b18}, C_{241Am9}, R_{y18}, R_{x18}) * e^{(-\lambda_{243Am} \cdot \Delta T_I)} * 1 / (K_1 * K_2 * K_3 * K_4 * K_5);$   
 $C_{241Am9} = \lambda_{241Pu} / (\lambda_{241Am} - \lambda_{241Pu}) * C_{241Pu} * (e^{(-\lambda_{241Pu} \cdot \Delta t_I)} - e^{(-\lambda_{241Am} \cdot \Delta t_I)}) * e^{(-\lambda_{241Am} \cdot \Delta T_{ITIMS});}$   
 {-----average amount content calculation-----}

Am-243 amount content with Am-241 spike by IDMS		
$C_{243Amchar} = (C_{243Am165} + C_{243Am83} + C_{243Am373} + C_{243Am111} + C_{243Am564} + C_{243Am345} + C_{243Am194} + C_{243Am284} + C_{243Am497}) / 9;$ <p>{-----certified (assigned) amount content-----}</p> $C_{243Amcert} = C_{243Amchar} + \delta C_{243Ambd} + \delta C_{243Amits} + \delta C_{243Amsts};$ <p>{-----consistency check-----}</p> $E_1 = C_{243Amchar} - C_{243Am165};$ $E_2 = C_{243Amchar} - C_{243Am456};$ $E_4 = C_{243Amchar} - C_{243Am83};$ $E_5 = C_{243Amchar} - C_{243Am373};$ $E_6 = C_{243Amchar} - C_{243Am517};$ $E_7 = C_{243Amchar} - C_{243Am111};$ $E_8 = C_{243Amchar} - C_{243Am309};$ $E_9 = C_{243Amchar} - C_{243Am54};$ $E_{10} = C_{243Amchar} - C_{243Am564};$ $E_{11} = C_{243Amchar} - C_{243Am134};$ $E_{12} = C_{243Amchar} - C_{243Am345};$ $E_{13} = C_{243Amchar} - C_{243Am194};$ $E_{15} = C_{243Amchar} - C_{243Am284};$ $E_{16} = C_{243Amchar} - C_{243Am508};$ $E_{17} = C_{243Amchar} - C_{243Am222};$ $E_{18} = C_{243Amchar} - C_{243Am497};$ $\ln_2 = \ln(2);$ $\lambda_{243Am} = \ln_2 / \tau_{243Am};$ $\lambda_{241Am} = \ln_2 / \tau_{241Am};$ $\lambda_{241Pu} = \ln_2 / \tau_{241Pu};$ <p>{calculated Am, 242mAm and 241Am amount contents using the certified 243Am amount content and the isotope amount fractions}</p> $C_{Amcert} = C_{243Amcert} / f_{dnorm_{243Amcert}};$ $C_{241Amcert} = C_{Amcert} * f_{dnorm_{241Amcert}};$ $C_{242mAmILC} = C_{Amcert} * f_{dnorm_{242mAmILC}};$ <p>{calculated Am, 242mAm and 241Am mass fractions using certified the 243Am amount content, the isotope mass fractions and the molar masses}</p> $\gamma_{243Amcert} = C_{243Amcert} * M_{243Am};$ $\gamma_{241Amcert} = C_{241Amcert} * M_{241Am};$ $\gamma_{242mAmILC} = C_{242mAmILC} * M_{242mAm};$ $\gamma_{Amcert} = C_{Amcert} * M_{dnorm_{Amcert}};$ <p>{calculated 243Am and 241Am mass fractions from high resolution gamma-ray and alpha counting measurements}</p>		
Date: 10/13/2017	File: Am-243 amount content with Am-241 spike by IDMS_edited.SMU	Page 3 of 29

Am-243 amount content with Am-241 spike by IDMS

$$\begin{aligned} \gamma_{243Am\gamma1} &= M_{243Am} * A_{243Am\gamma1} * (365.25 * 24 * 3600) / (\lambda_{243Am} * N_a); \\ \gamma_{241Am\gamma1} &= M_{241Am} * A_{241Am\gamma1} * (365.25 * 24 * 3600) / (\lambda_{241Am} * N_a); \\ \gamma_{243Am\gamma2} &= M_{243Am} * A_{243Am\gamma2} * (365.25 * 24 * 3600) / (\lambda_{243Am} * N_a); \\ \gamma_{241Am\gamma2} &= M_{241Am} * A_{241Am\gamma2} * (365.25 * 24 * 3600) / (\lambda_{241Am} * N_a); \\ \gamma_{243Am\alpha} &= M_{243Am} * A_{243Am\alpha} * (365.25 * 24 * 3600) / (\lambda_{243Am} * N_a); \\ \gamma_{241Am\alpha} &= M_{241Am} * A_{241Am\alpha} * (365.25 * 24 * 3600) / (\lambda_{241Am} * N_a); \end{aligned}$$

**List of Quantities:**

Quantity	Unit	Definition
C <sub>243Amcert</sub>	mol/g	certified <sup>243</sup> Am amount content
C <sub>241Amcert</sub>	mol/g	certified <sup>241</sup> Am amount content
C <sub>242mAmILC</sub>	mol/g	indicative <sup>242m</sup> Am amount content
C <sub>Amcert</sub>	mol/g	certified Am amount content
γ <sub>243Amcert</sub>	g/g	certified <sup>243</sup> Am mass fraction
γ <sub>241Amcert</sub>	g/g	certified <sup>241</sup> Am mass fraction
γ <sub>242mAmILC</sub>	g/g	indicative <sup>242m</sup> Am mass fraction (ILC)
γ <sub>Amcert</sub>	g/g	certified Am mass fraction
Mdnorm <sub>Amcert</sub>	mol/g	certified molar mass of Am
C <sub>243Amchar</sub>	mol/g	mean <sup>243</sup> Am amount content from characterisation (9 values)
C <sub>243Am165</sub>	mol/g	<sup>243</sup> Am amount content in the sample 165
C <sub>243Am456</sub>	mol/g	<sup>243</sup> Am amount content in the sample 465
C <sub>243Am83</sub>	mol/g	<sup>243</sup> Am amount content in the sample 83
C <sub>243Am373</sub>	mol/g	<sup>243</sup> Am amount content in the sample 373
C <sub>243Am517</sub>	mol/g	<sup>243</sup> Am amount content in the sample 517
C <sub>243Am11</sub>	mol/g	<sup>243</sup> Am amount content in the sample 11
C <sub>243Am309</sub>	mol/g	<sup>243</sup> Am amount content in the sample 309
C <sub>243Am54</sub>	mol/g	<sup>243</sup> Am amount content in the sample 54
C <sub>243Am564</sub>	mol/g	<sup>243</sup> Am amount content in the sample 564
C <sub>243Am134</sub>	mol/g	<sup>243</sup> Am amount content in the sample 134
C <sub>243Am345</sub>	mol/g	<sup>243</sup> Am amount content in the sample 345
C <sub>243Am194</sub>	mol/g	<sup>243</sup> Am amount content in the sample 194
C <sub>243Am284</sub>	mol/g	<sup>243</sup> Am amount content in the sample 284
C <sub>243Am508</sub>	mol/g	<sup>243</sup> Am amount content in the sample 508
C <sub>243Am222</sub>	mol/g	<sup>243</sup> Am amount content in the sample 222
C <sub>243Am497</sub>	mol/g	<sup>243</sup> Am amount content in the sample 497
γ <sub>243Amgamma1</sub>	g/g	<sup>243</sup> Am mass fraction from high resolution gamma-ray spectrometry (above ground)
γ <sub>241Amgamma1</sub>	g/g	<sup>241</sup> Am mass fraction from high resolution gamma-ray spectrometry (above ground)

Date: 10/13/2017

File: Am-243 amount content with Am-241 spike by IDMS\_edited.SMU

Page 4 of 29

## Am-243 amount content with Am-241 spike by IDMS

Quantity	Unit	Definition
$\gamma_{243Am\gamma 2}$	g/g	$^{243}\text{Am}$ mass fraction from high resolution gamma -ray spectrometry (HADES laboratory)
$\gamma_{241Am\gamma 2}$	g/g	$^{241}\text{Am}$ mass fraction from high resolution gamma-ray spectrometry (HADES laboratory)
$\gamma_{241Am\alpha}$	g/g	$^{243}\text{Am}$ mass fraction from alpha counting
$\gamma_{243Am\alpha}$	g/g	$^{241}\text{Am}$ mass fraction from alpha counting
$m_{x1}$	g	mass of sample in blend 165
$m_{x2}$	g	mass of sample in blend 465
$m_{x4}$	g	mass of sample in blend 83
$m_{x5}$	g	mass of sample in blend 373
$m_{x6}$	g	mass of sample in blend 517
$m_{x7}$	g	mass of sample in blend 11
$m_{x8}$	g	mass of sample in blend 309
$m_{x9}$	g	mass of sample in blend 54
$m_{x10}$	g	mass of sample in blend 564
$m_{x11}$	g	mass of sample in blend 134
$m_{x12}$	g	mass of sample in blend 345
$m_{x13}$	g	mass of sample in blend 194
$m_{x15}$	g	mass of sample in blend 284
$m_{x16}$	g	mass of sample in blend 508
$m_{x17}$	g	mass of sample in blend 222
$m_{x18}$	g	mass of sample in blend 497
$m_{y1}$	g	mass of spike in blend 165
$m_{y2}$	g	mass of spike in blend 465
$m_{y4}$	g	mass of spike in blend 83
$m_{y5}$	g	mass of spike in blend 373
$m_{y6}$	g	mass of spike in blend 517
$m_{y7}$	g	mass of spike in blend 11
$m_{y8}$	g	mass of spike in blend 309
$m_{y9}$	g	mass of spike in blend 54
$m_{y10}$	g	mass of spike in blend 564
$m_{y11}$	g	mass of spike in blend 134
$m_{y12}$	g	mass of spike in blend 345
$m_{y13}$	g	mass of spike in blend 194
$m_{y15}$	g	mass of spike in blend 284
$m_{y16}$	g	mass of spike in blend 508
$m_{y17}$	g	mass of spike in blend 222
$m_{y18}$	g	mass of spike in blend 497
$R_{b1}$	mol/mol	measured ratio ( $^{243}\text{Am}/^{241}\text{Am}$ ) of blend 165

Date: 10/13/2017

File: Am-243 amount content with Am-241 spike by IDMS\_edited.SMU

Page 5 of 29

## Am-243 amount content with Am-241 spike by IDMS

Quantity	Unit	Definition
R <sub>b2</sub>	mol/mol	measured ratio ( <sup>243</sup> Am/ <sup>241</sup> Am) of blend 465
R <sub>b4</sub>	mol/mol	measured ratio ( <sup>243</sup> Am/ <sup>241</sup> Am) of blend 83
R <sub>b5</sub>	mol/mol	measured ratio ( <sup>243</sup> Am/ <sup>241</sup> Am) of blend 373
R <sub>b6</sub>	mol/mol	measured ratio ( <sup>243</sup> Am/ <sup>241</sup> Am) of blend 517
R <sub>b7</sub>	mol/mol	measured ratio ( <sup>243</sup> Am/ <sup>241</sup> Am) of blend 11
R <sub>b8</sub>	mol/mol	measured ratio ( <sup>243</sup> Am/ <sup>241</sup> Am) of blend 309
R <sub>b9</sub>	mol/mol	measured ratio ( <sup>243</sup> Am/ <sup>241</sup> Am) of blend 54
R <sub>b10</sub>	mol/mol	measured ratio ( <sup>243</sup> Am/ <sup>241</sup> Am) of blend 564
R <sub>b11</sub>	mol/mol	measured ratio ( <sup>243</sup> Am/ <sup>241</sup> Am) of blend 134
R <sub>b12</sub>	mol/mol	measured ratio ( <sup>243</sup> Am/ <sup>241</sup> Am) of blend 345
R <sub>b13</sub>	mol/mol	measured ratio ( <sup>243</sup> Am/ <sup>241</sup> Am) of blend 194
R <sub>b15</sub>	mol/mol	measured ratio ( <sup>243</sup> Am/ <sup>241</sup> Am) of blend 284
R <sub>b16</sub>	mol/mol	measured ratio ( <sup>243</sup> Am/ <sup>241</sup> Am) of blend 508
R <sub>b17</sub>	mol/mol	measured ratio ( <sup>243</sup> Am/ <sup>241</sup> Am) of blend 222
R <sub>b18</sub>	mol/mol	measured ratio ( <sup>243</sup> Am/ <sup>241</sup> Am) of blend 497
R <sub>y1</sub>	mol/mol	<sup>243</sup> Am/ <sup>241</sup> Am amount ratio in spike for set A
R <sub>y2</sub>	mol/mol	<sup>243</sup> Am/ <sup>241</sup> Am amount ratio in spike for set A
R <sub>y4</sub>	mol/mol	<sup>243</sup> Am/ <sup>241</sup> Am amount ratio in spike for set B
R <sub>y5</sub>	mol/mol	<sup>243</sup> Am/ <sup>241</sup> Am amount ratio in spike for set C
R <sub>y6</sub>	mol/mol	<sup>243</sup> Am/ <sup>241</sup> Am amount ratio in spike for set C
R <sub>y7</sub>	mol/mol	<sup>243</sup> Am/ <sup>241</sup> Am amount ratio in spike for set D
R <sub>y8</sub>	mol/mol	<sup>243</sup> Am/ <sup>241</sup> Am amount ratio in spike for set D
R <sub>y9</sub>	mol/mol	<sup>243</sup> Am/ <sup>241</sup> Am amount ratio in spike for set E
R <sub>y10</sub>	mol/mol	<sup>243</sup> Am/ <sup>241</sup> Am amount ratio in spike for set E
R <sub>y11</sub>	mol/mol	<sup>243</sup> Am/ <sup>241</sup> Am amount ratio in spike for set F
R <sub>y12</sub>	mol/mol	<sup>243</sup> Am/ <sup>241</sup> Am amount ratio in spike for set F
R <sub>y13</sub>	mol/mol	<sup>243</sup> Am/ <sup>241</sup> Am amount ratio in spike for set G
R <sub>y15</sub>	mol/mol	<sup>243</sup> Am/ <sup>241</sup> Am amount ratio in spike for set H
R <sub>y16</sub>	mol/mol	<sup>243</sup> Am/ <sup>241</sup> Am amount ratio in spike for set H
R <sub>y17</sub>	mol/mol	<sup>243</sup> Am/ <sup>241</sup> Am amount ratio in spike for set I
R <sub>y18</sub>	mol/mol	<sup>243</sup> Am/ <sup>241</sup> Am amount ratio in spike for set I
R <sub>x1</sub>	mol/mol	measured <sup>243</sup> Am/ <sup>241</sup> Am amount ratio in sample 165
R <sub>x2</sub>	mol/mol	measured <sup>243</sup> Am/ <sup>241</sup> Am amount ratio in sample 465
R <sub>x4</sub>	mol/mol	measured <sup>243</sup> Am/ <sup>241</sup> Am amount ratio in sample 83
R <sub>x5</sub>	mol/mol	measured <sup>243</sup> Am/ <sup>241</sup> Am amount ratio in sample 373
R <sub>x6</sub>	mol/mol	measured <sup>243</sup> Am/ <sup>241</sup> Am amount ratio in sample 517
R <sub>x7</sub>	mol/mol	measured <sup>243</sup> Am/ <sup>241</sup> Am amount ratio in sample 11
R <sub>x8</sub>	mol/mol	measured <sup>243</sup> Am/ <sup>241</sup> Am amount ratio in sample 309

Date: 10/13/2017

File: Am-243 amount content with Am-241 spike by IDMS\_edited.SMU

Page 6 of 29

Generated with GUM Workbench Pro Version 2.4.1.458

## Am-243 amount content with Am-241 spike by IDMS

Quantity	Unit	Definition
R <sub>x9</sub>	mol/mol	measured <sup>243</sup> Am/ <sup>241</sup> Am amount ratio in sample 54
R <sub>x10</sub>	mol/mol	measured <sup>243</sup> Am/ <sup>241</sup> Am amount ratio in sample 564
R <sub>x11</sub>	mol/mol	measured <sup>243</sup> Am/ <sup>241</sup> Am amount ratio in sample 134
R <sub>x12</sub>	mol/mol	measured <sup>243</sup> Am/ <sup>241</sup> Am amount ratio in sample 345
R <sub>x13</sub>	mol/mol	measured <sup>243</sup> Am/ <sup>241</sup> Am amount ratio in sample 194
R <sub>x15</sub>	mol/mol	measured <sup>243</sup> Am/ <sup>241</sup> Am amount ratio in sample 284
R <sub>x16</sub>	mol/mol	measured <sup>243</sup> Am/ <sup>241</sup> Am amount ratio in sample 508
R <sub>x17</sub>	mol/mol	measured <sup>243</sup> Am/ <sup>241</sup> Am amount ratio in sample 222
R <sub>x18</sub>	mol/mol	measured <sup>243</sup> Am/ <sup>241</sup> Am amount ratio in sample 497
e		
Δt <sub>A</sub>	a	time for ingrowth of <sup>241</sup> Am for set A
Δt <sub>B</sub>	a	time for ingrowth of <sup>241</sup> Am for set B
Δt <sub>C</sub>	a	time for ingrowth of <sup>241</sup> Am for set C
Δt <sub>D</sub>	a	time for ingrowth of <sup>241</sup> Am for set D
Δt <sub>E</sub>	a	time for ingrowth of <sup>241</sup> Am for set E
Δt <sub>F</sub>	a	time for ingrowth of <sup>241</sup> Am for set F
Δt <sub>G</sub>	a	time for ingrowth of <sup>241</sup> Am for set G
Δt <sub>H</sub>	a	time for ingrowth of <sup>241</sup> Am for set H
Δt <sub>I</sub>	a	time for ingrowth of <sup>241</sup> Am for set I
ε <sub>1</sub>	mol/g	consistency check
ε <sub>2</sub>	mol/g	consistency check
ε <sub>4</sub>	mol/g	consistency check
ε <sub>5</sub>	mol/g	consistency check
ε <sub>6</sub>	mol/g	consistency check
ε <sub>7</sub>	mol/g	consistency check
ε <sub>8</sub>	mol/g	consistency check
ε <sub>9</sub>	mol/g	consistency check
ε <sub>10</sub>	mol/g	consistency check
ε <sub>11</sub>	mol/g	consistency check
ε <sub>12</sub>	mol/g	consistency check
ε <sub>13</sub>	mol/g	consistency check
ε <sub>15</sub>	mol/g	consistency check
ε <sub>16</sub>	mol/g	consistency check
ε <sub>17</sub>	mol/g	consistency check
ε <sub>18</sub>	mol/g	consistency check
ln <sub>2</sub>		
K <sub>1</sub>		evaporation factor
K <sub>2</sub>		evaporation factor

Date: 10/13/2017

File: Am-243 amount content with Am-241 spike by IDMS\_edited.SMU

Page 7 of 29

Generated with GUM Workbench Pro Version 2.4.1.458

Am-243 amount content with Am-241 spike by IDMS		
Quantity	Unit	Definition
$K_3$		evaporation factor
$K_4$		evaporation factor
$K_5$		evaporation factor
$C_{241Pu}$	mol/g	$^{241}Pu$ amount content in the purified $^{241}Pu$ solution on June 10, 2014 (time zero)
$\Delta T_A$	a	time difference between IDMS set A and reference date (1 Jan 2017)
$\Delta T_B$	a	time difference between IDMS set B and reference date (1 Jan 2017)
$\Delta T_C$	a	time difference between IDMS set C and reference date (1 Jan 2017)
$\Delta T_D$	a	time difference between IDMS set D and reference date (1 Jan 2017)
$\Delta T_E$	a	time difference between IDMS set E and reference date (1 Jan 2017)
$\Delta T_F$	a	time difference between IDMS set F and reference date (1 Jan 2017)
$\Delta T_G$	a	time difference between IDMS set G and reference date (1 Jan 2017)
$\Delta T_H$	a	time difference between IDMS set H and reference date (1 Jan 2017)
$\Delta T_I$	a	time difference between IDMS set I and reference date (1 Jan 2017)
$C_{241Am1}$	mol/g	$^{241}Am$ spike ingrown amount content for set A
$C_{241Am2}$	mol/g	$^{241}Am$ spike ingrown amount content for set B
$C_{241Am3}$	mol/g	$^{241}Am$ spike ingrown amount content for set C
$C_{241Am4}$	mol/g	$^{241}Am$ spike ingrown amount content for set D
$C_{241Am5}$	mol/g	$^{241}Am$ spike ingrown amount content for set E
$C_{241Am6}$	mol/g	$^{241}Am$ spike ingrown amount content for set F
$C_{241Am7}$	mol/g	$^{241}Am$ spike ingrown amount content for set G
$C_{241Am8}$	mol/g	$^{241}Am$ spike ingrown amount content for set H
$C_{241Am9}$	mol/g	$^{241}Am$ spike ingrown amount content for set I
$\Delta T_{ATIMS}$	a	time difference between $^{241}Am$ ingrowth and IDMS measurement for set A
$\Delta T_{BTIMS}$	a	time difference between $^{241}Am$ ingrowth and IDMS measurement for set B
$\Delta T_{CTIMS}$	a	time difference between $^{241}Am$ ingrowth and IDMS measurement for set C
$\Delta T_{DTIMS}$	a	time difference between $^{241}Am$ ingrowth and IDMS measurement for set D
$\Delta T_{ETIMS}$	a	time difference between $^{241}Am$ ingrowth and IDMS measurement for set E
$\Delta T_{FTIMS}$	a	time difference between $^{241}Am$ ingrowth and IDMS measurement for set F
$\Delta T_{GTIMS}$	a	time difference between $^{241}Am$ ingrowth and IDMS measurement for set G
$\Delta T_{HTIMS}$	a	time difference between $^{241}Am$ ingrowth and IDMS measurement for set H
$\Delta T_{ITIMS}$	a	time difference between $^{241}Am$ ingrowth and IDMS measurement for set I
$\tau_{241Am}$	a	half life of $^{241}Am$
Date: 10/13/2017	File: Am-243 amount content with Am-241 spike by IDMS_edited.SMU	Page 8 of 29

Generated with GUM Workbench Pro Version 2.4.1.458

Am-243 amount content with Am-241 spike by IDMS		
Quantity	Unit	Definition
$\tau_{243Am}$	a	half life of $^{243}Am$
$\tau_{241Pu}$	a	half life of $^{241}Pu$
$\lambda_{241Pu}$	$a^{-1}$	decay constant $^{241}Pu$
$\lambda_{241Am}$	$a^{-1}$	decay constant $^{241}Am$
$\lambda_{243Am}$	$a^{-1}$	decay constant $^{243}Am$
$fdnorm_{243Amcert}$	mol/mol	certified $^{243}Am$ amount fraction
$fdnorm_{241Amcert}$	mol/mol	certified $^{241}Am$ amount fraction
$fdnorm_{242mAmILC}$	mol/mol	indicative $^{242m}Am$ amount fraction (ILC)
$M_{243Am}$	mol/g	molar mass of $^{243}Am$
$M_{241Am}$	mol/g	molar mass of $^{241}Am$
$M_{242mAm}$	mol/g	molar mass of $^{242m}Am$
$N_a$	$mol^{-1}$	avogadro constant
$A_{243Am\gamma 1}$	Bq/Bq	activity measurement $^{243}Am$ by gamma spectrometry (above ground)
$A_{241Am\gamma 1}$	Bq/Bq	activity measurement $^{241}Am$ by gamma spectrometry (above ground)
$A_{241Am\gamma 2}$	Bq/Bq	activity measurement $^{241}Am$ by gamma spectrometry (HADES)
$A_{243Am\gamma 2}$	Bq/Bq	activity measurement $^{243}Am$ by gamma spectrometry (HADES)
$A_{243Am\alpha}$	Bq/Bq	activity measurement $^{243}Am$ by alpha spectrometry
$A_{241Am\alpha}$	Bq/Bq	activity measurement $^{241}Am$ by alpha spectrometry
$K_{ASTM}$	mol/mol	additional uncertainty component for $^{243}Am/^{241}Am$ ratio measurement according to ASTM standard (no standard available for TE mass bias correction)
$\delta C_{243Am\text{its}}$	mol/g	error term due to the long-term instability from stability study
$\delta C_{243Am\text{sts}}$	mol/g	error term due to the short-term instability from stability study
$\delta C_{243Am\text{bb}}$	mol/g	error term due to the between-unit variation from homogeneity assessment
$K_{QC}$	mol/mol	additional uncertainty component for $^{243}Am/^{241}Am$ ratio measurement from U and Pu QC charts (no standard available for TE mass bias correction)
<p><b>Mdnorm<sub>Amcert</sub>:</b> Import  Filename: ..\Am amount ratios\Am isotope amount ratios by TIMS.smu  Symbol: Mdnorm<sub>Amcert</sub></p> <p><b>m<sub>x1</sub>:</b> Type B normal distribution  Value: 2.54434 g  Expanded Uncertainty: 0.00010 g  Coverage Factor: 2</p> <p>E3871 weighing certificate</p> <p><b>m<sub>x2</sub>:</b> Type B normal distribution  Value: 2.67332 g  Expanded Uncertainty: 0.00029 g  Coverage Factor: 2</p> <p>E3871 weighing certificate</p>		
Date: 10/13/2017	File: Am-243 amount content with Am-241 spike by IDMS_edited.SMU	Page 9 of 29

Am-243 amount content with Am-241 spike by IDMS		
$m_{x4}$ :	Type B normal distribution Value: 2.66642 g Expanded Uncertainty: 0.00011 g Coverage Factor: 2	
E3871 weighing certificate		
$m_{x5}$ :	Type B normal distribution Value: 2.70530 g Expanded Uncertainty: 0.00010 g Coverage Factor: 2	
E3871 weighing certificate		
$m_{x6}$ :	Type B normal distribution Value: 2.68092 g Expanded Uncertainty: 0.00011 g Coverage Factor: 2	
E3871 weighing certificate		
$m_{x7}$ :	Type B normal distribution Value: 2.68865 g Expanded Uncertainty: 0.00011 g Coverage Factor: 2	
E3871 weighing certificate		
$m_{x8}$ :	Type B normal distribution Value: 2.70294 g Expanded Uncertainty: 0.00009 g Coverage Factor: 2	
E3871 weighing certificate		
$m_{x9}$ :	Type B normal distribution Value: 2.69368 g Expanded Uncertainty: 0.00009 g Coverage Factor: 2	
E3873 weighing certificate		
$m_{x10}$ :	Type B normal distribution Value: 2.70174 g Expanded Uncertainty: 0.00011 g Coverage Factor: 2	
E3873 weighing certificate		
$m_{x11}$ :	Type B normal distribution Value: 2.69410 g Expanded Uncertainty: 0.00011 g Coverage Factor: 2	
E3873 weighing certificate		
$m_{x12}$ :	Type B normal distribution Value: 2.70371 g Expanded Uncertainty: 0.00012 g Coverage Factor: 2	
E3873 weighing certificate		
Date: 10/13/2017	File: Am-243 amount content with Am-241 spike by IDMS_edited.SMU	Page 10 of 29

Am-243 amount content with Am-241 spike by IDMS		
$m_{x13}$ :	Type B normal distribution Value: 2.68537 g Expanded Uncertainty: 0.00010 g Coverage Factor: 2  E3873 weighing certificate	
$m_{x15}$ :	Type B normal distribution Value: 2.68993 g Expanded Uncertainty: 0.00009 g Coverage Factor: 2  E3873 weighing certificate	
$m_{x16}$ :	Type B normal distribution Value: 2.72778 g Expanded Uncertainty: 0.00009 g Coverage Factor: 2  E3873 weighing certificate	
$m_{x17}$ :	Type B normal distribution Value: 2.75064 g Expanded Uncertainty: 0.00012 g Coverage Factor: 2  E3873 weighing certificate	
$m_{x18}$ :	Type B normal distribution Value: 2.99959 g Expanded Uncertainty: 0.00009 g Coverage Factor: 2  E3873 weighing certificate	
$m_{y1}$ :	Type B normal distribution Value: 0.99222 g Expanded Uncertainty: 0.00008 g Coverage Factor: 2  E3871 weighing certificate	
$m_{y2}$ :	Type B normal distribution Value: 1.00713 g Expanded Uncertainty: 0.00009 g Coverage Factor: 2  E3871 weighing certificate	
$m_{y4}$ :	Type B normal distribution Value: 0.99805 g Expanded Uncertainty: 0.00008 g Coverage Factor: 2  E3871 weighing certificate	
$m_{y5}$ :	Type B normal distribution Value: 0.99975 g Expanded Uncertainty: 0.00009 g Coverage Factor: 2  E3871 weighing certificate	
Date: 10/13/2017	File: Am-243 amount content with Am-241 spike by IDMS_edited.SMU	Page 11 of 29

Am-243 amount content with Am-241 spike by IDMS		
<b>m<sub>y6</sub>:</b>	Type B normal distribution Value: 1.02458 g Expanded Uncertainty: 0.00008 g Coverage Factor: 2  E3871 weighing certificate	
<b>m<sub>y7</sub>:</b>	Type B normal distribution Value: 1.00827 g Expanded Uncertainty: 0.00008 g Coverage Factor: 2  E3871 weighing certificate	
<b>m<sub>y8</sub>:</b>	Type B normal distribution Value: 1.01395 g Expanded Uncertainty: 0.00010 g Coverage Factor: 2  E3871 weighing certificate	
<b>m<sub>y9</sub>:</b>	Type B normal distribution Value: 1.00165 g Expanded Uncertainty: 0.00008 g Coverage Factor: 2  E3873 weighing certificate	
<b>m<sub>y10</sub>:</b>	Type B normal distribution Value: 1.09047 g Expanded Uncertainty: 0.00010 g Coverage Factor: 2  E3873 weighing certificate	
<b>m<sub>y11</sub>:</b>	Type B normal distribution Value: 1.01220 g Expanded Uncertainty: 0.00013 g Coverage Factor: 2  E3873 weighing certificate	
<b>m<sub>y12</sub>:</b>	Type B normal distribution Value: 1.05167 g Expanded Uncertainty: 0.00011 g Coverage Factor: 2  E3873 weighing certificate	
<b>m<sub>y13</sub>:</b>	Type B normal distribution Value: 1.02945 g Expanded Uncertainty: 0.00008 g Coverage Factor: 2  E3873 weighing certificate	
<b>m<sub>y15</sub>:</b>	Type B normal distribution Value: 1.00438 g Expanded Uncertainty: 0.00010 g Coverage Factor: 2  E3873 weighing certificate	
Date: 10/13/2017	File: Am-243 amount content with Am-241 spike by IDMS_edited.SMU	Page 12 of 29

Am-243 amount content with Am-241 spike by IDMS		
<b>m<sub>y16</sub>:</b>	Type B normal distribution Value: 1.00332 g Expanded Uncertainty: 0.00010 g Coverage Factor: 2	
E3873 weighing certificate		
<b>m<sub>y17</sub>:</b>	Type B normal distribution Value: 1.00590 g Expanded Uncertainty: 0.00008 g Coverage Factor: 2	
E3873 weighing certificate		
<b>m<sub>y18</sub>:</b>	Type B normal distribution Value: 0.99681 g Expanded Uncertainty: 0.00008 g Coverage Factor: 2	
E3873 weighing certificate		
<b>R<sub>b1</sub>:</b>	Type B normal distribution Value: 1.7813 mol/mol Expanded Uncertainty: 0.0013 mol/mol Coverage Factor: 2	
internal report 3448-1		
<b>R<sub>b2</sub>:</b>	Type B normal distribution Value: 1.82739 mol/mol Expanded Uncertainty: 0.00039 mol/mol Coverage Factor: 2	
internal report 3448-1		
<b>R<sub>b4</sub>:</b>	Type B normal distribution Value: 1.79365 mol/mol Expanded Uncertainty: 0.00052 mol/mol Coverage Factor: 2	
internal report 3448-2		
<b>R<sub>b5</sub>:</b>	Type B normal distribution Value: 1.78638 mol/mol Expanded Uncertainty: 0.00030 mol/mol Coverage Factor: 2	
internal report 3448-3		
<b>R<sub>b6</sub>:</b>	Type B normal distribution Value: 1.74157 mol/mol Expanded Uncertainty: 0.00048 mol/mol Coverage Factor: 2	
internal report 3448-3		
<b>R<sub>b7</sub>:</b>	Type B normal distribution Value: 1.71172 mol/mol Expanded Uncertainty: 0.00010 mol/mol Coverage Factor: 2	
internal report 3448-4		
Date: 10/13/2017	File: Am-243 amount content with Am-241 spike by IDMS_edited.SMU	Page 13 of 29

Am-243 amount content with Am-241 spike by IDMS		
<b>R<sub>b8</sub>:</b>	Type B normal distribution Value: 1.71177 mol/mol Expanded Uncertainty: 0.00077 mol/mol Coverage Factor: 2	
internal report 3448-4		
<b>R<sub>b9</sub>:</b>	Type B normal distribution Value: 1.64570 mol/mol Expanded Uncertainty: 0.00033 mol/mol Coverage Factor: 2	
internal report 3553-1		
<b>R<sub>b10</sub>:</b>	Type B normal distribution Value: 1.54375 mol/mol Expanded Uncertainty: 0.00032 mol/mol Coverage Factor: 2	
internal report 3553-1		
<b>R<sub>b11</sub>:</b>	Type B normal distribution Value: 1.610977 mol/mol Expanded Uncertainty: 0.000089 mol/mol Coverage Factor: 2	
internal report 3553-2		
<b>R<sub>b12</sub>:</b>	Type B normal distribution Value: 1.56733 mol/mol Expanded Uncertainty: 0.00021 mol/mol Coverage Factor: 2	
internal report 3553-2		
<b>R<sub>b13</sub>:</b>	Type B normal distribution Value: 1.54773 mol/mol Expanded Uncertainty: 0.00021 mol/mol Coverage Factor: 2	
internal report 3553-3		
<b>R<sub>b15</sub>:</b>	Type B normal distribution Value: 1.55925 mol/mol Expanded Uncertainty: 0.00028 mol/mol Coverage Factor: 2	
internal report 3553-4		
<b>R<sub>b16</sub>:</b>	Type B normal distribution Value: 1.57787 mol/mol Expanded Uncertainty: 0.00014 mol/mol Coverage Factor: 2	
internal report 3553-4		
<b>R<sub>b17</sub>:</b>	Type B normal distribution Value: 1.56394 mol/mol Expanded Uncertainty: 0.00026 mol/mol Coverage Factor: 2	
internal report 3553-5		
Date: 10/13/2017	File: Am-243 amount content with Am-241 spike by IDMS_edited.SMU	Page 14 of 29

Am-243 amount content with Am-241 spike by IDMS		
$R_{D18}$ :	Type B normal distribution Value: 1.68474 mol/mol Expanded Uncertainty: 0.00069 mol/mol Coverage Factor: 2	
internal report 3553-5		
$R_{y1}$ :	Type B normal distribution Value: 0 mol/mol Expanded Uncertainty: 0 mol/mol Coverage Factor: 1	
$R_{y2}$ :	Type B normal distribution Value: 0 mol/mol Expanded Uncertainty: 0 mol/mol Coverage Factor: 1	
$R_{y4}$ :	Type B normal distribution Value: 0 mol/mol Expanded Uncertainty: 0 mol/mol Coverage Factor: 1	
$R_{y5}$ :	Type B normal distribution Value: 0 mol/mol Expanded Uncertainty: 0 mol/mol Coverage Factor: 1	
$R_{y6}$ :	Type B normal distribution Value: 0 mol/mol Expanded Uncertainty: 0 mol/mol Coverage Factor: 1	
$R_{y7}$ :	Type B normal distribution Value: 0 mol/mol Expanded Uncertainty: 0 mol/mol Coverage Factor: 1	
$R_{y8}$ :	Type B normal distribution Value: 0 mol/mol Expanded Uncertainty: 0 mol/mol Coverage Factor: 1	
$R_{y9}$ :	Type B normal distribution Value: 0 mol/mol Expanded Uncertainty: 0 mol/mol Coverage Factor: 1	
$R_{y10}$ :	Type B normal distribution Value: 0 mol/mol Expanded Uncertainty: 0 mol/mol Coverage Factor: 1	
Date: 10/13/2017	File: Am-243 amount content with Am-241 spike by IDMS_edited.SMU	Page 15 of 29

Am-243 amount content with Am-241 spike by IDMS		
$R_{y11}$ :	Type B normal distribution Value: 0 mol/mol Expanded Uncertainty: 0 mol/mol Coverage Factor: 1	
$R_{y12}$ :	Type B normal distribution Value: 0 mol/mol Expanded Uncertainty: 0 mol/mol Coverage Factor: 1	
$R_{y13}$ :	Type B normal distribution Value: 0 mol/mol Expanded Uncertainty: 0 mol/mol Coverage Factor: 1	
$R_{y15}$ :	Type B normal distribution Value: 0 mol/mol Expanded Uncertainty: 0 mol/mol Coverage Factor: 1	
$R_{y16}$ :	Type B normal distribution Value: 0 mol/mol Expanded Uncertainty: 0 mol/mol Coverage Factor: 1	
$R_{y17}$ :	Type B normal distribution Value: 0 mol/mol Expanded Uncertainty: 0 mol/mol Coverage Factor: 1	
$R_{y18}$ :	Type B normal distribution Value: 0 mol/mol Expanded Uncertainty: 0 mol/mol Coverage Factor: 1	
$R_{x1}$ :	Type B normal distribution Value: 7.3393 mol/mol Expanded Uncertainty: 0.0018 mol/mol Coverage Factor: 2	
internal report 3448-1		
$R_{x2}$ :	Type B normal distribution Value: 7.3359 mol/mol Expanded Uncertainty: 0.0034 mol/mol Coverage Factor: 2	
internal report 3448-1		
$R_{x4}$ :	Type B normal distribution Value: 7.3354 mol/mol Expanded Uncertainty: 0.0025 mol/mol Coverage Factor: 2	
internal report 3448-2		
Date: 10/13/2017	File: Am-243 amount content with Am-241 spike by IDMS_edited.SMU	Page 16 of 29

Am-243 amount content with Am-241 spike by IDMS		
<b>R<sub>x5</sub>:</b>	Type B normal distribution Value: 7.3358 mol/mol Expanded Uncertainty: 0.0015 mol/mol Coverage Factor: 2  internal report 3448-3	
<b>R<sub>x6</sub>:</b>	Type B normal distribution Value: 7.3357 mol/mol Expanded Uncertainty: 0.0018 mol/mol Coverage Factor: 2  internal report 3448-3	
<b>R<sub>x7</sub>:</b>	Type B normal distribution Value: 7.3340 mol/mol Expanded Uncertainty: 0.0012 mol/mol Coverage Factor: 2  internal report 3448-4	
<b>R<sub>x8</sub>:</b>	Type B normal distribution Value: 7.3360 mol/mol Expanded Uncertainty: 0.0015 mol/mol Coverage Factor: 2  internal report 3448-4	
<b>R<sub>x9</sub>:</b>	Type B normal distribution Value: 7.33975 mol/mol Expanded Uncertainty: 0.00092 mol/mol Coverage Factor: 2  internal report 3553-1	
<b>R<sub>x10</sub>:</b>	Type B normal distribution Value: 7.34015 mol/mol Expanded Uncertainty: 0.00057 mol/mol Coverage Factor: 2  internal report 3553-1	
<b>R<sub>x11</sub>:</b>	Type B normal distribution Value: 7.3410 mol/mol Expanded Uncertainty: 0.0012 mol/mol Coverage Factor: 2  internal report 3553-2 (same as 345), vial 134 for IA was discarded	
<b>R<sub>x12</sub>:</b>	Type B normal distribution Value: 7.3410 mol/mol Expanded Uncertainty: 0.0012 mol/mol Coverage Factor: 2  internal report 3553-2	
<b>R<sub>x13</sub>:</b>	Type B normal distribution Value: 7.3391 mol/mol Expanded Uncertainty: 0.0011 mol/mol Coverage Factor: 2  internal report 3553-3	
Date: 10/13/2017	File: Am-243 amount content with Am-241 spike by IDMS_edited.SMU	Page 17 of 29

Am-243 amount content with Am-241 spike by IDMS		
<b>R<sub>x15</sub>:</b>	Type B normal distribution Value: 7.33848 mol/mol Expanded Uncertainty: 0.00052 mol/mol Coverage Factor: 2	
internal report 3553-4		
<b>R<sub>x16</sub>:</b>	Type B normal distribution Value: 7.34258 mol/mol Expanded Uncertainty: 0.00088 mol/mol Coverage Factor: 2	
internal report 3553-4		
<b>R<sub>x17</sub>:</b>	Type B normal distribution Value: 7.3370 mol/mol Expanded Uncertainty: 0.0015 mol/mol Coverage Factor: 2	
internal report 3553-5		
<b>R<sub>x18</sub>:</b>	Type B normal distribution Value: 7.34034 mol/mol Expanded Uncertainty: 0.00077 mol/mol Coverage Factor: 2	
internal report 3553-5		
<b>e:</b>	Constant Value: 2.7182818284590	
<b>Δ<sub>A</sub>:</b>	Type B rectangular distribution Value: 1.637235 a Halfwidth of Limits: 0.001780 a	
10 June 2014, 29 January 2016		
<b>Δ<sub>B</sub>:</b>	Type B rectangular distribution Value: 1.691992 a Halfwidth of Limits: 0.001780 a	
10 June 2014, 18 February 2016		
<b>Δ<sub>C</sub>:</b>	Type B rectangular distribution Value: 1.724846 a Halfwidth of Limits: 0.001780 a	
10 June 2014, 1 March 2016		
<b>Δ<sub>D</sub>:</b>	Type B rectangular distribution Value: 1.801506 a Halfwidth of Limits: 0.001780 a	
10 June 2014, 29 March 2016		
<b>Δ<sub>E</sub>:</b>	Type B rectangular distribution Value: 1.916496 a Halfwidth of Limits: 0.001780 a	
10 June 2014, 20 May 2016		
Date: 10/13/2017	File: Am-243 amount content with Am-241 spike by IDMS_edited.SMU	Page 18 of 29

Am-243 amount content with Am-241 spike by IDMS		
<b><math>\Delta t_f</math>:</b>	Type B rectangular distribution Value: 1.952088 a Halfwidth of Limits: 0.001780 a 10 June 2014, 3 June 2016	
<b><math>\Delta t_G</math>:</b>	Type B rectangular distribution Value: 2.015058 a Halfwidth of Limits: 0.001780 a 10 June 2014, 24 July 2016	
<b><math>\Delta t_H</math>:</b>	Type B rectangular distribution Value: 2.050650 a Halfwidth of Limits: 0.001780 a 10 June 2014, 7 July 2016	
<b><math>\Delta t_i</math>:</b>	Type B rectangular distribution Value: 2.088980 a Halfwidth of Limits: 0.001780 a 10 June 2014, 22 July 2016	
<b><math>K_1</math>:</b>	Import Filename: ..\..\Processing\Pu-241 inhouse spike characterisation\Determination of the Pu-241 amount content by IDMS using IRMM-049d spike.smu Symbol: $K_{evap1}$	
<b><math>K_2</math>:</b>	Import Filename: ..\..\Processing\Pu-241 inhouse spike characterisation\Determination of the Pu-241 amount content by IDMS using IRMM-049d spike.smu Symbol: $K_{evap2}$	
<b><math>K_3</math>:</b>	Import Filename: ..\..\Processing\Pu-241 inhouse spike characterisation\Determination of the Pu-241 amount content by IDMS using IRMM-049d spike.smu Symbol: $K_{evap3}$	
<b><math>K_4</math>:</b>	Import Filename: ..\..\Processing\Pu-241 inhouse spike characterisation\Determination of the Pu-241 amount content by IDMS using IRMM-049d spike.smu Symbol: $K_{evap4}$	
<b><math>K_5</math>:</b>	Import Filename: ..\..\Processing\Pu-241 inhouse spike characterisation\Determination of the Pu-241 amount content by IDMS using IRMM-049d spike.smu Symbol: $K_{evap5}$	
<b><math>c_{241Pu}</math>:</b>	Import Filename: ..\..\Processing\Pu-241 inhouse spike characterisation\Determination of the Pu-241 amount content by IDMS using IRMM-049d spike.smu Symbol: $c_{241Pu}$	
<b><math>\Delta T_A</math>:</b>	Type B rectangular distribution Value: 0.908966 a Halfwidth of Limits: 0.001369 a	
Date: 10/13/2017	File: Am-243 amount content with Am-241 spike by IDMS_edited.SMU	Page 19 of 29

Am-243 amount content with Am-241 spike by IDMS		
4 February 2016, 1 January 2017		
$\Delta T_B$ :	Type B rectangular distribution Value: 0.851472 a Halfwidth of Limits: 0.001369 a	
25 February 2016, 1 January 2017		
$\Delta T_C$ :	Type B rectangular distribution Value: 0.813142 a Halfwidth of Limits: 0.001369 a	
10 March 2016, 1 January 2017		
$\Delta T_D$ :	Type B rectangular distribution Value: 0.736482 a Halfwidth of Limits: 0.001369 a	
7 April 2016, 1 January 2017		
$\Delta T_E$ :	Type B rectangular distribution Value: 0.618754 a Halfwidth of Limits: 0.001369 a	
20 May 2016, 1 January 2017		
$\Delta T_F$ :	Type B rectangular distribution Value: 0.580424 a Halfwidth of Limits: 0.001369 a	
3 June 2016, 1 January 2017		
$\Delta T_G$ :	Type B rectangular distribution Value: 0.522930 a Halfwidth of Limits: 0.001369 a	
24 June 2016, 1 January 2017		
$\Delta T_H$ :	Type B rectangular distribution Value: 0.487337 a Halfwidth of Limits: 0.001369 a	
7 July 2016, 1 January 2017		
$\Delta T_I$ :	Type B rectangular distribution Value: 0.446270 a Halfwidth of Limits: 0.001369 a	
22 July 2016, 1 January 2017		
$\Delta T_{ATIMS}$ :	Type B rectangular distribution Value: 0.016427 a Halfwidth of Limits: 0.001369 a	
29 January, 4 February 2016 (IDMS)		
$\Delta T_{BTIMS}$ :	Type B rectangular distribution Value: 0.019165 a Halfwidth of Limits: 0.001369 a	
18 February, 25 February 2016 (IDMS)		
Date: 10/13/2017	File: Am-243 amount content with Am-241 spike by IDMS_edited.SMU	Page 20 of 29

Am-243 amount content with Am-241 spike by IDMS		
$\Delta T_{CTIMS}$ :	Type B rectangular distribution Value: 0.024641 a Halfwidth of Limits: 0.001369 a 1 March, 10 March 2016 (IDMS)	
$\Delta T_{DTIMS}$ :	Type B rectangular distribution Value: 0.024641 a Halfwidth of Limits: 0.001369 a 29 March, 7 April 2016 (IDMS)	
$\Delta T_{ETIMS}$ :	Type B rectangular distribution Value: 0.027379 a Halfwidth of Limits: 0.001369 a 10 May, 20 May 2016 (IDMS)	
$\Delta T_{FTIMS}$ :	Type B rectangular distribution Value: 0.030116 a Halfwidth of Limits: 0.001369 a 23 May, 3 June 2016 (IDMS)	
$\Delta T_{GTIMS}$ :	Type B rectangular distribution Value: 0.024641 a Halfwidth of Limits: 0.001369 a 15 June, 24 June 2016 (IDMS)	
$\Delta T_{HTIMS}$ :	Type B rectangular distribution Value: 0.024641 a Halfwidth of Limits: 0.001369 a 28 June, 7 July 2016 (IDMS)	
$\Delta T_{ITIMS}$ :	Type B rectangular distribution Value: 0.027379 a Halfwidth of Limits: 0.001369 a 12 July, 22 July 2016 (IDMS)	
$t_{241Am}$ :	Type B normal distribution Value: 432.6 a Expanded Uncertainty: 0.6 a Coverage Factor: 1 <a href="http://www.nucleide.org/DDEP_WG/DDEPdata.htm">http://www.nucleide.org/DDEP_WG/DDEPdata.htm</a> "last checked on March 20, 2017"	
$t_{243Am}$ :	Type B normal distribution Value: 7367 a Expanded Uncertainty: 23 a Coverage Factor: 1 <a href="http://www.nucleide.org/DDEP_WG/DDEPdata.htm">http://www.nucleide.org/DDEP_WG/DDEPdata.htm</a> "last checked on March 20, 2017"	
$t_{241Pu}$ :	Type B normal distribution Value: 14.325 a Expanded Uncertainty: 0.024 a Coverage Factor: 2 R. Wellum, A. Verbruggen, R. Kessel, A new evaluation of the half-life of 241Pu, J. Anal. At. Spectros., 24, 801-807, 2009	
Date: 10/13/2017	File: Am-243 amount content with Am-241 spike by IDMS_edited.SMU	Page 21 of 29

Am-243 amount content with Am-241 spike by IDMS		
<b>fdnorm<sub>243Amcert</sub></b> :	Import Filename: ..\Am amount ratios\Am isotope amount ratios by TIMS.smu Symbol: fdnorm <sub>243Amcert</sub>	
<b>fdnorm<sub>241Amcert</sub></b> :	Import Filename: ..\Am amount ratios\Am isotope amount ratios by TIMS.smu Symbol: fdnorm <sub>241Amcert</sub>	
<b>fdnorm<sub>242mAmILC</sub></b> :	Import Filename: ..\Am amount ratios\Am isotope amount ratios by TIMS.smu Symbol: fdnorm <sub>242mAmILC</sub>	
<b>M<sub>243Am</sub></b> :	Type B normal distribution Value: 243.0613813 mol/g Expanded Uncertainty: 0.0000024 mol/g Coverage Factor: 1	
Wang et al. (The AME 2012 atomic mass evaluation (II). Tables, Graphs and References, Chinese Physics C, Vol. 36, No. 12, 1603-2014, 2012).		
<b>M<sub>241Am</sub></b> :	Type B normal distribution Value: 241.0568293 mol/g Expanded Uncertainty: 0.0000019 mol/g Coverage Factor: 1	
Wang et al. (The AME 2012 atomic mass evaluation (II). Tables, Graphs and References, Chinese Physics C, Vol. 36, No. 12, 1603-2014, 2012).		
<b>M<sub>242mAm</sub></b> :	Type B normal distribution Value: 242.0595494 mol/g Expanded Uncertainty: 0.0000019 mol/g Coverage Factor: 1	
Wang et al. (The AME 2012 atomic mass evaluation (II). Tables, Graphs and References, Chinese Physics C, Vol. 36, No. 12, 1603-2014, 2012).		
<b>N<sub>a</sub></b> :	Constant Value: 6.023·10 <sup>23</sup> mol <sup>-1</sup>	
<b>A<sub>243Amgamma1</sub></b> :	Type B normal distribution Value: 10140 Bq/Bq Expanded Uncertainty: 100 Bq/Bq Coverage Factor: 1	
gamma measurements by Timos, reference 1 Jan 2017		
<b>A<sub>241Amgamma1</sub></b> :	Type B normal distribution Value: 237400 Bq/Bq Expanded Uncertainty: 130 Bq/Bq Coverage Factor: 1	
gamma measurements by Timos, reference 1 January 2017		
<b>A<sub>241Amgamma2</sub></b> :	Type B normal distribution Value: 23510 Bq/Bq Expanded Uncertainty: 260 Bq/Bq Coverage Factor: 1	
gamma measurements by Faidra, HADES reference 1 Jan 2017		
Date: 10/13/2017	File: Am-243 amount content with Am-241 spike by IDMS_edited.SMU	Page 22 of 29

Am-243 amount content with Am-241 spike by IDMS		
<b>A<sub>243Amgamma2</sub></b> :	Type B normal distribution Value: 10100 Bq/Bq Expanded Uncertainty: 210 Bq/Bq Coverage Factor: 1	
gamma measurements by Faidra, HADES reference 1 Jan 2017		
<b>A<sub>243Aalpha</sub></b> :	Type B normal distribution Value: 10230 Bq/Bq Expanded Uncertainty: 70 Bq/Bq Coverage Factor: 1	
alpha measurement by Maria, 1 january 2017		
<b>A<sub>241Aalpha</sub></b> :	Type B normal distribution Value: 23720 Bq/Bq Expanded Uncertainty: 70 Bq/Bq Coverage Factor: 1	
alpha measurement by Maria, 1 january 2017		
<b>K<sub>ASTM</sub></b> :	Type B normal distribution Value: 1 mol/mol Expanded Uncertainty: 0.00033 mol/mol Coverage Factor: 2	
0.05 % for spanning 3 mass units (ASTM standard), for 241/243 ratio is then 0.033 % (k = 2)		
<b>δ<sub>c243AmIs</sub></b> :	Type B normal distribution Value: 0 mol/g Expanded Uncertainty: 0 mol/g Coverage Factor: 2	
<b>δ<sub>c243Amsts</sub></b> :	Type B normal distribution Value: 0 mol/g Expanded Uncertainty: 0 mol/g Coverage Factor: 2	
<b>δ<sub>c243Ambb</sub></b> :	Type B normal distribution Value: 0 mol/g Expanded Uncertainty: 0.001663·10 <sup>-9</sup> mol/g Coverage Factor: 1	
sbb, rel = 0.0292 % (k = 1, soft CRM), average value from homogeneity study c243Am = 5.696438 nmol/g, sbb = 0.001344 nmol/g		
<b>K<sub>QC</sub></b> :	Type B normal distribution Value: 1 mol/mol Expanded Uncertainty: 0.00013 mol/mol Coverage Factor: 2	
additional uncertainty contribution for <sup>241</sup> Am/ <sup>243</sup> Am ratio from U/Pu QC charts (no Am isotopic standard available) 0.02% for 3 mass units (0.013% for <sup>241</sup> Am/ <sup>243</sup> Am ratio)		
Date: 10/13/2017	File: Am-243 amount content with Am-241 spike by IDMS_edited.SMU	Page 23 of 29

Am-243 amount content with Am-241 spike by IDMS

**Input Correlation:**

	K <sub>1</sub>	K <sub>2</sub>	K <sub>3</sub>	K <sub>4</sub>	K <sub>5</sub>	C <sub>241Pu</sub>
K <sub>1</sub>	1					0.0368
K <sub>2</sub>		1				0.0383
K <sub>3</sub>			1			0.0241
K <sub>4</sub>				1		0.0242
K <sub>5</sub>					1	0.0265
C <sub>241Pu</sub>	0.0368	0.0383	0.0241	0.0242	0.0265	1

	Mdnorm <sub>Amcert</sub>	fdnorm <sub>243Amcert</sub>	fdnorm <sub>241Amcert</sub>	fdnorm <sub>242mAmILC</sub>
Mdnorm <sub>Amcert</sub>	1	0.9984	-0.9984	-0.0163
fdnorm <sub>243Amcert</sub>	0.9984	1	-0.9987	-0.0414
fdnorm <sub>241Amcert</sub>	-0.9984	-0.9987	1	-0.0087
fdnorm <sub>242mAmILC</sub>	-0.0163	-0.0414	-0.0087	1

## Am-243 amount content with Am-241 spike by IDMS

## Interim Results:

Quantity	Value	Standard Uncertainty
C <sub>243Am456</sub>	5.69148·10 <sup>-9</sup> mol/g	6.18·10 <sup>-12</sup> mol/g
C <sub>243Am517</sub>	5.69574·10 <sup>-9</sup> mol/g	6.10·10 <sup>-12</sup> mol/g
C <sub>243Am309</sub>	5.69806·10 <sup>-9</sup> mol/g	6.16·10 <sup>-12</sup> mol/g
C <sub>243Am54</sub>	5.69703·10 <sup>-9</sup> mol/g	5.86·10 <sup>-12</sup> mol/g
C <sub>243Am134</sub>	5.69927·10 <sup>-9</sup> mol/g	5.80·10 <sup>-12</sup> mol/g
C <sub>243Am508</sub>	5.69500·10 <sup>-9</sup> mol/g	5.71·10 <sup>-12</sup> mol/g
C <sub>243Am222</sub>	5.69911·10 <sup>-9</sup> mol/g	5.71·10 <sup>-12</sup> mol/g
ε <sub>1</sub>	1.59·10 <sup>-12</sup> mol/g	4.03·10 <sup>-12</sup> mol/g
ε <sub>2</sub>	4.48·10 <sup>-12</sup> mol/g	3.37·10 <sup>-12</sup> mol/g
ε <sub>4</sub>	2.47·10 <sup>-12</sup> mol/g	3.28·10 <sup>-12</sup> mol/g
ε <sub>5</sub>	1.04·10 <sup>-12</sup> mol/g	3.13·10 <sup>-12</sup> mol/g
ε <sub>6</sub>	220·10 <sup>-15</sup> mol/g	3.25·10 <sup>-12</sup> mol/g
ε <sub>7</sub>	-550·10 <sup>-15</sup> mol/g	2.96·10 <sup>-12</sup> mol/g
ε <sub>8</sub>	-2.10·10 <sup>-12</sup> mol/g	3.41·10 <sup>-12</sup> mol/g
ε <sub>9</sub>	-1.07·10 <sup>-12</sup> mol/g	2.90·10 <sup>-12</sup> mol/g
ε <sub>10</sub>	-2.55·10 <sup>-12</sup> mol/g	2.88·10 <sup>-12</sup> mol/g
ε <sub>11</sub>	-3.30·10 <sup>-12</sup> mol/g	2.78·10 <sup>-12</sup> mol/g
ε <sub>12</sub>	-1.23·10 <sup>-12</sup> mol/g	2.79·10 <sup>-12</sup> mol/g
ε <sub>13</sub>	140·10 <sup>-15</sup> mol/g	2.71·10 <sup>-12</sup> mol/g
ε <sub>15</sub>	490·10 <sup>-15</sup> mol/g	2.70·10 <sup>-12</sup> mol/g
ε <sub>16</sub>	970·10 <sup>-15</sup> mol/g	2.66·10 <sup>-12</sup> mol/g
ε <sub>17</sub>	-3.14·10 <sup>-12</sup> mol/g	2.67·10 <sup>-12</sup> mol/g
ε <sub>18</sub>	-1.40·10 <sup>-12</sup> mol/g	2.92·10 <sup>-12</sup> mol/g
C <sub>241Am1</sub>	6.19091·10 <sup>-9</sup> mol/g	6.55·10 <sup>-12</sup> mol/g
C <sub>241Am2</sub>	6.38929·10 <sup>-9</sup> mol/g	6.68·10 <sup>-12</sup> mol/g
C <sub>241Am3</sub>	6.50802·10 <sup>-9</sup> mol/g	6.76·10 <sup>-12</sup> mol/g
C <sub>241Am4</sub>	6.78442·10 <sup>-9</sup> mol/g	6.94·10 <sup>-12</sup> mol/g
C <sub>241Am5</sub>	7.19700·10 <sup>-9</sup> mol/g	7.23·10 <sup>-12</sup> mol/g
C <sub>241Am6</sub>	7.32420·10 <sup>-9</sup> mol/g	7.31·10 <sup>-12</sup> mol/g
C <sub>241Am7</sub>	7.54881·10 <sup>-9</sup> mol/g	7.47·10 <sup>-12</sup> mol/g
C <sub>241Am8</sub>	7.67541·10 <sup>-9</sup> mol/g	7.56·10 <sup>-12</sup> mol/g
C <sub>241Am9</sub>	7.81147·10 <sup>-9</sup> mol/g	7.65·10 <sup>-12</sup> mol/g
λ <sub>241Pu</sub>	0.0483872 a <sup>-1</sup>	40.5·10 <sup>-6</sup> a <sup>-1</sup>
λ <sub>241Am</sub>	1.60228·10 <sup>-3</sup> a <sup>-1</sup>	2.22·10 <sup>-6</sup> a <sup>-1</sup>
λ <sub>243Am</sub>	94.088·10 <sup>-6</sup> a <sup>-1</sup>	294·10 <sup>-9</sup> a <sup>-1</sup>

Date: 10/13/2017

File: Am-243 amount content with Am-241 spike by IDMS\_edited.SMU

Page 25 of 29

Generated with GUM Workbench Pro Version 2.4.1.458

## Am-243 amount content with Am-241 spike by IDMS

## Uncertainty Budgets:

 $C_{243Amchar}$ : mean  $^{243}Am$  amount content from characterisation (9 values)

Quantity	Value	Standard Uncertainty	Distribution	Sensitivity Coefficient	Uncertainty Contribution	Index
$m_{x1}$	2.5443400 g	$50.0 \cdot 10^{-6}$ g	normal	$-250 \cdot 10^{-12}$	$-12 \cdot 10^{-15}$ mol/g	0.0 %
$m_{x4}$	2.6664200 g	$55.0 \cdot 10^{-6}$ g	normal	$-240 \cdot 10^{-12}$	$-13 \cdot 10^{-15}$ mol/g	0.0 %
$m_{x5}$	2.7053000 g	$50.0 \cdot 10^{-6}$ g	normal	$-230 \cdot 10^{-12}$	$-12 \cdot 10^{-15}$ mol/g	0.0 %
$m_{x7}$	2.6886500 g	$55.0 \cdot 10^{-6}$ g	normal	$-240 \cdot 10^{-12}$	$-13 \cdot 10^{-15}$ mol/g	0.0 %
$m_{x10}$	2.7017400 g	$55.0 \cdot 10^{-6}$ g	normal	$-230 \cdot 10^{-12}$	$-13 \cdot 10^{-15}$ mol/g	0.0 %
$m_{x12}$	2.7037100 g	$60.0 \cdot 10^{-6}$ g	normal	$-230 \cdot 10^{-12}$	$-14 \cdot 10^{-15}$ mol/g	0.0 %
$m_{x13}$	2.6853700 g	$50.0 \cdot 10^{-6}$ g	normal	$-240 \cdot 10^{-12}$	$-12 \cdot 10^{-15}$ mol/g	0.0 %
$m_{x15}$	2.6899300 g	$45.0 \cdot 10^{-6}$ g	normal	$-240 \cdot 10^{-12}$	$-11 \cdot 10^{-15}$ mol/g	0.0 %
$m_{x18}$	2.9995900 g	$45.0 \cdot 10^{-6}$ g	normal	$-210 \cdot 10^{-12}$	$-9.5 \cdot 10^{-15}$ mol/g	0.0 %
$m_{y1}$	0.9922200 g	$40.0 \cdot 10^{-6}$ g	normal	$640 \cdot 10^{-12}$	$26 \cdot 10^{-15}$ mol/g	0.0 %
$m_{y4}$	0.9980500 g	$40.0 \cdot 10^{-6}$ g	normal	$630 \cdot 10^{-12}$	$25 \cdot 10^{-15}$ mol/g	0.0 %
$m_{y5}$	0.9997500 g	$45.0 \cdot 10^{-6}$ g	normal	$630 \cdot 10^{-12}$	$28 \cdot 10^{-15}$ mol/g	0.0 %
$m_{y7}$	1.0082700 g	$40.0 \cdot 10^{-6}$ g	normal	$630 \cdot 10^{-12}$	$25 \cdot 10^{-15}$ mol/g	0.0 %
$m_{y10}$	1.0904700 g	$50.0 \cdot 10^{-6}$ g	normal	$580 \cdot 10^{-12}$	$29 \cdot 10^{-15}$ mol/g	0.0 %
$m_{y12}$	1.0516700 g	$55.0 \cdot 10^{-6}$ g	normal	$600 \cdot 10^{-12}$	$33 \cdot 10^{-15}$ mol/g	0.0 %
$m_{y13}$	1.0294500 g	$40.0 \cdot 10^{-6}$ g	normal	$610 \cdot 10^{-12}$	$25 \cdot 10^{-15}$ mol/g	0.0 %
$m_{y15}$	1.0043800 g	$50.0 \cdot 10^{-6}$ g	normal	$630 \cdot 10^{-12}$	$32 \cdot 10^{-15}$ mol/g	0.0 %
$m_{y18}$	0.9968100 g	$40.0 \cdot 10^{-6}$ g	normal	$640 \cdot 10^{-12}$	$25 \cdot 10^{-15}$ mol/g	0.0 %
$R_{b1}$	1.781300 mol/mol	$650 \cdot 10^{-5}$ mol/mol	normal	$470 \cdot 10^{-12}$	$300 \cdot 10^{-15}$ mol/g	0.4 %
$R_{b4}$	1.793650 mol/mol	$260 \cdot 10^{-5}$ mol/mol	normal	$470 \cdot 10^{-12}$	$120 \cdot 10^{-15}$ mol/g	0.0 %
$R_{b5}$	1.786380 mol/mol	$150 \cdot 10^{-5}$ mol/mol	normal	$470 \cdot 10^{-12}$	$70 \cdot 10^{-15}$ mol/g	0.0 %
$R_{b7}$	1.7117200 mol/mol	$50.0 \cdot 10^{-6}$ mol/mol	normal	$480 \cdot 10^{-12}$	$24 \cdot 10^{-15}$ mol/g	0.0 %
$R_{b10}$	1.543750 mol/mol	$160 \cdot 10^{-5}$ mol/mol	normal	$520 \cdot 10^{-12}$	$83 \cdot 10^{-15}$ mol/g	0.0 %
$R_{b12}$	1.567330 mol/mol	$105 \cdot 10^{-5}$ mol/mol	normal	$510 \cdot 10^{-12}$	$54 \cdot 10^{-15}$ mol/g	0.0 %
$R_{b13}$	1.547730 mol/mol	$105 \cdot 10^{-5}$ mol/mol	normal	$520 \cdot 10^{-12}$	$54 \cdot 10^{-15}$ mol/g	0.0 %
$R_{b15}$	1.559250 mol/mol	$140 \cdot 10^{-5}$ mol/mol	normal	$520 \cdot 10^{-12}$	$72 \cdot 10^{-15}$ mol/g	0.0 %
$R_{b18}$	1.684740 mol/mol	$345 \cdot 10^{-5}$ mol/mol	normal	$490 \cdot 10^{-12}$	$170 \cdot 10^{-15}$ mol/g	0.1 %
$R_{y1}$	0.0 mol/mol	0.0 mol/mol	normal	0.0	0.0 mol/g	0.0 %
$R_{y4}$	0.0 mol/mol	0.0 mol/mol	normal	0.0	0.0 mol/g	0.0 %

Date: 10/13/2017

File: Am-243 amount content with Am-241 spike by IDMS\_edited.SMU

Page 26 of 29

Generated with GUM Workbench Pro Version 2.4.1.458

Am-243 amount content with Am-241 spike by IDMS						
Quantity	Value	Standard Uncertainty	Distribution	Sensitivity Coefficient	Uncertainty Contribution	Index
R <sub>y5</sub>	0.0 mol/mol	0.0 mol/mol	normal	0.0	0.0 mol/g	0.0 %
R <sub>y7</sub>	0.0 mol/mol	0.0 mol/mol	normal	0.0	0.0 mol/g	0.0 %
R <sub>y10</sub>	0.0 mol/mol	0.0 mol/mol	normal	0.0	0.0 mol/g	0.0 %
R <sub>y12</sub>	0.0 mol/mol	0.0 mol/mol	normal	0.0	0.0 mol/g	0.0 %
R <sub>y13</sub>	0.0 mol/mol	0.0 mol/mol	normal	0.0	0.0 mol/g	0.0 %
R <sub>y15</sub>	0.0 mol/mol	0.0 mol/mol	normal	0.0	0.0 mol/g	0.0 %
R <sub>y18</sub>	0.0 mol/mol	0.0 mol/mol	normal	0.0	0.0 mol/g	0.0 %
R <sub>x1</sub>	7.339300 mol/mol	900·10 <sup>-6</sup> mol/mol	normal	-28·10 <sup>-12</sup>	-25·10 <sup>-15</sup> mol/g	0.0 %
R <sub>x4</sub>	7.33540 mol/mol	1.25·10 <sup>-3</sup> mol/mol	normal	-28·10 <sup>-12</sup>	-35·10 <sup>-15</sup> mol/g	0.0 %
R <sub>x5</sub>	7.335800 mol/mol	750·10 <sup>-6</sup> mol/mol	normal	-28·10 <sup>-12</sup>	-21·10 <sup>-15</sup> mol/g	0.0 %
R <sub>x7</sub>	7.334000 mol/mol	600·10 <sup>-6</sup> mol/mol	normal	-26·10 <sup>-12</sup>	-16·10 <sup>-15</sup> mol/g	0.0 %
R <sub>x10</sub>	7.340150 mol/mol	285·10 <sup>-6</sup> mol/mol	normal	-23·10 <sup>-12</sup>	-6.5·10 <sup>-15</sup> mol/g	0.0 %
R <sub>x12</sub>	7.341000 mol/mol	600·10 <sup>-6</sup> mol/mol	normal	-23·10 <sup>-12</sup>	-14·10 <sup>-15</sup> mol/g	0.0 %
R <sub>x13</sub>	7.339100 mol/mol	550·10 <sup>-6</sup> mol/mol	normal	-23·10 <sup>-12</sup>	-13·10 <sup>-15</sup> mol/g	0.0 %
R <sub>x15</sub>	7.338480 mol/mol	260·10 <sup>-6</sup> mol/mol	normal	-23·10 <sup>-12</sup>	-6.0·10 <sup>-15</sup> mol/g	0.0 %
R <sub>x18</sub>	7.340340 mol/mol	385·10 <sup>-6</sup> mol/mol	normal	-26·10 <sup>-12</sup>	-9.9·10 <sup>-15</sup> mol/g	0.0 %
e	2.718281828459					
Δt <sub>A</sub>	1.63724 a	1.03·10 <sup>-3</sup> a	rectangular	370·10 <sup>-12</sup>	380·10 <sup>-15</sup> mol/g	0.5 %
Δt <sub>B</sub>	1.69199 a	1.03·10 <sup>-3</sup> a	rectangular	360·10 <sup>-12</sup>	370·10 <sup>-15</sup> mol/g	0.5 %
Δt <sub>C</sub>	1.72485 a	1.03·10 <sup>-3</sup> a	rectangular	350·10 <sup>-12</sup>	360·10 <sup>-15</sup> mol/g	0.5 %
Δt <sub>D</sub>	1.80151 a	1.03·10 <sup>-3</sup> a	rectangular	340·10 <sup>-12</sup>	350·10 <sup>-15</sup> mol/g	0.4 %
Δt <sub>E</sub>	1.91650 a	1.03·10 <sup>-3</sup> a	rectangular	310·10 <sup>-12</sup>	320·10 <sup>-15</sup> mol/g	0.4 %
Δt <sub>F</sub>	1.95209 a	1.03·10 <sup>-3</sup> a	rectangular	310·10 <sup>-12</sup>	320·10 <sup>-15</sup> mol/g	0.4 %
Δt <sub>G</sub>	2.01506 a	1.03·10 <sup>-3</sup> a	rectangular	300·10 <sup>-12</sup>	310·10 <sup>-15</sup> mol/g	0.4 %
Δt <sub>H</sub>	2.05065 a	1.03·10 <sup>-3</sup> a	rectangular	290·10 <sup>-12</sup>	300·10 <sup>-15</sup> mol/g	0.3 %
Δt <sub>I</sub>	2.08898 a	1.03·10 <sup>-3</sup> a	rectangular	290·10 <sup>-12</sup>	300·10 <sup>-15</sup> mol/g	0.3 %
K <sub>1</sub>	0.9984064	12.1·10 <sup>-6</sup>		-5.7·10 <sup>-9</sup>	-69·10 <sup>-15</sup> mol/g	0.0 %
K <sub>2</sub>	0.9996192	12.6·10 <sup>-6</sup>		-5.7·10 <sup>-9</sup>	-72·10 <sup>-15</sup> mol/g	0.0 %
K <sub>3</sub>	0.9991598	13.2·10 <sup>-6</sup>		-5.7·10 <sup>-9</sup>	-75·10 <sup>-15</sup> mol/g	0.0 %
K <sub>4</sub>	0.9992664	13.3·10 <sup>-6</sup>		-3.2·10 <sup>-9</sup>	-42·10 <sup>-15</sup> mol/g	0.0 %
K <sub>5</sub>	0.9997767	14.5·10 <sup>-6</sup>		-1.3·10 <sup>-9</sup>	-18·10 <sup>-15</sup> mol/g	0.0 %
Date: 10/13/2017	File: Am-243 amount content with Am-241 spike by IDMS_edited.SMU				Page 27 of 29	

Am-243 amount content with Am-241 spike by IDMS						
Quantity	Value	Standard Uncertainty	Distribution	Sensitivity Coefficient	Uncertainty Contribution	Index
$C_{241Pu}$	$81.3936 \cdot 10^{-9}$ mol/g	$26.7 \cdot 10^{-12}$ mol/g		0.070	$1.9 \cdot 10^{-12}$ mol/g	13.2 %
$\Delta T_A$	0.908966 a	$790 \cdot 10^{-6}$ a	rectangular	$-60 \cdot 10^{-15}$	$-47 \cdot 10^{-18}$ mol/g	0.0 %
$\Delta T_B$	0.851472 a	$790 \cdot 10^{-6}$ a	rectangular	$-60 \cdot 10^{-15}$	$-47 \cdot 10^{-18}$ mol/g	0.0 %
$\Delta T_C$	0.813142 a	$790 \cdot 10^{-6}$ a	rectangular	$-60 \cdot 10^{-15}$	$-47 \cdot 10^{-18}$ mol/g	0.0 %
$\Delta T_D$	0.736482 a	$790 \cdot 10^{-6}$ a	rectangular	$-60 \cdot 10^{-15}$	$-47 \cdot 10^{-18}$ mol/g	0.0 %
$\Delta T_E$	0.618754 a	$790 \cdot 10^{-6}$ a	rectangular	$-60 \cdot 10^{-15}$	$-47 \cdot 10^{-18}$ mol/g	0.0 %
$\Delta T_F$	0.580424 a	$790 \cdot 10^{-6}$ a	rectangular	$-60 \cdot 10^{-15}$	$-47 \cdot 10^{-18}$ mol/g	0.0 %
$\Delta T_G$	0.522930 a	$790 \cdot 10^{-6}$ a	rectangular	$-60 \cdot 10^{-15}$	$-47 \cdot 10^{-18}$ mol/g	0.0 %
$\Delta T_H$	0.487337 a	$790 \cdot 10^{-6}$ a	rectangular	$-60 \cdot 10^{-15}$	$-47 \cdot 10^{-18}$ mol/g	0.0 %
$\Delta T_I$	0.446270 a	$790 \cdot 10^{-6}$ a	rectangular	$-60 \cdot 10^{-15}$	$-47 \cdot 10^{-18}$ mol/g	0.0 %
$\Delta T_{ATIMS}$	0.016427 a	$790 \cdot 10^{-6}$ a	rectangular	$-1.0 \cdot 10^{-12}$	$-800 \cdot 10^{-18}$ mol/g	0.0 %
$\Delta T_{BTIMS}$	0.019165 a	$790 \cdot 10^{-6}$ a	rectangular	$-1.0 \cdot 10^{-12}$	$-800 \cdot 10^{-18}$ mol/g	0.0 %
$\Delta T_{CTIMS}$	0.024641 a	$790 \cdot 10^{-6}$ a	rectangular	$-1.0 \cdot 10^{-12}$	$-800 \cdot 10^{-18}$ mol/g	0.0 %
$\Delta T_{DTIMS}$	0.024641 a	$790 \cdot 10^{-6}$ a	rectangular	$-1.0 \cdot 10^{-12}$	$-800 \cdot 10^{-18}$ mol/g	0.0 %
$\Delta T_{ETIMS}$	0.027379 a	$790 \cdot 10^{-6}$ a	rectangular	$-1.0 \cdot 10^{-12}$	$-800 \cdot 10^{-18}$ mol/g	0.0 %
$\Delta T_{FTIMS}$	0.030116 a	$790 \cdot 10^{-6}$ a	rectangular	$-1.0 \cdot 10^{-12}$	$-800 \cdot 10^{-18}$ mol/g	0.0 %
$\Delta T_{GTIMS}$	0.024641 a	$790 \cdot 10^{-6}$ a	rectangular	$-1.0 \cdot 10^{-12}$	$-800 \cdot 10^{-18}$ mol/g	0.0 %
$\Delta T_{HTIMS}$	0.024641 a	$790 \cdot 10^{-6}$ a	rectangular	$-1.0 \cdot 10^{-12}$	$-800 \cdot 10^{-18}$ mol/g	0.0 %
$\Delta T_{ITIMS}$	0.027379 a	$790 \cdot 10^{-6}$ a	rectangular	$-1.0 \cdot 10^{-12}$	$-800 \cdot 10^{-18}$ mol/g	0.0 %
$\tau_{241Am}$	432.600 a	0.600 a	normal	$21 \cdot 10^{-15}$	$12 \cdot 10^{-15}$ mol/g	0.0 %
$\tau_{243Am}$	7367.0 a	23.0 a	normal	$48 \cdot 10^{-18}$	$1.1 \cdot 10^{-15}$ mol/g	0.0 %
$\tau_{241Pu}$	14.3250 a	0.0120 a	normal	$-380 \cdot 10^{-12}$	$-4.6 \cdot 10^{-12}$ mol/g	78.5 %
$K_{ASTM}$	1.000000 mol/mol	$165 \cdot 10^{-6}$ mol/mol	normal	$5.7 \cdot 10^{-9}$	$940 \cdot 10^{-15}$ mol/g	3.3 %
$K_{QC}$	1.000000 mol/mol	$65.0 \cdot 10^{-6}$ mol/mol	normal	$5.7 \cdot 10^{-9}$	$370 \cdot 10^{-15}$ mol/g	0.5 %
$C_{243Amchar}$	$5.69597 \cdot 10^{-9}$ mol/g	$5.14 \cdot 10^{-12}$ mol/g				
Date: 10/13/2017	File: Am-243 amount content with Am-241 spike by IDMS_edited.SMU				Page 28 of 29	

Am-243 amount content with Am-241 spike by IDMS

Results:


Quantity	Value	Expanded Uncertainty	Coverage factor	Coverage
C <sub>243Amcert</sub>	5.696·10 <sup>-9</sup> mol/g	11·10 <sup>-12</sup> mol/g	2.00	manual
C <sub>241Amcert</sub>	775.4·10 <sup>-12</sup> mol/g	1.5·10 <sup>-12</sup> mol/g	2.00	manual
C <sub>242mAmILC</sub>	782·10 <sup>-15</sup> mol/g	14·10 <sup>-15</sup> mol/g	2.00	manual
C <sub>Amcert</sub>	6.472·10 <sup>-9</sup> mol/g	12·10 <sup>-12</sup> mol/g	2.00	manual
γ <sub>243Amcert</sub>	1.3845·10 <sup>-6</sup> g/g	2.6·10 <sup>-9</sup> g/g	2.00	manual
γ <sub>241Amcert</sub>	186.92·10 <sup>-9</sup> g/g	360·10 <sup>-12</sup> g/g	2.00	manual
γ <sub>242mAmILC</sub>	189.3·10 <sup>-12</sup> g/g	3.3·10 <sup>-12</sup> g/g	2.00	manual
γ <sub>Amcert</sub>	1.5716·10 <sup>-6</sup> g/g	3.0·10 <sup>-9</sup> g/g	2.00	manual
C <sub>243Amchar</sub>	5.6960·10 <sup>-9</sup> mol/g	5.1·10 <sup>-12</sup> mol/g	1.00	manual
C <sub>243Am165</sub>	5.694·10 <sup>-9</sup> mol/g	13·10 <sup>-12</sup> mol/g	2.00	manual
C <sub>243Am83</sub>	5.693·10 <sup>-9</sup> mol/g	12·10 <sup>-12</sup> mol/g	2.00	manual
C <sub>243Am373</sub>	5.695·10 <sup>-9</sup> mol/g	12·10 <sup>-12</sup> mol/g	2.00	manual
C <sub>243Am11</sub>	5.697·10 <sup>-9</sup> mol/g	12·10 <sup>-12</sup> mol/g	2.00	manual
C <sub>243Am564</sub>	5.699·10 <sup>-9</sup> mol/g	12·10 <sup>-12</sup> mol/g	2.00	manual
C <sub>243Am345</sub>	5.697·10 <sup>-9</sup> mol/g	12·10 <sup>-12</sup> mol/g	2.00	manual
C <sub>243Am194</sub>	5.696·10 <sup>-9</sup> mol/g	12·10 <sup>-12</sup> mol/g	2.00	manual
C <sub>243Am284</sub>	5.695·10 <sup>-9</sup> mol/g	11·10 <sup>-12</sup> mol/g	2.00	manual
C <sub>243Am497</sub>	5.697·10 <sup>-9</sup> mol/g	12·10 <sup>-12</sup> mol/g	2.00	manual
γ <sub>243Amgamma1</sub>	1.372·10 <sup>-5</sup> g/g	28·10 <sup>-9</sup> g/g	2.00	manual
γ <sub>241Amgamma1</sub>	1.8713·10 <sup>-6</sup> g/g	5.6·10 <sup>-9</sup> g/g	2.00	manual
γ <sub>243Amgamma2</sub>	1.367·10 <sup>-5</sup> g/g	57·10 <sup>-9</sup> g/g	2.00	manual
γ <sub>241Amgamma2</sub>	185.3·10 <sup>-9</sup> g/g	4.1·10 <sup>-9</sup> g/g	2.00	manual
γ <sub>241Amalpha</sub>	187.0·10 <sup>-9</sup> g/g	1.2·10 <sup>-9</sup> g/g	2.00	manual
γ <sub>243Amalpha</sub>	1.385·10 <sup>-5</sup> g/g	21·10 <sup>-9</sup> g/g	2.00	manual

## Annex E Results of the characterisation for IRMM-0243


**Table E1** Results for the amount content of  $^{243}\text{Am}$ ,  $n(^{241}\text{Am})/n(^{243}\text{Am})$  and  $n(^{241}\text{Am})/n(^{243}\text{Am})$  amount ratios and their expanded uncertainties (coverage factor,  $k = 2$ ). The reference date is January 1, 2017.

Unit	$^{243}\text{Am}$ amount content [ $10^{-9}$ mol/g]	$n(^{241}\text{Am})/n(^{243}\text{Am})$ [mol/mol]	$n(^{242\text{m}}\text{Am})/n(^{243}\text{Am})$ [mol/mol]
165	$5.694 \pm 0.013$	$0.136066 \pm 0.000076$	$0.0001531 \pm 0.0000063$
83	$5.693 \pm 0.012$	$0.136150 \pm 0.000083$	$0.0001516 \pm 0.0000046$
373	$5.695 \pm 0.012$	$0.136152 \pm 0.000074$	$0.0001543 \pm 0.0000044$
11	$5.697 \pm 0.012$	$0.136200 \pm 0.000071$	$0.0001490 \pm 0.0000047$
564	$5.699 \pm 0.012$	$0.136110 \pm 0.000069$	$0.0001487 \pm 0.0000038$
345	$5.697 \pm 0.012$	$0.136102 \pm 0.000071$	$0.0001492 \pm 0.0000056$
194	$5.696 \pm 0.011$	$0.136150 \pm 0.000071$	$0.0001520 \pm 0.0000035$
284	$5.695 \pm 0.011$	$0.136168 \pm 0.000069$	$0.0001497 \pm 0.0000038$
497	$5.697 \pm 0.012$	$0.136141 \pm 0.000069$	$0.0001482 \pm 0.0000034$
Mean (GUM)	$5.696 \pm 0.010$	$0.136138 \pm 0.000051$	$0.0001506 \pm 0.0000015$
Rel. U ( $k = 2$ )	0.18 %	0.038 %	1.0 %


**Figure E1** The amount content of  $^{243}\text{Am}$  from the characterisation by IDMS on selected ampoules. 'Error bars' represent the expanded uncertainty (coverage factor,  $k = 2$ ). The red data point is the mean value.



**Figure E2**  $n(^{241}\text{Am})/n(^{243}\text{Am})$  amount ratio from the characterisation by TIMS on selected ampoules. 'Error bars' represent the expanded uncertainty (coverage factor,  $k = 2$ ). The red data point is the mean value.




**Figure E3**  $n(^{242\text{m}}\text{Am})/n(^{243}\text{Am})$  amount ratio from the characterisation by TIMS on selected ampoules. 'Error bars' represent the expanded uncertainty (coverage factor,  $k = 2$ ). The red data point is the mean value from the TIMS analysis but not the indicative value on the IRMM-0243 certificate.




**Annex F** Results of the radiometric verification measurements for IRMM-0243

**Figure F1** Results of the independent alpha-particle (triangles) and gamma-ray (diamonds and squares) measurements for the mass fractions of  $^{241}\text{Am}$  and  $^{243}\text{Am}$  and  $n(^{241}\text{Am})/n(^{243}\text{Am})$  amount ratios expressed as relative differences from the certified value. 'Error bars' show the relative expanded uncertainty ( $k = 2$ ). Red dotted lines show the relative expanded uncertainty ( $k = 2$ ) of the respective certified values.



**Annex G** Results for the  $n(^{242m}\text{Am})/n(^{243}\text{Am})$  amount ratio from the  $^{243}\text{Am}$  ILC

**Figure G1** Results for the  $n(^{242m}\text{Am})/n(^{243}\text{Am})$  amount ratios with associated measurement uncertainty from the nine participating laboratories. The green lines show the assigned indicative value (DerSimonian-Laird weighted mean). The red lines show the characterisation value measured by JRC- Geel.



## **GETTING IN TOUCH WITH THE EU**

### **In person**

All over the European Union there are hundreds of Europe Direct information centres. You can find the address of the centre nearest you at: <http://europa.eu/contact>

### **On the phone or by email**

Europe Direct is a service that answers your questions about the European Union. You can contact this service:

- by freephone: 00 800 6 7 8 9 10 11 (certain operators may charge for these calls),
- at the following standard number: +32 22999696, or
- by electronic mail via: <http://europa.eu/contact>

## **FINDING INFORMATION ABOUT THE EU**

### **Online**

Information about the European Union in all the official languages of the EU is available on the Europa website at: <http://europa.eu>

### **EU publications**

You can download or order free and priced EU publications from EU Bookshop at: <http://bookshop.europa.eu>. Multiple copies of free publications may be obtained by contacting Europe Direct or your local information centre (see <http://europa.eu/contact>).

## JRC Mission

As the science and knowledge service of the European Commission, the Joint Research Centre's mission is to support EU policies with independent evidence throughout the whole policy cycle.



**EU Science Hub**

[ec.europa.eu/jrc](https://ec.europa.eu/jrc)



@EU\_ScienceHub



EU Science Hub - Joint Research Centre



Joint Research Centre



EU Science Hub

