




PROFICIENCY TESTING SCHEME, Measurements of inorganic gaseous pollutants (SO₂, CO, O₃, NO and NO₂) in filtered ambient air, 18-21 March 2024 Ispra, Italy

*European Commission harmonisation programme for air quality
measurements*

Barbiere, M.G., Tarricone, C., Borowiak, A.

2024



This document is a publication by the Joint Research Centre (JRC), the European Commission's science and knowledge service. It aims to provide evidence-based scientific support to the European policymaking process. The contents of this publication do not necessarily reflect the position or opinion of the European Commission. Neither the European Commission nor any person acting on behalf of the Commission is responsible for the use that might be made of this publication. For information on the methodology and quality underlying the data used in this publication for which the source is neither Eurostat nor other Commission services, users should contact the referenced source. The designations employed and the presentation of material on the maps do not imply the expression of any opinion whatsoever on the part of the European Union concerning the legal status of any country, territory, city or area or of its authorities, or concerning the delimitation of its frontiers or boundaries.

Contact information

Name: Maurizio Barbieri
Address: Via E. Fermi 2749 Ispra (VA) Italy
Email : maurizio-giuseppe.barbieri@ec.europa.eu
Tel.: +39 0332783057

EU Science Hub

<https://joint-research-centre.ec.europa.eu>

JRC139409

EUR 40125

PDF ISBN 978-92-68-22422-9 ISSN 1831-9424 doi:10.2760/1771834 KJ-01-24-163-EN-N

Luxembourg: Publications Office of the European Union, 2024

© European Union, 2024




The reuse policy of the European Commission documents is implemented by the Commission Decision 2011/833/EU of 12 December 2011 on the reuse of Commission documents (OJ L 330, 14.12.2011, p. 39). Unless otherwise noted, the reuse of this document is authorised under the Creative Commons Attribution 4.0 International (CC BY 4.0) licence (<https://creativecommons.org/licenses/by/4.0/>). This means that reuse is allowed provided appropriate credit is given and any changes are indicated.

For any use or reproduction of photos or other material that is not owned by the European Union permission must be sought directly from the copyright holders.

How to cite this report: European Commission, Joint Research Centre, Barbieri, M.G., Tarricone, C. and Borowiak, A., *PROFICIENCY TESTING SCHEME, Measurements of inorganic gaseous pollutants (SO₂, CO, O₃, NO and NO₂) in filtered ambient air, 18-21 March 2024 Ispra, Italy*, Publications Office of the European Union, Luxembourg, 2024, <https://data.europa.eu/doi/10.2760/1771834>, JRC139409.

Contents

Abstract.....	3
Acknowledgements.....	4
1 Introduction.....	5
2 Proficiency test organisation.....	6
3 Participants.....	6
4 Preparation of test mixtures.....	8
5 The evaluation of laboratory's measurement proficiency.....	10
5.1 z-score - z'-score.....	10
5.2 En-score.....	19
6 Conclusions.....	27
References.....	31
List of abbreviations.....	34
Bureau International des Poids et Mesures.....	34
Mathematical Symbols.....	35
List of figures.....	36
List of tables.....	37
Annex A. Reference values.....	38
Homogeneity.....	40
Annex B: Results reported by participants.....	42
Annex C: Reproducibility.....	63
1 Introduction.....	63
2 Results.....	65
3 Conclusions.....	70
Annex D: Confidentiality.....	71
Annex E: Accreditation certificates.....	72



EUROPEAN REFERENCE LABORATORY FOR AIR POLLUTION - ERLAP

PROFICIENCY TESTING SCHEME

Measurement of inorganic gaseous pollutants
(SO₂, CO, O₃, NO and NO₂) in filtered ambient air
(18-21 March 2024, Ispra-Italy)

European Commission harmonisation programme
for air quality measurements

Barbiere M.G., Tarricone C., Borowiak A.



PTP N° 0018 P

Abstract

Within the harmonisation programme of Air Quality monitoring in Europe the European Reference Laboratory of Air Pollution (ERLAP) organises Proficiency Tests (PT).

From **18-21 March 2024**, eleven Laboratories of AQUILA (Network of European Air Quality Reference Laboratories, including ERLAP) gathered in Ispra, Italy, for a laboratory comparison exercise in Ispra (IT) to assess their proficiency in analysing inorganic gaseous air pollutants (NO, NO₂, SO₂, CO and O₃), covered by the European Air Quality Directive 2008/50 EC [1] and its last amendments 2015/1480/EC [42].

Ecomaks laboratory did not participate to O₃ and ACES laboratory to SO and CO₂ measurements, respectively.

The proficiency evaluation, where each participant's bias was compared to two criteria (z/z'-score and En-score), provides information on the current situation and capabilities to the European Commission and can be used by participants in their quality control system. In agreement with all participants an evaluation of SO₂ concentration measured as interference during NO_x measurement has been carried out.

Evaluation based on z/z'-scores indicated that **89.2%** of the reported results were satisfactory. In contrast, consideration of the En-score revealed that **78.8%** of all values met satisfactory criteria. Overall, the performance for both indicators was below the average typically achieved in exercises of this type.

Considering the reproducibility evaluation, the results among AQUILA participants at the highest generated concentration levels are 9.9% for CO, 12% for NO, 11% for NO₂, 4.5% for O₃, and 11% for SO₂, all of which fall within the objectives derived from criteria set by EN standards. However, NO, NO₂ and SO₂ show a poor performance, in comparison to the reference value of this PT.

Acknowledgements

Thank you for the participation to:

MARLIERE Fabrice, NATHALIE Bocquet, DÉZSI Viktor, PÓLAY Gábor, DELANEY Lin, RUANE Cathal, BAKRADZE Georgijs, PODSKOCIJS Valters, DAUGE Franck, MARSTEEN Leif, FRACZKOWSKI Tomasz, PINDEL Andrzej, CASIMIRO Susana, MARTINS Ana, BOSKOVIC Stefan, DURICIC Aleksandar, IVANC Natalija, MIKLAVŽINA Iztok, GRAHAM Emelie, PFALLER Stefan.

	NAME	VERSION	DATE
Editing	A. BOROWIAK (PT Coordinator) M. BARBIERE (Technician/QO) C. TARRICONE (Technician)	FINAL REV. 0	30/07/2024
Review	JRC Editorial Review Board (proof reading check with minor remarks)	FINAL REV. 1	03/10/2024
Approval	F. PEKAR (Head of Unit)	FINAL REV. 1	20/09/2024
Authorization	P. SZYMANSKI (Director)	FINAL REV. 1	22/11/2024

1 Introduction

The Directive 2008/50/EC [1] and its last amendments 2015/1480/EC [42] on ambient air quality and cleaner air for Europe sets a framework for a harmonised air quality assessment in Europe.

One important objective of the Directive [1] is that the ambient air quality shall be assessed based on common methods and criteria. It deals with the air pollutants sulphur dioxide (SO₂), nitrogen dioxide (NO₂) and nitrogen monoxide (NO), particulate matter, lead, benzene, carbon monoxide (CO) and ozone (O₃). Among others it specifies the reference methods for measurements and Data Quality Objectives (DQOs) for the accuracy of measurements.

The European Commission (EC) has supported the development and publication of reference measurement methods for CO [2], SO₂ [3], NO-NO₂ [4] and O₃ [5] as European standards. Appropriate calibration methods [6], [7] and [8] have been standardised by the International Organization for Standardization (ISO).

As foreseen in the Air Quality Directive [1, 42], the European Reference Laboratory of Air Pollution (ERLAP) of the Directorate for Energy, Mobility and Climate at the Joint Research Centre (JRC) organises Proficiency Tests (PT) to assess and improve the status of comparability of measurements of National Reference Laboratories (NRL) of the Member States of the European Union.

The World Health Organization Collaborating Centre for Air Quality Management and Air Pollution Control, Berlin (WHO CC) is carrying out similar activities since 1994 [9] [10], [24], [31], [35], [38], [45] and [50], but with a view to obtaining harmonised air quality data for health related studies. Their programme integrates within the WHO EURO region, which includes public health institutes and other national institutes - especially from Central Eastern Europe, Caucasus and countries from Central Asia.

Starting in 2004, it has been decided to bring together the efforts of both the JRC-ERLAP and WHO CC and to coordinate activities as far as possible, with a view to optimise resources and improve international harmonisation.

This report deals with the PT that took place in the period **18-21 of March 2024** in Ispra (IT).

Since 1990 ERLAP has organised PT in order to evaluate the comparability of measurements carried out by NRLs and promote information exchange among the expert laboratories. Recently, a more systematic approach has been adopted, in agreement with the Network of National Reference Laboratories for Air Quality (AQUILA) [11], aiming to both provide an alert mechanism for the requirements of the EC legislation and support the implementation of quality schemes by NRLs.

ERLAP is accredited to the standard ISO 17043 for the measurement of these air pollutants and EN standard methods used during this proficiency test (EN 14211:2012-NO/NO₂, EN 14212:2012-SO₂, EN 14625:2012-O₃, EN 14626:2012-CO) as proved by the certificates in Annex E.

The methodology for the organisation of PT was developed by ERLAP in collaboration with AQUILA and is described in a paper on the organisation of laboratory comparison exercises for gaseous air pollutants [12].

This evaluation scheme was adopted by AQUILA in December 2008 and has been applied to all PT since then. The evaluation scheme implements the z/z'-score method [13] as performance indicator. According to the above-mentioned document [12], NRLs with an unsatisfactory performance in the z/z'-score evaluation (one unsatisfactory or two questionable results per parameter) are asked to repeat their participation in the following PT in order to demonstrate remediation measures [12].

In addition, considering that the evaluation scheme should be useful to participants for accreditation according to ISO 17025, they are requested to include their measurement uncertainty. Hence, participants' results (measurement values and uncertainties) are also compared to the reference values applying the En-score method [13].

Besides the proficiency of participating laboratories, the reproducibility of standardised measurement methods [14], [15] and [16] are evaluated as well. This group evaluation is a useful indicator of trends in measurement quality over different proficiency tests.

2 Proficiency test organisation

The PT was announced in **September 2023** to the members of the AQUILA network and the WHO CC representative. Registration was opened in **November 2023** and closed the first week of **March 2024**.

Every participant, together with the registration confirmation, received a detailed protocol with all the necessary information about the PT. Each laboratory was required to bring their own measurement instruments, data acquisition equipment and travelling standards (to be used for calibrations or checks during the PT).

The participants were invited to arrive on **Monday, 18th March 2024**, for the installation of their equipment. The calibration of NO_x and O₃ analysers was carried out the morning of the following day and the generation of NO_x and O₃ gas mixtures started at 11:00 on Tuesday.

The calibration of SO₂ and CO analysers was carried out on Wednesday afternoon and the generation of CO and SO₂ gas mixtures started at 20:00 of the same day.

The test gases generation and measurements finished on Thursday at 9:00.

3 Participants

All participants (Table 1) were organisations dealing with the routine ambient air monitoring or institutions involved in environmental or public health protection. The national representatives came from Sweden, Portugal, Slovenia, Serbia, Ireland, Poland, Hungary, France, Latvia and Norway.

Table 1: List of participating organizations.

Laboratory Acronym	Complete laboratory name	Country
ACES	Stockholm University	SE
APA	Agencia Portuguesa Do Ambiente	PT
ARSO	Slovenian Environment Agency	SI
Ecomaks	Ecomaks Solutions	RS
EPA	Environmental Protection Agency	IE
ERLAP	European Reference Laboratory for Air Pollution	IT
GIOS	Chief Inspectorate of Environmental Protection	PL
HMS	Hungarian Meteorological Service	HU
INERIS	Institut National de l'Environnement Industriel et des Risques	FR
LVGMC	Latvian Environment, Geology and Meteorology Centre	LV
NILU	Norwegian Institute for Air Research	NO

Source: JRC 2024

The following Table 2 reports the manufacturer and model of the instrumentations used by every participant during the Proficiency Test, including those used in the calculation of the reference values. The list contains technical information and cannot be considered as an implicit or explicit endorsement by the organisers of any specific instrumentation. This list is used to identify presence of clusters during the data evaluation.

ACES laboratory didn't participate to the testing exercise for CO and SO₂ and Ecomaks laboratory for O₃ measurements.

Table 2: List of instruments used by participants.

Laboratory	Analyser SO ₂	Analyser CO	Analyser NO	Analyser O ₃
ACES	/	/	Environnement SA AS32M	Environnement SA 0342e
APA	Thermo FS 43i	Thermo FS 48i	Thermo FS 42i	Thermo FS 49i
ARSO	Horiba APSA 370	Horiba APMA 370	Horiba APNA 370	Thermo FS 49C
Ecomaks	Teledyne API T100	Teledyne API T300	Teledyne API T200	/
EPA	Teledyne API T100	Teledyne API T300	Teledyne API T200	Teledyne API T400
ERLAP	Thermo FS 43 i-TLE	Horiba APMA 370	Thermo FS 42i	Thermo FS 49i
GIOS	Horiba APSA 370	Horiba APMA 370	Teledyne API T200	Horiba APOA-370
HMS	Thermo FS 43 i-TLE	Thermo FS 48i	Thermo FS 42i	Thermo FS TEI 49 i
INERIS	Teledyne API T100	Horiba APMA 370	Horiba APNA 370	Thermo FS TEI 49 i
LVGMC	Teledyne API T100	Teledyne API T300	Teledyne API T200	Teledyne API T400
NILU	Horiba APSA 370	Horiba APMA 370	Horiba APNA 370	Horiba APOA-370

Source: JRC 2024

4 Preparation of test mixtures

The ERLAP PT facility has been described in several reports [17], [18]. During this PT, gas mixtures were prepared for SO₂, CO, O₃, NO and NO₂ at concentration levels around limit values, critical levels and assessment thresholds set by the European Air Quality Directive [1].

Table 3: Sequence program of generated test gases with indicative pollutant concentrations.

Run	Component	Day	Start time	Duration	NO	NO ₂	O ₃	SO ₂	CO
				h	nmol/mol				μmol/mol
	/	1st	09:00		Installation				
	/	2nd	08:00	3	Calibration				
0	NO-NO ₂ -O ₃	2nd	11:00	1	Zero air				
1	NO-NO ₂	2nd	12:00	2	200	0			
2	NO-NO ₂	2nd	14:00	2	190	10			
1	O ₃	2nd	16:00	2			60		
3	NO-NO ₂	2nd	18:00	2	30	0			
4	NO-NO ₂	2nd	20:00	2	20	10			
2	O ₃	2nd	22:00	2			20		
5	NO-NO ₂	3rd	00:00	2	60	0			
6	NO-NO ₂	3rd	02:00	2	45	25			
3	O ₃	3rd	04:00	2			180		
7	NO-NO ₂	3rd	06:00	2	480	7		X	
8	NO-NO ₂	3rd	08:00	2	280	200		X	
4	O ₃	3rd	10:00	2			90		
9	NO-NO ₂	3rd	12:00	2	300	4		X	
10	NO-NO ₂	3rd	14:00	2	180	120		X	
5	O ₃	3rd	16:00	2			120		
	/	3rd	18:00	2	Calibration				
0	CO-SO ₂	3rd	20:00	1				Zero air	
1	CO-SO ₂	3rd	21:00	2				120	0.7
2	CO-SO ₂	3rd	23:00	2				35	9
3	CO-SO ₂	4th	01:00	2				5	1
4	CO-SO ₂	4th	03:00	2				15	2
5	CO-SO ₂	4th	05:00	2				50	5
		4th	07:00	2	Zero air (not to be reported)				
		4th	09:00	END					

Source: JRC 2024

The sequence program of generated test concentration requested is given in Table 3. In addition, the PT provider, in agreement with the participants, requested also the concentrations of NO₂ and SO₂ indicated in Table 3 with an X.

The test mixtures were prepared by gas dilution from cylinders containing high concentrations of NO, SO₂, or CO using thermal mass flow controllers [8]. O₃ was added using an ozone generator and NO₂ was produced applying the gas phase titration method [19] in a condition of NO excess or as a possible presence of NO₂ in NO cylinders.

The participants were required to report three half-hour-mean measurements for each concentration level (run). Zero value concentration levels were generated for one hour and was requested to be reported one half-hour-mean measurement.

5 The evaluation of laboratory's measurement proficiency

To evaluate the participant's measurement proficiency, the methodology described in ISO 13528 [13] was applied and measurement results of ERLAP were used as the reference values [12].

The traceability of ERLAP's measurement results and the reference values list are presented in Annex A.

In the following proficiency evaluations, the uncertainty of test gas homogeneity (Annex A) was added to the uncertainties of ERLAP's measurement results.

In Annex B for each participant and for each run, the values of measurement submitted, the mean value, the uncertainties, σ_{pt} calculated, z-score or z'-score (z/z'-score) and En-score are reported. The data are reported in a table and a graphical format, with all values plotted against the established reference value.

As described in the AQUILA document 37 [12], the proficiency of the participants was assessed by calculating two performance indicators (z/z'-score and En-score).

The performance indicators z/z'-score verify if the difference between the participants measured value and the reference value remains within the limits of a common criterion. The selection of the appropriate indicator, z-score or z'-score, depends on whether the uncertainty of the reference value (u_{ref}) meets the criterion $u_{ref} < 0,3\sigma_{pt}$ from ISO 13528 (paragraph 9.2) [13]. When the criterion is met the z-score value is applied, in the other case z'-score is used. The second performance indicator (En-score) verifies if the difference between the participants measured values and reference value remains within the limits of a criterion calculated, from the uncertainty of the participant measurement result and the uncertainty of the reference value.

During the evaluation a process of detection of exceptional errors (error during typing, slip in performing the measurement or the calculation, wrong averaging interval, malfunction of instrumentation, etc.) is applied. In this procedure a data consistency evaluation is carried out as described in ISO 13528 [13].

Before the release of the draft report, laboratories showing clear blunders are requested to investigate the cause of discrepancies and are allowed to correct their results in case of identification of exceptional errors.

After blunders (exceptional error) data are considered definitive, and z/z'-scores calculation is used to estimate the participant performance.

5.1 z-score - z'-score

The z/z'- score performance indicators are calculated according to ISO 13528 (par. 9.4 and 9.5) [13] as:

$$\text{Equation 1}$$
$$z = \frac{(x_{lab} - X_{ref})}{\sqrt{\sigma_{pt}^2}} = \frac{(x_{lab} - X_{ref})}{\sqrt{(a \cdot X_{ref} + b)^2}}$$

$$\text{Equation 2}$$
$$z' = \frac{(x_{lab} - X_{ref})}{\sqrt{\sigma_{pt}^2 + u_{ref}^2}} = \frac{(x_{lab} - X_{ref})}{\sqrt{(a \cdot X_{ref} + b)^2 + u_{ref}^2}}$$

z = z-score

z' = z'-score

x_{lab} = participant average values

X_{ref} = reference value

u_{ref} = uncertainty of the reference value

σ_{pt} = Standard deviation for proficiency assessment

a = slope (see table 4)

b = intercept (see table 4)

In the NO₂, SO₂, CO and O₃ EN Standards [2, 3, 4, 5] the uncertainties for calibration gases used in ongoing quality control are prescribed. In fact, it is stated that maximum permitted expanded uncertainty for calibration gases at the calibration point (75% of certification range) is 5% and that 'zero gas' shall not give instrument reading higher than the detection limit. The 'standard deviation for proficiency assessment' (σ_{pt}) is derived in a fitness-for-purpose manner from requirements given in the EN standards, where in place of detection limits criteria, the specifications for purity of zero gas used in type approval as defined in EN Standards are taken.

Over the whole measurement range σ_{pt} is calculated by linear interpolation between the uncertainty of 2.5% at the calibration point (75% of certification range) and the uncertainty at zero level concentration ("b").

Table 4: Standard deviation for proficiency assessment (σ_{pt}).

Gas	a	b (nmol/mol)
SO ₂	0.022	1
CO	0.024	100
O ₃	0.020	1
NO	0.024	1
NO ₂	0.028	1.4

Source: JRC 2024

σ_{pt} is a linear function of concentration (c) with parameters in Table 4 identified “**a**” as slope and “**b**” as intercept.

Equation 3

$$\sigma_{pt} = (a \cdot X_{ref}) + b$$

σ_{pt} = Standard deviation for proficiency assessment

a = slope see table 4

X_{ref} = reference value

b = intercept see table 4

The parameters in table 4 have been discussed and agreed with the AQUILA network through a document named N37 [12].

The assessment of results in z/z' -score evaluation is made according to the following criteria:

$|z/z'| \leq 2$ are considered satisfactory.

$2 < |z/z'| < 3$ are considered questionable.

$|z/z'| \geq 3$ are considered unsatisfactory. Scores falling in this range are very unusual and are taken as evidence that an anomaly has occurred that should be investigated and corrected.

According to z/z' -score calculation, values between 2 and 3 are considered questionable and they deserve a specific check.

In this report, within all the results, many values were found unsatisfactory or questionable as reported in table 5.

The overall performance of all participants related to the z/z' -score parameters is shown in Figures 1 to 5 as colored graphical matrix where on the x axis are shown the runs (concentration levels) and the y-axis the participants are plotted. The figures, through a bi-dimensional graphical representation, identifies, with different colors according to the criteria explained above, the participant's values of z/z' -score for each concentration level analysed during the PT.


The values of z/z' -score calculated for each participant are reported in the matrix cells.

Table 5: Unsatisfactory/questionable results according to z/z'-score.

Laboratory	Component	Run	z/z'-score value	z/z'-score evaluation
APA	CO	2	-2.1	questionable
ACES	NO	1	-5.6	unsatisfactory
APA	NO	1	-2.1	questionable
ACES	NO	2	-5.4	unsatisfactory
APA	NO	2	-2.0	questionable
ACES	NO	3	-3.3	unsatisfactory
ACES	NO	4	-2.4	questionable
ACES	NO	5	-4.0	unsatisfactory
LVGMC	NO	5	-2.4	questionable
ACES	NO	6	-3.4	unsatisfactory
ACES	NO	7	-3.2	unsatisfactory
LVGMC	NO	7	-2.4	questionable
ACES	NO	8	-3.8	unsatisfactory
LVGMC	NO	8	-2.5	questionable
ACES	NO	9	-4.2	unsatisfactory
LVGMC	NO	9	-3.0	unsatisfactory
ACES	NO	10	-4.3	unsatisfactory
LVGMC	NO	10	-2.7	questionable
INERIS	NO ₂	1	-2.3	questionable
INERIS	NO ₂	2	-3.0	unsatisfactory
APA	NO ₂	7	-2.1	questionable
EPA	NO ₂	7	-5.0	unsatisfactory
GIOS	NO ₂	7	-3.0	questionable
INERIS	NO ₂	7	-4.2	unsatisfactory
EPA	NO ₂	8	-2.1	questionable
EPA	NO ₂	9	-4.0	unsatisfactory
INERIS	NO ₂	9	-2.5	questionable
EPA	NO ₂	10	-2.0	questionable
ACES	O ₃	0	-8.2	unsatisfactory
ACES	O ₃	1	-2.2	questionable
ACES	O ₃	2	-5.1	unsatisfactory
APA	SO ₂	1	-2.9	questionable
LVGMC	SO ₂	1	3.8	unsatisfactory
LVGMC	SO ₂	2	2.2	questionable
APA	SO ₂	5	-2.0	questionable
LVGMC	SO ₂	5	2.9	questionable

Source: JRC 2024

Figure 1: Graphical matrix of z/z'-score evaluations of CO measurements in $\mu\text{mol/mol}$



Source: JRC 2024

Underestimated unsatisfactory results are marked in red, in yellow underestimated questionable results, in green satisfactory results, in orange overestimated questionable results and in brown overestimated unsatisfactory results.


z/z'-score values are given for each participant and each tested concentration level (run).

For this pollutant, the graphical representation shows a general good performance of every participant for all concentration levels generated.

ACES laboratory didn't participate to the measurement of this pollutant.

Only 1 measurement of APA was found questionable.

Figure 2: Graphical matrix of z/z'-score evaluations of NO measurements in nmol/mol



Source: JRC 2024

Underestimated unsatisfactory results are marked in red, in yellow underestimated questionable results, in green satisfactory results, in orange overestimated questionable results and in brown overestimated unsatisfactory results.


z/z'-score values are given for each participant and each tested concentration level (run).

The graphical depiction for this pollutant indicates an overall satisfactory performance, with the exception of three participants.

LVGMC recorded three results that raised concerns and one that was deemed unsatisfactory, ACES exhibited generally inferior performance across most concentration levels. A thorough investigation into the issue should be initiated, beginning with a verification of the standard gas utilized in the calibration process.

Regarding APA, there were two results that were considered questionable.

Figure 3: Graphical matrix of z/z'-score evaluations of NO₂ measurements in nmol/mol



Source: JRC 2024

Underestimated unsatisfactory results are marked in red, in yellow underestimated questionable results, in green satisfactory results, in orange overestimated questionable results and in brown overestimated unsatisfactory results.

z/z'-score values are given for each participant and each tested concentration level (run).

For this pollutant, the graphical representation shows the majority of results as satisfactory for all concentration levels generated.


Only one measurement of APA was found questionable.

For EPA two values questionable and two unsatisfactory were identified for this pollutant measurement.

Only one measurement of GIOS was found questionable.

For INERIS two values questionable and two unsatisfactory were identified for this pollutant measurement.

Figure 4: Graphical matrix of z/z'-score evaluations of O₃ measurements in nmol/mol



Source: JRC 2024


Underestimated results are highlighted in red for unsatisfactory outcomes and in yellow for questionable ones. Conversely, results meeting expectations are displayed in green, while those that exceeded expectations in a questionable manner are in orange, and clearly excessive unsatisfactory results are in brown.

z/z'-score values are given for each participant and each tested concentration level (run).

The visual depiction of the data indicates that nearly all participants performed well across all generated concentration levels for this pollutant. However, it should be noted that ACES yielded one questionable result and two unsatisfactory results.

Ecomaks laboratory didn't participate in measuring this particular pollutant.

Figure 5: Graphical matrix of z/z'-score evaluations of SO₂ measurements in nmol/mol



Source: JRC 2024

Underestimated unsatisfactory results are marked in red, in yellow underestimated questionable results, in green satisfactory results, in orange overestimated questionable results and in brown overestimated unsatisfactory results.

z/z'-score values are given for each participant and each tested concentration level (run).

ACES laboratory didn't participate to the measurement of this pollutant.

The graphical data for this pollutant illustrates that the majority of participants achieved satisfactory results across all the concentration levels produced. However, APA reported two questionable value and LVGMC, two of their recorded values were deemed questionable and one was considered unsatisfactory.

5.2 En-score

In order to evaluate the participant's ability to have results close to the reference values within their reported uncertainties, the En-score parameters (En) were calculated according to:

$$\text{Equation 4}$$

$$En = \frac{x_{lab} - X_{ref}}{\sqrt{U_{lab}^2 + U_{ref}^2}}$$

En = En-score

x_{lab} = participant average values

X_{ref} = reference value

U_{lab} = expanded uncertainty of the participants

U_{ref} = expanded uncertainty of the reference value

The overall performance of all participants related to the En-score parameters is shown in Figures 6 to 15 as colored graphical matrix where on the x-axis are shown the run (concentration levels) and the y-axis the participants are plotted. The figures, through a bi-dimensional graphical representation, identifies, with different colors, the participant's values of En-score for each concentration level analysed during the PT.

Results with |En-score|<1 are considered satisfactory all the others are unsatisfactory.

The values of En-score calculated for each participant during the PT are reported in the matrix cells. At the bottom of each figure, the legend explains with different colors the identification of the results: in red are marked underestimated unsatisfactory results, in green satisfactory results and in brown overestimated unsatisfactory results. The En evaluations presented in Figures 6, 8, 10, 12, and 14 highlight some unsatisfactory results for different parameters and concentrations, as reported in table 6.

In Annex B, from table 11 to 20 are reported all En values calculated for each participant, parameter and concentration level.

Table 6: Unsatisfactory results according to En-score.

Laboratory	Component	Run	En-score	En evaluation
APA	CO	0	-7.9	unsatisfactory
HMS	CO	0	-1.8	unsatisfactory
APA	CO	1	-7.4	unsatisfactory
APA	CO	2	-3.1	unsatisfactory
HMS	CO	2	-1.5	unsatisfactory
APA	CO	3	-4.0	unsatisfactory
HMS	CO	3	-1.5	unsatisfactory
APA	CO	4	-3.7	unsatisfactory
EPA	CO	4	-1.0	unsatisfactory
HMS	CO	4	-1.7	unsatisfactory
APA	CO	5	-2.5	unsatisfactory
HMS	CO	5	-1.7	unsatisfactory
INERIS	CO	5	-1.1	unsatisfactory
ACES	NO	1	-9.9	unsatisfactory
APA	NO	1	-3.1	unsatisfactory
ACES	NO	2	-9.7	unsatisfactory
APA	NO	2	-3.0	unsatisfactory
ACES	NO	3	-3.5	unsatisfactory
APA	NO	3	-1.4	unsatisfactory
ACES	NO	4	-2.2	unsatisfactory
APA	NO	4	-1.1	unsatisfactory
ACES	NO	5	-5.4	unsatisfactory
APA	NO	5	-1.5	unsatisfactory

Laboratory	Component	Run	En-score	En evaluation
LVGMC	NO	5	-1.4	unsatisfactory
ACES	NO	6	-3.9	unsatisfactory
APA	NO	6	-1.1	unsatisfactory
LVGMC	NO	6	-1.3	unsatisfactory
ACES	NO	7	-5.4	unsatisfactory
APA	NO	7	-1.7	unsatisfactory
EPA	NO	7	-1.0	unsatisfactory
ACES	NO	8	-6.5	unsatisfactory
APA	NO	8	-1.9	unsatisfactory
ARSO	NO	8	1.1	unsatisfactory
EPA	NO	8	-1.0	unsatisfactory
LVGMC	NO	8	-1.0	unsatisfactory
ACES	NO	9	-7.3	unsatisfactory
APA	NO	9	-2.0	unsatisfactory
EPA	NO	9	-1.1	unsatisfactory
LVGMC	NO	9	-1.3	unsatisfactory
ACES	NO	10	-7.5	unsatisfactory
APA	NO	10	-2.0	unsatisfactory
EPA	NO	10	-1.0	unsatisfactory
LVGMC	NO	10	-1.2	unsatisfactory
EPA	NO ₂	1	-1.2	unsatisfactory
INERIS	NO ₂	1	-2.0	unsatisfactory
EPA	NO ₂	2	-1.9	unsatisfactory
INERIS	NO ₂	2	-2.9	unsatisfactory
EPA	NO ₂	5	-1.2	unsatisfactory
ACES	NO ₂	6	-2.0	unsatisfactory
EPA	NO ₂	6	-1.5	unsatisfactory
APA	NO ₂	7	-1.3	unsatisfactory
EPA	NO ₂	7	-3.3	unsatisfactory
GIOS	NO ₂	7	-1.9	unsatisfactory
INERIS	NO ₂	7	-2.7	unsatisfactory
ACES	NO ₂	8	-1.6	unsatisfactory
APA	NO ₂	8	-1.5	unsatisfactory
EPA	NO ₂	8	-2.1	unsatisfactory
INERIS	NO ₂	8	-1.5	unsatisfactory
APA	NO ₂	9	-1.2	unsatisfactory
EPA	NO ₂	9	-3.1	unsatisfactory
GIOS	NO ₂	9	-1.4	unsatisfactory
INERIS	NO ₂	9	-1.9	unsatisfactory
ACES	NO ₂	10	-2.9	unsatisfactory
APA	NO ₂	10	-1.7	unsatisfactory
EPA	NO ₂	10	-2.2	unsatisfactory
INERIS	NO ₂	10	-1.3	unsatisfactory
ACES	O ₃	0	-3.4	unsatisfactory
ACES	O ₃	1	-1.9	unsatisfactory
INERIS	O ₃	1	1.1	unsatisfactory
ACES	O ₃	2	-3.1	unsatisfactory
ACES	O ₃	3	1.5	unsatisfactory
INERIS	O ₃	4	1.0	unsatisfactory
INERIS	O ₃	5	1.0	unsatisfactory
APA	SO ₂	1	-2.4	unsatisfactory
LVGMC	SO ₂	1	1.6	unsatisfactory
APA	SO ₂	2	-1.9	unsatisfactory
LVGMC	SO ₂	2	1.4	unsatisfactory

Laboratory	Component	Run	En-score	En evaluation
APA	SO ₂	4	-1.8	unsatisfactory
LVGMC	SO ₂	4	1.3	unsatisfactory
APA	SO ₂	5	-2.1	unsatisfactory
LVGMC	SO ₂	5	1.6	unsatisfactory

Source: JRC 2024


The differences between individual participants' values and reference values (bias $X_{lab}-X_{ref}$) are evaluated and presented for each concentration level in the Figures 7, 9, 11, 13, 15.

The expanded uncertainties reported by participants are shown in the graphs for each run as error bar.

These plots represent also the En-score evaluations where, considering the En criterion ($|En-score| < 1$), all results are satisfactory if the error bars cross the red line representing the reference value.


Reported standard uncertainties of the participants, larger than the "standard deviation for proficiency assessments" (σ_{pt}), are considered not fit-for-purpose (see Annex B).

Figure 6: En-score for CO measurements in matrix view



Source: JRC 2024

Figure 7: Bias of participant's CO measurement results in $\mu\text{mol/mol}$




Source: JRC 2024

ACES laboratory didn't participate to the measurement of this pollutant.


During the En-score evaluation of CO, only 1 measurement of EPA and INERIS was found above the limit while for HMS all values beside one are unsatisfactory and for APA all values are out of the limits. APA reported a small uncertainty and general lower results compared to the assigned value as shown in Fig.7.

Figure 8: En-score for NO measurements in nmol/mol in matrix view



Source: JRC 2024


Figure 9: Bias of participant's NO measurement results in nmol/mol



Source: JRC 2024


For this pollutant, four laboratories displayed numerous unsatisfactory values during the En-score evaluation. ACES and APA underestimated the measurements in almost all runs except for the zero level. ARSO reported a value that was slightly above the limit. EPA slightly underestimated four values. Five of the values submitted by LVGMC were found to be unsatisfactory.

Figure 12: En-score for O₃ measurements in nmol/mol in matrix view



Source: JRC 2024

Figure 13: Bias of participant's O₃ measurement results in nmol/mol




Source: JRC 2024

Ecomaks did not participate to the measurement of this pollutant.


During the En-score evaluation, for this pollutant, both graphical representations show a general good performance of almost every participant, only ACES underestimated three values and overestimated one value. Three results of INERIS were slightly found overestimated.

Figure 14: En-score for SO₂ measurements in nmol/mol in matrix view



Source: JRC 2024

Figure 15: Bias of participant's SO₂ measurement results in nmol/mol



Source: JRC 2024

During the En-score evaluation, the graphical data for this pollutant illustrates that the majority of participants achieved satisfactory results across all the concentration levels produced. In the case of LVGMC, four of their recorded values were found unsatisfactory while APA submitted four values that were found underestimated.

6 Conclusions

The proficiency evaluation scheme has assessed the participants' measured values and their evaluated uncertainties through calculating two performance indicators (z/z' -score and En-score).

z/z' -scores calculation is used to estimate the participant performance.

Statistically unsatisfactory results obtained at this stage are not attributed to extraordinary errors, but rather to significant differences in participants' standard operating procedures.

The precision of standardised measurement methods reported in Annex C are calculated using the whole data pool submitted.

In this exercise **89.2%** of the results in the z/z' -score evaluations are satisfactory.

Regarding the En-score parameter **78.8%** of the results are found satisfactory and **21.2%** unsatisfactory according to the limits explained in paragraph 5.2.

As in previous PT, the adopted evaluation for high concentrations were the standard deviations for proficiency assessment, deriving from the European Standards' uncertainty requirements.

The reproducibility standard deviation obtained and described in Annex C, like previous PT [20], [21], [22], [23], [24], [25], [33], [34], [35], [36], [37], [38], [39], [40], [41], [43], [44], [45], [46], [47], [48], [49], [50], [51], [52], [53], [54] and [55] is comparable to the mentioned criteria. On the other hand, the uncertainty criteria for zero levels were those set in AQUILA's position paper [12].

The following figure 16 is reporting a histogram representation of the overview of the performance indicator z/z' -score results obtained in this PT. All charts indicate the results of each participant for every component analysed at all concentrations generated.

Generally, the overall performance of all laboratories is good and no unsatisfactory results are identified for z/z' -score indicator while a small number of values were found unsatisfactory for the En-score.

The result of this PT is in line with the performances of previous years as shown by the following Table 7 where the results of z/z' -score in all PT organized in JRC Ispra site from 2005 till 2024 are presented.

Figure 16: Overview on z/z'-score final evaluation

z/z'-score overview



Source: JRC 2024

Table 7: z/z'-score summary results of PT organized in Ispra from 2005 to 2024.

PT	Site	Satisfactory (%)	Questionable (%)	Unsatisfactory (%)
Jun_05	Ispra (IT)	94.7	2.3	3.0
Jun_07	Ispra (IT)	97.8	1.9	0.30
Apr_08	Ispra (IT)	93.8	2.1	4.1
Oct_08_I	Ispra (IT)	92.9	4.2	2.9
Oct_08_II	Ispra (IT)	97.0	3.0	0.00
Oct_09	Ispra (IT)	98.2	1.8	0.00
Jun_10	Ispra (IT)	97.0	3.0	0.00
Sep_11	Ispra (IT)	99.4	0.30	0.30
Oct_11	Ispra (IT)	98.7	1.3	0.00
Jun_12	Ispra (IT)	100	0.00	0.00
Sep_13	Ispra (IT)	100	0.00	0.00
Oct_13	Ispra (IT)	99.3	0.70	0.00
May_14	Ispra (IT)	98.1	0.70	1.1
Oct_15_I	Ispra (IT)	99.4	0.60	0.00
Oct_15_II	Ispra (IT)	93.7	4.1	2.2
Jun_16	Ispra (IT)	100	0.00	0.00
Jun_17_I	Ispra (IT)	98.9	0.70	0.40
Jun_17_II	Ispra (IT)	96.2	1.9	1.9
Jun_18	Ispra (IT)	100	0.00	0.00
May_19_I	Ispra (IT)	98.7	1.3	0.00
May_19_II	Ispra (IT)	97.5	2.5	0.00
Mar-22(I)	Ispra (IT)	97.2	1.1	1.7
Mar_22_II	Ispra (IT)	99.3	0.70	0.00
Apr_22_III	Ispra (IT)	91.5	3.2	5.3
Mar_23	Ispra (IT)	100	0.00	0.00
Mar_24_I	Ispra (IT)	98.2	1.8	0.00
Mar_24_II	Ispra (IT)	89.2	5.7	5.1

Source: JRC 2024

During this PT, in agreement with all participants, an extra set of values has been asked to be reported. In order to verify the impact of the interference of NO/NO₂ during SO₂ measurement, all participants quantified sulphur dioxide while, in the sampling tube, was generated high concentration of NO and NO₂.

In table 8 are summarised the readings of all laboratories and the correspondent instrument used. In annex E of the EN 14212 standard is described the type approval required for the instrument used for SO₂ analysis. In table E.1 of this document are listed the relevant performance characteristics and criteria for SO₂ measurement in presence of NO and NO₂ (for NO with concentration 500 nmol/mol SO₂ reading ≤ 5.0 nmol/mol, for NO₂ with concentration 200 nmol/mol SO₂ reading ≤ 5.0 nmol/mol). This test is pure indicative and doesn't have the goal of reaching any conclusion about any of the instrumentation used.

Table 8: SO₂ measurement during NO gas generation.

Run	NO ref value	ERLAP1	ERLAP2	INERIS	EPA	GIOS	HMS	NILU	LVGMC	ARSO	APA	ECOMAKS
SO ₂ mol/mol												
7	472	2.9	2.9	1.0	8.3	3.5	2.1	2.8	1.3	2.7	2.2	2.3
7	473	2.9	2.9	1.1	8.3	3.5	2.1	2.8	1.3	2.7	2.1	2.1
7	473	2.9	2.8	1.2	8.3	3.5	2.0	2.8	1.5	2.8	2.2	2.3
8	250	1.7	1.8	0.51	4.4	2.1	0.80	1.3	0.70	1.4	1.00	1.8
8	251	1.7	1.7		4.4	2.0	0.90	1.3	0.76	1.4	0.90	2.0
8	251	1.6	1.7	0.29	4.4	2.0	0.90	1.4	0.81	1.4	0.90	1.9
9	296	1.9	2.0	0.68	5.2	2.2	1.1	1.7	0.94	1.7	1.1	2.0
9	296	1.9	2.0	0.52	5.1	2.3	1.1	1.7	0.89	1.7	1.1	2.2
9	297	1.9	2.0	0.43	5.1	2.3	1.1	1.7	0.80	1.7	1.2	2.0
10	163	1.1	1.4	0.21	2.8	1.3	0.20	0.80	0.43	0.90	0.40	1.3
10	164	1.1	1.4	-0.06	2.8	1.3	0.40	0.80	0.48	0.90	0.50	1.1
10	164	1.1	1.3	0.04	2.8	1.3	0.30	0.80	0.53	0.90	0.40	1.2
		Thermo FS 43i/TLE	Horiba APSA-370	Teledyne API T100	Teledyne API T100	Horiba APSA-370	Thermo FS 43i	Horiba APSA-370	Teledyne API T100	Horiba APSA 370	Thermo FS 43i	Teledyne API T100

Source: JRC 2024

References

1. EC Directive 2008/50/EC of the European Parliament and of the Council of 21 May 2008 on ambient air quality and cleaner air for Europe, L 152, 11.06.2008
2. CEN, EN 14626:2012, ambient air quality - Standard method for the measurement of the concentration of carbon monoxide by non-dispersive infrared spectroscopy
3. CEN, EN 14212:2012, ambient air quality - Standard method for the measurement of the concentration of sulphur dioxide by ultraviolet fluorescence
4. CEN, EN 14211:2012, ambient air quality - Standard method for the measurement of the concentration of nitrogen dioxide and nitrogen monoxide by chemiluminescence
5. CEN, EN 14625:2012, ambient air quality - Standard method for the measurement of the concentration of ozone by ultraviolet photometry
6. ISO 6143:2001, Gas analysis - Comparison methods for determining and checking the composition of calibration gas mixtures
7. ISO 6144:2003, Gas analysis - Preparation of calibration gas mixtures - Static volumetric method
8. ISO 6145-7:2001, Gas analysis - Preparation of calibration gas mixtures using dynamic volumetric methods - Part 7: Thermal mass-flow controllers
9. Mücke H.-G., Air quality management in the WHO European Region - Results of a quality assurance and control programme on air quality monitoring (1994-2004), Environment International, EI-01718. 2008
10. Mücke H.-G. et al., European Intercomparison workshop on air quality monitoring vol.4 - Measuring NO, NO₂, O₃ and SO₂ - Air Hygiene Report 13, WHO Collaboration Centre for Air Quality Management and Air Pollution Control, ISSN 0938 - 9822. 2000
11. AQUILA Network of the National Reference Laboratories of the Member States [online]: Info available: <https://ec.europa.eu/jrc/en/aquila>.
12. AQUILA POSITION PAPER N. 37, Protocol for intercomparison exercise. Organisation of intercomparison exercises for gaseous air pollution for EU national air quality reference laboratories and laboratories of the WHO EURO region, 2008 Available: https://ec.europa.eu/jrc/sites/jrcsh/files/aquila-n_37-intercomparison-exercise_protocol_2008.pdf
13. ISO 13528:2022, Statistical methods for use in proficiency testing by inter-laboratory comparison
14. ISO 5725-1:1994, Accuracy (trueness and precision) of measurement methods and results - Part 1: General principles and definitions
15. ISO 5725-2:2019, Accuracy (trueness and precision) of measurement methods and results - Part 2: Basic method for the determination of repeatability and reproducibility of a standard measurement method
16. ISO 5725-6:1994, Accuracy (trueness and precision) of measurement methods and results - Part 6: Use in practice of accuracy values
17. De Saeger E. et al., Harmonisation of Directive 92/72/EEC on air pollution by ozone, EUR 17662, 1997
18. De Saeger E. et al., European comparison of Nitrogen Dioxide calibration methods, EUR 17661, 1997
19. ISO 15337:2009, Ambient air - Gas phase titration - Calibration of analysers for ozone
20. Kapus M. et al. The evaluation of the Intercomparison Exercise for SO₂, CO, O₃, NO and NO₂ carried out in June 2007 in Ispra. JRC scientific and technical reports. EUR 23804. 2009
21. Kapus M. et al. The evaluation of the Intercomparison Exercise for SO₂, CO, O₃, NO and NO₂ - April 2008. JRC scientific and technical reports. EUR 23805. 2009
22. Kapus M. et al. The evaluation of the Intercomparison Exercise for SO₂, CO, O₃, NO and NO₂ 6-9 October 2008. JRC scientific and technical reports. EUR 23806. 2009
23. Kapus M. et al. The evaluation of the Intercomparison Exercise for SO₂, CO, O₃, NO and NO₂ 13-16 October 2008. JRC scientific and technical reports. EUR 23807. 2009

24. Belis C. A. et al. The evaluation of the Interlaboratory comparison Exercise for SO₂, CO, O₃, NO and NO₂ Langen 20-25 September 2009. EUR 24376. 2010
25. Belis C. A. et al. The evaluation of the Interlaboratory comparison Exercise for SO₂, CO, O₃, NO and NO₂ 19-22 October 2009. EUR 24476. 2010
26. Viallon J. et al. Final report, on-going key comparison BIPM.QM-K1: Ozone at ambient level, comparison with JRC, 2008. Metrologia 46 08017, 2009. doi: 10.1088/0026-1394/46/1A/08017
27. Viallon, J., et al. International comparison CCQM-P28: Ozone at ambient level, Metrologia 43, Tech. Suppl., 08010, 2006. doi:10.1088/0026-1394/43/1A/08010
28. Tanimoto, H., et al. Intercomparison of ultraviolet photometry and gas-phase titration techniques for ozone reference standards at ambient levels, Journal of Geophysical Research, vol. 111, D16313, 2006. doi:10.1029/2005JD006983
29. ISO, Guide to the expression of uncertainty in measurements, Geneva, 1995, ISBN 92-67-10188-9
30. VDI 2449 Part3: 2001, Measurement methods test criteria- General method for the determination of the uncertainty of calibratable measurement methods.
31. Mücke H-G, et al. European Intercomparison Workshops on Air Quality Monitoring. Vol. 2 – Measuring of CO, NO, NO₂ and O₃ – Air Hygiene Report 9. Berlin, Germany: WHO Collaborating Centre for Air Quality Management and Air Pollution Control; 1996. ISSN 0938-9822.
32. ISO 17043:2023, Conformity assessment - General requirements for the competence of proficiency testing providers
33. Barbieri M. et al. The evaluation of the Interlaboratory Comparison Exercise for SO₂, CO, O₃, NO and NO₂ Ispra 14-17 June 2010. EUR 24943. 2011
34. Barbieri M. et al. Evaluation of the Laboratory Comparison Exercise for SO₂, CO, O₃, NO and NO₂, 11th-14th June 2012 Ispra. EUR 25536. 2012
35. Barbieri M. et al. Evaluation of the Laboratory Comparison Exercise for SO₂, CO, O₃, NO and NO₂, Langen 23rd-28th October 2011. EUR 25387. 2012
36. Barbieri M. et al. Evaluation of the Laboratory Comparison Exercise for SO₂, CO, O₃, NO and NO₂, 03rd-06th October 2011 Ispra. EUR 25386. 2012
37. Barbieri M. et al. Evaluation of the Laboratory Comparison Exercise for SO₂, CO, O₃, NO and NO₂, 26th-29th September 2011 Ispra. EUR 25385. 2012
38. Barbieri M., Lagler F., Mücke H.G., Wirtz K. and Stummer V. Evaluation of the Laboratory Comparison Exercise for NO, NO₂, SO₂, CO, and O₃ Langen (D) 1st-6th September 2013. EUR 26578. 2014
39. Barbieri M., Lagler F., Evaluation of the Laboratory Comparison Exercise for SO₂, CO, O₃, NO and NO₂ 30st September-3rd October 2013 Ispra. EUR 26604. 2014
40. Barbieri M., Lagler F., Evaluation of the Laboratory Comparison Exercise for SO₂, CO, O₃, NO and NO₂ 7st-10th October 2013 Ispra. EUR 26639. 2014
41. Barbieri M., Lagler F., Evaluation of the Laboratory Comparison Exercise for SO₂, CO, O₃, NO and NO₂ 19th-22nd May 2014 Ispra. EUR 27199. 2014
42. EC COMMISSION DIRECTIVE (EU) 2015/1480 of 28 August 2015 (L226/4) amending several annexes to Directives 2004/107/EC and 2008/50/EC of the European Parliament and of the Council laying down the rules concerning reference methods, data validation and location of sampling points for the assessment of ambient air quality
43. Lagler F., Barbieri M., Borowiak A. Evaluation of the Laboratory Comparison Exercise for SO₂, CO, O₃, NO and NO₂ 12th-15th October 2015 Ispra. EUR 28097. 2016
44. Barbieri M., Lagler F., Borowiak A. Evaluation of the Laboratory Comparison Exercise for SO₂, CO, O₃, NO and NO₂ 19th-23rd October 2015 Ispra. EUR 28047. 2016
45. Barbieri M., Lagler F., Mücke H.G., Wirtz K. and Stummer V. Evaluation of the Laboratory Comparison Exercise for NO, NO₂, SO₂, CO, and O₃ Langen (D) 4th-9th October 2015. EUR 27918. 2016
46. Barbieri M., Lagler F., Borowiak A. Evaluation of the Laboratory Comparison Exercise for SO₂, CO, O₃, NO and NO₂ 13-16 June 2016, Ispra. EUR 28610. 2017

47. Barbieri M., Lagler F., Borowiak A. Evaluation of the Laboratory Comparison Exercise for SO₂, CO, O₃, NO and NO₂ 19-22 June 2017, Ispra. EUR 29268. 2017
48. Barbieri M., Lagler F., Borowiak A. Evaluation of the Laboratory Comparison Exercise for SO₂, CO, O₃, NO and NO₂ 26-29 June 2017, Ispra. EUR 29271. 2017
49. Barbieri M., Lagler F., Borowiak A. Evaluation of the Laboratory Comparison Exercise for SO₂, CO, O₃, NO and NO₂ 4-7 June 2018, Ispra. EUR 29671. 2018
50. Barbieri M., Lagler F., Mücke H.G., Wirtz K. and Stummer V. Evaluation of the Laboratory Comparison Exercise for NO, NO₂, SO₂, CO, and O₃ Langen (D) 2-7 September 2018. EUR 29694. 2018
51. Barbieri M., Lagler F., Borowiak A. Evaluation of the Inter-Laboratory Comparison exercise for SO₂, CO, O₃, NO and NO₂ (13-16 May 2019, Ispra). EUR 29898. 2019
52. Barbieri M., Lagler F., Borowiak A. Evaluation of the Inter-Laboratory Comparison exercise for SO₂, CO, O₃, NO and NO₂ (20-23 May 2019, Ispra), Ispra. EUR 29896. 2019
53. Barbieri M., Lagler F., Tarricone C., Borowiak A., PROFICIENCY TESTING SCHEME, Measurements of inorganic gaseous pollutants (SO₂, CO, O₃, NO and NO₂) in filtered ambient air, 21-24 March 2022 Ispra, Italy. EUR 31293 EN
54. Barbieri M., Lagler F., Tarricone C., Borowiak A., PROFICIENCY TESTING SCHEME, Measurements of inorganic gaseous pollutants (SO₂, CO, O₃, NO and NO₂) in filtered ambient air, 28-31 March 2022 Ispra, Italy. EUR 31349
55. Barbieri M., Lagler F., Tarricone C., Borowiak A., PROFICIENCY TESTING SCHEME, Measurements of inorganic gaseous pollutants (SO₂, CO, O₃, NO and NO₂) in filtered ambient air, 04-07 April 2022 Ispra, Italy. EUR 31348
56. Barbieri M., Lagler F., Tarricone C., Borowiak A., PROFICIENCY TESTING SCHEME, Measurements of inorganic gaseous pollutants (SO₂, CO, O₃, NO and NO₂) in filtered ambient air, 27-30 March 2023 Ispra, Italy. EUR 31694

List of abbreviations

AQUILA	Network of National Reference Laboratories for Air Quality
BIPM	Bureau International des Poids et Mesures
CEN	European Committee for Standardization
CO	Carbon monoxide
CRM	Certified Reference Material
DQO	Data Quality Objective
ERLAP	European Reference Laboratory for Air Pollution
EC	European Commission
GPT	Gas Phase Titration
PT	Proficiency Test
ISO	International Organization for Standardization
JRC	Joint Research Centre
NO	Nitrogen monoxide
NO ₂	Nitrogen dioxide
NO _x	The oxides of nitrogen, the sum of NO and NO ₂
NRL	National Reference Laboratory
O ₃	Ozone
PT	Proficiency Test
SO ₂	Sulphur dioxide
VDI	Verein Deutscher Ingenieure
WHO-CC	World Health Organization Collaborating Centre for Air Quality Management and Air Pollution Control, Berlin

Mathematical Symbols

α	converter efficiency (EN 14211)
E_c	Converter efficiency for NO analyser
E_n	En-score statistic (ISO 13528)
r	repeatability limit (ISO 5725)
R	reproducibility limit (ISO 5725)
σ_{pt}	standard deviation for proficiency assessment (ISO 13528)
x^*	robust average (Annex C ISO 13528)
s^*	robust standard deviation (Annex C ISO 13528)
s_R	estimate of reproducibility variance (ISO 5725)
U_{ref}	expanded uncertainty of the reference value (ISO 13528)
U_{lab}	expanded uncertainty of the participant's value (ISO 13528)
u_{ref}	standard uncertainty of the reference value (ISO 13528)
X_{ref}	reference value (ISO 13528)
x_{lab}	average of 3 values reported by the participant i (for each parameter and concentration level)
z'	z' -score statistic (ISO 13528)
z	z -score statistic (ISO 13528)

List of figures

Figure 1: Graphical matrix of z/z'-score evaluations of CO measurements in $\mu\text{mol/mol}$	14
Figure 2: Graphical matrix of z/z'-score evaluations of NO measurements in nmol/mol	15
Figure 3: Graphical matrix of z/z'-score evaluations of NO ₂ measurements in nmol/mol	16
Figure 4: Graphical matrix of z/z'-score evaluations of O ₃ measurements in nmol/mol	17
Figure 5: Graphical matrix of z/z'-score evaluations of SO ₂ measurements in nmol/mol	18
Figure 6: En-score for CO measurements in matrix view.....	22
Figure 7: Bias of participant's CO measurement results in $\mu\text{mol/mol}$	22
Figure 8: En-score for NO measurements in nmol/mol in matrix view.....	23
Figure 9: Bias of participant's NO measurement results in nmol/mol	23
Figure 10: En-score for NO ₂ measurements in nmol/mol in matrix view.....	24
Figure 11: Bias of participant's NO ₂ measurement results in nmol/mol	24
Figure 12: En-score for O ₃ measurements in nmol/mol in matrix view.....	25
Figure 13: Bias of participant's O ₃ measurement results in nmol/mol	25
Figure 14: En-score for SO ₂ measurements in nmol/mol in matrix view.....	26
Figure 15: Bias of participant's SO ₂ measurement results in nmol/mol	26
Figure 16: Overview on z/z'-score final evaluation.....	28
Figure 17: Graphical report of results of ACES (Sweden).....	43
Figure 18: Graphical report of results of APA (Portugal).....	45
Figure 19: Graphical report of results of ARSO (Slovenia).....	47
Figure 20: Graphical report of results of EPA (Ireland).....	49
Figure 21: Graphical report of results of Ecomaks (Serbia).....	51
Figure 22: Graphical report of results of GIOS (Poland).....	53
Figure 23: Graphical report of results of HMS (Hungary).....	55
Figure 24: Graphical report of results of INERIS (France).....	57
Figure 25: Graphical report of results of LVGMC (Latvia).....	59
Figure 26: Graphical report of results of NILU (Norway).....	61
Figure 27: R of CO standard measurement method as a function of concentration.....	65
Figure 28: R of NO standard measurement method as a function of concentration.....	66
Figure 29: R of NO ₂ standard measurement method as a function of concentration.....	67
Figure 30: R of O ₃ standard measurement method as a function of concentration.....	68
Figure 31: R of SO ₂ standard measurement method as a function of concentration.....	69

List of tables

Table 1: List of participating organizations.....	6
Table 2: List of instruments used by participants.....	7
Table 3: Sequence program of generated test gases with indicative pollutant concentrations.....	8
Table 4: Standard deviation for proficiency assessment (σ_{pt}).....	12
Table 5: Unsatisfactory/questionable results according to z/z'-score.....	13
Table 6: Unsatisfactory results according to En-score.....	19
Table 7: z/z'-score summary results of PT organized in Ispra from 2005 to 2024.....	29
Table 8: SO ₂ measurement during NO gas generation.....	30
Table 9: Reference values (X).....	39
Table 10: Homogeneity test.....	41
Table 11: Reported values, uncertainties and performance evaluation results of ACES (Sweden)...	44
Table 12: Reported values, uncertainties and performance evaluation results of APA (Portugal)....	46
Table 13: Reported values, uncertainties and performance evaluation results of ARSO (Slovenia). 48	
Table 14: Reported values, uncertainties and performance evaluation results of EPA (Ireland).....	50
Table 15: Reported values, uncertainties and performance evaluation results of Ecomaks (Serbia).	52
Table 16: Reported values, uncertainties and performance evaluation results of GIOS (Poland).....	54
Table 17: Reported values, uncertainties and performance evaluation results of HMS (Hungary)...	56
Table 18: Reported values, uncertainties and performance evaluation results of INERIS (France)...	58
Table 19: Reported values, uncertainties and performance evaluation results of LVGMC (Latvia)..	60
Table 20: Reported values, uncertainties and performance evaluation results of NILU (Norway)...	62
Table 21: Critical values of t used in the reproducibility (R) evaluation.....	63

Annex A. Reference values

The reference values of tested concentration levels (run) were derived, as an average of two set of instruments, from ERLAP's measurements, which are calibrated against the certified reference values of CRMs and are traceable to international standards. In this perspective the reference values are defined as described in the ISO 13528 [13].

To foster its reference function ERLAP is participating regularly to key comparisons of the Gas Analysis Working Group within the framework of BIPM's CCQM and it is yearly confirming the compliance to standard to ISO 17043, through an annual audit run by the competent Italian accreditation body (see annex E).

During this PT ERLAP's SO₂, CO and NO analysers were calibrated according to the methodology described in the standard ISO 6143 [6]. Reference gas mixtures were produced from the primary reference materials (produced and certified by NMi Van Swinden Laboratorium) by dynamic dilution method using mass flow controllers [8].

All flows were verified with a certified molbloc/molbox1 system. For O₃ measurements, the analysers were calibrated using the JRC SRP42 primary standard (constructed by NIST), which has been compared to BIPM primary standard [26]. The photometer absorption cross section uncertainty (1.06%) was included in the uncertainty budget [27], [28].

The calibration functions parameters and uncertainties associated were carried out using VDI 2449 Part3 [30]. The contribution to the uncertainty budget " u_{ref} " include repeatability, certified reference material standard uncertainty, drift and homogeneity. For extending calibration from the NO to NO₂ channel of NO_x analyser, the GPT test was performed to establish the efficiency of NO₂-converter and its contribution is added to the uncertainty.

In table 9 the reference values are reported together with their uncertainties.

Table 9: Reference values (X)

Component	Run	Ref value	Ref unc com	Ref exp unc	sigmaPT	Unit	z or z'
CO	0	-0.00	0.01	0.02	0.10	µmol/mol	z
CO	1	0.62	0.01	0.02	0.11	µmol/mol	z
CO	2	8.8	0.09	0.18	0.31	µmol/mol	z
CO	3	0.92	0.01	0.03	0.12	µmol/mol	z
CO	4	1.9	0.02	0.04	0.15	µmol/mol	z
CO	5	4.8	0.05	0.10	0.22	µmol/mol	z
NO	0	0.23	0.72	1.5	1.0	nmol/mol	z'
NO	1	200	1.3	2.6	5.8	nmol/mol	z
NO	2	191	1.2	2.5	5.6	nmol/mol	z
NO	3	31	0.75	1.5	1.7	nmol/mol	z'
NO	4	21	0.73	1.5	1.5	nmol/mol	z'
NO	5	60	0.80	1.6	2.4	nmol/mol	z'
NO	6	41	0.76	1.5	2.0	nmol/mol	z'
NO	7	473	2.6	5.2	12	nmol/mol	z
NO	8	251	1.5	3.1	7.0	nmol/mol	z
NO	9	296	1.7	3.5	8.1	nmol/mol	z
NO	10	164	1.1	2.3	4.9	nmol/mol	z
NO ₂	0	-0.04	0.74	1.5	1.4	nmol/mol	z'
NO ₂	1	0.42	1.0	2.0	1.4	nmol/mol	z'
NO ₂	2	9.9	1.0	2.0	1.7	nmol/mol	z'
NO ₂	3	0.13	0.76	1.5	1.4	nmol/mol	z'
NO ₂	4	9.4	0.76	1.5	1.7	nmol/mol	z'
NO ₂	5	0.06	0.78	1.5	1.4	nmol/mol	z'
NO ₂	6	29	0.80	1.6	2.2	nmol/mol	z'
NO ₂	7	7.6	1.8	3.6	1.6	nmol/mol	z'
NO ₂	8	228	2.1	4.2	7.8	nmol/mol	z
NO ₂	9	4.3	1.3	2.6	1.5	nmol/mol	z'
NO ₂	10	136	1.4	2.8	5.2	nmol/mol	z
O ₃	0	0.28	0.26	0.52	1.0	nmol/mol	z
O ₃	1	65	0.64	1.3	2.3	nmol/mol	z
O ₃	2	21	0.29	0.57	1.4	nmol/mol	z
O ₃	3	196	1.9	3.8	4.9	nmol/mol	z'
O ₃	4	96	0.92	1.8	2.9	nmol/mol	z'
O ₃	5	130	1.2	2.5	3.6	nmol/mol	z'
SO ₂	0	0.19	0.50	1.0	1.00	nmol/mol	z'
SO ₂	1	126	0.97	1.9	3.8	nmol/mol	z
SO ₂	2	41	0.58	1.2	1.9	nmol/mol	z'
SO ₂	3	4.8	0.64	1.3	1.1	nmol/mol	z'
SO ₂	4	20	0.53	1.1	1.4	nmol/mol	z'
SO ₂	5	56	0.64	1.3	2.2	nmol/mol	z

Source: JRC 2024

Homogeneity

This type of PT for inorganic gases foresees a generation in real time of gas mixture that is analysed immediately. The gas is not stored nor manipulated before the participant's measurement. In this condition, the stability and homogeneity verification of test item is not fully applicable.

ERLAP verify the homogeneity of the mixture gases, using a ozone mixture because is the most reactive gas. The readings of two identical analysers (Thermo Fisher Scientific 49i), calibrated against an SRP photometer primary standard, are compared, while an ozone concentration is running in the sampling line. The sampling line has 20 positions with 2 connection ports in each position. An analysers measures ozone at the port 20a, the last of the sampling line, and the other one sequentially in the other ports.

The test is considered acceptable when the absolute difference between the initial measurement and all the others is below 0.5% (during this PT equivalent to 2,18 nmol/mol). Assuming a rectangular distribution, it constitutes an uncertainty of 0.3% of each concentration level. This test is performed before each inorganic gas PT organised by ERLAP in Ispra.

The homogeneity results are reported in Table 10.

Table 10: Homogeneity test.

Ref Analyzer (port 20a)	Bench position under test	Diff. between analysers	<0.5%?
nmol/mol		nmol/mol	
436	20b	0.00	Yes
436	1a	-1.00	Yes
435	1b	0.00	Yes
436	2a	0.00	Yes
436	2b	0.50	Yes
437	3a	1.00	Yes
437	3b	1.00	Yes
436	4a	0.80	Yes
438	4b	0.30	Yes
437	5a	0.60	Yes
438	5b	0.40	Yes
438	6a	-1.1	Yes
438	6b	-1.00	Yes
438	7a	-1.00	Yes
438	7b	0.20	Yes
438	8a	-0.40	Yes
438	8b	-0.40	Yes
439	9a	0.40	Yes
439	9b	0.40	Yes
438	10a	-0.40	Yes
438	10b	-0.50	Yes
438	11a	-0.50	Yes
438	11b	0.50	Yes
439	12a	0.40	Yes
439	12b	0.40	Yes
439	13a	0.30	Yes
439	13b	0.30	Yes
439	14a	0.20	Yes
439	14b	0.20	Yes
438	15a	1.00	Yes
439	15b	0.40	Yes
439	16a	0.40	Yes
438	16b	-0.30	Yes
438	17a	1.7	Yes
439	17b	1.3	Yes
439	18a	0.30	Yes
438	18b	0.50	Yes
439	19a	1.3	Yes
439	19b	0.30	Yes
439	20a	0.70	Yes
439	20b	1.4	Yes
440	21a	1.00	Yes
439	21b	0.60	Yes
439	22a	-0.40	Yes
439	22b	0.20	Yes
439	22c	0.40	Yes
439	22d	-0.60	Yes
439	23a	-2.0	Yes
438	23b	1.00	Yes
439	23c	-1.1	Yes
438	23d	-0.50	Yes

Source: JRC 2024

Annex B: Results reported by participants

In this annex are reported all results submitted by the participants and they are presented both as graphs and tables.

In the graph, from the left and top down, for each pollutant is represented the submitted measurement result compared to the reference value, the z/z'-score and En-score calculated and compared to criteria explained in paragraph 5.

In the table are reported all values obtained from the data evaluation of this PT.

From the left "Runs" indicate all gas mixture concentrations generated. In the column "**Lab mean**" is shown the average of the 3 results (Step 1,2,3) representing a measurement of 30 minutes each that every laboratory is asked to report.


The combined and expanded uncertainties of the participant are listed in column "**Lab unc com**" and "**Lab unc exp**", expressed in mol/mol units.

"sigmaPT" column shows the results of the calculation described in paragraph 5.1 using the parameter in Table 4.

"Ref. Val." is showing the calculated reference values.

In column "z/z'-score" and "En-score" are listed the performance indicators values obtained by each participant. "z/z' selection" column indicates for each pollutant's run if it was used z or z' according to the criteria explained in paragraph 5.

Figure 17: Graphical report of results of ACES (Sweden).




Source: JRC 2024

Table 11: Reported values, uncertainties and performance evaluation results of ACES (Sweden).

ACES													
Component	Run	Lab mean	Step 1	Step 2	Step 3	Lab unc com	Lab unc exp	unit	sigmaPT	Ref. Val.	z/z'-score	z/z'-selection	En-score
NO	0	0.40	0.40			0.50	1.00	nmol/mol	1.0	0.23	0.14	z'	0.10
NO	1	168	168	168	167	1.00	2.0	nmol/mol	5.8	200	-5.6	z	-9.9
NO	2	160	160	160	161	0.96	1.9	nmol/mol	5.6	191	-5.4	z	-9.7
NO	3	24	24	24	25	0.50	1.00	nmol/mol	1.7	31	-3.3	z'	-3.5
NO	4	17	17	17	17	0.50	1.00	nmol/mol	1.5	21	-2.4	z'	-2.2
NO	5	50	50	50	50	0.50	1.00	nmol/mol	2.4	60	-4.0	z'	-5.4
NO	6	34	34	34	34	0.50	1.00	nmol/mol	2.0	41	-3.4	z'	-3.9
NO	7	433	432	433	433	2.6	5.2	nmol/mol	12	473	-3.2	z	-5.4
NO	8	224	224	224	224	1.3	2.7	nmol/mol	7.0	251	-3.8	z	-6.5
NO	9	262	262	262	263	1.6	3.1	nmol/mol	8.1	296	-4.2	z	-7.3
NO	10	142	142	142	143	0.85	1.7	nmol/mol	4.9	164	-4.3	z	-7.5
NO ₂	0	-0.11	-0.11			0.50	1.00	nmol/mol	1.4	-0.04	-0.04	z'	-0.04
NO ₂	1	2.2	2.4	2.2	2.0	0.50	1.00	nmol/mol	1.4	0.42	1.0	z'	0.78
NO ₂	2	10	10.0	11	10.0	0.50	1.00	nmol/mol	1.7	9.9	0.20	z'	0.18
NO ₂	3	0.41	0.38	0.42	0.42	0.50	1.00	nmol/mol	1.4	0.13	0.17	z'	0.15
NO ₂	4	7.8	7.7	7.8	8.0	0.50	1.00	nmol/mol	1.7	9.4	-0.88	z'	-0.88
NO ₂	5	0.32	0.38	0.23	0.34	0.50	1.00	nmol/mol	1.4	0.06	0.16	z'	0.14
NO ₂	6	25	25	25	25	0.50	1.00	nmol/mol	2.2	29	-1.6	z'	-2.0
NO ₂	7	9.5	9.9	9.4	9.2	0.50	1.00	nmol/mol	1.6	7.6	0.80	z'	0.51
NO ₂	8	220	221	220	220	1.3	2.6	nmol/mol	7.8	228	-1.00	z	-1.6
NO ₂	9	5.4	5.5	5.2	5.5	0.50	1.00	nmol/mol	1.5	4.3	0.56	z'	0.41
NO ₂	10	127	127	126	127	0.76	1.5	nmol/mol	5.2	136	-1.8	z	-2.9
O ₃	0	-8.0	-8.0			1.2	2.4	nmol/mol	1.0	0.28	-8.2	z	-3.4
O ₃	1	60	60	60	61	1.2	2.3	nmol/mol	2.3	65	-2.2	z	-1.9
O ₃	2	14	14	14	14	1.2	2.3	nmol/mol	1.4	21	-5.1	z	-3.1
O ₃	3	203	202	203	204	1.3	2.5	nmol/mol	4.9	196	1.3	z'	1.5
O ₃	4	96	96	96	96	1.2	2.3	nmol/mol	2.9	96	-0.09	z'	-0.10
O ₃	5	127	127	127	128	1.2	2.3	nmol/mol	3.6	130	-0.75	z'	-0.84

Source: JRC 2024

Figure 18: Graphical report of results of APA (Portugal).




Source: JRC 2024

Table 12: Reported values, uncertainties and performance evaluation results of APA (Portugal).

APA													
Component	Run	Lab mean	Step 1	Step 2	Step 3	Lab unc com	Lab unc exp	unit	sigmaPT	Ref. Val.	z/z'-score	z/z'-selection	En-score
CO	0	-0.20	-0.20			0.01	0.02	µmol/mol	0.10	-0.00	-2.0	z	-7.9
CO	1	0.40	0.39	0.40	0.41	0.01	0.02	µmol/mol	0.11	0.62	-1.9	z	-7.4
CO	2	8.1	8.1	8.1	8.1	0.06	0.12	µmol/mol	0.31	8.8	-2.1	z	-3.1
CO	3	0.78	0.78	0.78	0.78	0.01	0.02	µmol/mol	0.12	0.92	-1.1	z	-4.0
CO	4	1.7	1.7	1.7	1.7	0.01	0.03	µmol/mol	0.15	1.9	-1.3	z	-3.7
CO	5	4.5	4.5	4.5	4.6	0.03	0.07	µmol/mol	0.22	4.8	-1.4	z	-2.5
NO	0	-0.01	-0.01			0.49	0.97	nmol/mol	1.0	0.23	-0.19	z'	-0.14
NO	1	188	188	188	188	1.5	3.0	nmol/mol	5.8	200	-2.1	z	-3.1
NO	2	179	179	179	180	1.4	2.9	nmol/mol	5.6	191	-2.0	z	-3.0
NO	3	28	28	28	28	0.53	1.1	nmol/mol	1.7	31	-1.4	z'	-1.4
NO	4	19	19	19	19	0.51	1.00	nmol/mol	1.5	21	-1.2	z'	-1.1
NO	5	57	57	57	57	0.65	1.3	nmol/mol	2.4	60	-1.2	z'	-1.5
NO	6	39	39	39	39	0.57	1.1	nmol/mol	2.0	41	-1.00	z'	-1.1
NO	7	458	457	458	458	3.5	6.9	nmol/mol	12	473	-1.2	z	-1.7
NO	8	241	241	241	242	1.9	3.8	nmol/mol	7.0	251	-1.3	z	-1.9
NO	9	285	285	285	285	2.2	4.4	nmol/mol	8.1	296	-1.4	z	-2.0
NO	10	157	157	157	157	1.3	2.5	nmol/mol	4.9	164	-1.3	z	-2.0
NO ₂	0	0.13	0.13			0.49	0.97	nmol/mol	1.4	-0.04	0.11	z'	0.10
NO ₂	1	-1.2	-1.1	-1.1	-1.3	0.49	0.98	nmol/mol	1.4	0.42	-0.91	z'	-0.70
NO ₂	2	7.8	7.7	7.9	7.9	0.49	0.98	nmol/mol	1.7	9.9	-1.1	z'	-0.93
NO ₂	3	-0.10	-0.11	-0.09	-0.09	0.49	0.97	nmol/mol	1.4	0.13	-0.14	z'	-0.13
NO ₂	4	8.8	8.8	8.8	8.9	0.49	0.98	nmol/mol	1.7	9.4	-0.33	z'	-0.33
NO ₂	5	-0.34	-0.31	-0.35	-0.36	0.49	0.97	nmol/mol	1.4	0.06	-0.25	z'	-0.22
NO ₂	6	27	27	27	27	0.53	1.1	nmol/mol	2.2	29	-0.74	z'	-0.90
NO ₂	7	2.5	2.9	2.5	2.1	0.50	0.99	nmol/mol	1.6	7.6	-2.1	z'	-1.3
NO ₂	8	220	220	220	220	1.7	3.4	nmol/mol	7.8	228	-1.0	z	-1.5
NO ₂	9	1.1	1.4	0.90	1.00	0.49	0.98	nmol/mol	1.5	4.3	-1.6	z'	-1.2
NO ₂	10	130	130	130	130	1.1	2.2	nmol/mol	5.2	136	-1.1	z	-1.7
O ₃	0	0.59	0.59			0.12	0.23	nmol/mol	1.0	0.28	0.31	z	0.56
O ₃	1	65	64	65	65	1.5	3.0	nmol/mol	2.3	65	-0.31	z	-0.22
O ₃	2	21	21	21	21	0.50	1.00	nmol/mol	1.4	21	-0.19	z	-0.24
O ₃	3	190	189	190	190	4.4	8.8	nmol/mol	4.9	196	-1.2	z'	-0.68
O ₃	4	94	94	94	94	2.2	4.3	nmol/mol	2.9	96	-0.75	z'	-0.49
O ₃	5	127	127	127	127	2.9	5.9	nmol/mol	3.6	130	-0.83	z'	-0.50
SO ₂	0	-0.47	-0.47			0.33	0.67	nmol/mol	1.00	0.19	-0.59	z'	-0.54
SO ₂	1	115	115	115	115	2.0	3.9	nmol/mol	3.8	126	-2.9	z	-2.4
SO ₂	2	37	37	37	37	0.71	1.4	nmol/mol	1.9	41	-1.8	z'	-1.9
SO ₂	3	3.9	3.1	4.2	4.4	0.51	1.0	nmol/mol	1.1	4.8	-0.74	z'	-0.58
SO ₂	4	18	18	18	18	0.46	0.92	nmol/mol	1.4	20	-1.6	z'	-1.8
SO ₂	5	51	51	51	51	0.92	1.8	nmol/mol	2.2	56	-2.0	z	-2.1

Source: JRC 2024

Figure 19: Graphical report of results of ARSO (Slovenia).




Source: JRC 2024

Table 13: Reported values, uncertainties and performance evaluation results of ARSO (Slovenia).

ARSO													
Component	Run	Lab mean	Step 1	Step 2	Step 3	Lab unc com	Lab unc exp	unit	sigmaPT	Ref. Val.	z/z'-score	z/z' selection	En-score
CO	0	0.01	0.01			0.04	0.09	µmol/mol	0.10	-0.00	0.13	z	0.14
CO	1	0.64	0.64	0.64	0.64	0.05	0.09	µmol/mol	0.11	0.62	0.17	z	0.20
CO	2	8.8	8.8	8.8	8.8	0.15	0.31	µmol/mol	0.31	8.8	0.12	z	0.10
CO	3	0.94	0.94	0.94	0.94	0.05	0.09	µmol/mol	0.12	0.92	0.20	z	0.26
CO	4	1.9	1.9	1.9	1.9	0.06	0.11	µmol/mol	0.15	1.9	0.05	z	0.06
CO	5	4.8	4.8	4.8	4.8	0.09	0.18	µmol/mol	0.22	4.8	-0.14	z	-0.15
NO	0	-1.00	-1.00			0.67	1.3	nmol/mol	1.0	0.23	-0.99	z'	-0.63
NO	1	205	205	205	205	3.7	7.4	nmol/mol	5.8	200	0.83	z	0.61
NO	2	195	195	195	196	3.5	7.0	nmol/mol	5.6	191	0.82	z	0.62
NO	3	31	31	31	31	1.5	3.0	nmol/mol	1.7	31	0.24	z'	0.13
NO	4	21	21	21	21	1.5	3.0	nmol/mol	1.5	21	-0.00	z'	-0.00
NO	5	62	62	62	62	1.7	3.3	nmol/mol	2.4	60	0.72	z'	0.50
NO	6	42	42	42	42	1.6	3.1	nmol/mol	2.0	41	0.41	z'	0.25
NO	7	491	490	491	491	8.9	18	nmol/mol	12	473	1.4	z	0.96
NO	8	262	261	262	262	4.7	9.4	nmol/mol	7.0	251	1.6	z	1.1
NO	9	306	306	306	306	5.5	11	nmol/mol	8.1	296	1.2	z	0.83
NO	10	170	169	170	170	3.1	6.2	nmol/mol	4.9	164	1.2	z	0.92
NO ₂	0	-0.42	-0.42			0.60	1.2	nmol/mol	1.4	-0.04	-0.24	z'	-0.20
NO ₂	1	0.64	0.55	0.57	0.80	0.60	1.2	nmol/mol	1.4	0.42	0.13	z'	0.09
NO ₂	2	11	11	10.0	11	0.70	1.4	nmol/mol	1.7	9.9	0.37	z'	0.30
NO ₂	3	-0.33	-0.18	-0.31	-0.50	0.60	1.2	nmol/mol	1.4	0.13	-0.29	z'	-0.24
NO ₂	4	9.3	9.3	9.3	9.4	0.70	1.4	nmol/mol	1.7	9.4	-0.06	z'	-0.05
NO ₂	5	-0.40	-0.28	-0.53	-0.40	0.60	1.2	nmol/mol	1.4	0.06	-0.29	z'	-0.24
NO ₂	6	29	29	29	29	0.70	1.4	nmol/mol	2.2	29	0.11	z'	0.13
NO ₂	7	7.8	8.0	7.7	7.6	0.70	1.4	nmol/mol	1.6	7.6	0.08	z'	0.05
NO ₂	8	237	237	237	237	5.3	11	nmol/mol	7.8	228	1.1	z	0.75
NO ₂	9	3.8	3.9	3.9	3.7	0.70	1.4	nmol/mol	1.5	4.3	-0.23	z'	-0.15
NO ₂	10	140	141	140	140	3.1	6.2	nmol/mol	5.2	136	0.84	z	0.64
O ₃	0	0.61	0.61			0.85	1.7	nmol/mol	1.0	0.28	0.33	z	0.19
O ₃	1	67	67	67	67	1.2	2.4	nmol/mol	2.3	65	0.70	z	0.59
O ₃	2	22	22	22	22	0.83	1.7	nmol/mol	1.4	21	0.51	z	0.40
O ₃	3	198	198	198	198	4.3	8.6	nmol/mol	4.9	196	0.36	z'	0.20
O ₃	4	98	98	98	98	1.9	3.7	nmol/mol	2.9	96	0.56	z'	0.41
O ₃	5	132	132	132	132	2.7	5.4	nmol/mol	3.6	130	0.48	z'	0.31
SO ₂	0	0.09	0.09			0.86	1.7	nmol/mol	1.00	0.19	-0.09	z'	-0.05
SO ₂	1	133	133	133	132	3.7	7.4	nmol/mol	3.8	126	1.8	z	0.90
SO ₂	2	43	42	43	43	1.4	2.8	nmol/mol	1.9	41	1.1	z'	0.69
SO ₂	3	5.0	4.2	5.3	5.4	0.87	1.7	nmol/mol	1.1	4.8	0.09	z'	0.05
SO ₂	4	21	22	21	21	1.00	2.0	nmol/mol	1.4	20	0.56	z'	0.38
SO ₂	5	58	58	58	58	1.8	3.6	nmol/mol	2.2	56	1.1	z	0.64

Source: JRC 2024

Figure 20: Graphical report of results of EPA (Ireland).




Source: JRC 2024

Table 14: Reported values, uncertainties and performance evaluation results of EPA (Ireland).

EPA													
Component	Run	Lab mean	Step 1	Step 2	Step 3	Lab unc com	Lab unc exp	unit	sigma PT	Ref. Val.	z/z'-score	z/z' selection	En-score
CO	0	0.00	0.00	0.00	0.00	0.00	0.00	µmol/mol	0.10	-0.00	0.03	z	0.16
CO	1	0.60	0.60	0.60	0.60	0.01	0.01	µmol/mol	0.11	0.62	-0.18	z	-0.82
CO	2	8.7	8.7	8.7	8.7	0.10	0.19	µmol/mol	0.31	8.8	-0.20	z	-0.24
CO	3	0.90	0.90	0.90	0.90	0.01	0.02	µmol/mol	0.12	0.92	-0.12	z	-0.45
CO	4	1.8	1.9	1.8	1.8	0.02	0.04	µmol/mol	0.15	1.9	-0.41	z	-1.0
CO	5	4.8	4.8	4.8	4.8	0.05	0.01	µmol/mol	0.22	4.8	-0.14	z	-0.30
NO	0	0.30	0.30			0.00	0.01	nmol/mol	1.0	0.23	0.06	z'	0.05
NO	1	194	194	194	194	3.1	6.0	nmol/mol	5.8	200	-1.1	z	-0.95
NO	2	185	185	185	185	3.0	5.7	nmol/mol	5.6	191	-1.0	z	-0.93
NO	3	29	29	29	29	0.47	0.90	nmol/mol	1.7	31	-0.82	z'	-0.89
NO	4	20	20	20	20	0.32	0.62	nmol/mol	1.5	21	-0.60	z'	-0.63
NO	5	58	58	58	58	0.92	1.8	nmol/mol	2.4	60	-0.84	z'	-0.90
NO	6	40	40	40	40	0.63	1.2	nmol/mol	2.0	41	-0.53	z'	-0.58
NO	7	458	457	458	458	7.3	14	nmol/mol	12	473	-1.2	z	-1.0
NO	8	242	241	242	243	3.9	7.5	nmol/mol	7.0	251	-1.2	z	-1.0
NO	9	286	286	286	286	4.6	8.9	nmol/mol	8.1	296	-1.3	z	-1.1
NO	10	158	158	158	158	2.5	4.9	nmol/mol	4.9	164	-1.1	z	-1.0
NO ₂	0	0.30	0.30			0.00	0.01	nmol/mol	1.4	-0.04	0.22	z'	0.23
NO ₂	1	-2.1	-1.7	-2.2	-2.5	-0.03	-0.07	nmol/mol	1.4	0.42	-1.5	z'	-1.2
NO ₂	2	6.1	6.2	6.1	5.9	0.10	0.19	nmol/mol	1.7	9.9	-2.0	z'	-1.9
NO ₂	3	-0.72	-0.65	-0.75	-0.75	-0.01	-0.02	nmol/mol	1.4	0.13	-0.53	z'	-0.56
NO ₂	4	8.2	8.3	8.0	8.2	0.13	0.25	nmol/mol	1.7	9.4	-0.69	z'	-0.83
NO ₂	5	-1.8	-1.9	-1.8	-1.7	-0.03	-0.06	nmol/mol	1.4	0.06	-1.2	z'	-1.2
NO ₂	6	26	26	26	26	0.41	0.80	nmol/mol	2.2	29	-1.2	z'	-1.5
NO ₂	7	-4.6	-4.4	-4.8	-4.6	-0.07	-0.14	nmol/mol	1.6	7.6	-5.0	z'	-3.3
NO ₂	8	212	212	212	212	3.4	6.6	nmol/mol	7.8	228	-2.1	z	-2.1
NO ₂	9	-3.7	-3.9	-3.4	-3.7	-0.06	-0.11	nmol/mol	1.5	4.3	-4.0	z'	-3.1
NO ₂	10	125	125	125	126	2.0	3.9	nmol/mol	5.2	136	-2.0	z	-2.2
O ₃	0	0.15	0.10	0.15	0.20	0.00	0.01	nmol/mol	1.0	0.28	-0.12	z	-0.24
O ₃	1	65	64	65	65	1.3	2.7	nmol/mol	2.3	65	-0.31	z	-0.24
O ₃	2	21	21	21	21	0.43	0.87	nmol/mol	1.4	21	-0.19	z	-0.26
O ₃	3	199	199	199	200	4.0	8.2	nmol/mol	4.9	196	0.61	z'	0.36
O ₃	4	98	98	98	98	2.0	4.0	nmol/mol	2.9	96	0.56	z'	0.39
O ₃	5	133	132	133	133	2.7	5.4	nmol/mol	3.6	130	0.65	z'	0.42
SO ₂	0	-0.10	-0.10	-0.10	-0.10	-0.00	-0.00	nmol/mol	1.00	0.19	-0.26	z'	-0.29
SO ₂	1	128	127	128	128	2.6	5.2	nmol/mol	3.8	126	0.52	z	0.35
SO ₂	2	41	41	41	41	0.82	1.7	nmol/mol	1.9	41	0.28	z'	0.27
SO ₂	3	4.7	3.8	5.1	5.2	0.09	0.19	nmol/mol	1.1	4.8	-0.13	z'	-0.13
SO ₂	4	21	21	21	21	0.41	0.85	nmol/mol	1.4	20	0.18	z'	0.21
SO ₂	5	57	57	57	57	1.1	2.3	nmol/mol	2.2	56	0.54	z	0.45

Source: JRC 2024

Figure 21: Graphical report of results of Ecomaks (Serbia).




Source: JRC 2024

Table 15: Reported values, uncertainties and performance evaluation results of Ecomaks (Serbia).

Ecomaks													
Component	Run	Lab mean	Step 1	Step 2	Step 3	Lab unc com	Lab unc exp	unit	sigmaPT	Ref. Val.	z/z'-score	z/z'-selection	En-score
CO	0	0.10	0.10			0.09	0.18	μmol/mol	0.10	-0.00	1.0	z	0.57
CO	1	0.72	0.74	0.72	0.71	0.10	0.21	μmol/mol	0.11	0.62	0.89	z	0.48
CO	2	8.8	8.8	8.8	8.8	0.30	0.60	μmol/mol	0.31	8.8	0.12	z	0.06
CO	3	0.99	0.99	0.99	0.99	0.11	0.22	μmol/mol	0.12	0.92	0.61	z	0.34
CO	4	1.9	1.9	1.9	1.9	0.15	0.30	μmol/mol	0.15	1.9	0.05	z	0.02
CO	5	4.9	4.9	4.9	4.9	0.20	0.40	μmol/mol	0.22	4.8	0.32	z	0.17
NO	0	0.50	0.50			1.2	2.4	nmol/mol	1.0	0.23	0.22	z'	0.10
NO	1	196	197	196	196	7.0	14	nmol/mol	5.8	200	-0.66	z	-0.27
NO	2	186	186	186	186	6.5	13	nmol/mol	5.6	191	-0.85	z	-0.36
NO	3	30	30	30	30	1.5	3.0	nmol/mol	1.7	31	-0.29	z'	-0.16
NO	4	20	20	20	20	1.5	3.0	nmol/mol	1.5	21	-0.60	z'	-0.30
NO	5	58	58	58	58	2.5	5.0	nmol/mol	2.4	60	-0.84	z'	-0.41
NO	6	40	40	40	40	2.0	4.0	nmol/mol	2.0	41	-0.53	z'	-0.26
NO	7	457	457	457	457	14	29	nmol/mol	12	473	-1.3	z	-0.53
NO	8	242	242	242	242	8.0	16	nmol/mol	7.0	251	-1.2	z	-0.52
NO	9	283	283	283	284	9.5	19	nmol/mol	8.1	296	-1.6	z	-0.68
NO	10	157	157	157	156	5.5	11	nmol/mol	4.9	164	-1.4	z	-0.62
NO ₂	0	0.09	0.09			0.70	1.4	nmol/mol	1.4	-0.04	0.08	z'	0.07
NO ₂	1	0.98	0.95	1.00	1.00	0.75	1.5	nmol/mol	1.4	0.42	0.32	z'	0.22
NO ₂	2	10.0	10.0	10.0	10.0	1.5	3.0	nmol/mol	1.7	9.9	0.03	z'	0.02
NO ₂	3	0.29	0.34	0.34	0.19	0.71	1.4	nmol/mol	1.4	0.13	0.10	z'	0.08
NO ₂	4	9.2	9.1	9.2	9.3	1.0	2.1	nmol/mol	1.7	9.4	-0.13	z'	-0.09
NO ₂	5	0.11	0.10	0.14	0.10	0.70	1.4	nmol/mol	1.4	0.06	0.03	z'	0.03
NO ₂	6	28	28	28	28	2.0	4.0	nmol/mol	2.2	29	-0.31	z'	-0.17
NO ₂	7	7.2	7.7	7.1	6.9	1.00	2.0	nmol/mol	1.6	7.6	-0.14	z'	-0.08
NO ₂	8	220	221	220	220	9.0	18	nmol/mol	7.8	228	-1.00	z	-0.42
NO ₂	9	3.6	3.6	3.6	3.6	0.85	1.7	nmol/mol	1.5	4.3	-0.34	z'	-0.22
NO ₂	10	130	131	130	130	6.0	12	nmol/mol	5.2	136	-1.1	z	-0.46
SO ₂	0	0.54	0.54			0.66	1.3	nmol/mol	1.00	0.19	0.31	z'	0.21
SO ₂	1	119	118	119	119	4.5	9.0	nmol/mol	3.8	126	-1.9	z	-0.77
SO ₂	2	39	39	39	39	2.0	4.0	nmol/mol	1.9	41	-0.80	z'	-0.38
SO ₂	3	5.1	4.2	5.5	5.6	0.95	1.9	nmol/mol	1.1	4.8	0.20	z'	0.11
SO ₂	4	20	20	20	20	1.5	3.0	nmol/mol	1.4	20	-0.30	z'	-0.15
SO ₂	5	53	53	53	53	2.5	5.0	nmol/mol	2.2	56	-1.1	z	-0.50

Source: JRC 2024

Figure 22: Graphical report of results of GIOS (Poland).




Source: JRC 2024

Table 16: Reported values, uncertainties and performance evaluation results of GIOS (Poland).

GIOS													
Component	Run	Lab mean	Step 1	Step 2	Step 3	Lab unc com	Lab unc exp	unit	sigmaPT	Ref. Val.	z/z'-score	z/z'-selection	En-score
CO	0	0.00	0.00			0.00	0.01	µmol/mol	0.10	-0.00	0.04	z	0.19
CO	1	0.63	0.63	0.63	0.63	0.06	0.12	µmol/mol	0.11	0.62	0.08	z	0.07
CO	2	8.8	8.8	8.8	8.8	0.17	0.34	µmol/mol	0.31	8.8	0.12	z	0.10
CO	3	0.93	0.93	0.93	0.93	0.06	0.12	µmol/mol	0.12	0.92	0.12	z	0.12
CO	4	1.9	1.9	1.9	1.9	0.07	0.14	µmol/mol	0.15	1.9	0.05	z	0.05
CO	5	4.8	4.8	4.8	4.8	0.10	0.21	µmol/mol	0.22	4.8	-0.14	z	-0.13
NO	0	0.10	0.10			0.50	1.00	nmol/mol	1.0	0.23	-0.10	z'	-0.07
NO	1	201	201	201	201	3.7	7.3	nmol/mol	5.8	200	0.14	z	0.10
NO	2	191	191	191	191	3.5	7.0	nmol/mol	5.6	191	0.04	z	0.03
NO	3	30	30	30	30	0.79	1.6	nmol/mol	1.7	31	-0.29	z'	-0.25
NO	4	21	21	21	21	0.69	1.4	nmol/mol	1.5	21	-0.00	z'	-0.00
NO	5	59	59	59	59	1.2	2.4	nmol/mol	2.4	60	-0.45	z'	-0.40
NO	6	40	40	40	40	0.92	1.8	nmol/mol	2.0	41	-0.53	z'	-0.47
NO	7	480	480	480	480	8.7	17	nmol/mol	12	473	0.59	z	0.40
NO	8	251	251	251	251	4.6	9.1	nmol/mol	7.0	251	0.06	z	0.05
NO	9	298	298	298	298	5.4	11	nmol/mol	8.1	296	0.20	z	0.14
NO	10	163	163	163	163	3.0	6.0	nmol/mol	4.9	164	-0.12	z	-0.09
NO ₂	0	0.11	0.11			0.50	1.00	nmol/mol	1.4	-0.04	0.10	z'	0.09
NO ₂	1	0.22	0.23	0.24	0.20	0.50	1.00	nmol/mol	1.4	0.42	-0.11	z'	-0.09
NO ₂	2	9.9	9.8	9.9	9.9	0.61	1.2	nmol/mol	1.7	9.9	-0.03	z'	-0.03
NO ₂	3	0.09	0.16	0.11	0.00	0.50	1.00	nmol/mol	1.4	0.13	-0.03	z'	-0.02
NO ₂	4	9.4	9.4	9.3	9.4	0.61	1.2	nmol/mol	1.7	9.4	-0.04	z'	-0.04
NO ₂	5	0.01	0.02	0.00	0.01	0.50	1.00	nmol/mol	1.4	0.06	-0.03	z'	-0.03
NO ₂	6	29	29	29	29	0.83	1.7	nmol/mol	2.2	29	0.11	z'	0.12
NO ₂	7	0.30	0.41	0.28	0.21	0.50	1.00	nmol/mol	1.6	7.6	-3.0	z'	-1.9
NO ₂	8	229	229	229	229	4.8	9.5	nmol/mol	7.8	228	0.11	z	0.08
NO ₂	9	0.39	0.28	0.44	0.44	0.50	1.00	nmol/mol	1.5	4.3	-2.0	z'	-1.4
NO ₂	10	135	135	135	135	2.8	5.7	nmol/mol	5.2	136	-0.18	z	-0.15
O ₃	0	0.10	0.10			0.50	1.00	nmol/mol	1.0	0.28	-0.17	z	-0.16
O ₃	1	66	66	66	66	1.1	2.3	nmol/mol	2.3	65	0.27	z	0.23
O ₃	2	21	21	21	21	0.93	1.8	nmol/mol	1.4	21	-0.19	z	-0.14
O ₃	3	195	195	195	195	2.3	4.6	nmol/mol	4.9	196	-0.21	z'	-0.19
O ₃	4	96	96	96	96	1.4	2.8	nmol/mol	2.9	96	-0.09	z'	-0.09
O ₃	5	130	130	130	130	1.7	3.4	nmol/mol	3.6	130	-0.05	z'	-0.04
SO ₂	0	0.10	0.10			0.50	1.00	nmol/mol	1.00	0.19	-0.08	z'	-0.06
SO ₂	1	125	125	125	125	2.3	4.7	nmol/mol	3.8	126	-0.20	z	-0.15
SO ₂	2	40	40	40	40	0.93	1.9	nmol/mol	1.9	41	-0.29	z'	-0.26
SO ₂	3	5.2	5.1	5.2	5.3	0.59	1.2	nmol/mol	1.1	4.8	0.27	z'	0.20
SO ₂	4	21	21	21	21	0.69	1.4	nmol/mol	1.4	20	0.35	z'	0.31
SO ₂	5	56	56	56	56	1.2	2.3	nmol/mol	2.2	56	0.20	z	0.17

Source: JRC 2024

Figure 23: Graphical report of results of HMS (Hungary).




Source: JRC 2024

Table 17: Reported values, uncertainties and performance evaluation results of HMS (Hungary).

HMS													
Component	Run	Lab mean	Step 1	Step 2	Step 3	Lab unc com	Lab unc exp	unit	sigmaPT	Ref. Val.	z/z'-score	z/z'-selection	En-score
CO	0	-0.05	-0.05			0.01	0.01	µmol/mol	0.10	-0.00	-0.43	z	-1.8
CO	1	0.60	0.60	0.60	0.60	0.01	0.02	µmol/mol	0.11	0.62	-0.18	z	-0.74
CO	2	8.4	8.4	8.4	8.4	0.08	0.17	µmol/mol	0.31	8.8	-1.2	z	-1.5
CO	3	0.86	0.86	0.86	0.87	0.01	0.02	µmol/mol	0.12	0.92	-0.43	z	-1.5
CO	4	1.8	1.8	1.8	1.8	0.02	0.04	µmol/mol	0.15	1.9	-0.64	z	-1.7
CO	5	4.6	4.6	4.6	4.6	0.05	0.09	µmol/mol	0.22	4.8	-1.1	z	-1.7
NO	0	-0.86	-0.86			0.75	1.5	nmol/mol	1.0	0.23	-0.88	z'	-0.52
NO	1	201	201	201	201	2.0	4.1	nmol/mol	5.8	200	0.14	z	0.17
NO	2	192	192	192	192	2.0	3.9	nmol/mol	5.6	191	0.22	z	0.27
NO	3	31	31	30	31	0.80	1.6	nmol/mol	1.7	31	0.06	z'	0.05
NO	4	21	21	21	21	0.78	1.6	nmol/mol	1.5	21	-0.00	z'	-0.00
NO	5	60	60	60	60	0.94	1.9	nmol/mol	2.4	60	-0.06	z'	-0.06
NO	6	41	41	41	41	0.85	1.7	nmol/mol	2.0	41	-0.06	z'	-0.06
NO	7	474	474	474	474	4.5	9.0	nmol/mol	12	473	0.10	z	0.12
NO	8	252	251	252	252	2.5	5.0	nmol/mol	7.0	251	0.16	z	0.19
NO	9	298	298	298	298	2.9	5.8	nmol/mol	8.1	296	0.20	z	0.23
NO	10	165	164	165	165	1.7	3.4	nmol/mol	4.9	164	0.22	z	0.26
NO ₂	0	-0.10	-0.10			0.14	0.28	nmol/mol	1.4	-0.04	-0.03	z'	-0.04
NO ₂	1	0.78	0.91	0.73	0.70	0.14	0.28	nmol/mol	1.4	0.42	0.21	z'	0.17
NO ₂	2	10.0	10.0	10.0	10.0	0.22	0.45	nmol/mol	1.7	9.9	0.03	z'	0.03
NO ₂	3	0.12	0.11	0.14	0.11	0.14	0.28	nmol/mol	1.4	0.13	-0.01	z'	-0.01
NO ₂	4	9.5	9.5	9.5	9.5	0.22	0.43	nmol/mol	1.7	9.4	0.03	z'	0.04
NO ₂	5	0.06	0.13	0.09	-0.04	0.14	0.28	nmol/mol	1.4	0.06	0.00	z'	0.00
NO ₂	6	29	29	29	29	0.50	1.1	nmol/mol	2.2	29	0.11	z'	0.14
NO ₂	7	7.0	7.3	7.1	6.7	0.19	0.38	nmol/mol	1.6	7.6	-0.22	z'	-0.14
NO ₂	8	230	230	230	230	4.0	8.0	nmol/mol	7.8	228	0.24	z	0.21
NO ₂	9	3.7	3.8	3.7	3.5	0.15	0.31	nmol/mol	1.5	4.3	-0.31	z'	-0.24
NO ₂	10	137	137	137	137	2.4	4.8	nmol/mol	5.2	136	0.20	z	0.19
O ₃	0	0.47	0.47			0.30	0.60	nmol/mol	1.0	0.28	0.19	z	0.25
O ₃	1	65	65	65	66	0.94	1.9	nmol/mol	2.3	65	-0.02	z	-0.02
O ₃	2	21	21	21	21	0.41	0.83	nmol/mol	1.4	21	-0.19	z	-0.27
O ₃	3	196	196	196	197	2.7	5.4	nmol/mol	4.9	196	0.04	z'	0.03
O ₃	4	96	96	96	96	1.3	2.7	nmol/mol	2.9	96	-0.09	z'	-0.09
O ₃	5	130	130	130	130	1.8	3.6	nmol/mol	3.6	130	-0.05	z'	-0.04
SO ₂	0	-0.52	-0.52			0.39	0.78	nmol/mol	1.00	0.19	-0.63	z'	-0.55
SO ₂	1	123	123	123	123	3.6	7.1	nmol/mol	3.8	126	-0.73	z	-0.37
SO ₂	2	40	40	40	40	1.2	2.4	nmol/mol	1.9	41	-0.29	z'	-0.22
SO ₂	3	5.0	4.9	5.0	5.1	0.41	0.82	nmol/mol	1.1	4.8	0.12	z'	0.10
SO ₂	4	20	20	20	20	0.68	1.4	nmol/mol	1.4	20	-0.30	z'	-0.27
SO ₂	5	55	55	55	55	1.6	3.2	nmol/mol	2.2	56	-0.25	z	-0.16

Source: JRC 2024

Figure 24: Graphical report of results of INERIS (France).




Source: JRC 2024

Table 18: Reported values, uncertainties and performance evaluation results of INERIS (France).

INERIS													
Component	Run	Lab mean	Step 1	Step 2	Step 3	Lab unc com	Lab unc exp	unit	sigmaPT	Ref. Val.	z/z'-score	z/z' selection	En-score
CO	0	-0.08	-0.08			0.12	0.24	µmol/mol	0.10	-0.00	-0.82	z	-0.34
CO	1	0.48	0.48	0.48	0.48	0.11	0.22	µmol/mol	0.11	0.62	-1.2	z	-0.64
CO	2	8.7	8.7	8.7	8.7	0.14	0.28	µmol/mol	0.31	8.8	-0.20	z	-0.19
CO	3	0.76	0.76	0.75	0.76	0.11	0.22	µmol/mol	0.12	0.92	-1.3	z	-0.71
CO	4	1.7	1.7	1.7	1.7	0.10	0.20	µmol/mol	0.15	1.9	-1.3	z	-0.94
CO	5	4.6	4.6	4.6	4.6	0.09	0.18	µmol/mol	0.22	4.8	-1.1	z	-1.1
NO	0	0.31	0.31			1.2	2.4	nmol/mol	1.0	0.23	0.06	z'	0.03
NO	1	197	197	197	197	2.4	4.8	nmol/mol	5.8	200	-0.55	z	-0.58
NO	2	188	188	188	188	2.3	4.6	nmol/mol	5.6	191	-0.49	z	-0.53
NO	3	30	30	30	30	1.2	2.4	nmol/mol	1.7	31	-0.29	z'	-0.19
NO	4	21	21	21	21	1.2	2.4	nmol/mol	1.5	21	-0.00	z'	-0.00
NO	5	60	60	60	60	1.3	2.6	nmol/mol	2.4	60	-0.06	z'	-0.05
NO	6	41	41	41	41	1.2	2.5	nmol/mol	2.0	41	-0.06	z'	-0.04
NO	7	468	467	468	468	5.6	11	nmol/mol	12	473	-0.41	z	-0.41
NO	8	250	250	250	250	2.9	5.9	nmol/mol	7.0	251	-0.08	z	-0.08
NO	9	293	293	293	293	3.4	6.9	nmol/mol	8.1	296	-0.42	z	-0.44
NO	10	163	163	163	163	2.0	4.1	nmol/mol	4.9	164	-0.12	z	-0.13
NO ₂	0	0.28	0.28			0.00	0.01	nmol/mol	1.4	-0.04	0.21	z'	0.22
NO ₂	1	-3.6	-3.9	-3.6	-3.4	0.18	0.35	nmol/mol	1.4	0.42	-2.3	z'	-2.0
NO ₂	2	4.0	6.0	0.00	6.1	6.1	0.28	nmol/mol	1.7	9.9	-3.0	z'	-2.9
NO ₂	3	-0.12	-0.04	-0.07	-0.26	0.01	0.02	nmol/mol	1.4	0.13	-0.16	z'	-0.17
NO ₂	4	9.0	8.9	9.0	9.0	0.14	0.29	nmol/mol	1.7	9.4	-0.26	z'	-0.30
NO ₂	5	-1.1	-1.1	-1.1	-1.00	0.04	0.08	nmol/mol	1.4	0.06	-0.70	z'	-0.73
NO ₂	6	27	27	27	27	0.45	0.90	nmol/mol	2.2	29	-0.74	z'	-0.95
NO ₂	7	-2.7	-2.9	-2.7	-2.4	0.47	0.94	nmol/mol	1.6	7.6	-4.2	z'	-2.7
NO ₂	8	213	213	213	213	4.6	9.2	nmol/mol	7.8	228	-1.9	z	-1.5
NO ₂	9	-0.65	-0.70	-0.73	-0.53	0.31	0.63	nmol/mol	1.5	4.3	-2.5	z'	-1.9
NO ₂	10	128	129	128	128	2.6	5.3	nmol/mol	5.2	136	-1.5	z	-1.3
O ₃	0	1.1	1.1			1.2	2.3	nmol/mol	1.0	0.28	0.82	z	0.35
O ₃	1	68	68	68	68	1.00	2.1	nmol/mol	2.3	65	1.1	z	1.1
O ₃	2	23	23	23	23	1.1	2.2	nmol/mol	1.4	21	1.2	z	0.76
O ₃	3	202	201	202	202	2.6	5.1	nmol/mol	4.9	196	1.0	z'	0.87
O ₃	4	99	99	99	100	1.1	2.3	nmol/mol	2.9	96	0.99	z'	1.0
O ₃	5	134	134	134	134	1.4	2.8	nmol/mol	3.6	130	1.00	z'	1.0
SO ₂	0	-0.62	-0.62			1.2	2.4	nmol/mol	1.00	0.19	-0.72	z'	-0.31
SO ₂	1	124	124	124	124	1.5	3.0	nmol/mol	3.8	126	-0.47	z	-0.49
SO ₂	2	40	40	40	40	1.00	2.0	nmol/mol	1.9	41	-0.29	z'	-0.25
SO ₂	3	4.7	3.8	5.0	5.3	1.2	2.3	nmol/mol	1.1	4.8	-0.12	z'	-0.06
SO ₂	4	20	20	20	20	1.1	2.2	nmol/mol	1.4	20	-0.30	z'	-0.19
SO ₂	5	55	55	55	55	1.00	2.0	nmol/mol	2.2	56	-0.25	z	-0.24

Source: JRC 2024

Figure 25: Graphical report of results of LVGMC (Latvia).




Source: JRC 2024

Table 19: Reported values, uncertainties and performance evaluation results of LVGMC (Latvia).

LVGMC													
Component	Run	Lab mean	Step 1	Step 2	Step 3	Lab unc com	Lab unc exp	unit	sigmaPT	Ref. Val.	z/z'-score	z/z'-selection	En-score
CO	0	-0.01	-0.01			0.12	0.24	µmol/mol	0.10	-0.00	-0.07	z	-0.03
CO	1	0.55	0.55	0.54	0.55	0.12	0.24	µmol/mol	0.11	0.62	-0.64	z	-0.31
CO	2	8.4	8.4	8.4	8.4	0.30	0.60	µmol/mol	0.31	8.8	-1.2	z	-0.58
CO	3	0.84	0.84	0.83	0.84	0.12	0.24	µmol/mol	0.12	0.92	-0.64	z	-0.32
CO	4	1.8	1.8	1.8	1.8	0.13	0.26	µmol/mol	0.15	1.9	-0.64	z	-0.35
CO	5	4.6	4.6	4.6	4.6	0.19	0.38	µmol/mol	0.22	4.8	-1.1	z	-0.59
NO	0	0.29	0.29			0.63	1.3	nmol/mol	1.0	0.23	0.05	z'	0.03
NO	1	203	204	202	202	7.2	14	nmol/mol	5.8	200	0.43	z	0.17
NO	2	191	194	190	189	6.8	14	nmol/mol	5.6	191	0.04	z	0.02
NO	3	28	28	28	28	1.2	2.3	nmol/mol	1.7	31	-1.4	z'	-0.93
NO	4	19	19	19	19	0.92	1.8	nmol/mol	1.5	21	-1.2	z'	-0.86
NO	5	54	54	54	54	2.0	4.0	nmol/mol	2.4	60	-2.4	z'	-1.4
NO	6	37	37	37	37	1.5	2.9	nmol/mol	2.0	41	-1.9	z'	-1.3
NO	7	443	441	444	444	16	32	nmol/mol	12	473	-2.4	z	-0.92
NO	8	233	232	233	233	8.3	17	nmol/mol	7.0	251	-2.5	z	-1.0
NO	9	272	270	272	273	10.0	19	nmol/mol	8.1	296	-3.0	z	-1.3
NO	10	150	150	150	151	5.4	11	nmol/mol	4.9	164	-2.7	z	-1.2
NO ₂	0	0.00	0.00			0.87	1.7	nmol/mol	1.4	-0.04	0.03	z'	0.02
NO ₂	1	1.8	1.7	2.0	1.8	0.87	1.8	nmol/mol	1.4	0.42	0.81	z'	0.52
NO ₂	2	10	9.4	11	11	0.94	1.9	nmol/mol	1.7	9.9	0.27	z'	0.19
NO ₂	3	0.20	0.16	0.25	0.20	0.87	1.7	nmol/mol	1.4	0.13	0.05	z'	0.03
NO ₂	4	8.8	8.8	8.7	8.8	0.92	1.8	nmol/mol	1.7	9.4	-0.37	z'	-0.28
NO ₂	5	0.33	0.33	0.34	0.32	0.87	1.7	nmol/mol	1.4	0.06	0.17	z'	0.12
NO ₂	6	27	27	27	27	1.2	2.5	nmol/mol	2.2	29	-0.74	z'	-0.59
NO ₂	7	10.0	10.0	10.0	10.0	0.93	1.9	nmol/mol	1.6	7.6	1.00	z'	0.59
NO ₂	8	226	226	226	226	7.4	15	nmol/mol	7.8	228	-0.27	z	-0.14
NO ₂	9	5.2	5.3	5.4	5.0	0.89	1.8	nmol/mol	1.5	4.3	0.48	z'	0.30
NO ₂	10	133	132	133	133	4.4	8.8	nmol/mol	5.2	136	-0.63	z	-0.36
O ₃	0	0.03	0.03			1.3	2.6	nmol/mol	1.0	0.28	-0.24	z	-0.09
O ₃	1	65	65	65	65	1.5	3.0	nmol/mol	2.3	65	-0.17	z	-0.12
O ₃	2	21	21	21	21	1.3	2.6	nmol/mol	1.4	21	-0.19	z	-0.10
O ₃	3	196	195	196	196	2.6	5.1	nmol/mol	4.9	196	-0.09	z'	-0.07
O ₃	4	96	96	96	96	1.7	3.4	nmol/mol	2.9	96	-0.09	z'	-0.08
O ₃	5	130	130	130	130	2.0	3.9	nmol/mol	3.6	130	-0.05	z'	-0.04
SO ₂	0	-0.30	-0.30			0.48	0.00	nmol/mol	1.00	0.19	-0.44	z'	-0.48
SO ₂	1	140	140	140	140	4.2	8.4	nmol/mol	3.8	126	3.8	z	1.6
SO ₂	2	45	45	45	45	1.4	2.9	nmol/mol	1.9	41	2.2	z'	1.4
SO ₂	3	5.4	5.3	5.3	5.5	0.51	1.00	nmol/mol	1.1	4.8	0.40	z'	0.32
SO ₂	4	23	23	23	23	0.83	1.7	nmol/mol	1.4	20	1.6	z'	1.3
SO ₂	5	62	62	62	62	1.9	3.8	nmol/mol	2.2	56	2.9	z	1.6

Source: JRC 2024

Figure 26: Graphical report of results of NILU (Norway).



Source: JRC 2024

Table 20: Reported values, uncertainties and performance evaluation results of NILU (Norway).

NILU													
Component	Run	Lab mean	Step 1	Step 2	Step 3	Lab unc com	Lab unc exp	unit	sigmaPT	Ref. Val.	z/z'-score	z/z'-selection	En-score
CO	0	0.00		0.00	0.00	0.06	0.12	µmol/mol	0.10	-0.00	0.03	z	0.02
CO	1	0.64	0.64	0.64	0.64	0.06	0.13	µmol/mol	0.11	0.62	0.17	z	0.14
CO	2	9.0	9.0	9.0	9.0	0.30	0.60	µmol/mol	0.31	8.8	0.76	z	0.38
CO	3	0.94	0.94	0.94	0.94	0.07	0.13	µmol/mol	0.12	0.92	0.20	z	0.19
CO	4	1.9	1.9	1.9	1.9	0.10	0.20	µmol/mol	0.15	1.9	0.05	z	0.03
CO	5	4.9	4.9	4.9	4.9	0.20	0.30	µmol/mol	0.22	4.8	0.32	z	0.22
NO	0	0.02		0.05	0.00	0.58	1.2	nmol/mol	1.0	0.23	-0.17	z'	-0.11
NO	1	202	202	202	202	5.0	10.0	nmol/mol	5.8	200	0.31	z	0.18
NO	2	192	192	192	192	5.0	10.0	nmol/mol	5.6	191	0.22	z	0.12
NO	3	31	31	31	31	1.00	2.0	nmol/mol	1.7	31	0.24	z'	0.18
NO	4	21	21	21	21	1.00	2.0	nmol/mol	1.5	21	-0.00	z'	-0.00
NO	5	61	61	61	61	2.0	3.0	nmol/mol	2.4	60	0.33	z'	0.25
NO	6	42	42	42	42	1.00	2.0	nmol/mol	2.0	41	0.41	z'	0.35
NO	7	480	480	480	480	12	25	nmol/mol	12	473	0.59	z	0.28
NO	8	255	255	255	256	7.0	13	nmol/mol	7.0	251	0.68	z	0.36
NO	9	301	300	301	301	8.0	16	nmol/mol	8.1	296	0.52	z	0.26
NO	10	166	166	166	166	4.0	9.0	nmol/mol	4.9	164	0.49	z	0.26
NO ₂	0	0.10		0.10	0.10	0.58	1.2	nmol/mol	1.4	-0.04	0.09	z'	0.08
NO ₂	1	0.10	0.10	0.10	0.10	0.58	1.2	nmol/mol	1.4	0.42	-0.18	z'	-0.14
NO ₂	2	9.4	9.3	9.3	9.5	0.60	1.3	nmol/mol	1.7	9.9	-0.29	z'	-0.23
NO ₂	3	0.17	0.10	0.10	0.30	0.58	1.2	nmol/mol	1.4	0.13	0.02	z'	0.02
NO ₂	4	8.9	8.9	8.9	9.0	0.60	1.3	nmol/mol	1.7	9.4	-0.28	z'	-0.25
NO ₂	5	0.57	0.50	0.60	0.60	0.58	1.2	nmol/mol	1.4	0.06	0.32	z'	0.26
NO ₂	6	28	28	28	28	1.00	1.00	nmol/mol	2.2	29	-0.31	z'	-0.39
NO ₂	7	6.0	6.3	6.0	5.8	0.60	1.2	nmol/mol	1.6	7.6	-0.63	z'	-0.40
NO ₂	8	228	229	228	228	6.0	12	nmol/mol	7.8	228	0.03	z	0.02
NO ₂	9	3.0	3.2	3.0	2.9	0.60	1.2	nmol/mol	1.5	4.3	-0.63	z'	-0.44
NO ₂	10	136	136	135	136	4.0	7.0	nmol/mol	5.2	136	-0.06	z	-0.04
O ₃	0	0.10		0.10	0.10	0.78	1.6	nmol/mol	1.0	0.28	-0.17	z	-0.11
O ₃	1	67	67	67	67	2.0	5.0	nmol/mol	2.3	65	0.70	z	0.31
O ₃	2	22	22	22	22	1.00	1.00	nmol/mol	1.4	21	0.51	z	0.63
O ₃	3	199	199	199	199	7.0	14	nmol/mol	4.9	196	0.55	z'	0.20
O ₃	4	98	98	98	98	3.0	7.0	nmol/mol	2.9	96	0.56	z'	0.24
O ₃	5	132	132	132	132	5.0	9.0	nmol/mol	3.6	130	0.48	z'	0.19
SO ₂	0	0.15		0.15	0.15	0.59	1.2	nmol/mol	1.00	0.19	-0.04	z'	-0.03
SO ₂	1	127	127	128	127	7.0	14	nmol/mol	3.8	126	0.42	z	0.11
SO ₂	2	41	41	41	40	2.0	5.0	nmol/mol	1.9	41	0.04	z'	0.02
SO ₂	3	5.0		5.0	5.1	0.70	1.3	nmol/mol	1.1	4.8	0.16	z'	0.11
SO ₂	4	20	20	20	20	1.00	3.0	nmol/mol	1.4	20	-0.30	z'	-0.15
SO ₂	5	55	55	55	55	3.0	6.0	nmol/mol	2.2	56	-0.25	z	-0.09

Source: JRC 2024

Annex C: Reproducibility

1 Introduction

In this PT the standard ISO 5725 was used to confirm the test method's validity and reliability through the calculation of reproducibility.

The instruments used in these exercises have consistently demonstrated very high repeatability. Confirming observations made over the last 15 years, no significant variability has been noted among any of the participants. Consequently, due to the limited utility of this assessment, ERLAP has decided to omit repeatability from the report.

Reproducibility is defined as the closeness of the agreement between the results of measurements of the same measurand carried out under various conditions: different operators, different equipment, different equipment calibration etc.

During the PT the participants were invited to use, as much as possible, their own everyday work procedure.

The reproducibility limit (R) is a quantitative measure indicating the maximum difference in results that can be expected under reproducibility conditions for a given level of probability, usually 95%.

For the main purpose of monitoring trends between different PT undertaken by ERLAP, the precision of standardized SO₂, CO, O₃ and NO_x measurement methods [2], [3], [4] and [5] as implemented by NRLs, is evaluated. Due to the small number of participants the reproducibility results doesn't give an accurate response about the quality of the methods application during the PT. Applied methodology is described in ISO 5725-1, 5725-2 and 5725-6 [14], [15] and [16]. The precision experiment has involved a maximum of eleven laboratories, the actual number of laboratories (p) is listed in Table 21 according to the data reported.

Six concentration levels were tested for O₃, CO, SO₂ and NO₂, and eleven for NO. Performance tests for each participant are performed and results are reported in Annex B.

In ISO 5725 reproducibility (R) limits is defined. This limit is the value less than or equal to which the absolute difference between two test results, obtained under reproducibility conditions, may be expected to be with (1 - α) probability level.

The reproducibility standard deviation (SR) was calculated as the square root of sum of reproducibility of between-laboratory variance at the 95% confidence level (α). The reproducibility limit (R) is calculated using Equation 5 [16].

Equation 5

$$R = t_{v/\alpha} \cdot \sqrt{2} \cdot SR$$

R = reproducibility limit

t_{v/α} = t Student distribution value

SR = standard deviation of between-lab reproducibility.

The reproducibility standard deviation with (p-1) degrees of freedom (v) and the critical range Student factors, for R, corresponding to defined confidence level (α) and degree of freedom (v), are reported in Table 21.

Table 21: Critical values of t used in the reproducibility (R) evaluation.

parameter	run	p	t critical value 95% for R
CO	1	9	2.306
CO	2,3,4,5	10	2.262
NO	6	9	2.306
NO	7	11	2.228
NO	1,2,3,4,5,8,9,10	10	2.262
NO ₂	2	9	2.306
NO ₂	4	10	2.262
NO ₂	6,8,10	11	2.228
O ₃	1,2	9	2.306


O ₃	3,4,5	10	2.262
SO ₂	2	7	2.447
SO ₂	1,3,4,5	8	2.365

Source: JRC 2024

2 Results

The Reproducibility (R) limits of measurement methods are presented from Figure 27 to Figure 31. It is reported also the 'reproducibility from common criteria (R_Ref.)' calculated by substituting SR in Equation 5 with a 'standard deviation for proficiency assessment' (σ_{pt}) (see Table 4). Comparison between R and R_Ref. serves to indicate that σ_{pt} is realistic [13] or from the other point of view, that the general methodology implemented by NRLs is appropriate for σ_{pt} .


Figure 27: R of CO standard measurement method as a function of concentration.



Component	Run	Mean	R	R_ref	R%	unit
CO	0	-0.02	0.11	0.33		μmol/mol
CO	1	0.61	0.22	0.37	36	μmol/mol
CO	3	0.89	0.24	0.39	27	μmol/mol
CO	4	1.8	0.26	0.46	14	μmol/mol
CO	5	4.7	0.44	0.68	9.4	μmol/mol
CO	2	8.6	0.85	0.98	9.9	μmol/mol

Source: JRC 2024


Figure 28: R of NO standard measurement method as a function of concentration.



Component	Run	Mean	R	R_ref	R%	unit
NO	0	0.24	0.56	3.3		nmol/mol
NO	4	20	2.7	4.8	14	nmol/mol
NO	3	30	3.6	5.5	12	nmol/mol
NO	6	41	3.3	6.4	8.0	nmol/mol
NO	5	59	7.3	7.7	12	nmol/mol
NO	10	161	18	16	11	nmol/mol
NO	2	189	15	18	7.9	nmol/mol
NO	1	199	16	18	8.0	nmol/mol
NO	8	248	27	22	11	nmol/mol
NO	9	292	33	26	11	nmol/mol
NO	7	465	54	38	12	nmol/mol

Source: JRC 2024


Figure 29: R of NO₂ standard measurement method as a function of concentration.



Component	Run	Mean	R	R_ref	R%	unit
NO ₂	0	0.08	0.46	4.5		nmol/mol
NO ₂	4	9.1	1.3	5.3	14	nmol/mol
NO ₂	2	9.8	2.7	5.5	28	nmol/mol
NO ₂	6	28	4.2	6.8	15	nmol/mol
NO ₂	10	132	15	16	11	nmol/mol
NO ₂	8	224	24	24	11	nmol/mol

Source: JRC 2024


Figure 30: R of O₃ standard measurement method as a function of concentration.



Component	Run	Mean	R	R _{ref}	R%	unit
O ₃	0	0.38	1.1	3.3		nmol/mol
O ₃	2	21	2.3	4.7	11	nmol/mol
O ₃	1	66	3.9	7.6	5.9	nmol/mol
O ₃	4	97	4.9	9.4	5.1	nmol/mol
O ₃	5	131	7.1	12	5.4	nmol/mol
O ₃	3	198	9.0	16	4.5	nmol/mol

Source: JRC 2024

Figure 31: R of SO₂ standard measurement method as a function of concentration.



Component	Run	Mean	R	R _{ref}	R%	unit
SO ₂	0	-0.09	1.2	3.2		nmol/mol
SO ₂	3	5.0	0.73	3.6	15	nmol/mol
SO ₂	4	20	1.8	4.8	9.0	nmol/mol
SO ₂	2	40	2.4	6.5	6.0	nmol/mol
SO ₂	5	56	4.9	7.4	8.8	nmol/mol
SO ₂	1	126	14	13	11	nmol/mol

Source: JRC 2024

3 Conclusions

Before assessing reproducibility in this proficiency test (PT), outliers were identified and excluded using boxplot analysis.

A boxplot, or box and whisker diagram, is the tool used to depict the distribution of a dataset and to identify outliers among the participants' results. The plot highlights the lower quartile, median, and upper quartile, while the "whiskers" extend to the last data points not considered abnormal. Any data point that falls outside the 1.5 times the interquartile range (IQR) is considered an outlier, indicating a result significantly different from the norm.

The outliers were removed before proceeding with the reproducibility evaluation. At the highest concentration levels, the relative reproducibility limits are 9.9% for CO, 12% for NO, 11% for NO₂, 4.5% for O₃, and 11% for SO₂.

Graphical representations of reproducibility for CO and O₃ showed no intersection of the blue line (R) with the reference red line, signaling overall satisfactory performance for all pollutants measured. For NO, NO₂ and SO₂ the graphs show a poor performance of the measurement during this PT. As reported in Table 7 this proficiency test, when compared to nearly two decades of historical data, has regrettably been deemed the poorest in terms of the total number of z-score valid values.

Annex D: Confidentiality

Results of the PT are published according to the agreements included in the document AQUILA-N37 [12] approved by all NRL of the AQUILA network.

In order to ensure confidentiality of the laboratories information, ERLAP guarantees the submitted data as follows:

- Any administrative information provided by the laboratory is confidential and cannot be communicated to a third party.
- Access to ERLAP facilities is allowed only to members of the Unit JRC-C5 and authorized persons (cleaning staff, maintenance staff, safety and security staff etc.)
- Confidential passwords to access the web application for data submission are sent once the registration to PT is completed. Confidential passwords allow access to the WEB interface and to on-line questionnaire. Passwords are valid until the PT is closed. Laboratories can change their password online.

The form LAB-REC-2000 (Confidentiality involvement form) is asked to be signed by the participants during their first participation to a PT organized by ERLAP.

Annex E: Accreditation certificates



DL0018P/003

CERTIFICATO DI ACCREDITAMENTO Accreditation Certificate

ACCREDITAMENTO N.
ACCREDITATION N. **0018P REV. 03**

EMESSO DA
ISSUED BY **DIPARTIMENTO LABORATORI DI PROVA**

SI DICHIARA CHE
We DECLARE THAT **European Reference Laboratory for Air Pollution (ERLAP)
Clean Air and Climate Unit
Directorate C. Energy, Mobility and Climate
Joint Research Centre -European Commission**
Sede/Headquarters:
- Via E. Fermi 2749 - 21027 Ispra VA

IND-CAL-03 (REV. 06)

È CONFORME AI REQUISITI DELLA NORMA
UNI CEI EN ISO/IEC 17043:2010

MEETS THE REQUIREMENTS OF THE STANDARD
ISO/IEC 17043:2010

QUALE
Organizzatori di prove valutative interlaboratorio

AS
Proficiency Testing Provider

Data di 1ª emissione
1st issue date
17-01-2019

Data di revisione
Review date
03-02-2023

Data di scadenza
Expiring date
15-01-2027

L'accreditamento attesta la competenza, l'imparzialità e il costante e coerente funzionamento dell'Organizzazione relativamente al campo di accreditamento riportato nell'Elenco Schemi allegato al presente certificato di accreditamento.

Il presente certificato non è da ritenersi valido se non accompagnato dagli Elenchi Schemi e può essere sospeso o revocato o ridotto in qualsiasi momento nel caso di inadempienza accertata da parte di ACCREDIA.

La validità dell'accreditamento può essere verificata sul sito web (www.accredia.it) o richiesta al Dipartimento di competenza.

The accreditation attests competence, impartiality and consistent operation in performing laboratory activities, limited to the scope detailed in the attached Enclosure.

The present certificate is valid only if associated to the annexed Lists and can be suspended, withdrawn or reduced at any time in the event of non fulfillment as ascertained by ACCREDIA.

Confirmation of the validity of accreditation can be verified on the website (www.accredia.it) or by contacting the relevant Department.

Il QRcode consente di accedere direttamente al sito www.accredia.it per verificare la validità del certificato di accreditamento rilasciato al CAB.

La data di revisione riportata sul certificato corrisponde alla data di aggiornamento / di delibera del pertinente Comitato Settoriale di Accreditamento. L'atto di delibera, firmato dal Presidente di ACCREDIA, è scaricabile dal sito www.accredia.it, sezione "Documenti".

The QRcode links directly to the website www.accredia.it to check the validity of the accreditation certificate issued to the CAB.

The revision date shown on the certificate refers to the update / resolution date of the Sector Accreditation Committee. The Resolution, signed by the President of ACCREDIA, can be downloaded from the website www.accredia.it, "Documents" section.

ACCREDIA è l'Ente Unico nazionale di accreditamento designato dal governo italiano, in applicazione del Regolamento Europeo 765/2008.

ACCREDIA is the sole national Accreditation Body, appointed by the Italian government in compliance with the application of REGULATION (EC) No 765/2008.

ACCREDIA - Dipartimento Laboratori di prova

pag. 1/1

Sede operativa, legale e amministrativa: Via Guglielmo Saliceto, 7/9 | 00161 Roma - Italy
Tel. +39 06 8440991 | Fax +39 06 8841199
info@accredia.it | www.accredia.it | Partita IVA - Codice Fiscale 10566361001

Getting in touch with the EU

In person

All over the European Union there are hundreds of Europe Direct centres. You can find the address of the centre nearest you online (european-union.europa.eu/contact-eu/meet-us_en).

On the phone or in writing

Europe Direct is a service that answers your questions about the European Union. You can contact this service:

- by freephone: 00 800 6 7 8 9 10 11 (certain operators may charge for these calls),
- at the following standard number: +32 22999696,
- via the following form: european-union.europa.eu/contact-eu/write-us_en.

Finding information about the EU

Online

Information about the European Union in all the official languages of the EU is available on the Europa website (european-union.europa.eu).

EU publications

You can view or order EU publications at op.europa.eu/en/publications. Multiple copies of free publications can be obtained by contacting Europe Direct or your local documentation centre (european-union.europa.eu/contact-eu/meet-us_en).

EU law and related documents

For access to legal information from the EU, including all EU law since 1951 in all the official language versions, go to EUR-Lex (eur-lex.europa.eu).

EU open data

The portal data.europa.eu provides access to open datasets from the EU institutions, bodies and agencies. These can be downloaded and reused for free, for both commercial and non-commercial purposes. The portal also provides access to a wealth of datasets from European countries.

Science for policy

The Joint Research Centre (JRC) provides independent, evidence-based knowledge and science, supporting EU policies to positively impact society



EU Science Hub

[Joint-research-centre.ec.europa.eu](https://joint-research-centre.ec.europa.eu)



Publications Office
of the European Union